



Analytical intercomparison of heavy metals in precipitation 2001

Hilde Thelle Uggerud and Jan Erik Skjelmoen



NILU : EMEP/CCC-Report 2/2002
REFERENCE : O-7729
DATE : AUGUST 2002

**EMEP Co-operative Programme for Monitoring and Evaluation
of the Long-range Transmission of Air Pollutants
in Europe**

**Analytical intercomparison of
heavy metals in precipitation 2001**

Hilde Thelle Uggerud and Jan Erik Skjelmoen



Norwegian Institute for Air Research
P.O. Box 100, N-2027 Kjeller, Norway

Contents

	Page
1. Introduction.....	3
2. Organization of the intercomparison	3
3. Intercomparison samples	3
4. Presentation of data	4
5. Results	4
5.1 Chromium (Cr)	4
5.2 Nickel (Ni).....	4
5.3 Copper (Cu).....	5
5.4 Zinc (Zn).....	5
5.5 Arsenic (As).....	5
5.6 Cadmium (Cd).....	5
5.7 Lead (Pb)	6
6. Conclusions and further work	6
7. References.....	6
Appendix 1 Tables	8
Appendix 2 Figures	19

Analytical intercomparison of heavy metals in precipitation 2001

1. Introduction

Heavy metals were included in the EMEP's monitoring programme in 1999. 21 laboratories are reporting data to the heavy metal database. Since EMEP's measurement programme is based on individual national networks, different sampling and analytical methods are applied by the participating laboratories. In order to ensure data comparability, interlaboratory tests are organized by the Chemical Co-ordinating Centre (CCC) at the Norwegian Institute for Air Research. So far three intercomparisons have been arranged (Berg and Semb, 1995; Berg and Aas, 1999; Uggerud and Skjelmoen, 2000).

This report presents results from the fourth analytical intercomparison of heavy metals in precipitation, which was carried out during 2001. Seven heavy metals were included: Pb, Cd, Cu, Zn, As, Cr, and Ni.

2. Organization of the intercomparison

The samples for the fourth intercomparison were prepared and distributed to 45 laboratories. In addition to 16 EMEP laboratories, 13 laboratories connected to the ICP-forest measurement programme, 13 laboratories connected to WMO and 3 other laboratories also received samples. A total of 29 laboratories, 15 from the EMEP network, reported results within the end of October 2001. In accordance with the decision of the Steering Body of EMEP, the results are presented in such a way that the different laboratories are identified. Tables A.1.1a and A1.1b give the names of the participating laboratories together with the number used when presenting the results in tables and figures.

3. Intercomparison samples

The four synthetic precipitation samples distributed were made from multi-element standards traceable to NIST-standards. The multi-element standards were conserved with 2.5% HNO₃. The distributed synthetic precipitation samples contained Pb, Cd, Cu, Zn, As, Cr, and Ni in 0.5% HNO₃. Sample H1 and H2 contained concentrations similar to what is normally found in Southern Scandinavia. Sample H3 and H4 contained the elements in concentrations normally found in Central Europe.

All equipment in contact with the samples were soaked in 3% HNO₃ for 4 days. Preparation of the intercomparison samples was carried out in a clean room area.

4. Presentation of data

Tables A1.2-A1.8 present the reported results in decreasing order together with the laboratory numbers. The theoretical value, the number of results, the arithmetic mean value, the median, the standard deviation and the relative standard deviation are also given. In the first statistic run only values below detection limit were excluded. In the second run also outliers were excluded. The outliers were defined as values more than two standard deviations from the mean value in the first run.

In Figures A2.1-A2.7 the results are presented in plots showing the relative percentage deviation from expected value for each participating laboratory. There is one plot for every single sample.

The median calculated from the results reported from EMEP laboratories and other participating laboratories respectively, are compared to expected value in Figure A2.8.

A summary of the results is presented in Table A1.9. The results reported from each laboratory are divided in four percent intervals. The number of results reported by the laboratories in each per cent interval is also shown.

Table A1.10 gives information of the analytical techniques used by each laboratory.

5. Results

The analytical results from the intercomparison are presented in Figures A2.1-A2.8, Tables A1.2-A1.8 and Table A1.9. The results that were reported by the participating laboratories were generally in accordance with the theoretical values, with good agreement between the median from the second runs and the expected values.

5.1 Chromium (Cr)

A total of 25 laboratories reported values for Cr. Four laboratories reported eight values below detection limit for the low concentration samples. Seven laboratories reported results from the low concentration samples that deviated more than 25% from expected value, whereas ten laboratories reported results more than 15% away from expected value for the high concentration samples. The relative standard deviations for the low concentration samples were 23.2% and 32.2%, outliers excluded. For the high concentration samples the relative standard deviation were 10.6% and 12.7%, when outliers were excluded. This is the same as last year's results.

5.2 Nickel (Ni)

26 laboratories reported results for Ni. Six laboratories reported values below detection limit for the low concentration samples, whereas no such results were reported for the high concentration samples. For low concentration samples, results that deviated more than 25% from expected value, were reported by nine

laboratories. For high concentration samples, three laboratories reported results that deviated more than 15% from expected value. The relative standard deviations, for the two low concentration samples, were 103.8% and 48.1% when outliers were excluded. Compared to last year's result this might seem bad, but the concentrations in the samples distributed this year were lower than in last year's samples. For the high concentration samples, the relative standard deviations were 11% and 7.3%, which is an improvement compared to earlier intercomparisons.

5.3 Copper (Cu)

A total of 28 laboratories reported values for the determination of Cu. Three laboratories reported results from analysis of low concentration samples to be below detection limit. No such results were reported for the high concentration samples. Six laboratories reported values, for the low concentration samples, that deviated more than 25% from expected value. For the high concentration samples, six laboratories reported results that deviated more than 15%. The relative standard deviations for the two low concentration samples were 13.2% and 21.9%, when outliers were excluded. For the high concentration samples the standard deviations were 7.8% and 10.8% when outliers were excluded.

5.4 Zinc (Zn)

Results were obtained from 29 laboratories. For the low concentration samples, three laboratories reported values below detection limits. For both low and high concentration samples, eight laboratories reported values more than 25% and 15% away from expected value, respectively. The relative standard deviations were for the two low concentration samples 28.6% and 23.7%, and for the high concentration samples 18.6% and 18%, when outliers were excluded. For both sample groups, these results are not as good as those obtained in the last intercomparison.

5.5 Arsenic (As)

A total of 23 laboratories reported values for the determination of As. Five laboratories reported values below detection limit. For low concentration samples four laboratories reported values that deviated more than 25% from expected value, while six laboratories reported values that deviated more than 15% from expected value for the high concentration samples. The relative standard deviations for the two low concentration samples were 16.8% and 8.5%, when outliers were excluded. For the high concentration samples the relative standard deviations were 8.2% and 8.3%, when outliers were excluded. For both sample groups, this is an improvement compared to last year's results.

5.6 Cadmium (Cd)

Results for the determination of Cd were obtained from 27 laboratories. 11 laboratories reported values below detection limit. Six laboratories reported results for low concentration samples that deviated more than 25% from expected value. For the high concentration samples seven laboratories reported results that deviated more than 15%. Relative standard deviation for the two low concentration samples were 27.2% and 29.3%, outliers excluded. For the high

concentration samples, the relative standard deviations were 8.3% and 7.5%, outliers excluded. This is an improvement compared to last intercomparison.

5.7 Lead (Pb)

A total of 28 laboratories reported values for the determination of Pb. Two laboratories reported values to be below detection limit. Four laboratories reported results for the low concentration samples that deviated more than 25% from expected value. For high concentration samples six laboratories reported values that deviated more than 15%. Relative standard deviations for low concentration samples were 12% and 9.4%, when outliers were excluded. This is an improvement compared to last year. For the two high concentration samples, the relative standard deviations were 9.2% and 8.2%, when outliers were excluded. This is slightly worse than last year.

6. Conclusions and further work

The elements showed the following order of success: Pb>As>Cu>Cd>Cr>Zn>Ni.

For all the samples analysed the deviation from the theoretical value was calculated. The median deviations for the EMEP laboratories were below 6.3% and below 5.45% for the low- and high concentration samples, respectively. For low concentration samples this is a marked improvement. The median deviations for the other participating laboratories were below 6.9% for high concentration samples. This is about the same as last year. The median deviations for low concentration samples were (exclusive the median value for Ni in sample H1), below 34%. This is not as good as last year's result.

The field intercomparison on Hg in precipitation and air planned to be arranged in Germany during 2002, is postponed to spring 2003.

A new data flag system is in progress. For heavy metal data, the new data quality objectives differentiate between high and low concentrations, the accuracy in the laboratory should be better than 15% and 25% for high and low concentrations, respectively.

7. References

Berg, T., Hjellbrekke, A.-G. and Larsen, R. (2000) Heavy metals and POPs within Europe 1998. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 2/2000).

Berg, T. and Semb, A. (1995) Preliminary results from HELCOM-EMEP-PARCOM-AMAP analytical intercomparison of heavy metals in precipitation. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Note 1/95).

Berg, T. and Aas, W. (1999) Analytical intercomparison of heavy metals in precipitation 1999. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 8/2000).

Uggerud, H.Th. and Skjelmoen, J.E. (2001) Analytical intercomparison of heavy metals in precipitation 2000. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 12/2001).

Appendix 1

Tables

Table A1.1a: Participating laboratories in the EMEP network. The numbers in front are used in tables.

No	Laboratory identification
1	Federal Environmental Agency, Austria
3	Czech Hydrometeorological Institute, Czech Republic
5	Finnish Meteorological Institute, Finland
6	Laboratories Wolff, France
8	Umweltbundesamt, Germany
14	RIVM Laboratory of Inorganic Analytical Chemistry, The Netherlands
15	The Norwegian Institute for Air Research, Norway
16	Inst. Of Meteorology and Water Management, Poland
23	AEA Technology, National Environmental Techn. Centre, United Kingdom
26	Ontario Ministry of Environment, Canada
31	Slovak Hydrometeorological Institute, Slovakia
33	Latvian Hydrometeorological Agency, Latvia
34	Ministry of Health, Refit Saydam Hygiene Center, Turkey
36	Hydrometeorological Institute of Slovenia, Slovenia
38	Estonian Environmental Research Centre, Estonia

Table A1.1b: Participating laboratories outside the EMEP network. The number in front of the names is used in tables and figures.

No	Laboratory identification
105	Universität des Saarlandes Umweltforschungszentrum, Germany
109	Institut f. Bondenkunde und Waldernährung der Universität, Germany
112	Niedersächsische Forstliche Versuchsanstalt (N VF), Germany
115	Bayerische Landesanstalt f. Wald- und Forstwirtschaft, Germany
117	Sächsische Landesanstalt für Forsten, Germany
118	Forstliche Versuchs- und Forschungsanstalt Baden-Württemberg, Germany
119	Landesumweltamt (LUA)
121	Landesamt für Natur und Umwelt, Germany
125	Bayerisches Landesamt für Umweltschutz, Germany
127	Department of Chemistry, Jalan Sultan, Malaysia
128	Dubai Central Laboratory, United Arab Emirates
129	Ecole Nationale d'Ingenieurs de Sfax, Tunisie
132	Comision Chilena De Energia Nuclear, Chile
134	Wissenschaftszentrum für Ernährung, Landnutzung und Umwelt, Germany

Table A1.2: Analytical results for Cr in synthetic precipitation samples.

<p>Cr</p> <p>SAMPLE NO.: H1 THEORETICAL VALUE 0.400 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 23 ARITHMETIC MEAN VALUE: 0.665 MEDIAN: 0.410 STANDARD DEVIATION: 0.845 REL. ST. DEVIATION (%): 127.006</p> <p>RUN 2: NUMBER OF LABORATORIES: 18 ARITHMETIC MEAN VALUE: 0.474 MEDIAN: 0.410 STANDARD DEVIATION: 0.153 REL. ST. DEVIATION (%): 32.222</p> <p>RESULTS IN DECREASING ORDER: 125 4.100 UNUSED 132 < 1.05 6 < 1.0 38 < 1.0 117 < 1.0 112 0.820 115 0.770 3 0.730 109 0.570 14 0.500 127 0.500 8 0.430 1 0.410 15 0.410</p> <p>121 0.410 23 0.405 16 0.400 36 0.400 105 0.400 26 0.390 119 0.390 31 0.355 5 0.250</p> <p>"UNUSED": DATA UNUSED IN RUN 2</p>	<p>Cr</p> <p>SAMPLE NO.: H2 THEORETICAL VALUE 0.600 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 24 ARITHMETIC MEAN VALUE: 0.726 MEDIAN: 0.610 STANDARD DEVIATION: 0.335 REL. ST. DEVIATION (%): 46.125</p> <p>RUN 2: NUMBER OF LABORATORIES: 19 ARITHMETIC MEAN VALUE: 0.659 MEDIAN: 0.610 STANDARD DEVIATION: 0.153 REL. ST. DEVIATION (%): 23.188</p> <p>RESULTS IN DECREASING ORDER: 129 2.000 UNUSED 132 < 1.05 6 < 1.0 38 < 1.0 117 < 1.0 3 1.000 112 0.960 115 0.940 105 0.700 127 0.700 109 0.670 15 0.620 23 0.620 8 0.610</p> <p>119 0.610 125 0.609 14 0.600 16 0.600 36 0.600 1 0.590 26 0.590 121 0.590 31 0.526 5 0.380</p> <p>"UNUSED": DATA UNUSED IN RUN 2</p>
<p>Cr</p> <p>SAMPLE NO.: H3 THEORETICAL VALUE 8.500 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 25 ARITHMETIC MEAN VALUE: 8.200 MEDIAN: 8.450 STANDARD DEVIATION: 1.461 REL. ST. DEVIATION (%): 17.819</p> <p>RUN 2: NUMBER OF LABORATORIES: 23 ARITHMETIC MEAN VALUE: 8.229 MEDIAN: 8.450 STANDARD DEVIATION: 0.876 REL. ST. DEVIATION (%): 10.649</p> <p>RESULTS IN DECREASING ORDER: 129 12.000 UNUSED 26 8.390 38 10.200 125 8.360 3 9.200 15 8.300 23 9.110 36 8.100 112 8.980 118 7.800 115 8.960 6 7.570 8 8.530 109 7.500 14 8.500 127 7.500 16 8.500 132 7.050 31 8.499 5 6.430 119 8.490 105 6.400 1 8.450 117 3.740 UNUSED</p> <p>"UNUSED": DATA UNUSED IN RUN 2</p>	<p>Cr</p> <p>SAMPLE NO.: H4 THEORETICAL VALUE 1.500 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 25 ARITHMETIC MEAN VALUE: 1.695 MEDIAN: 1.500 STANDARD DEVIATION: 0.916 REL. ST. DEVIATION (%): 54.072</p> <p>RUN 2: NUMBER OF LABORATORIES: 24 ARITHMETIC MEAN VALUE: 1.516 MEDIAN: 1.500 STANDARD DEVIATION: 0.193 REL. ST. DEVIATION (%): 12.704</p> <p>RESULTS IN DECREASING ORDER: 129 6.000 UNUSED 16 1.500 115 1.900 105 1.500 112 1.830 121 1.500 117 1.830 132 1.480 3 1.800 26 1.440 14 1.700 36 1.400 23 1.610 109 1.400 119 1.560 125 1.400 38 1.550 118 1.300 31 1.523 127 1.300 8 1.510 5 1.170 1 1.500 6 1.170 15 1.500</p> <p>"UNUSED": DATA UNUSED IN RUN 2</p>

Table A1.3: Analytical results for Ni in synthetic precipitation samples.

<p>Ni SAMPLE NO.: H1 THEORETICAL VALUE 0.300 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 24 ARITHMETIC MEAN VALUE: 0.747 MEDIAN: 0.314 STANDARD DEVIATION: 0.998 REL. ST. DEVIATION (%): 133.623</p> <p>RUN 2: NUMBER OF LABORATORIES: 17 ARITHMETIC MEAN VALUE: 0.551 MEDIAN: 0.310 STANDARD DEVIATION: 0.572 REL. ST. DEVIATION (%): 103.806</p> <p>RESULTS IN DECREASING ORDER:</p> <table border="0"> <tbody> <tr><td>125</td><td>4.070</td><td>UNUSED</td><td>1</td><td>0.310</td></tr> <tr><td>105</td><td>2.400</td><td></td><td>121</td><td>0.310</td></tr> <tr><td>127</td><td>< 2</td><td></td><td></td><td></td></tr> <tr><td>132</td><td>< 1.35</td><td></td><td></td><td></td></tr> <tr><td>115</td><td>1.290</td><td></td><td>14</td><td>0.300</td></tr> <tr><td>38</td><td>1.100</td><td></td><td>26</td><td>0.300</td></tr> <tr><td>117</td><td>< 1</td><td></td><td></td><td></td></tr> <tr><td>119</td><td>< 0.9</td><td></td><td>34</td><td>0.300</td></tr> <tr><td>31</td><td>0.567</td><td></td><td>36</td><td>0.300</td></tr> <tr><td>3</td><td>0.550</td><td></td><td>8</td><td>0.280</td></tr> <tr><td>112</td><td>0.530</td><td></td><td>5</td><td>0.250</td></tr> <tr><td>6</td><td>< 0.5</td><td></td><td></td><td></td></tr> <tr><td>16</td><td>0.350</td><td></td><td>15</td><td>0.220</td></tr> <tr><td>23</td><td>0.318</td><td></td><td>109</td><td>< 0.1</td></tr> </tbody> </table> <p>"UNUSED": DATA UNUSED IN RUN 2</p>	125	4.070	UNUSED	1	0.310	105	2.400		121	0.310	127	< 2				132	< 1.35				115	1.290		14	0.300	38	1.100		26	0.300	117	< 1				119	< 0.9		34	0.300	31	0.567		36	0.300	3	0.550		8	0.280	112	0.530		5	0.250	6	< 0.5				16	0.350		15	0.220	23	0.318		109	< 0.1	<p>Ni SAMPLE NO.: H2 THEORETICAL VALUE 0.500 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 24 ARITHMETIC MEAN VALUE: 0.664 MEDIAN: 0.500 STANDARD DEVIATION: 0.452 REL. ST. DEVIATION (%): 68.057</p> <p>RUN 2: NUMBER OF LABORATORIES: 18 ARITHMETIC MEAN VALUE: 0.579 MEDIAN: 0.500 STANDARD DEVIATION: 0.278 REL. ST. DEVIATION (%): 48.061</p> <p>RESULTS IN DECREASING ORDER:</p> <table border="0"> <tbody> <tr><td>105</td><td>2.200</td><td>UNUSED</td><td>26</td><td>0.500</td></tr> <tr><td>127</td><td>< 2</td><td></td><td></td><td></td></tr> <tr><td>115</td><td>1.490</td><td></td><td>36</td><td>0.500</td></tr> <tr><td>132</td><td>< 1.35</td><td></td><td></td><td></td></tr> <tr><td>38</td><td>1.150</td><td></td><td>16</td><td>0.480</td></tr> <tr><td>117</td><td>< 1</td><td></td><td></td><td></td></tr> <tr><td>119</td><td>< 0.9</td><td></td><td></td><td></td></tr> <tr><td>31</td><td>0.649</td><td></td><td>8</td><td>0.460</td></tr> <tr><td>3</td><td>0.590</td><td></td><td>125</td><td>0.447</td></tr> <tr><td>112</td><td>0.560</td><td></td><td>15</td><td>0.430</td></tr> <tr><td>23</td><td>0.535</td><td></td><td>5</td><td>0.410</td></tr> <tr><td>1</td><td>0.510</td><td></td><td>34</td><td>0.400</td></tr> <tr><td>121</td><td>0.510</td><td></td><td>109</td><td>0.300</td></tr> <tr><td>6</td><td>< 0.5</td><td></td><td></td><td></td></tr> <tr><td>14</td><td>0.500</td><td></td><td></td><td></td></tr> </tbody> </table> <p>"UNUSED": DATA UNUSED IN RUN 2</p>	105	2.200	UNUSED	26	0.500	127	< 2				115	1.490		36	0.500	132	< 1.35				38	1.150		16	0.480	117	< 1				119	< 0.9				31	0.649		8	0.460	3	0.590		125	0.447	112	0.560		15	0.430	23	0.535		5	0.410	1	0.510		34	0.400	121	0.510		109	0.300	6	< 0.5				14	0.500			
125	4.070	UNUSED	1	0.310																																																																																																																																														
105	2.400		121	0.310																																																																																																																																														
127	< 2																																																																																																																																																	
132	< 1.35																																																																																																																																																	
115	1.290		14	0.300																																																																																																																																														
38	1.100		26	0.300																																																																																																																																														
117	< 1																																																																																																																																																	
119	< 0.9		34	0.300																																																																																																																																														
31	0.567		36	0.300																																																																																																																																														
3	0.550		8	0.280																																																																																																																																														
112	0.530		5	0.250																																																																																																																																														
6	< 0.5																																																																																																																																																	
16	0.350		15	0.220																																																																																																																																														
23	0.318		109	< 0.1																																																																																																																																														
105	2.200	UNUSED	26	0.500																																																																																																																																														
127	< 2																																																																																																																																																	
115	1.490		36	0.500																																																																																																																																														
132	< 1.35																																																																																																																																																	
38	1.150		16	0.480																																																																																																																																														
117	< 1																																																																																																																																																	
119	< 0.9																																																																																																																																																	
31	0.649		8	0.460																																																																																																																																														
3	0.590		125	0.447																																																																																																																																														
112	0.560		15	0.430																																																																																																																																														
23	0.535		5	0.410																																																																																																																																														
1	0.510		34	0.400																																																																																																																																														
121	0.510		109	0.300																																																																																																																																														
6	< 0.5																																																																																																																																																	
14	0.500																																																																																																																																																	
<p>Ni SAMPLE NO.: H3 THEORETICAL VALUE 6.700 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 26 ARITHMETIC MEAN VALUE: 7.548 MEDIAN: 6.630 STANDARD DEVIATION: 4.154 REL. ST. DEVIATION (%): 55.380</p> <p>RUN 2: NUMBER OF LABORATORIES: 25 ARITHMETIC MEAN VALUE: 6.729 MEDIAN: 6.600 STANDARD DEVIATION: 0.738 REL. ST. DEVIATION (%): 10.966</p> <p>RESULTS IN DECREASING ORDER:</p> <table border="0"> <tbody> <tr><td>129</td><td>28.000</td><td>UNUSED</td><td>1</td><td>6.600</td></tr> <tr><td>105</td><td>9.800</td><td></td><td>121</td><td>6.600</td></tr> <tr><td>38</td><td>7.420</td><td></td><td>132</td><td>6.520</td></tr> <tr><td>115</td><td>7.290</td><td></td><td>36</td><td>6.500</td></tr> <tr><td>31</td><td>7.245</td><td></td><td>109</td><td>6.500</td></tr> <tr><td>112</td><td>6.990</td><td></td><td>125</td><td>6.410</td></tr> <tr><td>16</td><td>6.900</td><td></td><td>127</td><td>6.400</td></tr> <tr><td>23</td><td>6.820</td><td></td><td>6</td><td>6.350</td></tr> <tr><td>14</td><td>6.800</td><td></td><td>15</td><td>6.200</td></tr> <tr><td>3</td><td>6.770</td><td></td><td>34</td><td>6.200</td></tr> <tr><td>8</td><td>6.740</td><td></td><td>118</td><td>6.000</td></tr> <tr><td>119</td><td>6.720</td><td></td><td>117</td><td>5.960</td></tr> <tr><td>26</td><td>6.660</td><td></td><td>5</td><td>5.840</td></tr> </tbody> </table> <p>"UNUSED": DATA UNUSED IN RUN 2</p>	129	28.000	UNUSED	1	6.600	105	9.800		121	6.600	38	7.420		132	6.520	115	7.290		36	6.500	31	7.245		109	6.500	112	6.990		125	6.410	16	6.900		127	6.400	23	6.820		6	6.350	14	6.800		15	6.200	3	6.770		34	6.200	8	6.740		118	6.000	119	6.720		117	5.960	26	6.660		5	5.840	<p>Ni SAMPLE NO.: H4 THEORETICAL VALUE 8.500 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 26 ARITHMETIC MEAN VALUE: 8.251 MEDIAN: 8.320 STANDARD DEVIATION: 0.751 REL. ST. DEVIATION (%): 9.097</p> <p>RUN 2: NUMBER OF LABORATORIES: 25 ARITHMETIC MEAN VALUE: 8.324 MEDIAN: 8.340 STANDARD DEVIATION: 0.612 REL. ST. DEVIATION (%): 7.342</p> <p>RESULTS IN DECREASING ORDER:</p> <table border="0"> <tbody> <tr><td>38</td><td>9.570</td><td></td><td>132</td><td>8.300</td></tr> <tr><td>34</td><td>9.400</td><td></td><td>26</td><td>8.280</td></tr> <tr><td>23</td><td>9.100</td><td></td><td>6</td><td>8.200</td></tr> <tr><td>115</td><td>9.050</td><td></td><td>15</td><td>8.100</td></tr> <tr><td>31</td><td>8.808</td><td></td><td>36</td><td>8.100</td></tr> <tr><td>3</td><td>8.740</td><td></td><td>109</td><td>8.100</td></tr> <tr><td>112</td><td>8.680</td><td></td><td>117</td><td>7.940</td></tr> <tr><td>14</td><td>8.600</td><td></td><td>125</td><td>7.900</td></tr> <tr><td>119</td><td>8.540</td><td></td><td>127</td><td>7.800</td></tr> <tr><td>16</td><td>8.470</td><td></td><td>118</td><td>7.600</td></tr> <tr><td>1</td><td>8.450</td><td></td><td>5</td><td>7.220</td></tr> <tr><td>121</td><td>8.450</td><td></td><td>105</td><td>6.800</td></tr> <tr><td>8</td><td>8.340</td><td></td><td>129</td><td>6.000 UNUSED</td></tr> </tbody> </table> <p>"UNUSED": DATA UNUSED IN RUN 2</p>	38	9.570		132	8.300	34	9.400		26	8.280	23	9.100		6	8.200	115	9.050		15	8.100	31	8.808		36	8.100	3	8.740		109	8.100	112	8.680		117	7.940	14	8.600		125	7.900	119	8.540		127	7.800	16	8.470		118	7.600	1	8.450		5	7.220	121	8.450		105	6.800	8	8.340		129	6.000 UNUSED															
129	28.000	UNUSED	1	6.600																																																																																																																																														
105	9.800		121	6.600																																																																																																																																														
38	7.420		132	6.520																																																																																																																																														
115	7.290		36	6.500																																																																																																																																														
31	7.245		109	6.500																																																																																																																																														
112	6.990		125	6.410																																																																																																																																														
16	6.900		127	6.400																																																																																																																																														
23	6.820		6	6.350																																																																																																																																														
14	6.800		15	6.200																																																																																																																																														
3	6.770		34	6.200																																																																																																																																														
8	6.740		118	6.000																																																																																																																																														
119	6.720		117	5.960																																																																																																																																														
26	6.660		5	5.840																																																																																																																																														
38	9.570		132	8.300																																																																																																																																														
34	9.400		26	8.280																																																																																																																																														
23	9.100		6	8.200																																																																																																																																														
115	9.050		15	8.100																																																																																																																																														
31	8.808		36	8.100																																																																																																																																														
3	8.740		109	8.100																																																																																																																																														
112	8.680		117	7.940																																																																																																																																														
14	8.600		125	7.900																																																																																																																																														
119	8.540		127	7.800																																																																																																																																														
16	8.470		118	7.600																																																																																																																																														
1	8.450		5	7.220																																																																																																																																														
121	8.450		105	6.800																																																																																																																																														
8	8.340		129	6.000 UNUSED																																																																																																																																														

Table A1.4: Analytical results for Cu in synthetic precipitation samples.

Cu SAMPLE NO.: H1 THEORETICAL VALUE 1.500 UNIT: ng /ml	Cu SAMPLE NO.: H2 THEORETICAL VALUE 1.200 UNIT: ng /ml
RUN 1: NUMBER OF LABORATORIES: 27 ARITHMETIC MEAN VALUE: 1.500 MEDIAN: 1.500 STANDARD DEVIATION: 0.433 REL. ST. DEVIATION (%): 28.891	RUN 1: NUMBER OF LABORATORIES: 27 ARITHMETIC MEAN VALUE: 1.140 MEDIAN: 1.170 STANDARD DEVIATION: 0.250 REL. ST. DEVIATION (%): 21.943
RUN 2: NUMBER OF LABORATORIES: 21 ARITHMETIC MEAN VALUE: 1.426 MEDIAN: 1.450 STANDARD DEVIATION: 0.188 REL. ST. DEVIATION (%): 13.211	RUN 2: NUMBER OF LABORATORIES: 24 ARITHMETIC MEAN VALUE: 1.140 MEDIAN: 1.170 STANDARD DEVIATION: 0.250 REL. ST. DEVIATION (%): 21.943
RESULTS IN DECREASING ORDER: 33 2.900 UNUSED 26 1.450 118 2.400 UNUSED 31 1.440 128 < 2 117 1.760 UNUSED 112 1.430 115 1.730 14 1.400 38 1.700 15 1.400 125 1.690 132 1.340 119 < 1.68 3 1.550 5 1.320 8 1.540 6 1.240 1 1.510 34 1.100 121 1.510 109 1.100 16 1.500 127 1.000 23 1.500 105 0.500 UNUSED 36 1.500	RESULTS IN DECREASING ORDER: 128 < 2 119 < 1.68 33 1.500 121 1.170 118 1.500 26 1.160 112 1.350 117 1.150 115 1.350 15 1.100 3 1.300 16 1.100 14 1.300 36 1.100 38 1.300 105 1.100 23 1.280 31 1.060 125 1.210 6 < 1.0 5 1.000 132 1.180 109 0.800 1 1.170 127 0.700 8 1.170 34 0.300
"UNUSED": DATA UNUSED IN RUN 2	"UNUSED": DATA UNUSED IN RUN 2
Cu SAMPLE NO.: H3 THEORETICAL VALUE 7.200 UNIT: ng /ml	Cu SAMPLE NO.: H4 THEORETICAL VALUE 10.000 UNIT: ng /ml
RUN 1: NUMBER OF LABORATORIES: 28 ARITHMETIC MEAN VALUE: 6.868 MEDIAN: 7.000 STANDARD DEVIATION: 0.983 REL. ST. DEVIATION (%): 14.310	RUN 1: NUMBER OF LABORATORIES: 28 ARITHMETIC MEAN VALUE: 9.180 MEDIAN: 9.710 STANDARD DEVIATION: 1.846 REL. ST. DEVIATION (%): 20.108
RUN 2: NUMBER OF LABORATORIES: 25 ARITHMETIC MEAN VALUE: 7.004 MEDIAN: 7.000 STANDARD DEVIATION: 0.553 REL. ST. DEVIATION (%): 7.890	RUN 2: NUMBER OF LABORATORIES: 27 ARITHMETIC MEAN VALUE: 9.477 MEDIAN: 9.720 STANDARD DEVIATION: 1.028 REL. ST. DEVIATION (%): 10.848
RESULTS IN DECREASING ORDER: 33 8.900 UNUSED 36 7.000 117 8.600 121 7.000 3 7.700 26 6.920 23 7.600 15 6.800 118 7.500 38 6.700 115 7.490 128 6.650 125 7.380 109 6.500 8 7.330 132 6.440 14 7.300 5 6.280 31 7.180 6 6.260 119 7.140 34 6.200 112 7.030 127 6.100 1 7.000 105 4.300 UNUSED 16 7.000 129 4.000 UNUSED	RESULTS IN DECREASING ORDER: 23 10.700 36 9.700 117 10.550 112 9.670 14 10.400 128 9.590 115 10.400 8 9.540 118 10.400 26 9.540 3 10.100 132 9.350 31 10.020 38 9.200 16 10.000 109 9.200 119 9.970 127 8.800 125 9.810 6 8.610 15 9.800 5 8.500 33 9.800 105 6.800 1 9.720 129 6.000 121 9.720 34 1.150 UNUSED
"UNUSED": DATA UNUSED IN RUN 2	"UNUSED": DATA UNUSED IN RUN 2

Table A1.5: Analytical results for Zn in synthetic precipitation samples.

Zn SAMPLE NO.: H1 THEORETICAL VALUE 7.000 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 28 ARITHMETIC MEAN VALUE: 6.895 MEDIAN: 6.680 STANDARD DEVIATION: 3.193 REL. ST. DEVIATION (%): 46.313 RUN 2: NUMBER OF LABORATORIES: 23 ARITHMETIC MEAN VALUE: 6.708 MEDIAN: 6.680 STANDARD DEVIATION: 1.918 REL. ST. DEVIATION (%): 28.600 RESULTS IN DECREASING ORDER: 119 < 30 15 6.700 127 < 25 1 6.680 129 18.000 UNUSED 121 6.680 125 12.900 36 6.500 38 < 10 5 6.290 33 8.700 14 6.000 8 8.300 118 5.900 23 7.980 132 5.740 34 7.800 117 5.510 115 7.700 112 5.310 26 7.240 127 4.800 3 7.100 105 4.400 31 7.000 6 2.150 16 6.900 109 0.100 UNUSED "UNUSED": DATA UNUSED IN RUN 2	Zn SAMPLE NO.: H2 THEORETICAL VALUE 6.500 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 28 ARITHMETIC MEAN VALUE: 7.040 MEDIAN: 6.000 STANDARD DEVIATION: 7.344 REL. ST. DEVIATION (%): 104.325 RUN 2: NUMBER OF LABORATORIES: 23 ARITHMETIC MEAN VALUE: 5.820 MEDIAN: 6.000 STANDARD DEVIATION: 1.381 REL. ST. DEVIATION (%): 23.720 RESULTS IN DECREASING ORDER: 129 42.000 UNUSED 14 6.000 119 < 30 31 6.000 127 < 25 34 5.800 38 < 10 5 5.600 23 8.700 36 5.400 33 7.500 132 5.330 8 7.100 118 5.100 26 7.080 117 5.080 125 6.990 112 4.930 115 6.870 105 4.800 3 6.800 127 4.300 15 6.400 16 3.970 1 6.180 6 1.780 121 6.180 109 0.100 UNUSED "UNUSED": DATA UNUSED IN RUN 2
Zn SAMPLE NO.: H3 THEORETICAL VALUE 110.000 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 29 ARITHMETIC MEAN VALUE: 94.855 MEDIAN: 104.000 STANDARD DEVIATION: 30.364 REL. ST. DEVIATION (%): 32.011 RUN 2: NUMBER OF LABORATORIES: 27 ARITHMETIC MEAN VALUE: 101.446 MEDIAN: 104.000 STANDARD DEVIATION: 18.918 REL. ST. DEVIATION (%): 18.648 RESULTS IN DECREASING ORDER: 8 123.000 23 103.000 129 122.000 33 102.000 125 115.000 5 101.000 3 113.000 118 100.800 112 112.000 16 100.400 119 112.000 134 100.000 6 110.500 132 97.600 14 110.000 36 97.000 31 110.000 127 94.800 115 110.000 117 92.240 128 110.000 38 42.700 26 108.000 105 37.000 15 107.000 34 11.650 UNUSED 1 104.000 109 0.100 UNUSED "UNUSED": DATA UNUSED IN RUN 2	Zn SAMPLE NO.: H4 THEORETICAL VALUE 130.000 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 29 ARITHMETIC MEAN VALUE: 114.021 MEDIAN: 123.000 STANDARD DEVIATION: 36.081 REL. ST. DEVIATION (%): 31.644 RUN 2: NUMBER OF LABORATORIES: 27 ARITHMETIC MEAN VALUE: 121.956 MEDIAN: 128.000 STANDARD DEVIATION: 21.956 REL. ST. DEVIATION (%): 18.003 RESULTS IN DECREASING ORDER: 8 156.000 121 123.000 23 154.000 118 120.900 119 134.000 16 120.200 3 133.000 134 120.000 112 133.000 5 120.000 14 132.000 33 119.000 6 131.500 36 116.000 15 131.000 132 114.000 115 131.000 127 113.800 125 131.000 117 110.100 26 130.000 38 65.300 31 130.000 105 43.000 128 130.000 34 13.710 UNUSED 129 128.000 109 0.110 UNUSED 1 123.000 "UNUSED": DATA UNUSED IN RUN 2

Table A1.6: Analytical results for As in synthetic precipitation samples.

<p>As SAMPLE NO.: H1 THEORETICAL VALUE 0.300 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 22 ARITHMETIC MEAN VALUE: 0.302 MEDIAN: 0.300 STANDARD DEVIATION: 0.060 REL. ST. DEVIATION (%): 19.945</p> <p>RUN 2: NUMBER OF LABORATORIES: 16 ARITHMETIC MEAN VALUE: 0.292 MEDIAN: 0.299 STANDARD DEVIATION: 0.048 REL. ST. DEVIATION (%): 16.423</p> <p>RESULTS IN DECREASING ORDER: 128 < 5 119 < 1.05 38 < 1 117 < 1 112 < 0.6 6 0.450 UNUSED 125 0.299 33 0.400 31 0.278 115 0.380 23 0.273 3 0.320 15 0.270 1 0.300 8 0.260 26 0.300 5 0.250 36 0.300 14 0.250 121 0.300 132 0.200 127 0.300 15 0.470</p> <p>"UNUSED": DATA UNUSED IN RUN 2</p>	<p>As SAMPLE NO.: H2 THEORETICAL VALUE 0.500 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 22 ARITHMETIC MEAN VALUE: 0.466 MEDIAN: 0.470 STANDARD DEVIATION: 0.072 REL. ST. DEVIATION (%): 15.466</p> <p>RUN 2: NUMBER OF LABORATORIES: 15 ARITHMETIC MEAN VALUE: 0.465 MEDIAN: 0.470 STANDARD DEVIATION: 0.039 REL. ST. DEVIATION (%): 8.476</p> <p>RESULTS IN DECREASING ORDER: 128 < 5 119 < 1.05 38 < 1 117 < 1 6 0.650 UNUSED 121 0.470 112 < 0.6 23 0.530 31 0.460 115 0.530 26 0.450 36 0.500 132 0.450 125 0.485 14 0.430 1 0.470 127 0.400 3 0.470 5 0.390 8 0.470 33 0.300 UNUSED 15 0.470</p> <p>"UNUSED": DATA UNUSED IN RUN 2</p>
<p>As SAMPLE NO.: H3 THEORETICAL VALUE 2.500 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 23 ARITHMETIC MEAN VALUE: 2.412 MEDIAN: 2.405 STANDARD DEVIATION: 0.293 REL. ST. DEVIATION (%): 12.155</p> <p>RUN 2: NUMBER OF LABORATORIES: 20 ARITHMETIC MEAN VALUE: 2.411 MEDIAN: 2.405 STANDARD DEVIATION: 0.199 REL. ST. DEVIATION (%): 8.262</p> <p>RESULTS IN DECREASING ORDER: 128 < 5 8 2.400 6 3.150 UNUSED 1 2.390 112 2.780 121 2.390 3 2.700 14 2.370 132 2.640 15 2.300 23 2.570 118 2.300 31 2.550 127 2.200 119 2.540 117 2.150 36 2.500 5 2.100 125 2.490 33 2.000 26 2.440 38 1.700 UNUSED 115 2.410 1 4.850</p> <p>"UNUSED": DATA UNUSED IN RUN 2</p>	<p>As SAMPLE NO.: H4 THEORETICAL VALUE 5.000 UNIT: ng /ml</p> <p>RUN 1: NUMBER OF LABORATORIES: 23 ARITHMETIC MEAN VALUE: 4.910 MEDIAN: 4.850 STANDARD DEVIATION: 0.596 REL. ST. DEVIATION (%): 12.142</p> <p>RUN 2: NUMBER OF LABORATORIES: 20 ARITHMETIC MEAN VALUE: 4.926 MEDIAN: 4.850 STANDARD DEVIATION: 0.411 REL. ST. DEVIATION (%): 8.334</p> <p>RESULTS IN DECREASING ORDER: 6 6.200 UNUSED 121 4.850 23 5.780 115 4.830 33 5.700 125 4.810 117 5.500 15 4.800 3 5.300 118 4.800 31 5.050 112 4.780 119 5.000 8 4.760 128 < 5 132 4.600 14 4.960 127 4.200 26 4.920 5 4.130 36 4.900 38 3.300 UNUSED 1 4.850</p> <p>"UNUSED": DATA UNUSED IN RUN 2</p>

Table A1.7: Analytical results for Cd in synthetic precipitation samples.

Cd SAMPLE NO.: H1 THEORETICAL VALUE 0.040 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 26 ARITHMETIC MEAN VALUE: 0.053 MEDIAN: 0.040 STANDARD DEVIATION: 0.040 REL. ST. DEVIATION (%): 74.600 RUN 2: NUMBER OF LABORATORIES: 15 ARITHMETIC MEAN VALUE: 0.043 MEDIAN: 0.040 STANDARD DEVIATION: 0.012 REL. ST. DEVIATION (%): 27.158 RESULTS IN DECREASING ORDER: 128 < 1 125 0.042 132 < 0.6 1 0.040 112 < 0.32 31 0.040 34 0.200 33 0.040 36 < 0.2 121 0.040 127 < 0.2 119 < 0.035 6 < 0.1 23 0.039 109 < 0.1 16 0.038 38 0.080 8 0.037 115 0.060 14 0.037 26 < 0.05 5 0.034 15 0.047 117 0.030 3 0.045 105 < 0.01 "UNUSED": DATA UNUSED IN RUN 2	Cd SAMPLE NO.: H2 THEORETICAL VALUE 0.060 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 27 ARITHMETIC MEAN VALUE: 0.091 MEDIAN: 0.060 STANDARD DEVIATION: 0.092 REL. ST. DEVIATION (%): 101.063 RUN 2: NUMBER OF LABORATORIES: 17 ARITHMETIC MEAN VALUE: 0.060 MEDIAN: 0.059 STANDARD DEVIATION: 0.018 REL. ST. DEVIATION (%): 29.252 RESULTS IN DECREASING ORDER: 128 < 1 16 0.060 132 < 0.6 26 0.060 112 < 0.32 121 0.060 105 0.300 UNUSED 8 0.059 34 0.300 UNUSED 125 0.059 36 < 0.2 31 0.058 127 < 0.2 33 0.054 38 0.120 14 0.052 6 < 0.1 26 < 0.05 109 < 0.1 5 0.048 115 0.080 119 0.046 23 0.067 15 0.040 3 0.064 117 0.040 1 0.060 "UNUSED": DATA UNUSED IN RUN 2
Cd SAMPLE NO.: H3 THEORETICAL VALUE 0.600 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 27 ARITHMETIC MEAN VALUE: 0.604 MEDIAN: 0.600 STANDARD DEVIATION: 0.063 REL. ST. DEVIATION (%): 10.443 RUN 2: NUMBER OF LABORATORIES: 26 ARITHMETIC MEAN VALUE: 0.595 MEDIAN: 0.600 STANDARD DEVIATION: 0.050 REL. ST. DEVIATION (%): 8.336 RESULTS IN DECREASING ORDER: 128 < 1 109 0.600 34 0.800 UNUSED 118 0.600 112 0.720 8 0.594 132 0.670 6 0.590 38 0.640 1 0.580 3 0.634 121 0.580 23 0.633 14 0.567 31 0.625 117 0.540 26 0.620 33 0.532 15 0.610 5 0.530 115 0.610 36 0.500 119 0.610 127 0.500 125 0.604 105 < 0.01 16 0.600 "UNUSED": DATA UNUSED IN RUN 2	Cd SAMPLE NO.: H4 THEORETICAL VALUE 0.900 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 27 ARITHMETIC MEAN VALUE: 0.875 MEDIAN: 0.900 STANDARD DEVIATION: 0.162 REL. ST. DEVIATION (%): 18.532 RUN 2: NUMBER OF LABORATORIES: 24 ARITHMETIC MEAN VALUE: 0.899 MEDIAN: 0.900 STANDARD DEVIATION: 0.067 REL. ST. DEVIATION (%): 7.484 RESULTS IN DECREASING ORDER: 128 < 1 16 0.900 34 1.100 26 0.900 118 1.100 109 0.900 23 1.010 1 0.880 132 1.010 121 0.880 8 0.989 125 0.866 3 0.979 117 0.830 31 0.953 36 0.800 112 0.950 127 0.800 119 0.930 33 0.785 115 0.920 5 0.780 15 0.910 105 0.400 UNUSED 38 0.910 14 0.380 UNUSED 6 0.900 "UNUSED": DATA UNUSED IN RUN 2

Table A1.8: Analytical results for Pb in synthetic precipitation samples.

Pb SAMPLE NO.: H1 THEORETICAL VALUE 1.300 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 27 ARITHMETIC MEAN VALUE: 1.250 MEDIAN: 1.240 STANDARD DEVIATION: 0.330 REL. ST. DEVIATION (%): 26.409 RUN 2: NUMBER OF LABORATORIES: 22 ARITHMETIC MEAN VALUE: 1.279 MEDIAN: 1.265 STANDARD DEVIATION: 0.153 REL. ST. DEVIATION (%): 11.978 RESULTS IN DECREASING ORDER: 128 < 20 119 1.220 118 2.200 UNUSED 15 1.200 115 1.630 16 1.200 112 1.560 33 1.200 34 1.500 36 1.200 3 1.450 23 1.180 31 1.399 5 1.150 125 1.360 132 1.130 6 1.300 117 1.040 8 1.300 38 < 1 14 1.300 109 1.000 1 1.290 127 0.500 UNUSED 121 1.290 105 0.400 UNUSED 26 1.240 "UNUSED": DATA UNUSED IN RUN 2	Pb SAMPLE NO.: H2 THEORETICAL VALUE 1.700 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 25 ARITHMETIC MEAN VALUE: 1.636 MEDIAN: 1.690 STANDARD DEVIATION: 0.279 REL. ST. DEVIATION (%): 17.034 RUN 2: NUMBER OF LABORATORIES: 23 ARITHMETIC MEAN VALUE: 1.704 MEDIAN: 1.700 STANDARD DEVIATION: 0.161 REL. ST. DEVIATION (%): 9.430 RESULTS IN DECREASING ORDER: 128 < 20 121 1.690 31 2.032 26 1.610 109 2.000 15 1.600 112 1.940 118 1.600 3 1.920 117 1.570 125 1.810 132 1.570 34 1.800 16 1.500 115 1.790 33 1.500 8 1.780 36 1.500 119 1.740 5 1.440 23 1.710 38 < 1 6 1.700 127 0.900 UNUSED 14 1.700 105 0.800 UNUSED 1 1.690 "UNUSED": DATA UNUSED IN RUN 2
Pb SAMPLE NO.: H3 THEORETICAL VALUE 45.000 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 28 ARITHMETIC MEAN VALUE: 43.617 MEDIAN: 44.750 STANDARD DEVIATION: 9.909 REL. ST. DEVIATION (%): 22.718 RUN 2: NUMBER OF LABORATORIES: 26 ARITHMETIC MEAN VALUE: 44.026 MEDIAN: 44.750 STANDARD DEVIATION: 4.096 REL. ST. DEVIATION (%): 9.241 RESULTS IN DECREASING ORDER: 132 71.900 UNUSED 8 44.700 128 52.850 1 44.600 117 48.200 121 44.600 112 47.300 3 43.540 6 47.200 26 43.200 115 47.200 15 43.000 23 46.100 16 42.600 119 45.900 36 42.500 31 45.680 127 41.600 118 45.600 5 41.000 125 45.400 105 40.500 14 45.100 38 36.600 33 44.900 129 30.000 34 44.800 109 4.700 UNUSED "UNUSED": DATA UNUSED IN RUN 2	Pb SAMPLE NO.: H4 THEORETICAL VALUE 30.000 UNIT: ng /ml RUN 1: NUMBER OF LABORATORIES: 28 ARITHMETIC MEAN VALUE: 28.634 MEDIAN: 29.650 STANDARD DEVIATION: 6.067 REL. ST. DEVIATION (%): 21.187 RUN 2: NUMBER OF LABORATORIES: 26 ARITHMETIC MEAN VALUE: 30.106 MEDIAN: 29.700 STANDARD DEVIATION: 2.461 REL. ST. DEVIATION (%): 8.175 RESULTS IN DECREASING ORDER: 128 35.690 8 29.600 132 35.200 3 29.450 34 35.200 26 29.400 23 32.100 125 29.200 118 31.900 16 28.500 115 31.800 36 28.400 117 31.700 6 28.150 112 31.600 15 28.000 31 30.370 127 28.000 1 29.800 105 27.000 14 29.800 5 27.000 121 29.800 38 25.700 33 29.700 129 16.000 UNUSED 119 29.700 109 3.000 UNUSED "UNUSED": DATA UNUSED IN RUN 2

Table A1.9: Percentage deviation from theoretical concentration value.

Element and percent interval	Lab. Identification The number in brackets are number of results reported in the particular percent interval by the laboratory		
	EMEP	Other	
Cr			
0-5%	9	5	1(4), 8(3), 14(2), 15(4), 16(4), 23(2), 26(4), 32(2), 36(3), 105(2), 119(4), 121(4), 125(2), 132(1)
5-15%	7	6	3(1), 6(1), 8(1), 14(1), 23(2), 31(2), 36(1), 109(3), 112(1), 115(1), 118(2), 125(1), 127(2),
15-25%	5	4	3(1), 5(2), 6(1), 14(1), 10(1), 112(1), 117(1), 127(2), 132(1)
>25%	2	7	3(2), 5(2), 105(1), 109(1), 112(2), 115(3), 117(1), 125(1), 129(3)
Ni			
0-5%	12	7	1(4), 3(2), 6(1), 8(2), 14(4), 5(1), 16(3), 23(1), 26(4), 31(1), 34(1), 36(4), 109(2), 112(2), 119(2), 121(4), 125(1), 127(1), 132(2)
5-15%	8	6	5(1), 6(1), 8(2), 15(2), 23(3), 31(1), 34(2), 38(2), 112(1), 115(2), 117(2), 118(2), 125(2), 127(1),
15-25%	3	1	5(3), 16(1), 34(1), 105(1)
>25%	3	6	3(1), 31(2), 38(2), 105(3), 109(1), 112(1), 115(2), 125(1), 129(2)
Cu			
0-5%	11	8	1(4), 3(2), 8(4), 14(2), 15(1), 16(3), 23(1), 26(4), 31(3), 33(1), 36(3), 112(3), 115(2), 117(2), 118(2), 119(2), 121(4), 125(3), 132(2)
5-15%	10	7	3(2), 5(3), 6(2), 14(2), 15(3), 16(1), 23(3), 31(1), 36(1), 38(4), 105(1), 109(2), 112(1), 115(2), 125(1), 127(1), 132(2)
15-25%	4	3	5(1), 6(1), 33(2), 34(1), 117(2), 118(1), 127(1),
>25%	2	5	33(1), 34(3), 105(3), 109(2), 117(1), 127(2), 129(2)
Zn			
0-5%	8	6	1(2), 3(4), 6(2), 14(2), 15(4), 16(1), 26(3), 31(3), 112(2), 115(2), 119(2), 121(2), 125(2), 129(1)
5-15%	11	7	1(2), 5(4), 8(2), 14(2), 16(2), 23(2), 26(1), 31(1), 33(2), 34(2), 36(3), 115(2), 118(2), 121(2), 125(1), 129(1), 132(2), 134(2)
15-25%	4	5	8(2), 23(1), 33(2), 36(1), 112(2), 117(4), 118(2), 127(2), 132(2)
>25%	5	5	6(2), 16(1), 23(1), 34(2), 38(2), 105(4), 109(4), 125(1), 127(2), 129(2)
As			
0-5%	8	7	1(3), 8(2), 14(1), 15(1), 23(1), 26(3), 31(2), 36(4), 112(1), 115(2), 118(1), 119(2), 121(3), 125(4), 127(1)
5-15%	9	7	1(1), 3(4), 8(2), 14(2), 15(3), 23(2), 26(1), 31(2), 33(1), 112(1), 115(1), 117(2), 118(1), 121(1), 127(1), 132(3)
15-25%	5	1	5(4), 6(1), 14(1), 23(1), 33(1), 127(2)
>25%	3	2	6(3), 33(2), 38(2), 115(1), 132(1)
Cd			
0-5%	10	6	1(4), 6(2), 8(2), 15(2), 16(4), 23(1), 26(3), 31(3), 33(1), 38(1), 109(2), 115(2), 118(1), 119(2), 121(4), 125(4)
5-15%	9	4	3(4), 5(3), 8(2), 14(3), 23(3), 31(1), 33(3), 36(1), 38(1), 112(1), 117(2), 127(1), 133(1)
15-25%	4	5	5(1), 15(1), 34(1), 36(1), 112(1), 117(1), 118(1), 119(1), 127(1)
>25%	2	3	14(1), 15(1), 34(3), 105(2), 115(2), 117(1)
Pb			
0-5%	12	5	1(4), 3(2), 6(3), 8(4), 14(4), 15(2), 16(1), 23(2), 26(3), 31(2), 33(2), 34(1), 115(1), 118(1), 119(3), 121(4), 125(3)
5-15%	12	9	3(2), 5(3), 6(1), 15(2), 16(3), 23(2), 26(1), 31(1), 33(2), 34(1), 36(4), 38(1), 105(2), 112(3), 115(2), 117(3), 118(2), 119(1), 125(1), 127(2), 132(2)
15-25%	4	5	5(1), 31(1), 34(2), 38(1), 109(2), 112(1), 117(1), 128(2), 132(1)
>25%	0	7	105(2), 109(2), 115(1), 118(1), 127(2), 129(2), 132(1)

Table A1.10: Analytical techniques used at the participating laboratories for the different elements.

Lab. no.	Elements	Technique
1	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
3	Ni, Cd, Cu, Pb, Cr, As Zn	GF-AAS ICP-MS F-AAS
5	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
6	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
8	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
14	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
15	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
16	Cr, Ni, Cu, Cd, Pb Zn	GF-AAS F-AAS
23	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
26	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
31	Cr, Ni, Cu, Cd, Pb As Zn	GF-AAS HG-GF-AAS F-AAS
33	Cu, Cd, Pb Zn As	GF-AAS F-AAS HG-AAS
34	Ni, Cu	AAS
36	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
38	Cr, Ni, Cu, , Cd, Pb Zn	GF-AAS F-AAS
105	Cr, Ni, Cu, Zn, Cd, Pb	GF-AAS
109	Cr, Ni, Cu, Zn, Cd, Pb	GF-AAS
112	Cr, Ni, Cu, Zn, As, Cd, Pb	USN-ICP-MS
115	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
117	Cr, Ni, Cu , Zn As, Cd, Pb	USN-ICP-OES GF-AAS
118	Cu, Cd, Pb As ,Zn Cr, Ni	GF-AAS ICP-OES
119	Cr, Ni, Cu, As, Cd, Pb Zn	GF-AAS F-AAS
121	Cr, Ni, Cu, Cd, Pb Zn As	GF-AAS Voltammetry HG-AAS
125	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
127	Cr, Ni, Cu, Zn, As, Cd, Pb	ICP-MS
128	Cu, Cd, Pb As Zn	GF-AAS HG-AAS F-AAS
129	Cr, Ni, Cu, Zn, Cd, Pb	F-AAS (Polarized Zeeman)
132	Cr, Ni, Cu, Zn , Cd Pb As	USN-ICP-OES GF-AAS HG-AAS
134	Zn	ICP-OES

Appendix 2

Figures




Figure A2.1: Results from determination of Cr.




Figure A2.2: Results from determination of Ni.




Figure A2.3: Results from determination of Cu.




Figure A2.4: Results from determination of Zn.




Figure A2.5: Results from determination of As.




Figure A2.6: Results from determination of Cd.




Figure A2.7: Results from determination of Pb.




Figure A2.8: The median compared to theoretical value for low and high concentration samples, respectively.