


The twenty-third and twenty-fourth intercomparison of analytical methods within EMEP

Hilde Th. Uggerud and Anne-Gunn Hjellbrekke



NILU : EMEP/CCC-Report 5/2007
REFERENCE : O-7729
DATE : JULY 2008

**EMEP Co-operative Programme for Monitoring and Evaluation
of the Long-range Transmission of Air Pollutants
in Europe**

**The twenty-third and twenty-fourth
intercomparison of
analytical methods within EMEP**

Hilde Th. Uggerud and Anne-Gunn Hjellbrekke



Norwegian Institute for Air Research
P.O. Box 100, N-2027 Kjeller, Norway

Contents

	Page
1. Introduction.....	5
2. Organisation of the intercomparisons.....	5
2.1 The twenty-third intercomparison	5
2.2 The twenty-fourth intercomparison	6
3. Data handling	6
3.1 Data analysis.....	6
3.2 Bar plots.....	6
3.3 Youden plot	7
3.4 Estimating random and systematic errors from laboratory comparisons	8
3.4.1 Estimating random errors	8
3.4.2 Estimating systematic errors.....	10
4. Results – 23rd intercomparison	10
4.1 Sulphur dioxide in absorbing solution (A-samples).....	10
4.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples) ...	10
4.3 Nitrogen dioxide in absorbing solution (C-samples).....	11
4.4 Ammonia on impregnated filters (J-samples)	11
4.5 Precipitation (G-samples).....	11
4.5.1 Conductivity and ion balance	11
5. Results – 24th intercomparison.....	12
5.1 Sulphur dioxide in absorbing solution (A-samples).....	12
5.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples) ...	12
5.3 Nitrogen dioxide in absorbing solution (C-samples).....	13
5.4 Ammonia on impregnated filters (J-samples)	13
5.5 Precipitation (G-samples).....	13
5.5.1 Conductivity and ion balance	13
6. Summary.....	14
7. References	15
Appendix 1 Tables – 23rd intercomparison	17
Appendix 2 Figures – 23rd intercomparison.....	49
Appendix 3 Tables – 24rd intercomparison	81
Appendix 4 Figures – 24rd intercomparison.....	115

The twenty-third and twenty-fourth intercomparison of analytical methods within EMEP

1. Introduction

36 different laboratories in European countries are performing chemical analysis of air and precipitation samples within EMEP (Co-operative Programme for Monitoring and Evaluation of Long-range Transmission of Air Pollutants in Europe). Since the measurement programme is based on individual national networks, the participating laboratories apply different sampling and analytical methods. Most of the methods used are described in the manual for sampling and chemical analysis (EMEP, 1996).

In order to improve the data comparability and to get a picture of the different laboratories' performance, interlaboratory comparisons are organised by the Chemical Co-ordinating Centre (CCC) at the Norwegian Institute for Air Research (NILU). So far twenty-four intercomparisons have been arranged (Hanssen, 1988, 1990; Hanssen et al., 1983; Hanssen and Ladegård, 1984, 1985, 1987; Hanssen and Skjelmoen, 1992, 1994, 1995, 1996, 1997, 2001; Thrane, 1978, 1980a, 1980b, 1981; Uggerud et al., 2001, 2002, 2003, 2004; Hjellbrekke et al., 2005).

Since 2000 the laboratory intercomparisons within EMEP have also been open for participation of laboratories from other networks.

This report gives the results of the twenty-third interlaboratory test.

2. Organisation of the intercomparisons

2.1 The twenty-third intercomparison

The samples for the twenty-third intercomparison (see Table 2) were prepared and distributed to 91 laboratories in July 2005.

Most of the laboratories had returned their results to the CCC within one month after the deadline given as 15 October 2005. A total of 65 laboratories have returned their results. This includes 32 EMEP-laboratories.

The participating laboratories received the theoretical (expected) values by e-mail 18.11.2005. The laboratories were given the opportunity to compare their results with the expected ones, and give corrected values if obvious mistakes e.g. misprints had occurred. A few corrections were reported. In those cases the corrected values are used in this report. In accordance with the decision of the Steering Body of EMEP, the results are presented in such a way that the different laboratories are identified. Tables 3a and 3b give the names of the participating laboratories together with the numbers used when presenting the results in tables and figures.

Information received on the analytical methods used is given in Tables 4–7.

2.2 The twenty-fourth intercomparison

The samples for the twenty-fourth intercomparison were prepared and distributed to 90 laboratories in July 2006.

Most of the laboratories had returned their results to the CCC within one month after the deadline given as 15 October 2005. A total of 63 laboratories have returned their results. This includes 32 EMEP-laboratories.

The participating laboratories received the theoretical (expected) values by e-mail 13.12.2006. The laboratories were given the opportunity to compare their results with the expected ones, and give corrected values if obvious mistakes e.g. misprints had occurred. A few corrections were reported. In those cases the corrected values are used in this report. In accordance with the decision of the Steering Body of EMEP, the results are presented in such a way that the different laboratories are identified. Tables 3a and 3b give the names of the participating laboratories together with the numbers used when presenting the results in tables and figures.

Information received on the analytical methods used is given in Tables 34–38.

3. Data handling

The data reported for the twenty-third intercomparison from the participants are presented in Tables 8, 10, 12, 14 and 16–26.

The data reported for the twenty-fourth intercomparison from the participants are presented in Tables 39, 41, 43, 44, 46 and 48–58.

3.1 Data analysis

The reported values are presented in the tables in decreasing order together with the number of the laboratory. The expected (theoretical) value, the number of results, the arithmetic mean value, the median, the standard deviation and the relative standard deviation in percent are also given. After the first statistical run with all results included, the calculation was repeated with the outliers excluded. The outliers (unused) are defined as the results more than two standard deviations from the mean value in the first run.

The ratio between expected values (theoretical) to reported values, the ratio between measured to calculated conductivity and the ratio between equivalent concentrations of anions to equivalent concentrations cations, are presented in tables.

3.2 Bar plots

Bar-plots are used for the graphical presentation of the data. Figures 2–16 are showing the relative deviation from expected value for the different laboratories. There is one plot for each single sample.

Figure 17 gives median compared to expected value for the results reported by EMEP-laboratories and the other participating laboratories, respectively.

3.3 Youden plot

The Youden plot is a graphical method to analyse inter-laboratory data where the samples are ordered in pairs with similar concentrations. One plot is made for each pair of samples and gives results for all participating laboratories. The plots visualize both systematic and random errors.




Figure 1: Youden plot showing concentrations for a pair of samples (green), expected values (blue lines), average of measured values (dotted lines) and random and systematic errors (red lines)

The plot is draw as a scatter plot where each point represents a pair of concentrations for one laboratory. The expected values for the two samples are drawn as solid blue lines. The arithmetic average of the measured values excluding outliers are drawn as dotted lines. The solid lines divide the plot in four quadrants and a 45° reference line going through the intercept of the solid lines may be added.

If errors are due to random factors, the points will be evenly distributed around the mean value and situated in all four quadrants.

If systematic errors dominate, the results will be close to 45° reference line, and be situated in the upper right quadrant (overestimation) or lower left quadrant (underestimation).

Drawing a line from a given point perpendicular on the 45° reference line gives two line segments, one from the point to the intercept on the reference line (a), and one continuing from the intercept to the point representing the expected values (b). The lengths of these line segments are measures of the random and systematic errors respectively.

Ellipses with radii corresponding to the data quality objectives (DQO, Table 1) are added in each plot. The data points are colour coded depending on the magnitude of errors as given in Table 1.

In Figures 18–33 the reported data are presented in Youden plots.

Table 1: Youden plot parameters.


Radii	Components
10%	SO ₂ in abs.sol, NO ₂ in abs.sol.
20%	SO ₂ , HNO ₃ and NH ₃ in impregnated filter
Radii = DQO	Components
10% accuracy or better	SO ₄ ²⁻ , NO ₃ ⁻
15% accuracy or better	NH ₄ ⁺ , Cl ⁻ , Ca ²⁺ , K ⁺ , Mg ²⁺ , Na ⁺ , cond, H+ (from pH)
0.1 units	pH
Criteria	Colour
Within 0.5*DQO	Blue
Within DQO	Green
Within 2*DQO	Orange
> 2*DQO	Red

3.4 Estimating random and systematic errors from laboratory comparisons

Table 30 presents relative random and relative systematic errors obtained by the different laboratories in the analysis of each parameter in the precipitation samples. The calculation method and assumptions used are given in Chapter 3.4.1 and Chapter 3.4.2.

3.4.1 Estimating random errors

Systematic errors or bias in the laboratory analyses give a constant shift in the results from the expected ones at a particular concentration level. It is assumed that laboratories taking part in comparisons will obtain results near the expected ones when this bias is removed, and that the differences between expected and obtained results more often will be close to zero than not. A triangular distribution, based upon this assumption, can be used to quantify the random errors in the laboratory results (Eurachem/CITAC, 2000).



The triangle distribution is symmetric with a baseline 2a. The height in the triangle will be 1/a when the triangle area equals 1. The standard uncertainty is given by

$$u(x) = \frac{a}{\sqrt{6}} \quad (1)$$

and more than 95% of the data will be within $\pm 2 \cdot u(x)$. The distance from -a to a (i.e. 2a) is called the range. When applied on the laboratory comparison results, the range equals the distance between the largest and smallest of the four differences between expected and found concentrations. As long as the bias can be assumed to be constant for the samples in the comparison of a specific component, it cannot have an effect on the distance corresponding to 2a. The bias may be dependent upon the concentrations, but can be considered approximate constant for the concentrations used here in the comparison of the main components in precipitation, since the differences between the concentrations are small.

L and T represent the laboratories' and the expected concentrations respectively, and D is the difference. The difference for the lowest concentration is

$$D_1 = L_1 - T_1 \quad (2)$$

and the differences are D_1, D_2, D_3, D_4 in increasing order.

The range is $D_4 - D_1$ and the standard uncertainty for the differences $u(D)$ becomes

$$u(D) = \frac{D_4 - D_1}{\sqrt{6}} \quad (3)$$

The average expected concentration T for the four samples is given by

$$T = \frac{T_1 + T_2 + T_3 + T_4}{4} \quad (4)$$

The relative standard uncertainty, RSD, for 4 samples is given by $\frac{u(D)}{T}$, or

$$RSD = \frac{2 \cdot (D_4 - D_1) \cdot 100}{\sqrt{6} \cdot (T_1 + T_2 + T_3 + T_4)} \% \quad (5)$$

and 95 per cent of the laboratory results in this comparison are expected to be within $\pm 2 \cdot RSD$.

If the data quality objectives (DQO) likewise are looked upon as 95 percentiles, then 95 per cent of the laboratory analytical results should not be more than 10 or 15 per cent from the correct values (10 per cent for S and N containing components and 15 per cent for other components).

Correspondingly, the values 2·RSD should therefore be less than 10 or 15 per cent in order to comply with the DQO.

3.4.2 Estimating systematic errors

An estimation of bias in single measurements requires a long data series, and four samples as we normally have in laboratory comparison, are merely able to give an indication of the bias or a very coarse estimate.

Coarse estimates have been performed here in the cases where the four samples had similar concentrations and where all four laboratory results were either higher or lower than the expected concentrations. The median of the differences D_i , as defined above, was taken as a measure of the bias, B, in these cases.

$$B = \text{median}[D_i] \quad (6)$$

A relative bias, RB, was also calculated based upon the average expected concentration T, as defined in (4).

$$RB = \frac{4 \cdot \text{median} [D_i]}{T_1 + T_2 + T_3 + T_4} \cdot 100\% \quad (7)$$

4. Results – 23rd intercomparison

4.1 Sulphur dioxide in absorbing solution (A-samples)

Four samples and one blank solution were distributed to the laboratories that use the hydrogen peroxide absorption solution method. The results are given in Table 8, Figures 2 and 18. For those laboratories that reported a blank value this has been subtracted from the reported results. The ratios of measured value to expected value are presented in Table 9.

The sulphate concentration in the sample solutions correspond to a SO₂ concentration in air of 2.33–7.79 µg S m⁻³, when 70 ml absorbing solution and 3.6 m³ sampling volume is used.

Values for SO₂ in absorbing solution were reported by 8 laboratories.

4.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples)

Five impregnated filter samples (including one blank) for determination of sulphur dioxide were analysed by 21 laboratories. The value reported for the blank filter was subtracted from the other values before the data were used.

The amount of sulphur on the distributed filters corresponds to air concentrations between 0.48–3.20 µg S m⁻³ when 25 m³ is sampled.

In addition to sulphate, nitrate was added to the same impregnated filters for determination of HNO₃-N. The value reported for the blank filter was subtracted from the other values before using the data.

The amount of nitrogen on the distributed filters corresponds to air concentrations between $0.46 \mu\text{g N m}^{-3}$ – $1.05 \mu\text{g N m}^{-3}$ when 25 m^3 sampling volume is used.

Sulphur dioxide results show prevalence of systematic versus random errors. The systematic error is most clearly for the low concentration samples. Even so, most values are within 20% of expected value.

A few reported values for nitric acid on impregnated filters are more than 20% away from expected value. Youden plot shows that for the low concentration filters a prevalence of systematic versus random. For the high concentration filters the reported results show good agreement with expected value.

The results are presented in Tables 10 and 12 and Figures 3, 4, 19 and 20.

4.3 Nitrogen dioxide in absorbing solution (C-samples)

Results from analysis of C-samples will not be reported this year due to error in the samples distributed.

4.4 Ammonia on impregnated filters (J-samples)

For the third time impregnated filters for determination of ammonia were distributed. Six impregnated filters inclusive two unidentified blank filters were sent to 27 laboratories. 18 laboratories have reported their analytical results. The two blank values reported by each laboratory were averaged and subtracted from the other values reported before the data were used. The results are shown in Table 14 and Figures 6 and 22.

The amount of nitrogen on the filters correspond to air concentrations between 0.36 – $1.60 \mu\text{g N m}^{-3}$, if 25 m^3 sampling volume is used.

Several reported values are more than 20% away from expected value. Youden plot shows that for the low concentration filters a considerable number of random errors are present. For the high concentration filters the reported result mostly lies within DQO.

4.5 Precipitation (G-samples)

Four precipitation samples were distributed and 2547 single results from 65 laboratories were reported. 108 results were identified as outliers. This is ~4% of the data, which is about the same as obtained earlier. It should be noted that 56% of the outliers are caused by seven laboratories. The results are presented in Tables 16–26 and Figures 7–16 and 23–33.

4.5.1 Conductivity and ion balance

In EMEP, conductivity measurements are mainly used for quality control reasons. When all the main ions in the precipitation have been measured, conductivity values are compared with values calculated from the reported results. Table 27 gives the ratios of the measured to the calculated values.

Low concentration ions do not contribute much to the sum of ionic conductivities. By looking at the ratio of measured to calculated conductivity, errors in determination of low concentration ions may not be revealed. To include low concentration ions in the quality control, ion balance control must be used. This ratio should be used as a tool in the quality control system for those laboratories that measure all main components. The ratios of equivalent concentrations of anions versus equivalent concentrations of cations are shown in Table 28.

The Youden plots of conductivity show mainly systematic error. This may be due to bad calibration of the instrument.

5. Results – 24th intercomparison

5.1 Sulphur dioxide in absorbing solution (A-samples)

Four samples and one blank solution were distributed to the laboratories that use the hydrogen peroxide absorption solution method. The results are given in Table 39, Figures 32 and 48. For those laboratories that reported a blank value this has been subtracted from the reported results. The ratios of measured value to expected value are presented in Table 40.

The sulphate concentration in the sample solutions correspond to a SO₂ concentration in air of 2.72–8.58 µg S m⁻³, when 70 ml absorbing solution and 3.6 m³ sampling volume is used.

Values for SO₂ in absorbing solution were reported by 7 laboratories.

5.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples)

Five impregnated filter samples (including one blank) for determination of sulphur dioxide were analysed by 19 laboratories. The value reported for the blank filter was subtracted from the other values before the data were used.

The amount of sulphur on the distributed filters corresponds to air concentrations between 0.40–2.88 µg S m⁻³ when 25 m³ is sampled.

In addition to sulphate, nitrate was added to the same impregnated filters for determination of HNO₃-N. The value reported for the blank filter was subtracted from the other values before using the data.

The amount of nitrogen on the distributed filters corresponds to air concentrations between 0.39 µg N m⁻³–1.31 µg N m⁻³ when 25 m³ sampling volume is used.

Sulphur dioxide results show prevalence of systematic versus random errors. The systematic error is most clearly for the low concentration samples. Even so, most values are within 20% of expected value.

A few reported results for low concentration of sulphur dioxide are more than 20% away from expected value. For the high concentration filters the reported results show good agreement with expected value.

A few reported values for nitric acid on impregnated filters are more than 20% away from expected value. Even so, most values are within 20% of expected value.

The results are presented in Tables 41 and 43 and Figures 33, 34, 49 and 50.

5.3 Nitrogen dioxide in absorbing solution (C-samples)

The four samples distributed are made to represent both absorption solutions and extracts from iodide-impregnated glass filters. The samples contained known amount of sodium nitrite diluted in water. In order to assure sample stability and to give the laboratories the opportunity to use the matrix they use in their daily routine, the distributed samples were to be diluted 1:10. The results should be reported as the diluted concentrations. The 10 times diluted samples correspond to air concentrations between $2.55\text{--}5.9 \mu\text{g NO}_2\text{-N m}^{-3}$, when 70 ml absorbing solution and 1.4 m^3 is sampled. When 4 ml extraction solution and 0.7 m^{-3} sampling volume is used, the samples correspond to air concentrations between $0.29\text{--}0.67 \mu\text{g NO}_2\text{-N m}^{-3}$.

The results are presented in Table 44 and Figures 35 and 51. The average ratio is presented in Table 42. 2 of 18 laboratories have an average ratio between 0.90 and 1.10, which is a quite satisfactory result.

5.4 Ammonia on impregnated filters (J-samples)

Six impregnated filters inclusive two unidentified blank filters were sent to 27 laboratories. 18 laboratories have reported their analytical results. The two blank values reported by each laboratory were averaged and subtracted from the other values reported before the data were used. The results are shown in Table 46 and Figures 36 and 52.

The amount of nitrogen on the filters correspond to air concentrations between $0.36\text{--}1.28 \mu\text{g N m}^{-3}$, if 25 m^3 sampling volume is used.

Several reported values are more than 20% away from expected value. Outliers are found in both low- and high concentration filters. 87 % of the reported results are within DQO.

5.5 Precipitation (G-samples)

Four precipitation samples were distributed and 2665 single results from 64 laboratories were reported. 121 results were identified as outliers. This is 4,5 % of the data, which is about the same as obtained earlier. The results are presented in Tables 48–58 and Figures 37–46 and 53–63.

5.5.1 Conductivity and ion balance

In EMEP, conductivity measurements are mainly used for quality control reasons. When all the main ions in the precipitation have been measured, conductivity values are compared with values calculated from the reported results. Table 59 gives the ratios of the measured to the calculated values.

Low concentration ions do not contribute much to the sum of ionic conductivities. By looking at the ratio of measured to calculated conductivity, errors in determination of low concentration ions may not be revealed. To include low concentration ions in the quality control, ion balance control must be used. This ratio should be used as a tool in the quality control system for those laboratories that measure all main components. The ratios of equivalent concentrations of anions versus equivalent concentrations of cations are shown in Table 60.

The Youden plots of conductivity show mainly systematic error. This may be due to bad calibration of the instrument.

6. Summary

A total of 68 laboratories participated in the twenty-third intercomparison. 32 of these laboratories are within the EMEP network.

For all the samples analysed, the deviations from theoretical values are calculated. Table 17 shows the median values compared to the expected values for all the parameters. For the EMEP laboratories the median deviations for both low- and high concentration samples are less than 3%. This is the same result as in the last intercomparison. For the other participants the median deviations for both low- and high concentration samples are less than 5%. This is slightly better than obtained in earlier.

As in earlier intercomparisons, outliers are defined as values that deviate more than two standard deviations from the mean value. Outliers occur for all samples and almost all parameters. Out of a total of 2892 single results, 122 are defined as outliers. This is 4% of the reported data, which is comparable to earlier intercomparisons.

The ratio of the median values to the theoretical values for all the parameters is presented in Table 29. All parameters show median values in good agreement with the theoretical values.

A total of 64 laboratories participated in the twenty-fourth intercomparison. 32 of these laboratories are within the EMEP network.

For all the samples analysed, the deviations from theoretical values are calculated. Figure 47 shows the median values compared to the expected values for all the parameters. For the EMEP laboratories the median deviations for both low- and high concentration samples are less than 6 %. This is an increase compared to earlier intercomparisons. For the other participants the median deviations for both low- and high concentration samples are less than 9%. This is also an increase compared to earlier intercomparisons.

As in earlier intercomparisons, outliers are defined as values that deviate more than two standard deviations from the mean value. Outliers occur for all samples and almost all parameters. Out of a total of 2983 single results, 137 are defined as outliers. This is 4,6% of the reported data, which is a slight increase compared to last year.

The ratio of the median values to the theoretical values for all the parameters is presented in Table 61. All parameters show median values in good agreement with the theoretical values.

7. References

- Eurachem/CITAC (2000) Quantifying uncertainty in analytical measurements.
2nd ed. URL: <http://www.measurementuncertainty.org/mu/QUAM2000-1.pdf>.
- EMEP (1996) Manual for sampling and chemical analysis. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 1/95) (Newest revision in 2001).
- Hanssen, J.E. (1988) The tenth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 3/88).
- Hanssen, J.E. (1990) The eleventh intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/90).
- Hanssen, J.E. and Ladegård, N.E. (1984) The seventh intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/84).
- Hanssen, J.E. and Ladegård, N.E. (1985) The eighth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 4/85).
- Hanssen, J.E. and Ladegård, N.E. (1987) The ninth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 4/87).
- Hanssen, J.E., Ladegård, N.E. and Thrane, K.E. (1983) The sixth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/83).
- Hanssen, J.E. and Skjelmoen, J.E. (1992) The twelfth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 7/92).
- Hanssen, J.E. and Skjelmoen, J.E. (1994) The thirteenth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/94).
- Hanssen, J.E. and Skjelmoen, J.E. (1995) The fourteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 3/95).

- Hanssen, J.E. and Skjelmoen, J.E. (1996) The fifteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 2/96).
- Hanssen, J.E. and Skjelmoen, J.E. (1997) The sixteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 2/97).
- Hanssen, J.E. and Skjelmoen, J.E. (2001) The seventeenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 10/2001).
- Hjellbrekke, A.-G., Uggerud, H.T., Hanssen, J.E. and Schaug, J. (2005) The twenty-second intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 8/2005).
- Thrane, K.E. (1978) Report on the first intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/78).
- Thrane, K.E. (1980a) Report on the second and third intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/80).
- Thrane, K.E. (1980b) Report on the fourth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 5/80).
- Thrane, K.E. (1981) The fifth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/81).
- Uggerud, H.T., Hanssen, J.E. and Skjelmoen, J.E. (2001) The eighteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 11/2001).
- Uggerud, H.T., Hanssen, J.E., Schaug, J. and Skjelmoen, J.E. (2002) The nineteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 1/2002).
- Uggerud, H.T., Hanssen, J.E., Schaug, J. and Skjelmoen, J.E. (2003) The twentieth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 8/2003).
- Uggerud, H.T., Hanssen, J.E., Schaug, J. and Hjellbrekke, A.-G. (2004) The twenty-first intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 6/2004).

Appendix 1

Tables – 23rd intercomparison

Table 2: Samples distributed for the twenty-third interlaboratory test.

A.	5 synthetic samples for determination of SO ₂ , consisting of 0.3% H ₂ O ₂ absorbing solution and containing different concentrations of sulphuric acid. One of the samples was an unidentified blank.
B.	5 KOH-impregnated Whatman 40 filters, comprising 1 blank and 4 filters to which different amounts of sulphuric acid and nitrate salt have been added.
C.	4 synthetic samples for determination of NO ₂ consisting of sodium nitrite diluted in water.
J.	6 Whatman 40 filters impregnated with 3% oxalic acid, comprising 2 blank and 4 filters to which different amounts of ammonium salt solution have been added.
G.	4 synthetic precipitation samples, containing SO ₄ ²⁻ , NO ₃ ⁻ , NH ₄ ⁺ , H ⁺ , Na ⁺ , Mg ²⁺ and Cl ⁻ , and Ca ²⁺ and K ⁺ .

Table 3a: EMEP laboratories participating in the twenty-third laboratory intercomparison. The numbers in front of the names are used in tables and figures.

Austria	(1)	Umweltbundesamt Zweigstelle Sud, Klagenfurt
Canada	(26)	The Ontario Ministry of the Environment, Laboratory Service Branch
Croatia	(35)	Meteorological and Hydrological Service of Croatia
Czech Republic	(3)	Czech Hydrometeorological Institute, Praha
Denmark	(4)	National Environmental Research Institute. Air Pollution Laboratory
Estonia	(38)	Estonian Environmental Research Centre, Tallinn
Finland	(5)	Finnish Meteorological Institute. Air Quality Department
France	(6)	SGS Multilab
Germany	(7)	IfE Leipzig GmbH, Umweltlabor
Germany	(8)	Umweltbundesamt, Messstelle Schauinsland
Hungary	(10)	Institute for Atmospheric Physics
Iceland	(11)	Idntæknistofnun Islands (Technological Inst. of Iceland)
Ireland	(12)	Met. Eirann, Dublin
Italy	(13)	C.N.R. Istituto Inquinamento Atmosferico
Italy	(30)	Join Research centre, Ispra
Latvia	(33)	Air Pollution Observation Laboratory
Lithuania	(32)	Atmospheric Pollution Research Laboratory, Institute of Physics, Vilnius
Netherlands	(14)	National Institute of Public Health and Environmental Protection (RIVM)
Norway	(15)	Norwegian Institute for Air Research (NILU)
Macedonia	(40)	Hydrometeorological Institute, Skopje
Poland	(16)	Institute of Meteorology and Water Management, Warsaw
Poland	(39)	Environmental Monitoring Laboratory, Institute of Environmental Protection
Portugal	(17)	Laboratorio Santo Andre
Romania	(18)	Research and Engineering Institute for Environment
Russian Federation	(22)	Institute of Global Climate and Ecology
Serbia and Montenegro	(24)	Rep. Hydrometeorological Institute of Serbia
Slovakia	(31)	Slovak Hydrometeorological Institute
Slovenia	(36)	Hydrometeorological Institute of Slovenia
Spain	(19)	Centro Nacional de Sanidad Ambiental
Sweden	(20)	Swedish Environmental Research Institute (IVL), Gothenburg
Switzerland	(21)	Swiss Federal Laboratories for Materials Testing (EMPA)
Turkey	(34)	Refik Saydam Institute, Ankara
United Kingdom	(23)	AEA Technology, National Environmental Technology Centre
United States of America	(27)	Illinois State Water Survey

Table 3b: Participating laboratories outside the EMEP network.

Germany	(104)	Hessige Landwirtschaftliche
Finland	(107)	The Finnish Forest Institute
Germany	(108)	Institut f. Bondenkunde und Standortlehre, Dredsten
Germany	(109)	Institut für Bondenkunde und Waldernährung
Germany	(110)	Thüringer Landesanstalt für Landwirtschaft (TTL), Jena
Germany	(112)	Niedersächsische Forstliche Versuchsanstalt (NVF)
Italy	(114)	C.N.R. Istituto Italiano di Idrobiologia
Germany	(115)	Bayerische Landesanstalt f. Wald- und Forstwirtschaft
Switzerland	(116)	Institute for Applied Plant Biology
Germany	(117)	Sächsische Landesanstalt für Forsten, Graupa
Germany	(118)	Forstliche Versuchs-und Forschungsanstalt
Germany	(120)	Landwirtschaftliche Untersuchungs- und Forschungsanstalt (LUFA)
Germany	(121)	Landesamt für Natur und Umwelt
Belgium	(124)	Laboratorium voor Bondenkunde, Gent
Germany	(125)	Bayerisches Landesamt für Umweltschutz, Augsburg
Italy	(126)	APPA Laboratorio Biologico Provinciale
Belarus	(133)	Institute for Problems of Natural Resources Use and Ecology
China	(135)	Hunan Research Institute of Environmental Protection Service
Denmark	(139)	Danish Centre for Forest, Landscape and Planning, Hørsholm
Luxembourg	(146)	Cellule de Recherche en Environment et Biotechnologies Public Research Center-Gabriel Lippmann
Netherlands	(148)	Alterra, Wageningen
Switzerland	(149)	WSL Zentrallabor, Birmensdorf
Spain	(150)	Fundaciòn Centro de Estudios Ambientale des Mediterrane (CEAM)
Norway	(152)	Norwegian Forest Research Institute, Ås
Slovenia	(153)	Slovenian Forestry Institute, Ljubljana
United Kingdom	(155)	Environmental Research Branch, Farnham
Greece	(156)	Laboratory of Forest Land and Biogeochemistry, Ilisia
Hungary	(157)	Ecological Laboratory of Forest Research Institute, Sárvár
Japan	(158)	Acid Deposition and Oxidant Research Center (ADOCRC), Niigata
Ireland	(160)	Coillte Research Laboratory, Newtownmountkennedy
Thailand	(164)	Pollution Control Department (PCD), Bangkok
Viet Nam	(165)	Institute of Meteorology and Hydrology, Ha Noi

Table 4: Analytical methods used at the participating laboratories for the determination of sulphur dioxide in absorbing solution (A).

Method	Laboratory
1. Ion chromatography	6, 15, 19, 21, 23

Table 5: Analytical methods used at the participating laboratories for the determination of sulphur dioxide on impregnated filters (B).

Method	Laboratory
1. Spectrophotometry	16
2. Ion chromatography	3, 4, 5, 8, 11, 12, 13, 15, 20, 22, 31, 32, 33, 34, 36, 38, 116, 131, 138, 158
3. Capillary Ion Analysis	39

Table 6: Analytical methods used at the participating laboratories for determination of nitric acid on impregnated filters (B).

Method	Laboratory
1. Reduction to nitrite	16
2. Ion chromatography	3, 4, 5, 8, 11, 13, 15, 20, 22, 31, 32, 33, 34, 36, 116, 131, 138, 158
3. Capillary Ion Analysis	39

Table 7: Analytical method for determination of ammonia on impregnated filters (J).

Method	Laboratory
1. Spectrophotometry	3, 4, 8, 10, 16, 19, 32, 33, 34, 39, 116
2. FIA	11
3. Ion chromatography	5, 13, 15, 20, 36, 131, 138, 158

Table 8: Analytical results for sulphur dioxide in absorbing solution.

SO₂-S in absorbing solution
 Sample no.: A1
 Theoretical value: 0.281
 Unit: µg S/ml

Run 1:

Number of laboratories: 8
 Arithmetic mean value: 91.478
 Median: 0.273
 Standard deviation 258.002
 Rel. st. deviation (%) 282.037

Run 2:

Number of laboratories: 7
 Arithmetic mean value: 0.261
 Median: 0.268
 Standard deviation 0.039
 Rel. st. deviation (%) 15.145

Results in decreasing order:

159	730.000 (*)	23	0.268
34	0.298	17	0.267
15	0.280	21	0.257
6	0.278	19	0.176

SO₂-S in absorbing solution
 Sample no.: A4
 Theoretical value: 0.401
 Unit: µg S/ml

Run 1:

Number of laboratories: 8
 Arithmetic mean value: 139.078
 Median: 0.378
 Standard deviation 392.312
 Rel. st. deviation (%) 282.080

Run 2:

Number of laboratories: 7
 Arithmetic mean value: 0.375
 Median: 0.376
 Standard deviation 0.023
 Rel. st. deviation (%) 6.252

Results in decreasing order:

159	1110.000 (*)	21	0.376
34	0.409	15	0.370
6	0.395	17	0.359
23	0.379	19	0.337

SO₂-S in absorbing solution
 Sample no.: A2
 Theoretical value: 0.120
 Unit: µg S/ml

Run 1:

Number of laboratories: 8
 Arithmetic mean value: 32.596
 Median: 0.113
 Standard deviation 91.885
 Rel. st. deviation (%) 281.896

Run 2:

Number of laboratories: 7
 Arithmetic mean value: 0.109
 Median: 0.107
 Standard deviation 0.019
 Rel. st. deviation (%) 17.005

Results in decreasing order:

159	260.000 (*)	23	0.107
34	0.137	21	0.101
6	0.122	17	0.095
15	0.120	19	0.082

SO₂-S in absorbing solution
 Sample no.: A5
 Theoretical value: 0.200
 Unit: µg S/ml

Run 1:

Number of laboratories: 8
 Arithmetic mean value: 61.422
 Median: 0.205
 Standard deviation 173.171
 Rel. st. deviation (%) 281.935

Run 2:

Number of laboratories: 7
 Arithmetic mean value: 0.197
 Median: 0.201
 Standard deviation 0.023
 Rel. st. deviation (%) 11.490

Results in decreasing order:

159	490.000 (*)	6	0.201
19	0.233	21	0.180
15	0.210	23	0.179
34	0.209	17	0.168

Table 9: The ratios of the theoretical values and the results found by the laboratories in the determination of sulphur dioxide in absorbing solutions.

Lab no	Measured / Expected value					
	Sample no					
	A1	A2	A4	A5		
6	0.92	0.85	0.94	0.91		
15	1.00	1.00	0.92	1.05		
17	0.95	0.79	0.90	0.84		
19	0.63	0.68	0.84	1.17		
21	0.91	0.84	0.94	0.90		
23	0.99	0.90	0.99	0.95		
34	0.95	0.89	0.95	0.90		
159	2598	2167	2768	2450		

Table 10: Analytical results for sulphur dioxide in impregnated filter.

SO₂-S on impregnated filter
 Sample no.: B1
 Theoretical value: 12.020
 Unit: ug S/filter

Run 1:

Number of laboratories: 21
 Arithmetic mean value: 13.563
 Median: 11.300
 Standard deviation 12.681
 Rel. st. deviation (%) 93.499

Run 2:

Number of laboratories: 20
 Arithmetic mean value: 10.867
 Median: 11.295
 Standard deviation 2.945
 Rel. st. deviation (%) 27.098

Results in decreasing order:

164	67.470	(*)	158	11.290
8	13.700		20	11.260
31	13.460		4	10.920
33	13.124		34	10.780
16	13.100		3	10.500
39	12.930		22	10.300
5	12.650		38	9.900
36	11.870		116	8.640
135	11.810		11	8.490
32	11.310		165	0.010
15	11.300			

SO₂-S on impregnated filter
 Sample no.: B2
 Theoretical value: 20.040
 Unit: ug S/filter

Run 1:

Number of laboratories: 21
 Arithmetic mean value: 18.262
 Median: 19.170
 Standard deviation 4.636
 Rel. st. deviation (%) 25.389

Run 2:

Number of laboratories: 20
 Arithmetic mean value: 19.174
 Median: 19.250
 Standard deviation 2.057
 Rel. st. deviation (%) 10.729

Results in decreasing order:

39	22.260		135	18.970
31	21.899		34	18.809
8	21.600		4	18.720
33	21.507		20	18.480
16	21.100		22	18.000
36	20.500		38	16.100
5	20.450		116	15.900
32	19.540		11	15.860
158	19.490		15	15.800
3	19.330		165	0.017 (*)
164	19.170			

SO₂-S on impregnated filter
 Sample no.: B4
 Theoretical value: 48.070
 Unit: ug S/filter

Run 1:

Number of laboratories: 21
 Arithmetic mean value: 44.651
 Median: 47.555
 Standard deviation 11.060
 Rel. st. deviation (%) 24.771

Run 2:

Number of laboratories: 20
 Arithmetic mean value: 46.882
 Median: 47.727
 Standard deviation 4.336
 Rel. st. deviation (%) 9.249

Results in decreasing order:

39	53.050		20	46.100
33	52.148		158	45.890
31	51.037		4	45.780
16	50.900		135	45.130
5	50.750		11	41.820
36	50.330		38	41.800
32	50.100		15	41.400
22	49.500		3	39.970
164	48.270		116	38.200
8	47.900		165	0.042 (*)
34	47.555			

SO₂-S on impregnated filter
 Sample no.: B5
 Theoretical value: 80.120
 Unit: ug S/filter

Run 1:

Number of laboratories: 21
 Arithmetic mean value: 76.417
 Median: 81.700
 Standard deviation 18.543
 Rel. st. deviation (%) 24.266

Run 2:

Number of laboratories: 20
 Arithmetic mean value: 80.234
 Median: 81.751
 Standard deviation 6.314
 Rel. st. deviation (%) 7.870

Results in decreasing order:

22	91.500		15	81.400
39	89.490		4	77.240
33	85.497		135	77.060
5	85.450		8	76.800
36	84.390		158	75.890
31	84.039		20	75.420
164	83.970		3	71.490
16	82.500		38	70.900
32	81.840		116	66.300
34	81.803		165	0.075 (*)
11	81.700			

Table 11: The ratios of the theoretical values and the results found by the laboratories in the determination of sulphur dioxide on impregnated filters. The reported results are corrected for blank value (B3).

Lab no	Measured / Expected value				Average	
	Sample No					
	B1	B2	B4	B5		
3	0.87	0.96	0.83	0,89	0,89	
4	0.91	0.93	0.95	0,96	0,94	
5	1.05	1.02	1.06	1,07	1,05	
8	1.14	1.08	1.00	0,96	1,04	
11	0.71	0.79	0.87	1,02	0,85	
15	0.94	0.79	0.86	1,02	0,90	
16	1.09	1.05	1.06	1,03	1,06	
20	0.94	0.92	0.96	0,94	0,94	
22	0.86	0.90	1.03	1,14	0,98	
31	1.12	1.09	1.06	1,05	1,08	
32	0.94	0.98	1.04	1,02	0,99	
33	1.09	1.07	1.08	1,07	1,08	
34	0.90	0.94	0.99	1,02	0,96	
36	0.99	1.02	1.05	1,05	1,03	
38	0.82	0.80	0.87	0,88	0,85	
39	1.08	1.11	1.10	1,12	1,10	
116	0.72	0.79	0.79	0,83	0,78	
135	0.98	0.95	0.94	0,96	0,96	
164	5.61	0.96	1.00	1,05	2,16	
158	0.94	0.97	0.95	0,95	0,95	

Table 12: Analytical results for nitric acid on impregnated filter.

HNO₃-N on impregnated filter
 Sample no.: B1
 Theoretical value: 11.450
 Unit: ug N/filter

Run 1:

Number of laboratories: 20
 Arithmetic mean value: 11.005
 Median: 11.175
 Standard deviation 1.735
 Rel. st. deviation (%) 15.764

Run 2:

Number of laboratories: 19
 Arithmetic mean value: 11.206
 Median: 11.190
 Standard deviation 1.524
 Rel. st. deviation (%) 13.595

Results in decreasing order:

11	13.880	20	11.160
39	13.010	4	10.980
36	12.710	15	10.900
5	12.670	32	10.640
8	12.300	3	10.600
31	12.162	135	10.480
16	11.840	22	9.200
158	11.710	34	8.343
165	11.200	33	7.947
164	11.190	116	7.180 (*)

HNO₃-N on impregnated filter
 Sample no.: B2
 Theoretical value: 14.720
 Unit: ug N/filter

Run 1:

Number of laboratories: 20
 Arithmetic mean value: 14.460
 Median: 14.555
 Standard deviation 2.059
 Rel. st. deviation (%) 14.242

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 14.432
 Median: 14.555
 Standard deviation 1.351
 Rel. st. deviation (%) 9.363

Results in decreasing order:

11	19.680 (*)	32	14.470
36	16.630	165	14.200
8	16.200	20	14.080
5	16.170	22	14.000
39	15.440	3	13.490
16	15.240	15	13.300
31	15.224	33	12.451
135	15.160	34	12.292
158	14.810	116	11.980
4	14.640	164	9.740 (*)

HNO₃-N on impregnated filter
 Sample no.: B4
 Theoretical value: 26.180
 Unit: ug N/filter

Run 1:

Number of laboratories: 20
 Arithmetic mean value: 25.950
 Median: 26.520
 Standard deviation 2.479
 Rel. st. deviation (%) 9.551

Run 2:

Number of laboratories: 19
 Arithmetic mean value: 26.338
 Median: 26.550
 Standard deviation 1.819
 Rel. st. deviation (%) 6.905

Results in decreasing order:

11	30.730	164	26.490
5	28.370	158	25.710
36	27.990	22	25.700
33	27.735	20	25.640
39	27.070	4	25.580
16	27.000	165	25.300
31	26.959	3	23.930
8	26.800	34	23.103
32	26.670	15	23.100
135	26.550	116	18.580 (*)

HNO₃-N on impregnated filter
 Sample no.: B5
 Theoretical value: 22.900
 Unit: ug N/filter

Run 1:

Number of laboratories: 20
 Arithmetic mean value: 22.229
 Median: 22.155
 Standard deviation 2.393
 Rel. st. deviation (%) 10.766

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 22.220
 Median: 22.155
 Standard deviation 1.593
 Rel. st. deviation (%) 7.168

Results in decreasing order:

11	28.040 (*)	32	22.140
5	25.070	158	22.110
39	24.230	15	22.100
36	23.850	20	21.400
8	23.400	4	21.060
16	23.240	165	20.900
164	23.190	22	20.800
135	23.000	34	19.743
31	22.810	33	18.751
3	22.170	116	16.580 (*)

Table 13: The ratios of the theoretical values and the results found by the laboratories in the determination of nitric acid on impregnated filters. The reported results are corrected for blank value (B3).

Lab No	Measured / Expected value				Average	
	Sample No					
	B1	B2	B4	B5		
3	0.93	0.92	0.91	0.97	0.93	
4	0.96	0.99	0.98	0.92	0.96	
5	1.11	1.10	1.08	1.09	1.10	
8	1.07	1.10	1.02	1.02	1.06	
11	1.21	1.34	1.17	1.22	1.24	
15	0.95	0.90	0.88	0.97	0.93	
16	1.03	1.04	1.03	1.01	1.03	
20	0.97	0.96	0.98	0.93	0.96	
22	0.80	0.95	0.98	0.91	0.91	
31	1.06	1.03	1.03	1.00	1.03	
32	0.93	0.98	1.02	0.97	0.97	
33	0.69	0.85	1.06	0.82	0.85	
34	0.73	0.84	0.88	0.86	0.83	
36	1.11	1.13	1.07	1.04	1.09	
39	1.14	1.05	1.03	1.06	1.07	
116	0.63	0.81	0.71	0.72	0.72	
135	0.92	1.03	1.01	1.00	0.99	
164	0.98	0.66	1.01	1.01	0.92	
158	1.02	1.01	0.98	0.97	0.99	

Table 14: Analytical results for ammonia on impregnated filter.

NH₃-N on impregnated filter
 Sample no.: J1
 Theoretical value: 20.050
 Unit: ug N/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 19.997
 Median: 19.895
 Standard deviation 2.485
 Rel. st. deviation (%) 12.425

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 20.377
 Median: 20.000
 Standard deviation 1.948
 Rel. st. deviation (%) 9.558

Results in decreasing order:

13	23.760	116	19.790
31	23.631	135	19.655
19	22.438	15	19.400
8	21.800	165	18.800
32	21.700	158	18.755
11	21.650	5	18.650
33	21.212	34	17.493
20	20.450	39	17.225
36	20.000	10	13.532 (*)

NH₃-N on impregnated filter
 Sample no.: J2
 Theoretical value: 12.030
 Unit: ug N/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 11.938
 Median: 11.655
 Standard deviation 1.739
 Rel. st. deviation (%) 14.568

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 11.677
 Median: 11.410
 Standard deviation 1.383
 Rel. st. deviation (%) 11.846

Results in decreasing order:

13	16.370 (*)	36	11.410
31	14.002	20	11.390
135	13.505	165	10.700
19	13.468	34	10.656
33	12.867	5	10.650
11	12.850	15	10.500
116	12.490	39	10.225
32	12.430	158	10.155
8	11.900	10	9.315

NH₃-N on impregnated filter
 Sample no.: J4
 Theoretical value: 9.020
 Unit: ug N/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 8.618
 Median: 8.402
 Standard deviation 2.132
 Rel. st. deviation (%) 24.741

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 8.227
 Median: 8.355
 Standard deviation 1.383
 Rel. st. deviation (%) 16.814

Results in decreasing order:

34	15.256 (*)	158	8.355
11	10.750	165	8.200
135	9.935	32	7.830
36	9.560	8	7.500
19	9.537	20	7.150
13	9.300	33	7.062
31	8.841	39	6.585
116	8.790	15	6.300
5	8.450	10	5.714

NH₃-N on impregnated filter
 Sample no.: J6
 Theoretical value: 40.100
 Unit: ug N/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 38.746
 Median: 39.213
 Standard deviation 3.903
 Rel. st. deviation (%) 10.074

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 39.363
 Median: 39.255
 Standard deviation 2.986
 Rel. st. deviation (%) 7.585

Results in decreasing order:

32	46.160	36	39.170
8	42.700	20	39.070
13	42.140	135	39.045
19	42.117	31	38.712
33	40.431	5	37.150
116	39.490	15	36.600
11	39.450	39	35.225
165	39.300	34	33.156
158	39.255	10	28.262 (*)

Table 15: The ratios of the theoretical values and the results found by the laboratories in the determination of ammonia on impregnated filters. The reported results are corrected for an average blank value (J2 and J5).

Lab No	Measured / Expected value				Average	
	Sample No					
	J1	J3	J4	J6		
5	0.93	0.89	0.94	0.93	0.92	
8	1.09	0.99	0.83	1.06	0.99	
10	0.67	0.77	0.63	0.70	0.70	
11	1.08	1.07	1.19	0.98	1.08	
13	1.19	1.36	1.03	1.05	1.16	
15	0.97	0.87	0.70	0.91	0.86	
19	1.12	1.12	1.06	1.05	1.09	
20	1.02	0.95	0.79	0.97	0.93	
31	1.18	1.16	0.98	0.97	1.07	
32	1.08	1.03	0.87	1.15	1.03	
33	1.06	1.07	0.78	1.01	0.98	
34	0.87	0.89	1.69	0.83	1.07	
36	1.00	0.95	1.06	0.98	1.00	
39	0.86	0.85	0.73	0.88	0.83	
116	0.99	1.04	0.97	0.98	1.00	
135	0.98	1.12	1.10	0.97	1.04	
158	0.94	0.84	0.93	0.98	0.92	
165	0.94	0.89	0.91	0.98	0.93	

Table 16: Analytical results for sulphate in precipitation samples.

Sulphate in precipitation				Sulphate in precipitation			
Sample no.: G1	Theoretical value:	1.558	Unit: µg/ml	Sample no.: G2	Theoretical value:	1.773	Unit: µg/ml
Run 1:	Number of laboratories:	65		Run 1:	Number of laboratories:	65	
Arithmetic mean value:	1.535			Arithmetic mean value:	1.743		
Median:	1.540			Median:	1.769		
Standard deviation	0.156			Standard deviation	0.179		
Rel. st. deviation (%)	10.133			Rel. st. deviation (%)	10.259		
Run 2:	Number of laboratories:	62		Run 2:	Number of laboratories:	64	
Arithmetic mean value:	1.549			Arithmetic mean value:	1.761		
Median:	1.540			Median:	1.769		
Standard deviation	0.070			Standard deviation	0.107		
Rel. st. deviation (%)	4.549			Rel. st. deviation (%)	6.068		
Results in decreasing order:				Results in decreasing order:			
139	2.050 (*)	110	1.540	107	2.050	114	1.760
124	1.770	114	1.540	139	2.030	38	1.760
107	1.720	125	1.540	110	2.000	6	1.760
148	1.710	36	1.539	148	1.910	20	1.754
133	1.633	156	1.537	124	1.900	152	1.750
13	1.630	10	1.537	133	1.885	12	1.750
35	1.627	20	1.536	18	1.866	36	1.748
39	1.626	26	1.532	116	1.850	26	1.742
7	1.619	32	1.524	13	1.850	117	1.740
15	1.610	17	1.522	32	1.846	115	1.730
3	1.602	38	1.520	7	1.836	158	1.730
104	1.600	158	1.520	35	1.835	17	1.728
150	1.600	115	1.510	146	1.830	125	1.720
116	1.600	120	1.510	3	1.818	22	1.719
21	1.594	11	1.505	39	1.816	165	1.711
5	1.593	135	1.505	21	1.815	160	1.700
33	1.590	121	1.505	5	1.811	112	1.700
146	1.581	165	1.502	34	1.811	153	1.695
118	1.580	152	1.500	15	1.810	121	1.695
14	1.580	160	1.500	104	1.810	135	1.676
34	1.574	164	1.490	150	1.810	120	1.670
23	1.570	112	1.490	23	1.800	30	1.670
8	1.570	155	1.490	16	1.798	155	1.651
18	1.567	31	1.483	14	1.790	164	1.650
27	1.561	153	1.473	8	1.790	126	1.632
24	1.560	30	1.460	24	1.790	31	1.631
1	1.560	19	1.425	118	1.780	149	1.616
4	1.557	126	1.422	11	1.772	19	1.568
16	1.554	149	1.420	10	1.770	109	1.534
22	1.551	109	1.334	1	1.770	156	1.487
117	1.550	108	1.030 (*)	33	1.770	108	1.480
12	1.540	157	0.690 (*)	4	1.770	157	0.600 (*)
6	1.540			27	1.769		
Sulphate in precipitation				Sulphate in precipitation			
Sample no.: G3	Theoretical value:	1.003	Unit: µg/ml	Sample no.: G4	Theoretical value:	1.166	Unit: µg/ml
Run 1:	Number of laboratories:	65		Run 1:	Number of laboratories:	65	
Arithmetic mean value:	0.997			Arithmetic mean value:	1.160		
Median:	1.000			Median:	1.170		
Standard deviation	0.100			Standard deviation	0.117		
Rel. st. deviation (%)	10.035			Rel. st. deviation (%)	10.065		
Run 2:	Number of laboratories:	61		Run 2:	Number of laboratories:	62	
Arithmetic mean value:	0.999			Arithmetic mean value:	1.163		
Median:	1.000			Median:	1.170		
Standard deviation	0.057			Standard deviation	0.063		
Rel. st. deviation (%)	5.749			Rel. st. deviation (%)	5.439		
Results in decreasing order:				Results in decreasing order:			
124	1.310 (*)	116	1.000	133	1.416 (*)	38	1.170
133	1.266 (*)	115	1.000	124	1.400 (*)	110	1.170
139	1.160	36	0.997	107	1.390	24	1.170
107	1.150	12	0.990	148	1.280	4	1.162
156	1.120	114	0.990	7	1.246	6	1.160
148	1.090	6	0.990	116	1.240	12	1.160
32	1.070	26	0.989	10	1.231	36	1.158
7	1.064	20	0.986	104	1.220	26	1.153
18	1.058	11	0.985	139	1.220	120	1.150
5	1.055	38	0.980	15	1.220	114	1.150
13	1.050	120	0.980	22	1.213	20	1.148
146	1.043	10	0.976	39	1.212	152	1.140
15	1.040	33	0.973	13	1.210	18	1.136
104	1.040	158	0.970	35	1.209	165	1.136
14	1.030	160	0.970	3	1.206	121	1.131
117	1.030	165	0.968	146	1.203	158	1.130
150	1.030	155	0.964	32	1.202	160	1.130
35	1.030	135	0.961	14	1.200	125	1.120
21	1.022	152	0.960	150	1.190	33	1.120
34	1.022	157	0.960	118	1.190	153	1.117
8	1.020	112	0.950	21	1.188	17	1.115
1	1.020	153	0.948	34	1.187	19	1.108
39	1.019	164	0.930	5	1.187	164	1.100
3	1.017	125	0.928	155	1.184	112	1.100
118	1.010	121	0.918	11	1.182	156	1.096
19	1.010	22	0.917	8	1.180	135	1.091
24	1.010	149	0.916	117	1.180	30	1.080
31	1.008	30	0.910	23	1.180	149	1.073
4	1.008	126	0.891	1	1.180	126	1.045
27	1.007	109	0.826	31	1.177	157	1.000
16	1.001	108	0.670 (*)	27	1.177	109	0.963
23	1.000	17	0.591 (*)	16	1.176	108	0.450 (*)
110	1.000			115	1.170		

Table 17: Analytical results for nitrate in precipitation samples.

Nitrate in precipitation				Nitrate in precipitation			
Sample no.: G1	Theoretical value:	0.452	Unit: µg/ml	Sample no.: G2	Theoretical value:	0.361	Unit: µg/ml
Run 1:				Run 1:			
Number of laboratories:	65			Number of laboratories:	65		
Arithmetic mean value:	0.484			Arithmetic mean value:	0.394		
Median:	0.450			Median:	0.360		
Standard deviation	0.215			Standard deviation	0.206		
Rel. st. deviation (%)	44.302			Rel. st. deviation (%)	52.242		
Run 2:				Run 2:			
Number of laboratories:	63			Number of laboratories:	63		
Arithmetic mean value:	0.450			Arithmetic mean value:	0.359		
Median:	0.450			Median:	0.360		
Standard deviation	0.056			Standard deviation	0.043		
Rel. st. deviation (%)	12.477			Rel. st. deviation (%)	11.878		
Results in decreasing order:				Results in decreasing order:			
121 1.988 (*) 110 0.450				121 1.598 (*) 133 0.360			
157 1.130 (*) 12 0.450				157 1.430 (*) 1 0.360			
139 0.620 23 0.450				14 0.460 110 0.360			
14 0.540 13 0.450				124 0.430 12 0.360			
124 0.540 3 0.449				109 0.430 20 0.359			
116 0.518 7 0.448				40 0.418 146 0.359			
19 0.505 22 0.446				11 0.404 16 0.357			
10 0.503 165 0.441				156 0.395 39 0.355			
11 0.502 158 0.440				10 0.391 3 0.355			
40 0.499 38 0.440				36 0.386 7 0.355			
36 0.490 34 0.440				34 0.382 152 0.350			
5 0.470 24 0.440				5 0.381 158 0.350			
155 0.468 133 0.440				139 0.380 13 0.350			
135 0.462 112 0.440				126 0.373 24 0.350			
35 0.460 114 0.440				115 0.372 114 0.350			
156 0.460 104 0.440				26 0.370 165 0.350			
15 0.460 152 0.440				108 0.370 150 0.350			
118 0.460 16 0.439				118 0.370 104 0.350			
126 0.460 33 0.439				15 0.370 107 0.345			
17 0.460 31 0.437				112 0.370 117 0.340			
6 0.460 125 0.434				17 0.370 153 0.340			
21 0.459 107 0.434				35 0.368 120 0.340			
39 0.459 150 0.430				155 0.366 160 0.340			
27 0.458 109 0.430				27 0.363 33 0.339			
26 0.457 117 0.430				21 0.363 22 0.338			
115 0.455 120 0.430				135 0.362 116 0.336			
8 0.453 153 0.420				8 0.362 125 0.330			
146 0.453 160 0.420				32 0.361 31 0.329			
20 0.453 149 0.407				4 0.361 149 0.328			
32 0.453 30 0.402				164 0.360 30 0.327			
1 0.452 108 0.370				23 0.360 19 0.325			
4 0.450 148 0.110				38 0.360 148 0.090			
164 0.450 6 0.360							
Nitrate in precipitation				Nitrate in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value:	0.615			Theoretical value:	0.643		
Unit: µg/ml				Unit: µg/ml			
Run 1:				Run 1:			
Number of laboratories:	65			Number of laboratories:	65		
Arithmetic mean value:	0.635			Arithmetic mean value:	0.672		
Median:	0.609			Median:	0.640		
Standard deviation	0.275			Standard deviation	0.287		
Rel. st. deviation (%)	43.284			Rel. st. deviation (%)	42.686		
Run 2:				Run 2:			
Number of laboratories:	64			Number of laboratories:	64		
Arithmetic mean value:	0.602			Arithmetic mean value:	0.638		
Median:	0.608			Median:	0.640		
Standard deviation	0.068			Standard deviation	0.074		
Rel. st. deviation (%)	11.354			Rel. st. deviation (%)	11.596		
Results in decreasing order:				Results in decreasing order:			
121 2.749 (*) 16 0.607				121 2.873 (*) 133 0.640			
157 0.740 7 0.605				157 0.860 24 0.640			
11 0.694 34 0.604				11 0.762 118 0.640			
10 0.682 135 0.603				10 0.708 104 0.640			
124 0.680 155 0.602				14 0.700 13 0.640			
14 0.670 104 0.600				36 0.695 7 0.639			
36 0.664 120 0.600				108 0.690 16 0.639			
5 0.655 133 0.600				35 0.675 26 0.637			
139 0.650 12 0.600				22 0.673 116 0.636			
35 0.639 152 0.600				124 0.670 34 0.631			
15 0.630 158 0.600				5 0.668 158 0.630			
19 0.625 114 0.600				39 0.667 139 0.630			
146 0.625 165 0.599				146 0.660 114 0.630			
39 0.625 17 0.597				155 0.660 38 0.630			
27 0.622 31 0.592				15 0.660 117 0.630			
21 0.621 150 0.590				27 0.656 165 0.629			
6 0.620 117 0.590				156 0.654 135 0.626			
38 0.620 125 0.588				21 0.652 107 0.625			
115 0.618 107 0.587				110 0.650 31 0.624			
4 0.616 116 0.584				164 0.650 150 0.620			
1 0.615 109 0.580				23 0.650 120 0.620			
126 0.615 108 0.580				115 0.648 17 0.620			
20 0.613 153 0.580				32 0.646 112 0.620			
8 0.612 33 0.574				1 0.644 125 0.616			
24 0.610 112 0.570				20 0.644 109 0.610			
32 0.610 149 0.565				19 0.643 40 0.601			
13 0.610 22 0.560				126 0.642 33 0.600			
118 0.610 160 0.560				8 0.642 153 0.600			
164 0.610 40 0.547				3 0.641 149 0.592			
110 0.610 30 0.541				12 0.640 160 0.580			
23 0.610 156 0.536				152 0.640 30 0.566			
26 0.609 148 0.140				4 0.640 148 0.150			
3 0.609 6 0.640							

Table 18: Analytical results for ammonium in precipitation samples.

Ammonium in precipitation				Ammonium in precipitation			
Sample no.: G1	Theoretical value:	0.257	Unit: µg/ml	Sample no.: G2	Theoretical value:	0.160	Unit: µg/ml
Run 1:	Number of laboratories:	65		Run 1:	Number of laboratories:	65	
Arithmetic mean value:	0.264			Arithmetic mean value:	0.169		
Median:	0.262			Median:	0.163		
Standard deviation	0.073			Standard deviation	0.057		
Rel. st. deviation (%)	27.810			Rel. st. deviation (%)	33.473		
Run 2:	Number of laboratories:	62		Run 2:	Number of laboratories:	62	
Arithmetic mean value:	0.259			Arithmetic mean value:	0.165		
Median:	0.262			Median:	0.162		
Standard deviation	0.040			Standard deviation	0.034		
Rel. st. deviation (%)	15.299			Rel. st. deviation (%)	20.746		
Results in decreasing order:				Results in decreasing order:			
31 0.594 (*) 135 0.261				18 0.457 (*) 135 0.161			
18 0.527 (*) 104 0.260				31 0.315 (*) 24 0.160			
112 0.350 114 0.260				112 0.260 109 0.160			
133 0.349 117 0.260				156 0.252 117 0.160			
11 0.324 109 0.260				133 0.248 114 0.160			
156 0.320 160 0.260				3 0.227 118 0.160			
3 0.312 146 0.259				150 0.210 124 0.160			
121 0.304 36 0.257				34 0.204 12 0.160			
152 0.300 10 0.255				11 0.196 16 0.160			
125 0.295 26 0.254				125 0.192 6 0.160			
110 0.290 1 0.251				152 0.190 158 0.160			
15 0.290 6 0.250				15 0.190 7 0.158			
164 0.290 115 0.250				121 0.187 36 0.158			
34 0.281 118 0.250				19 0.186 26 0.157			
13 0.280 24 0.250				13 0.180 10 0.157			
107 0.277 8 0.250				110 0.180 107 0.156			
116 0.275 35 0.247				164 0.180 115 0.153			
33 0.274 7 0.247				149 0.177 8 0.153			
32 0.274 5 0.246				33 0.172 35 0.152			
23 0.270 14 0.242				32 0.171 40 0.151			
19 0.270 153 0.240				116 0.171 165 0.151			
139 0.270 126 0.237				104 0.170 160 0.150			
108 0.270 16 0.236				23 0.170 126 0.146			
12 0.270 120 0.210				139 0.170 14 0.146			
124 0.270 157 0.200				108 0.170 153 0.140			
158 0.270 40 0.198				146 0.166 148 0.130			
165 0.267 148 0.190				21 0.166 120 0.110			
4 0.265 30 0.173				20 0.164 30 0.108			
39 0.265 22 0.167				1 0.164 157 0.090			
149 0.264 155 0.160				39 0.163 155 0.073			
21 0.264 150 0.137				5 0.163 22 0.060			
20 0.262 17 0.000 (*)				4 0.163 17 0.000 (*)			
27 0.262				27 0.163			
Ammonium in precipitation				Ammonium in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value:	0.385			Theoretical value:	0.401		
Unit: µg/ml				Unit: µg/ml			
Run 1:	Number of laboratories:	65		Run 1:	Number of laboratories:	65	
Arithmetic mean value:	0.402			Arithmetic mean value:	0.422		
Median:	0.383			Median:	0.405		
Standard deviation	0.118			Standard deviation	0.093		
Rel. st. deviation (%)	29.401			Rel. st. deviation (%)	22.088		
Run 2:	Number of laboratories:	64		Run 2:	Number of laboratories:	63	
Arithmetic mean value:	0.388			Arithmetic mean value:	0.408		
Median:	0.383			Median:	0.405		
Standard deviation	0.041			Standard deviation	0.045		
Rel. st. deviation (%)	10.587			Rel. st. deviation (%)	10.974		
Results in decreasing order:				Results in decreasing order:			
18 1.283 (*) 135 0.382				40 0.999 (*) 16 0.405			
3 0.484 1 0.381				18 0.716 (*) 4 0.405			
156 0.472 108 0.380				11 0.521 39 0.404			
112 0.470 110 0.380				133 0.501 5 0.403			
11 0.465 117 0.380				121 0.491 135 0.402			
121 0.462 153 0.380				112 0.490 1 0.401			
150 0.452 118 0.380				31 0.481 13 0.400			
152 0.440 16 0.380				150 0.468 6 0.400			
31 0.439 10 0.380				156 0.458 117 0.400			
125 0.421 149 0.380				19 0.444 109 0.400			
15 0.420 6 0.380				125 0.443 160 0.400			
133 0.419 24 0.380				152 0.440 108 0.400			
32 0.412 13 0.380				124 0.440 24 0.400			
158 0.410 35 0.379				15 0.440 8 0.398			
139 0.410 26 0.379				32 0.434 26 0.398			
164 0.410 8 0.377				139 0.430 35 0.397			
124 0.410 165 0.377				104 0.430 7 0.395			
19 0.409 34 0.376				17 0.427 118 0.390			
116 0.405 7 0.376				20 0.426 115 0.390			
20 0.402 160 0.370				116 0.424 165 0.388			
33 0.400 109 0.370				149 0.424 14 0.388			
104 0.400 14 0.368				3 0.423 153 0.380			
114 0.400 115 0.366				33 0.422 110 0.380			
5 0.393 120 0.360				23 0.420 126 0.378			
36 0.392 107 0.356				164 0.420 107 0.373			
23 0.390 126 0.353				12 0.420 120 0.370			
12 0.390 40 0.352				158 0.420 34 0.358			
21 0.390 155 0.317				36 0.415 157 0.330			
27 0.388 157 0.300				21 0.413 22 0.314			
39 0.387 22 0.293				27 0.410 148 0.310			
146 0.386 148 0.290				114 0.410 155 0.293			
4 0.384 30 0.258				146 0.406 30 0.265			
17 0.383				10 0.405			

Table 19: Analytical results for pH in precipitation samples.

pH in precipitation			
Sample no.: G1			
Theoretical value: 4.100			
Unit: pH-unit			
Run 1:		Run 1:	
Number of laboratories:	65	Number of laboratories:	65
Arithmetic mean value:	4.126	Arithmetic mean value:	4.054
Median:	4.130	Median:	4.070
Standard deviation	0.092	Standard deviation	0.081
Rel. st. deviation (%)	2.226	Rel. st. deviation (%)	2.004
Run 2:		Run 2:	
Number of laboratories:	58	Number of laboratories:	61
Arithmetic mean value:	4.138	Arithmetic mean value:	4.070
Median:	4.135	Median:	4.070
Standard deviation	0.047	Standard deviation	0.052
Rel. st. deviation (%)	1.142	Rel. st. deviation (%)	1.278
Results in decreasing order:			
156	4.420 (*)	116	4.130
153	4.310 (*)	104	4.130
124	4.290	160	4.130
19	4.244	12	4.120
150	4.200	126	4.120
110	4.200	16	4.120
26	4.190	36	4.120
14	4.190	33	4.120
30	4.190	155	4.120
8	4.190	21	4.120
112	4.190	39	4.110
35	4.180	165	4.110
22	4.180	135	4.110
148	4.180	11	4.110
15	4.180	146	4.110
115	4.170	3	4.102
20	4.170	164	4.100
120	4.160	24	4.100
158	4.160	13	4.100
152	4.160	6	4.090
117	4.150	149	4.090
27	4.150	139	4.090
108	4.150	7	4.080
118	4.150	121	4.060
114	4.150	133	4.050
34	4.142	1	4.050
109	4.140	23	4.010
107	4.140	40	3.940 (*)
4	4.140	18	3.940 (*)
31	4.140	17	3.900 (*)
5	4.140	157	3.860 (*)
32	4.130	38	3.820 (*)
10	4.130		
pH in precipitation			
Sample no.: G2			
Theoretical value: 4.040			
Unit: pH-unit			
Run 1:		Run 1:	
Number of laboratories:	65	Number of laboratories:	65
Arithmetic mean value:	4.054	Arithmetic mean value:	4.338
Median:	4.070	Median:	4.340
Standard deviation	0.081	Standard deviation	0.106
Rel. st. deviation (%)	2.004	Rel. st. deviation (%)	2.454
Run 2:		Run 2:	
Number of laboratories:	61	Number of laboratories:	60
Arithmetic mean value:	4.070	Arithmetic mean value:	4.353
Median:	4.070	Median:	4.350
Standard deviation	0.052	Standard deviation	0.057
Rel. st. deviation (%)	1.278	Rel. st. deviation (%)	1.317
Results in decreasing order:			
150	4.620 (*)	34	4.425
19	4.540	31	4.420
156	4.530	148	4.420
153	4.520	16	4.420
14	4.500	21	4.420
110	4.500	155	4.420
146	4.500	114	4.410
15	4.490	120	4.410
8	4.490	109	4.410
124	4.490	5	4.410
26	4.480	33	4.410
35	4.470	10	4.400
112	4.470	12	4.400
22	4.470	139	4.400
158	4.460	164	4.400
118	4.460	24	4.390
30	4.460	121	4.390
152	4.460	126	4.390
108	4.460	6	4.390
20	4.460	149	4.380
107	4.450	135	4.380
27	4.450	7	4.370
117	4.450	133	4.360
4	4.440	13	4.350
36	4.440	1	4.330
116	4.440	23	4.320
115	4.440	104	4.320
11	4.440	18	4.200 (*)
3	4.438	40	4.190 (*)
32	4.430	17	4.110 (*)
39	4.430	157	4.090 (*)
160	4.430	38	4.050 (*)
165	4.430		
pH in precipitation			
Sample no.: G4			
Theoretical value: 4.320			
Unit: pH-unit			
Run 1:		Run 1:	
Number of laboratories:	65	Number of laboratories:	65
Arithmetic mean value:	4.338	Arithmetic mean value:	4.338
Median:	4.340	Median:	4.340
Standard deviation	0.106	Standard deviation	0.106
Rel. st. deviation (%)	2.454	Rel. st. deviation (%)	2.454
Run 2:		Run 2:	
Number of laboratories:	60	Number of laboratories:	60
Arithmetic mean value:	4.353	Arithmetic mean value:	4.353
Median:	4.350	Median:	4.350
Standard deviation	0.057	Standard deviation	0.057
Rel. st. deviation (%)	1.317	Rel. st. deviation (%)	1.317
Results in decreasing order:			
156	4.700 (*)	120	4.340
160	4.520	109	4.340
30	4.470	117	4.340
150	4.470	21	4.340
14	4.451	31	4.340
153	4.440	32	4.340
146	4.420	36	4.340
124	4.410	5	4.340
26	4.410	34	4.338
8	4.410	104	4.330
110	4.400	12	4.330
112	4.400	6	4.330
22	4.400	165	4.320
35	4.390	39	4.310
158	4.390	121	4.310
118	4.380	24	4.310
146	4.380	3	4.304
152	4.380	139	4.300
108	4.380	133	4.300
20	4.380	133	4.300
115	4.370	135	4.300
116	4.370	126	4.300
158	4.370	149	4.300
117	4.370	164	4.300
4	4.360	7	4.290
108	4.360	13	4.270
148	4.360	23	4.250
111	4.360	1	4.240
114	4.360	18	4.180
16	4.360	40	4.080 (*)
107	4.350	17	4.050 (*)
33	4.350	157	4.030 (*)
155	4.350	38	3.960 (*)
10	4.340		

Table 20: Analytical results for strong acid calculated from pH.

Strong acid calculated from pH			
Sample no.: G1			
Theoretical value: 80.000			
Unit: µeg/l			
Run 1:	Number of laboratories: 65	Run 1:	Number of laboratories: 65
Number of laboratories: 65	Arithmetic mean value: 76.643	Arithmetic mean value: 90.021	Arithmetic mean value: 90.021
Median: 74.130	Standard deviation 18.189	Median: 85.110	Median: 85.110
Standard deviation 18.189	Rel. st. deviation (%) 23.732	Standard deviation 19.894	Standard deviation 19.894
Rel. st. deviation (%) 23.732		Rel. st. deviation (%) 22.100	Rel. st. deviation (%) 22.100
Run 2:	Number of laboratories: 59	Run 2:	Number of laboratories: 61
Number of laboratories: 59	Arithmetic mean value: 72.862	Arithmetic mean value: 85.726	Arithmetic mean value: 85.726
Median: 72.440	Standard deviation 8.454	Median: 85.110	Median: 85.110
Standard deviation 8.454	Rel. st. deviation (%) 11.603	Standard deviation 10.231	Standard deviation 10.231
Rel. st. deviation (%) 11.603		Rel. st. deviation (%) 11.935	Rel. st. deviation (%) 11.935
Results in decreasing order:			
38 151.360 (*) 32 74.130	157 173.780 (*) 146 85.110	157 173.780 (*) 146 85.110	17 158.490 (*) 32 85.110
157 138.040 (*) 4 72.440	38 154.880 (*) 165 85.110	17 158.490 (*) 32 85.110	38 154.880 (*) 165 85.110
17 125.890 (*) 5 72.440	18 134.900 (*) 116 85.110	18 134.900 (*) 116 85.110	18 134.900 (*) 116 85.110
18 114.820 (*) 31 72.440	40 117.490 36 85.110	40 117.490 36 85.110	40 117.490 36 85.110
40 114.820 (*) 109 72.440	1 109.650 5 85.110	1 109.650 5 85.110	1 109.650 5 85.110
23 97.720 107 72.440	23 104.710 4 83.180	23 104.710 4 83.180	23 104.710 4 83.180
1 89.130 34 72.110	133 104.710 11 83.180	133 104.710 11 83.180	133 104.710 11 83.180
133 89.130 118 70.790	164 100.000 108 83.180	164 100.000 108 83.180	164 100.000 108 83.180
121 87.100 108 70.790	139 95.500 117 83.180	139 95.500 117 83.180	139 95.500 117 83.180
7 83.180 27 70.790	126 95.500 148 83.180	126 95.500 148 83.180	126 95.500 148 83.180
149 81.280 114 70.790	7 95.500 118 81.280	7 95.500 118 81.280	7 95.500 118 81.280
139 81.280 117 70.790	13 95.500 27 81.280	13 95.500 27 81.280	13 95.500 27 81.280
6 81.280 152 69.180	149 95.500 20 81.280	149 95.500 20 81.280	149 95.500 20 81.280
164 79.430 158 69.180	104 93.330 115 81.280	104 93.330 115 81.280	104 93.330 115 81.280
24 79.430 120 69.180	121 93.330 152 81.280	121 93.330 152 81.280	121 93.330 152 81.280
13 79.430 115 67.610	135 93.330 107 81.280	135 93.330 107 81.280	135 93.330 107 81.280
3 79.070 20 67.610	24 93.330 22 79.430	24 93.330 22 79.430	24 93.330 22 79.430
39 77.620 15 66.070	6 91.200 35 79.430	6 91.200 35 79.430	6 91.200 35 79.430
11 77.620 35 66.070	39 91.200 158 79.430	39 91.200 158 79.430	39 91.200 158 79.430
165 77.620 148 66.070	160 91.200 110 79.430	160 91.200 110 79.430	160 91.200 110 79.430
146 77.620 22 66.070	3 90.780 14 77.620	3 90.780 14 77.620	3 90.780 14 77.620
135 77.620 8 64.570	12 89.130 112 75.860	12 89.130 112 75.860	12 89.130 112 75.860
12 75.860 14 64.570	33 89.130 26 75.860	33 89.130 26 75.860	33 89.130 26 75.860
21 75.860 30 64.570	34 88.310 15 75.860	34 88.310 15 75.860	34 88.310 15 75.860
16 75.860 112 64.570	109 87.100 124 74.130	109 87.100 124 74.130	109 87.100 124 74.130
36 75.860 26 64.570	10 87.100 8 74.130	10 87.100 8 74.130	10 87.100 8 74.130
126 75.860 150 63.100	114 87.100 19 70.790	114 87.100 19 70.790	114 87.100 19 70.790
155 75.860 110 63.100	16 87.100 150 69.180	16 87.100 150 69.180	16 87.100 150 69.180
33 75.860 19 57.020	155 87.100 153 64.570	155 87.100 153 64.570	155 87.100 153 64.570
104 74.130 124 51.290	31 87.100 156 63.100	31 87.100 156 63.100	31 87.100 156 63.100
116 74.130 153 48.980	21 87.100 30 63.100	21 87.100 30 63.100	21 87.100 30 63.100
160 74.130 156 38.020 (*)	120 85.110	120 85.110	120 85.110
10 74.130			
Strong acid calculated from pH			
Sample no.: G3			
Theoretical value: 40.000			
Unit: µeg/l			
Run 1:	Number of laboratories: 65	Run 1:	Number of laboratories: 65
Number of laboratories: 65	Arithmetic mean value: 39.958	Arithmetic mean value: 47.407	Arithmetic mean value: 47.407
Median: 37.150	Standard deviation 11.456	Median: 45.710	Standard deviation 13.836
Standard deviation 11.456	Rel. st. deviation (%) 28.669	Rel. st. deviation (%) 29.185	Rel. st. deviation (%) 29.185
Rel. st. deviation (%) 28.669		Rel. st. deviation (%) 29.185	Rel. st. deviation (%) 29.185
Run 2:	Number of laboratories: 60	Run 2:	Number of laboratories: 61
Number of laboratories: 60	Arithmetic mean value: 37.026	Arithmetic mean value: 44.364	Arithmetic mean value: 44.364
Median: 37.150	Standard deviation 4.506	Median: 44.670	Standard deviation 6.687
Standard deviation 4.506	Rel. st. deviation (%) 12.169	Rel. st. deviation (%) 15.074	Rel. st. deviation (%) 15.074
Rel. st. deviation (%) 12.169		Rel. st. deviation (%) 15.074	Rel. st. deviation (%) 15.074
Results in decreasing order:			
38 89.130 (*) 32 37.150	38 109.650 (*) 155 44.670	38 109.650 (*) 155 44.670	157 93.330 (*) 33 44.670
157 81.280 (*) 39 37.150	157 93.330 (*) 33 44.670	157 93.330 (*) 33 44.670	17 89.130 (*) 107 44.670
17 77.620 (*) 160 37.150	40 83.180 (*) 11 43.650	40 83.180 (*) 11 43.650	40 83.180 (*) 11 43.650
40 64.570 (*) 3 36.480	18 66.070 148 43.650	18 66.070 148 43.650	18 66.070 148 43.650
18 63.100 (*) 4 36.310	1 57.540 16 43.650	1 57.540 16 43.650	1 57.540 16 43.650
23 47.860 11 36.310	23 56.230 4 43.650	23 56.230 4 43.650	23 56.230 4 43.650
104 47.860 36 36.310	13 53.700 114 43.650	13 53.700 114 43.650	13 53.700 114 43.650
1 46.770 115 36.310	7 51.290 108 43.650	7 51.290 108 43.650	7 51.290 108 43.650
13 44.670 116 36.310	126 50.120 115 42.660	126 50.120 115 42.660	164 50.120 116 42.660
133 43.650 107 35.480	149 50.120 158 42.660	149 50.120 158 42.660	149 50.120 158 42.660
7 42.660 117 35.480	139 50.120 27 42.660	139 50.120 27 42.660	139 50.120 27 42.660
135 41.690 27 35.480	135 50.120 20 41.690	135 50.120 20 41.690	135 50.120 20 41.690
149 41.690 118 34.670	104 46.770 110 39.810	104 46.770 110 39.810	121 48.980 15 40.740
6 40.740 20 34.670	165 47.860 22 39.810	165 47.860 22 39.810	165 47.860 22 39.810
24 40.740 152 34.670	10 45.710 12 39.810	10 45.710 12 39.810	10 45.710 12 39.810
121 40.740 158 34.670	6 46.770 124 38.900	6 46.770 124 38.900	21 45.710 26 38.900
126 40.740 30 34.670	39 48.980 152 41.690	39 48.980 152 41.690	48 980 35 40.740
139 39.810 108 34.670	24 48.980 15 40.740	24 48.980 15 40.740	24 48.980 15 40.740
164 39.810 22 33.880	121 48.980 15 40.740	121 48.980 15 40.740	121 48.980 15 40.740
10 39.810 35 33.880	165 47.860 22 39.810	165 47.860 22 39.810	165 47.860 22 39.810
12 39.810 112 33.880	104 46.770 110 39.810	104 46.770 110 39.810	104 46.770 110 39.810
120 38.900 26 33.110	12 46.770 112 39.810	12 46.770 112 39.810	12 46.770 112 39.810
5 38.900 8 32.360	6 46.770 124 38.900	6 46.770 124 38.900	6 46.770 124 38.900
109 38.900 15 32.360	34 49.920 8 38.900	34 49.920 8 38.900	34 49.920 8 38.900
33 38.900 124 32.360	21 45.710 26 38.900	21 45.710 26 38.900	21 45.710 26 38.900
114 38.900 146 31.620	5 45.710 14 38.020	5 45.710 14 38.020	5 45.710 14 38.020
21 38.020 110 31.620	36 45.710 153 36.310	36 45.710 153 36.310	36 45.710 153 36.310
16 38.020 14 31.620	10 45.710 19 35.400	10 45.710 19 35.400	10 45.710 19 35.400
31 38.020 153 30.200	109 45.710 150 33.880	109 45.710 150 33.880	109 45.710 150 33.880
155 38.020 156 29.510	120 45.710 30 33.880	120 45.710 30 33.880	120 45.710 30 33.880
148 38.020 19 28.840	117 45.710 160 30.200	117 45.710 160 30.200	117 45.710 160 30.200
34 37.580 150 23.990	31 45.710 156 19.950	31 45.710 156 19.950	31 45.710 156 19.950
165 37.150	32 45.710	32 45.710	32 45.710

Table 21: Analytical results for chloride in precipitation samples.

Chloride in precipitation Sample no.: G1 Theoretical value: 0.162 Unit: µg/ml				Chloride in precipitation Sample no.: G2 Theoretical value: 0.185 Unit: µg/ml			
Run 1: Number of laboratories: 61 Arithmetic mean value: 0.162 Median: 0.162 Standard deviation 0.145 Rel. st. deviation (%) 70.067				Run 1: Number of laboratories: 63 Arithmetic mean value: 0.189 Median: 0.189 Standard deviation 0.186 Rel. st. deviation (%) 80.857			
Run 2: Number of laboratories: 57 Arithmetic mean value: 0.176 Median: 0.160 Standard deviation 0.085 Rel. st. deviation (%) 48.447				Run 2: Number of laboratories: 58 Arithmetic mean value: 0.179 Median: 0.185 Standard deviation 0.063 Rel. st. deviation (%) 35.312			
Results in decreasing order: 157 0.730 (*) 116 0.161 133 0.711 (*) 104 0.160 124 0.630 (*) 149 0.160 110 0.500 (*) 118 0.160 156 0.473 6 0.160				Results in decreasing order: 110 0.900 (*) 116 0.187 124 0.890 (*) 32 0.187 133 0.794 (*) 21 0.183 157 0.760 (*) 27 0.182 108 0.750 (*) 3 0.181 112 < 0.370 119 < 0.310 152 0.300 165 0.181			
108 < 0.470 21 0.160 112 < 0.370 135 0.337 13 0.160 118 0.327 27 0.156 1200 0.310 165 0.155 119 < 0.310 36 0.265 8 0.155 5 0.260 3 0.155 139 0.250 34 0.154 115 0.250 7 0.153 109 < 0.250 117 0.220 153 0.150 152 0.200 114 0.150 125 < 0.200 155 0.193 164 0.150 158 0.190 12 0.150 24 0.180 15 0.150 30 0.178 20 0.141 11 0.177 150 0.139				135 0.265 8 0.176 155 0.262 33 0.175 11 0.252 148 0.170 18 0.251 164 0.170 109 < 0.250 117 0.250 12 0.170 146 0.228 114 0.170 125 < 0.200 118 0.210 4 0.169 158 0.210 1 0.167 5 0.201 20 0.164 38 0.200 120 0.160 24 0.200 107 0.160 30 0.199 150 0.155 39 0.198 10 0.136 126 0.195 160 0.130 149 0.194 14 0.130 35 0.192 139 0.120 26 0.191 22 0.117 125 < 0.100 104 0.190 23 0.080 6 0.180 17 0.030 13 0.180 109 0.000 153 0.190 112 0.000 16 0.190 125 0.000 34 0.189			
Chloride in precipitation Sample no.: G3 Theoretical value: 0.510 Unit: µg/ml				Chloride in precipitation Sample no.: G4 Theoretical value: 0.533 Unit: µg/ml			
Run 1: Number of laboratories: 65 Arithmetic mean value: 0.519 Median: 0.496 Standard deviation 0.173 Rel. st. deviation (%) 33.402				Run 1: Number of laboratories: 65 Arithmetic mean value: 0.554 Median: 0.526 Standard deviation 0.177 Rel. st. deviation (%) 31.961			
Run 2: Number of laboratories: 60 Arithmetic mean value: 0.486 Median: 0.492 Standard deviation 0.068 Rel. st. deviation (%) 14.052				Run 2: Number of laboratories: 61 Arithmetic mean value: 0.527 Median: 0.526 Standard deviation 0.073 Rel. st. deviation (%) 13.940			
Results in decreasing order: 133 1.297 (*) 30 0.495 108 1.120 (*) 164 0.490 157 1.030 (*) 38 0.490 110 1.000 (*) 13 0.490 124 0.830 158 0.490 115 0.558 32 0.489 18 0.554 34 0.489 36 0.553 165 0.485 146 0.547 31 0.480 5 0.543 118 0.480 135 0.532 125 0.480 39 0.524 116 0.475 117 0.520 11 0.471 155 0.514 126 0.471 26 0.512 153 0.470 138 0.510 120 0.460 24 0.510 112 0.460 12 0.510 1 0.456 104 0.510 33 0.455 4 0.508 21 0.449 35 0.506 150 0.448 27 0.503 10 0.439 149 0.502 107 0.438 3 0.501 109 0.430 16 0.501 19 0.422 20 0.501 148 0.410 21 0.500 14 0.410 6 0.500 160 0.400 114 0.500 23 0.390 152 0.500 156 0.338 15 0.500 22 0.253 7 0.498 17 0.165 (*) 8 0.496				Results in decreasing order: 133 1.548 (*) 16 0.526 108 1.070 (*) 34 0.524 157 1.060 (*) 8 0.523 124 0.830 20 0.522 110 0.800 158 0.520 139 0.670 6 0.520 10 0.617 32 0.519 18 0.604 30 0.518 115 0.590 4 0.517 36 0.576 165 0.515 155 0.569 31 0.514 11 0.565 13 0.510 39 0.563 156 0.506 118 0.560 125 0.500 5 0.554 152 0.500 117 0.550 153 0.500 146 0.547 120 0.500 26 0.546 126 0.498 104 0.540 33 0.495 24 0.540 112 0.490 135 0.536 107 0.479 3 0.535 1 0.479 116 0.535 121 0.476 35 0.533 150 0.463 27 0.532 19 0.450 21 0.532 14 0.440 114 0.530 23 0.420 38 0.530 109 0.420 12 0.530 148 0.410 15 0.530 22 0.405 164 0.530 160 0.370 7 0.529 17 0.193 (*) 149 0.526			

Table 22: Analytical results for sodium in precipitation samples.

Sodium in precipitation			
Sample no.: G1			
Theoretical value:			0.234
Unit: $\mu\text{g/ml}$			
Run 1:			
Number of laboratories:	64	Number of laboratories:	64
Arithmetic mean value:	0.240	Arithmetic mean value:	0.271
Median:	0.240	Median:	0.274
Standard deviation	0.047	Standard deviation	0.044
Rel. st. deviation (%)	19.745	Rel. st. deviation (%)	16.070
Run 2:			
Number of laboratories:	59	Number of laboratories:	62
Arithmetic mean value:	0.238	Arithmetic mean value:	0.272
Median:	0.240	Median:	0.274
Standard deviation	0.029	Standard deviation	0.032
Rel. st. deviation (%)	12.076	Rel. st. deviation (%)	11.751
Results in decreasing order:			
23 0.410 (*) 15 0.240	23 0.440 (*) 21 0.274		
5 0.361 (*) 40 0.240	126 0.358 33 0.273		
121 0.350 (*) 21 0.240	155 0.341 3 0.271		
116 0.320 34 0.237	121 0.340 164 0.270		
156 0.290 33 0.236	156 0.320 13 0.270		
157 0.290 7 0.234	31 0.319 15 0.270		
155 0.285 8 0.233	116 0.311 7 0.267		
133 0.280 164 0.230	157 0.310 8 0.266		
31 0.275 114 0.230	133 0.300 32 0.264		
4 0.271 20 0.228	158 0.300 1 0.263		
158 0.270 1 0.227	16 0.295 135 0.262		
126 0.267 165 0.225	149 0.294 20 0.261		
125 0.262 135 0.225	6 0.290 114 0.260		
14 0.259 32 0.223	160 0.290 165 0.260		
117 0.256 124 0.220	120 0.290 35 0.251		
16 0.256 12 0.220	19 0.290 109 0.250		
149 0.252 104 0.220	34 0.290 110 0.250		
6 0.250 139 0.216	125 0.287 104 0.250		
153 0.250 10 0.215	4 0.286 12 0.250		
120 0.250 35 0.214	14 0.283 24 0.250		
160 0.250 109 0.213	117 0.282 139 0.248		
19 0.250 38 0.210	153 0.280 10 0.245		
112 0.250 110 0.210	11 0.280 38 0.244		
13 0.250 24 0.210	112 0.280 107 0.241		
27 0.244 107 0.210	5 0.279 124 0.240		
36 0.242 150 0.203	27 0.279 150 0.230		
33 0.242 148 0.200	36 0.278 22 0.212		
30 0.241 118 0.180	30 0.277 118 0.210		
146 0.240 22 0.166	40 0.277 148 0.210		
11 0.240 115 0.161	26 0.276 115 0.200		
39 0.240 108 0.130 (*)	39 0.275 108 0.190		
26 0.240 17 0.080 (*)	146 0.274 17 0.100 (*)		
Sodium in precipitation			
Sample no.: G3			
Theoretical value:			0.588
Unit: $\mu\text{g/ml}$			
Run 1:			
Number of laboratories:	64	Number of laboratories:	64
Arithmetic mean value:	0.575	Arithmetic mean value:	0.631
Median:	0.589	Median:	0.647
Standard deviation	0.076	Standard deviation	0.093
Rel. st. deviation (%)	13.244	Rel. st. deviation (%)	14.669
Run 2:			
Number of laboratories:	61	Number of laboratories:	62
Arithmetic mean value:	0.588	Arithmetic mean value:	0.645
Median:	0.590	Median:	0.648
Standard deviation	0.044	Standard deviation	0.050
Rel. st. deviation (%)	7.477	Rel. st. deviation (%)	7.825
Results in decreasing order:			
133 0.700 27 0.589	157 0.780 22 0.647		
126 0.682 19 0.588	133 0.750 33 0.646		
155 0.681 26 0.586	126 0.745 19 0.646		
23 0.670 117 0.585	158 0.720 16 0.645		
149 0.660 8 0.583	149 0.716 26 0.643		
158 0.650 114 0.580	4 0.715 8 0.642		
156 0.640 164 0.580	120 0.710 20 0.641		
120 0.640 20 0.579	125 0.694 7 0.640		
22 0.632 31 0.578	156 0.690 114 0.640		
4 0.624 135 0.575	14 0.679 135 0.633		
125 0.618 7 0.575	34 0.670 3 0.630		
16 0.616 3 0.573	36 0.667 13 0.630		
14 0.611 13 0.570	31 0.665 164 0.630		
153 0.610 1 0.567	35 0.665 1 0.627		
121 0.610 12 0.560	6 0.660 10 0.621		
157 0.610 24 0.560	121 0.660 12 0.620		
5 0.609 165 0.552	40 0.660 109 0.616		
34 0.605 104 0.550	112 0.660 104 0.610		
36 0.605 109 0.550	15 0.660 110 0.610		
112 0.600 10 0.547	23 0.660 24 0.610		
35 0.600 110 0.540	5 0.659 165 0.600		
160 0.600 38 0.535	146 0.657 148 0.600		
15 0.600 139 0.535	32 0.657 139 0.599		
30 0.597 148 0.530	39 0.654 150 0.597		
39 0.597 107 0.529	30 0.654 38 0.597		
146 0.594 150 0.511	27 0.653 107 0.586		
32 0.593 115 0.504	153 0.650 115 0.574		
21 0.591 124 0.500	11 0.650 124 0.550		
40 0.590 118 0.480	160 0.650 118 0.530		
11 0.590 17 0.390 (*)	155 0.649 17 0.450		
33 0.590 108 0.340 (*)	117 0.648 108 0.210		
6 0.590 116 0.189 (*)	21 0.648 116 0.189 (*)		

Table 23: Analytical results for magnesium in precipitation samples.

Magnesium in precipitation
Sample no.: G1
Theoretical value: 0.093
Unit: $\mu\text{g}/\text{ml}$

Run 1:
Number of laboratories: 63
Arithmetic mean value: 0.090
Median: 0.093
Standard deviation 0.016
Rel. st. deviation (%) 18.227

Run 2:
Number of laboratories: 61
Arithmetic mean value: 0.091
Median: 0.093
Standard deviation 0.010
Rel. st. deviation (%) 10.762

Results in decreasing order:
35 0.144 (*) 26 0.092
17 0.110 13 0.092
120 0.110 135 0.092
164 0.110 157 0.090
133 0.103 114 0.090
5 0.102 24 0.090
125 0.101 124 0.090
34 0.101 11 0.090
112 0.100 121 0.090
158 0.100 36 0.089
153 0.100 109 0.089
156 0.100 104 0.087
15 0.100 118 0.085
160 0.100 1 0.083
3 0.099 116 0.083
146 0.099 4 0.082
12 0.098 22 0.082
150 0.098 20 0.081
149 0.098 139 0.080
117 0.096 110 0.080
14 0.096 115 0.080
16 0.095 6 0.080
21 0.095 148 0.080
10 0.095 155 0.077
40 0.095 38 0.076
7 0.094 31 0.075
27 0.094 165 0.074
19 0.094 107 0.072
33 0.094 23 0.070
8 0.093 126 0.070
39 0.093 108 0.000 (*)
30 0.093

Magnesium in precipitation
Sample no.: G3
Theoretical value: 0.111
Unit: $\mu\text{g}/\text{ml}$

Run 1:
Number of laboratories: 63
Arithmetic mean value: 0.108
Median: 0.111
Standard deviation 0.020
Rel. st. deviation (%) 18.176

Run 2:
Number of laboratories: 61
Arithmetic mean value: 0.109
Median: 0.111
Standard deviation 0.011
Rel. st. deviation (%) 10.251

Results in decreasing order:
35 0.177 (*) 153 0.110
17 0.130 34 0.110
120 0.130 157 0.110
164 0.130 24 0.110
149 0.126 11 0.110
158 0.120 14 0.107
112 0.120 109 0.106
114 0.120 1 0.105
15 0.120 13 0.105
133 0.119 118 0.104
3 0.118 20 0.104
116 0.118 36 0.102
150 0.118 104 0.100
146 0.118 110 0.100
135 0.118 124 0.100
125 0.116 139 0.100
5 0.116 121 0.100
117 0.115 6 0.100
21 0.115 148 0.100
33 0.114 160 0.100
155 0.114 156 0.100
12 0.114 40 0.099
19 0.113 165 0.099
10 0.113 4 0.099
8 0.113 115 0.092
27 0.113 38 0.092
22 0.112 126 0.090
26 0.112 107 0.089
16 0.111 23 0.080
30 0.111 31 0.075
7 0.111 108 0.000 (*)
39 0.111

Magnesium in precipitation
Sample no.: G2
Theoretical value: 0.087
Unit: $\mu\text{g}/\text{ml}$

Run 1:
Number of laboratories: 63
Arithmetic mean value: 0.085
Median: 0.088
Standard deviation 0.018
Rel. st. deviation (%) 21.112

Run 2:
Number of laboratories: 60
Arithmetic mean value: 0.085
Median: 0.088
Standard deviation 0.010
Rel. st. deviation (%) 11.992

Results in decreasing order:
40 0.155 (*) 165 0.088
35 0.128 (*) 8 0.088
164 0.110 26 0.087
120 0.100 7 0.087
17 0.100 39 0.087
15 0.100 30 0.087
149 0.099 13 0.084
133 0.094 104 0.083
3 0.092 36 0.083
12 0.092 118 0.082
150 0.091 109 0.082
125 0.091 124 0.080
33 0.091 110 0.080
160 0.090 139 0.080
157 0.090 6 0.080
24 0.090 14 0.080
16 0.090 121 0.080
158 0.090 148 0.080
146 0.090 11 0.080
156 0.090 1 0.078
114 0.090 4 0.077
112 0.090 20 0.074
153 0.090 115 0.072
10 0.089 31 0.072
135 0.089 38 0.071
155 0.089 116 0.069
5 0.089 22 0.065
19 0.088 107 0.064
21 0.088 23 0.060
34 0.088 126 0.050
117 0.088 108 0.000 (*)
27 0.088

Magnesium in precipitation
Sample no.: G4
Theoretical value: 0.124
Unit: $\mu\text{g}/\text{ml}$

Run 1:
Number of laboratories: 63
Arithmetic mean value: 0.122
Median: 0.123
Standard deviation 0.020
Rel. st. deviation (%) 16.658

Run 2:
Number of laboratories: 61
Arithmetic mean value: 0.123
Median: 0.123
Standard deviation 0.011
Rel. st. deviation (%) 9.006

Results in decreasing order:
35 0.179 (*) 7 0.123
164 0.150 22 0.122
17 0.150 30 0.121
160 0.140 24 0.120
120 0.140 11 0.120
15 0.140 139 0.120
149 0.139 156 0.120
146 0.133 6 0.120
135 0.132 121 0.120
3 0.132 14 0.119
133 0.132 165 0.118
114 0.130 109 0.117
112 0.130 13 0.117
158 0.130 118 0.116
153 0.130 36 0.115
150 0.130 31 0.114
125 0.130 20 0.113
16 0.129 40 0.112
10 0.128 4 0.111
21 0.128 104 0.110
5 0.127 1 0.110
116 0.126 110 0.110
155 0.126 124 0.110
19 0.126 148 0.110
117 0.126 157 0.110
8 0.125 126 0.107
27 0.125 38 0.103
12 0.124 115 0.103
33 0.124 107 0.093
26 0.124 108 0.000 (*)
39 0.123

Table 24: Analytical results for calcium in precipitation samples.

Calcium in precipitation			
Sample no.: G1			
Theoretical value: 0.192			
Unit: µg/ml			
Run 1:		Run 1:	
Number of laboratories:	64	Number of laboratories:	64
Arithmetical mean value:	0.201	Arithmetical mean value:	0.247
Median:	0.195	Median:	0.232
Standard deviation	0.054	Standard deviation	0.148
Rel. st. deviation (%)	26.938	Rel. st. deviation (%)	59.881
Run 2:		Run 2:	
Number of laboratories:	62	Number of laboratories:	63
Arithmetical mean value:	0.193	Arithmetical mean value:	0.228
Median:	0.194	Median:	0.232
Standard deviation	0.026	Standard deviation	0.031
Rel. st. deviation (%)	13.554	Rel. st. deviation (%)	13.480
Results in decreasing order:			
40	0.542 (*)	27	0.194
126	0.350 (*)	7	0.194
156	0.260	30	0.193
109	0.244	4	0.190
114	0.240	165	0.190
157	0.240	36	0.190
153	0.240	24	0.190
35	0.239	110	0.190
164	0.220	12	0.190
121	0.220	118	0.190
112	0.220	14	0.189
3	0.217	19	0.188
5	0.215	116	0.187
158	0.210	104	0.180
16	0.209	11	0.180
34	0.208	23	0.180
125	0.203	1	0.180
155	0.203	107	0.174
133	0.202	124	0.170
8	0.202	160	0.170
146	0.202	32	0.167
26	0.200	22	0.165
13	0.200	115	0.163
15	0.200	135	0.161
21	0.200	20	0.161
6	0.200	108	0.160
120	0.200	139	0.160
149	0.199	148	0.160
33	0.199	17	0.160
117	0.198	38	0.146
39	0.197	150	0.137
10	0.196	31	0.125
Calcium in precipitation			
Sample no.: G3			
Theoretical value: 0.268			
Unit: µg/ml			
Run 1:		Run 1:	
Number of laboratories:	64	Number of laboratories:	64
Arithmetical mean value:	0.275	Arithmetical mean value:	0.297
Median:	0.270	Median:	0.291
Standard deviation	0.051	Standard deviation	0.052
Rel. st. deviation (%)	18.686	Rel. st. deviation (%)	17.579
Run 2:		Run 2:	
Number of laboratories:	60	Number of laboratories:	62
Arithmetical mean value:	0.269	Arithmetical mean value:	0.297
Median:	0.269	Median:	0.291
Standard deviation	0.033	Standard deviation	0.039
Rel. st. deviation (%)	12.350	Rel. st. deviation (%)	13.061
Results in decreasing order:			
32	0.489 (*)	13	0.270
40	0.407 (*)	7	0.267
133	0.398 (*)	33	0.267
139	0.350	117	0.265
35	0.335	135	0.265
157	0.330	146	0.265
164	0.320	27	0.264
109	0.313	30	0.262
114	0.310	34	0.261
158	0.310	118	0.260
153	0.310	19	0.260
156	0.310	4	0.260
3	0.308	120	0.260
149	0.303	107	0.259
112	0.300	14	0.258
116	0.299	36	0.258
16	0.296	11	0.250
22	0.295	110	0.250
5	0.283	20	0.244
26	0.283	104	0.240
125	0.280	108	0.240
21	0.280	17	0.240
6	0.280	165	0.236
10	0.279	126	0.230
8	0.278	23	0.220
39	0.276	148	0.220
155	0.275	150	0.219
1	0.271	115	0.218
24	0.270	124	0.200
15	0.270	160	0.200
121	0.270	38	0.194
12	0.270	31	0.132 (*)
Calcium in precipitation			
Sample no.: G4			
Theoretical value: 0.291			
Unit: µg/ml			
Run 1:		Run 1:	
Number of laboratories:	64	Number of laboratories:	64
Arithmetical mean value:	0.297	Arithmetical mean value:	0.297
Median:	0.291	Median:	0.291
Standard deviation	0.052	Standard deviation	0.052
Rel. st. deviation (%)	17.579	Rel. st. deviation (%)	17.579
Run 2:		Run 2:	
Number of laboratories:	62	Number of laboratories:	62
Arithmetical mean value:	0.297	Arithmetical mean value:	0.297
Median:	0.291	Median:	0.291
Standard deviation	0.039	Standard deviation	0.039
Rel. st. deviation (%)	13.061	Rel. st. deviation (%)	13.061
Results in decreasing order:			
32	0.519 (*)	7	0.291
114	0.400	13	0.290
22	0.384	12	0.290
133	0.373	4	0.290
157	0.360	118	0.290
35	0.351	1	0.289
156	0.350	27	0.288
126	0.348	135	0.288
40	0.345	36	0.287
164	0.340	19	0.286
3	0.335	34	0.285
158	0.330	155	0.284
153	0.330	30	0.282
16	0.330	160	0.280
149	0.326	120	0.280
6	0.320	24	0.280
112	0.320	110	0.280
109	0.314	23	0.280
116	0.311	17	0.280
125	0.309	11	0.270
10	0.308	20	0.267
5	0.304	139	0.260
117	0.294	107	0.255
6	0.304	165	0.250
21	0.302	104	0.250
33	0.301	150	0.244
15	0.300	124	0.240
8	0.300	115	0.239
121	0.300	31	0.236
24	0.294	148	0.230
117	0.294	165	0.210
10	0.292	38	0.205
146	0.292	108	0.120 (*)

Table 25: Analytical results for potassium in precipitation samples.

Potassium in precipitation
Sample no.: G1
Theoretical value: 0.326
Unit: $\mu\text{g/ml}$

Run 1:
Number of laboratories: 64
Arithmetic mean value: 0.318
Median: 0.320
Standard deviation 0.056
Rel. st. deviation (%) 17.518

Run 2:
Number of laboratories: 60
Arithmetic mean value: 0.315
Median: 0.320
Standard deviation 0.037
Rel. st. deviation (%) 11.736

Results in decreasing order:
110 0.550 (*) 133 0.320
156 0.490 (*) 121 0.320
31 0.427 26 0.319
120 0.380 8 0.319
112 0.380 7 0.317
126 0.377 135 0.316
109 0.374 3 0.312
14 0.368 36 0.311
21 0.354 12 0.310
16 0.350 1 0.310
5 0.342 158 0.310
160 0.340 22 0.307
124 0.330 107 0.306
15 0.330 38 0.303
24 0.330 4 0.300
13 0.330 11 0.300
19 0.330 104 0.300
125 0.330 155 0.291
27 0.329 40 0.291
39 0.328 34 0.290
146 0.327 150 0.290
32 0.325 114 0.280
117 0.324 165 0.265
33 0.323 118 0.260
10 0.323 23 0.260
35 0.323 139 0.260
20 0.322 149 0.255
157 0.320 116 0.246
148 0.320 17 0.230
153 0.320 30 0.210
164 0.320 115 0.199 (*)
6 0.320 108 0.190 (*)

Potassium in precipitation
Sample no.: G3
Theoretical value: 0.204
Unit: $\mu\text{g/ml}$

Run 1:
Number of laboratories: 64
Arithmetic mean value: 0.208
Median: 0.200
Standard deviation 0.067
Rel. st. deviation (%) 32.366

Run 2:
Number of laboratories: 62
Arithmetic mean value: 0.198
Median: 0.200
Standard deviation 0.040
Rel. st. deviation (%) 20.245

Results in decreasing order:
110 0.510 (*) 12 0.200
116 0.507 (*) 13 0.200
156 0.330 31 0.199
149 0.300 20 0.198
109 0.294 3 0.197
112 0.250 125 0.197
16 0.246 8 0.197
126 0.243 22 0.196
120 0.240 1 0.194
153 0.230 36 0.191
121 0.220 114 0.190
124 0.220 158 0.190
21 0.217 135 0.190
10 0.215 157 0.190
5 0.213 34 0.190
19 0.212 6 0.190
11 0.210 38 0.183
146 0.210 107 0.180
15 0.210 133 0.180
14 0.208 4 0.180
32 0.208 104 0.180
27 0.206 150 0.180
40 0.206 155 0.175
26 0.205 165 0.171
33 0.204 23 0.170
117 0.203 118 0.160
39 0.203 139 0.160
35 0.201 17 0.130
7 0.200 30 0.127
24 0.200 108 0.110
164 0.200 148 0.090
160 0.200 115 0.089

Potassium in precipitation
Sample no.: G2
Theoretical value: 0.306
Unit: $\mu\text{g/ml}$

Run 1:
Number of laboratories: 64
Arithmetic mean value: 0.310
Median: 0.299
Standard deviation 0.099
Rel. st. deviation (%) 31.910

Run 2:
Number of laboratories: 62
Arithmetic mean value: 0.296
Median: 0.298
Standard deviation 0.046
Rel. st. deviation (%) 15.658

Results in decreasing order:
110 0.960 (*) 3 0.299
126 0.534 (*) 20 0.298
156 0.460 135 0.297
149 0.428 7 0.296
109 0.388 1 0.292
31 0.388 36 0.291
120 0.360 160 0.290
112 0.350 11 0.290
16 0.348 158 0.290
21 0.330 12 0.290
14 0.321 107 0.287
5 0.318 38 0.285
124 0.310 40 0.282
153 0.310 104 0.280
15 0.310 157 0.280
24 0.310 34 0.278
121 0.310 35 0.271
19 0.310 150 0.270
13 0.310 114 0.270
27 0.310 10 0.270
39 0.307 4 0.270
33 0.307 22 0.267
146 0.306 155 0.255
116 0.305 118 0.250
117 0.302 139 0.250
125 0.301 23 0.250
8 0.300 165 0.249
26 0.300 148 0.230
164 0.300 108 0.220
32 0.300 17 0.210
6 0.300 30 0.196
133 0.300 115 0.186

Potassium in precipitation
Sample no.: G4
Theoretical value: 0.163
Unit: $\mu\text{g/ml}$

Run 1:
Number of laboratories: 63
Arithmetic mean value: 0.167
Median: 0.160
Standard deviation 0.078
Rel. st. deviation (%) 46.669

Run 2:
Number of laboratories: 62
Arithmetic mean value: 0.158
Median: 0.160
Standard deviation 0.037
Rel. st. deviation (%) 23.379

Results in decreasing order:
116 0.703 (*) 125 0.159
156 0.270 7 0.158
110 0.250 1 0.157
109 0.239 3 0.157
112 0.210 20 0.156
153 0.210 8 0.155
120 0.190 6 0.150
157 0.190 133 0.150
126 0.188 158 0.150
16 0.180 104 0.150
10 0.179 160 0.150
35 0.179 164 0.150
31 0.177 36 0.150
21 0.173 13 0.150
22 0.172 38 0.145
40 0.171 135 0.141
146 0.171 150 0.140
15 0.170 4 0.140
121 0.170 148 0.140
117 0.170 11 0.140
124 0.170 34 0.140
19 0.169 165 0.138
32 0.164 107 0.136
27 0.164 155 0.134
26 0.163 114 0.130
5 0.162 118 0.130
14 0.162 139 0.120
33 0.161 30 0.105
23 0.160 17 0.100
24 0.160 108 0.040
39 0.160 115 0.028
12 0.160

Table 26: Analytical results for conductivity in precipitation samples.

Conductivity in precipitation			
Sample no.: G1		Conductivity in precipitation	
Theoretical value:	41.800	Sample no.: G2	
Unit: $\mu\text{S}/\text{cm}$		Theoretical value:	46.300
Run 1:		Run 1:	
Number of laboratories:	63	Number of laboratories:	63
Arithmetic mean value:	47.331	Arithmetic mean value:	52.793
Median:	40.600	Median:	44.800
Standard deviation	44.749	Standard deviation	52.754
Rel. st. deviation (%)	94.546	Rel. st. deviation (%)	99.925
Run 2:		Run 2:	
Number of laboratories:	61	Number of laboratories:	61
Arithmetic mean value:	40.128	Arithmetic mean value:	44.475
Median:	40.540	Median:	44.800
Standard deviation	3.462	Standard deviation	3.332
Rel. st. deviation (%)	8.628	Rel. st. deviation (%)	7.492
Results in decreasing order:			
104 376.000 (*) 20	40.540	104 446.000 (*) 6	44.800
110 158.000 (*) 157	40.500	110 167.000 (*) 126	44.641
139 51.000 158	40.300	139 55.000 150	44.600
149 47.025 112	40.200	149 51.990 16	44.600
150 44.600 16	40.200	18 48.200 36	44.600
124 44.000 39	40.100	115 47.280 11	44.500
18 43.200 35	40.100	5 47.200 158	44.500
14 43.000 11	40.000	108 47.200 35	44.400
115 42.650 114	40.000	117 47.000 112	44.300
31 42.500 156	40.000	12 47.000 8	44.100
117 42.500 4	39.900	14 47.000 22	44.100
30 42.500 6	39.600	19 47.000 13	44.070
19 42.400 8	39.500	31 46.900 39	43.900
108 42.400 164	39.200	17 46.500 157	43.900
3 42.300 22	39.000	21 46.360 7	43.800
5 42.200 7	39.000	165 46.300 164	43.500
12 42.200 146	38.900	27 46.200 146	43.200
24 42.000 107	38.900	15 46.000 23	43.000
120 42.000 148	38.700	24 46.000 107	42.800
21 41.840 118	38.000	120 46.000 34	42.300
17 41.750 23	38.000	30 45.900 160	42.000
27 41.600 126	37.569	10 45.900 118	42.000
10 41.400 160	37.000	20 45.770 153	41.720
36 41.300 153	36.960	33 45.700 38	41.500
15 41.200 34	36.100	3 45.700 148	41.400
165 41.200 38	35.700	114 45.700 109	40.970
32 41.200 109	34.270	156 45.700 124	39.000
33 41.130 135	33.000	121 45.500 135	36.300
1 40.900 133	32.000	32 45.300 133	36.000
121 40.900 40	31.000	1 45.200 116	35.960
13 40.670 116	29.420	155 45.200 40	35.000
155 40.600		4 44.800	
Conductivity in precipitation			
Sample no.: G3		Conductivity in precipitation	
Theoretical value:	28.200	Sample no.: G4	
Unit: $\mu\text{S}/\text{cm}$		Theoretical value:	32.200
Run 1:		Run 1:	
Number of laboratories:	63	Number of laboratories:	63
Arithmetic mean value:	33.861	Arithmetic mean value:	38.164
Median:	27.900	Median:	31.900
Standard deviation	37.900	Standard deviation	39.559
Rel. st. deviation (%)	111.926	Rel. st. deviation (%)	103.655
Run 2:		Run 2:	
Number of laboratories:	61	Number of laboratories:	61
Arithmetic mean value:	27.578	Arithmetic mean value:	31.678
Median:	27.900	Median:	31.700
Standard deviation	1.988	Standard deviation	2.374
Rel. st. deviation (%)	7.208	Rel. st. deviation (%)	7.495
Results in decreasing order:			
104 306.000 (*) 4	27.900	104 325.000 (*) 6	31.700
110 145.000 (*) 11	27.900	110 147.000 (*) 1	31.700
139 36.000 33	27.900	139 41.000 36	31.620
149 31.256 13	27.850	160 38.000 165	31.500
10 29.700 153	27.530	40 36.000 4	31.500
18 29.700 6	27.500	149 35.634 158	31.400
148 29.700 158	27.400	14 34.000 13	31.300
5 29.300 16	27.400	31 33.300 35	31.300
31 29.100 112	27.300	5 33.300 109	31.300
14 29.000 150	27.200	19 33.000 112	31.300
15 28.900 8	27.200	15 32.900 11	31.200
17 28.750 7	27.200	18 32.800 107	31.100
115 28.680 36	27.110	108 32.700 126	31.050
3 28.600 157	27.100	146 32.600 39	31.000
12 28.500 156	27.000	12 32.600 8	31.000
34 28.500 126	26.961	27 32.600 156	31.000
108 28.500 39	26.900	21 32.510 23	31.000
117 28.400 107	26.800	3 32.500 150	31.000
21 28.300 35	26.400	30 32.400 17	31.000
20 28.220 22	26.200	115 32.350 153	30.720
19 28.200 146	26.200	121 32.300 16	30.700
30 28.200 124	26.000	10 32.300 34	30.100
165 28.100 160	26.000	155 32.200 164	29.800
155 28.100 109	25.900	20 32.120 22	29.800
121 28.100 164	25.400	33 32.100 118	29.000
114 28.100 118	25.000	7 32.100 38	28.800
32 28.050 38	24.900	32 32.100 116	28.030
27 28.000 116	24.460	24 32.000 148	28.000
24 28.000 40	23.000	120 32.000 135	26.100
120 28.000 135	22.800	117 32.000 133	26.000
23 28.000 133	22.000	114 32.000 124	26.000
1 27.900		157 31.900	

Table 27: Ratio of the measured to the calculated conductivity in synthetic precipitation samples (G1-G4).

Lab.No.	Measured / Calculated value				Remarks
	G1	G2	G3	G4	
1	0.91	0.86	0.92	0.90	
3	1.01	0.98	1.04	0.98	
4	1.02	1.04	1.04	1.03	
5	1.06	1.07	1.03	1.05	
6	0.94	0.97	0.97	1.00	
7	0.90	0.92	0.93	0.95	
8	1.08	1.10	1.07	1.07	
10	1.04	1.03	1.06	1.00	
11	0.97	1.02	1.02	0.98	
12	1.05	1.04	1.02	1.03	
13	0.97	0.92	0.93	0.91	
14	1.17	1.13	1.14	1.17	
15	1.10	1.12	1.11	1.08	
16	1.00	0.99	0.99	0.99	
17	0.73	0.68	0.76	0.69	
18	0.84	0.81	0.84	0.97	NO ₃ ⁻ .Mg ²⁺ .Na ⁺ . Ca ²⁺ and K ⁺ are missing
19	1.28	1.26	1.16	1.20	Cl ⁻ values < LOD
20	1.09	1.09	1.08	1.07	
21	1.03	1.03	1.02	1.03	
22	1.08	1.09	1.07	1.02	
23	0.79	0.85	0.92	0.88	
24	1.01	0.98	0.99	0.99	
26					Cond. is missing
27	1.08	1.09	1.05	1.07	
30	1.21	1.31	1.15	1.27	
31	1.05	1.05	1.08	1.05	
32	1.05	1.04	1.01	1.01	
33	1.01	1.01	1.02	1.05	
34	0.92	0.93	1.05	0.96	
35	1.07	1.05	0.99	1.03	
36	1.02	1.01	1.00	1.00	
38	0.55	0.62	0.58	0.56	NH ₄ ⁺ is missing
39	0.97	0.95	0.98	0.94	
40	0.66	0.70	0.75	0.88	SO ₄ ²⁻ is missing
104	9.46	9.55	9.92	10.18	
107	0.98	0.99	1.01	0.99	
108	1.20	1.11	1.17	1.24	
109	0.92	0.95	0.99	1.06	
110	4.27	3.65	5.45	4.92	
112	1.10	1.09	1.04	1.06	
114	1.04	1.03	1.01	1.04	
115	1.16	1.13	1.09	1.09	
116	0.73	0.82	0.92	0.91	
117	1.10	1.09	1.07	1.02	
118	0.99	0.99	0.97	0.97	
120	1.11	1.07	1.01	1.03	
121	0.79	0.87	0.72	0.73	
124	1.29	0.92	0.94	0.85	
125					Cond. is missing
126	0.94	0.94	0.97	0.95	
133	0.68	0.68	0.66	0.70	
135	0.80	0.78	0.80	0.80	
139	1.12	1.13	1.23	1.23	
146	0.94	0.98	1.02	1.07	
148	1.11	0.99	1.22	1.01	

Table 27, cont.

Lab.No.	Measured / Calculated value				Remarks
	G1	G2	G3	G4	
149	1.14	1.10	1.10	1.10	
150	1.27	1.16	1.21	1.14	
152					Cond. is missing
153	1.21	1.15	1.13	1.11	
155	1.02	1.03	1.04	1.05	
156	1.39	1.26	1.07	1.36	
157	0.65	0.58	0.62	0.64	
158	1.06	1.06	1.04	1.03	
160	0.94	0.93	0.99	1.51	
164	0.94	0.89	0.91	0.90	
165	1.02	1.07	1.06	1.00	

Table 28: Ratio of equivalent concentration of anions to the equivalent concentration of cation measured in synthetic precipitation samples.

Lab.No.	Measured value / calculated value					Remarks
	G1	G2	G3	G4	Average	
1	0.95	0.90	0.96	0.95	0.94	
3	0.98	0.98	0.96	0.98	0.98	
4	1.06	1.07	1.04	1.02	1.04	
5	1.05	1.07	1.04	1.03	1.05	
6	0.99	0.99	1.00	0.99	0.99	
7	1.01	1.00	1.01	1.02	1.01	
8	1.14	1.16	1.08	1.08	1.11	
10	1.07	1.05	1.02	1.10	1.06	
11	1.00	1.10	1.02	1.06	1.04	
12	1.02	1.01	0.99	1.00	1.01	
13	1.02	0.99	0.99	0.99	1.00	
14	1.17	1.17	1.11	1.11	1.14	
15	1.11	1.12	1.06	1.05	1.09	
16	1.02	1.02	0.98	1.01	1.01	
17	0.75	0.69	0.57	0.71	0.68	NO ₃ ⁻ and Cl ⁻ for G2 are missing
18	0.70	0.74	0.53	0.75	0.68	NO ₃ ⁻ . K ⁺ . Ca ²⁺ . Na ⁺ and Mg ²⁺ are missing
19	1.11	0.98	1.08	1.03	1.05	
20	1.11	1.10	1.05	1.05	1.07	
21	1.04	1.04	1.01	1.02	1.03	
22	1.24	1.18	0.94	1.09	1.11	
23	0.84	0.89	0.92	0.91	0.89	
24	1.02	1.00	1.01	1.02	1.01	
26	1.12	1.12	1.05	1.06	1.09	
27	1.07	1.07	1.04	1.05	1.06	
30	1.14	1.25	1.05	1.13	1.14	
31	0.86	0.89	1.06	0.98	0.95	
32	1.10	1.16	1.02	1.00	1.07	
33	1.03	1.00	0.95	0.96	0.98	
34	1.04	1.02	1.03	1.04	1.03	
35	1.11	1.12	1.01	1.03	1.07	
36	1.07	1.08	1.07	1.05	1.07	
38	0.72	0.76	0.89	0.84	0.80	NH ₄ ⁺ is missing
39	1.05	1.02	1.04	1.03	1.03	
40						SO ₄ ²⁻ is missing
104	1.08	1.02	0.98	1.06	1.03	
107	1.15	1.26	1.16	1.20	1.19	
108	0.96	1.15	1.26	1.22	1.15	
109	0.93	0.95	0.94	0.94	0.94	Cl ⁻ values < LOD
110	1.16	1.21	1.16	1.16	1.17	
112	0.96	0.98	0.91	0.94	0.95	Cl ⁻ values < LOD
114	1.05	1.01	0.97	0.98	1.00	
115	1.19	1.18	1.16	1.17	1.18	
116	1.09	1.08	1.06	1.10	1.08	
117	1.06	1.05	1.05	1.02	1.04	
118	1.14	1.14	1.11	1.13	1.13	
120	1.08	0.99	0.96	0.98	1.00	
121	1.57	1.49	2.09	2.04	1.80	Cl ⁻ values < LOD
124	1.60	1.45	1.43	1.33	1.45	
125						pH is missing
126	0.91	0.87	0.94	0.89	0.90	
133	1.00	1.02	1.15	1.17	1.08	

Table 28, cont.

Lab.No.	Measured value / calculated value					Remarks
	G1	G2	G3	G4	Average	
135	1.05	0.98	0.97	0.95	0.99	
139	1.37	1.11	1.09	1.06	1.16	
146	1.02	1.08	1.11	1.07	1.07	
148	1.03	1.05	0.87	0.88	0.96	Cl ⁻ values < LOD
149	0.90	0.85	0.86	0.89	0.87	
150	1.29	1.21	1.15	1.10	1.19	
152						Na ⁺ . K ⁺ . Ca ²⁺ and Mg ²⁺ are missing
153	1.20	1.17	1.00	1.02	1.10	
155	1.07	1.03	1.00	1.13	1.06	
156	1.33	1.00	0.96	1.11	1.10	
157	0.75	0.72	0.89	0.84	0.80	
158	1.04	1.06	0.96	0.97	1.00	
160	0.98	0.96	0.99	1.05	1.00	
164	0.94	0.86	0.92	0.93	0.91	
165	1.01	1.05	1.05	1.04	1.04	

Table 29: The ratio of the median values to the theoretical values for all parameters and samples.

Parameter	Sample No	Median / Expected
SO ₄ -S	G1	0.99
	G2	1.00
	G3	1.00
	G4	1.00
NO ₃ -N	G1	1.00
	G2	1.00
	G3	0.99
	G4	1.00
NH ₄ -N	G1	1.02
	G2	1.02
	G3	0.99
	G4	1.01
pH (calc. from H ⁺)	G1	0.93
	G2	0.94
	G3	0.93
	G4	0.95
H	G1	0.93
	G2	0.95
	G3	0.96
	G4	0.96
Mg	G1	1.00
	G2	1.01
	G3	1.00
	G4	1.00
Na	G1	1.03
	G2	1.01
	G3	1.00
	G4	1.00
Cl	G1	1.02
	G2	1.03
	G3	0.97
	G4	0.99
Ca	G1	1.02
	G2	1.01
	G3	1.01
	G4	1.00
K	G1	0.98
	G2	0.98
	G3	0.98
	G4	0.98
Cond	G1	0.97
	G2	0.97
	G3	0.99
	G4	0.98

Table 30: Relative random and systematic errors obtained by the different laboratories in the analysis of each parameter in the precipitation samples.

Lab. no.	Na ⁺		Cl ⁻		Ca ²⁺		K ⁺		Cond.	
	Random error %	Systematic error %								
1	1	1	0	0	1	-1	3	-9	7	14
3	1	3	0	-1	10	20	1	6	3	-2
4	0	0	0	0	1	1	1	-11	3	-9
5	1	3	2	4	3	1	3	4	4	-7
6	0	-1	1	0	1	-1	4	-9	2	0
7	1	5	0	-1	1	-2	1	0	1	5
8	0	1	0	0	1	-2	1	1	6	-19
10	3	-1	3	11	1	-1	1	2	4	-6
11	2	-1	6	12	11	24	2	-3	4	-6
12	1	-1	1	0	3	3	2	4	2	-3
13	1	4	1	-1	4	3	2	-4	4	6
14	1	2	3	14	1	-5	4	-4	4	-19
15	1	3	1	2	0	11	4	9	6	-17
16	1	0	1	-1	3	-1	2	2	2	-7
17	11	-3	3	-1	5	-1	5	17	18	67
18	4	2			85	102			16	45
19	6	-7	7	1	4	8	0	1	7	-26
20	0	-1	0	0	3	4	2	-11	4	-13
21	1	2	1	1	1	2	1	3	2	-5
22	4	-2	7	-3	2	-30	9	-6	5	-16
23	1	1	1	0	2	4	18	-24	6	16
24	0	0	1	-2	1	-1	3	-2	1	1
26	1	-1	1	0	0	-1	1	0	6	-19
27	0	0	1	1	1	1	0	1	4	-11
30	1	-7	3	-12	11	-35	1	0	15	-23
31	5	-3	1	-4	38	39	10	-16	4	-6
32	3	4	1	0	3	7			3	-7
33	2	-1	2	-6	1	5	2	2	2	-5
34	1	1	3	-2	12	2	5	0	4	-5
35	1	4	2	3	1	-2	10	51	5	-15
36	1	-1	2	8	2	1	2	-6	3	-6
38	1	-1	1	-1			2	-17	14	96
39	2	3	2	2	1	1	0	0	2	-2
40			10	0	89	-7	31	-4	7	46
104	1	3	1	-2	4	4	4	-8	9	0
107	4	14	1	-3	7	-5	4	-22	5	-9
108	13	-31	10	-3	2	1			3	-11
109	2	-16	8	-5	2	0	1	-5	4	-6
110	7	0	1	0	7	2	3	-12	5	-16
112	1	-5	4	-3	2	30	2	6	6	-18
114	0	-1	0	-2	2	2	5	4	5	-7
115	2	-2	1	1	2	-3	3	-16	5	-12
116	2	4	8	-3	2	6	10	-4	2	-9
117	2	0	1	-4	1	0	1	2	4	-10
118	1	1	1	0	1	-2	1	-7	3	-12
120	3	-3	1	-4	3	-13	2	16	6	-7
121	1	-5	78	354	9	21	3	-5	4	2
124	5	16	5	13	5	6	4	-9	13	-21
125	2	-4	1	-5	1	12	2	5		
126	1	-9	1	1	2	-7	8	-21	5	2
133	6	13	1	-1	9	30	1	8	7	10
135	2	-5	2	-1	1	0	4	4	3	2
139	13	15	1	1	3	6	4	-9	2	3
146	1	3	1	1	1	1	2	6	4	-10
148	2	9	18	-79	9	-26	3	-12	8	-10
149	2	-8	1	-9	4	4	4	13	1	3
150	1	2	1	-4	25	19	1	5	5	-25
152	1	-3	1	-2	3	14			3	-13
153	1	-5	2	-6	2	-6	3	4	13	-30
155	4	-4	2	2	5	-31	7	2	2	-6
156	12	-3	9	2	5	25	7	0	20	-44
157	34	-38	74	86	4	-23	7	-2	25	80
158	0	-3	0	-2	3	5	2	6	5	-12
160	1	-3	3	-8	2	-2	11	5	11	-7
164	2	-5	1	0	2	7	4	20	5	1
165	1	-3	0	-2	3	-3	8	-9	4	-4

Table 30, cont.

Lab. no.	SO ₄ ²⁻		NO ₃ ⁻		NH ₄ ⁺		Mg ²⁺		H ⁺ calc	
	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %
1	1	-3	4	-11	3	-3	2	-5	1	-2
3	2	-2	1	-2	3	14	1	-3	1	1
4	5	8	3	-3	3	-1	2	-10	2	-3
5	11	4	10	8	2	6	3	4	1	3
6	2	3	2	-2	3	7	1	-4	2	-3
7	1	-1	1	-2	1	0	1	-3	3	-5
8	0	-1	1	-3	0	4	0	-3	1	-5
10	2	-6	18	-11	2	5	9	2	2	0
11	1	1	12	7	1	-8	5	-8	2	-4
12	1	-5	2	-2	2	0	2	-4	0	1
13	3	-2	3	-3	2	2	3	0	2	-3
14	2	6	6	-21	2	-1	7	4	1	3
15	1	2	1	-2	1	2	0	2	1	1
16	3	5	2	-1	4	12	4	13	1	-4
17	4	-42	25	-71	3	-12	5	-34	2	0
18			14	20					1	4
19	2	2	28	6	1	-2	1	2	1	2
20	0	-2	1	-5	3	-11	1	-3	1	-1
21	0	1	1	-1	1	4	3	7	0	0
22	11	-7	31	-28	24	0	8	-5	1	-6
23	15	29	2	-32	6	-9	10	-18	4	0
24	1	-6	2	3	2	0	1	0	0	0
26	1	0	1	2	1	6	1	-1		
27	1	2	1	-1	1	0	0	1	1	0
30	0	2	4	0	2	-2	9	-37	1	0
31	5	7	4	-4	13	-30	17	19	1	2
32	2	0	3	-2	45	40	2	0	1	-1
33	0	0	7	-7	2	2	1	0	1	-1
34	2	4	3	-2	4	2	4	-10	7	-8
35	4	-1	1	0	9	22	8	-1	1	-5
36	1	3	7	21	1	-2	0	-6	1	0
38	3	-9	4	1	7	-28	1	-8	3	-11
39	0	2	2	4	1	2	1	0	1	-4
40	1	1			183	100	7	-4	17	-22
104	2	-7	1	1	5	-12	2	-10	134	844
107	3	-10	6	-13	5	-11	1	-9	3	-6
108	33	-40	35	159	32	-12	8	-43	1	1
109	2	-6			5	18	7	32	7	-10
110	3	-7	53	119	3	-4	93	106	6	314
112	1	3	30	20	1	12	2	19	1	-3
114	1	-2	1	-3	11	20	5	-14	2	-1
115	1	-17	6	21	4	-18	3	-49	1	2
116	51	-41	4	0	10	3	101	60	10	-20
117	2	1	6	11	1	1	2	-1	1	1
118	6	-19	7	3	1	-1	5	-20	1	-9
120	5	8	23	-8	3	-2	4	18	1	-1
121	10	10	3	-18	5	2	4	2	1	-1
124	8	-14	48	113	8	-19	2	2	10	-11
125	3	7	8	-2	1	5	2	-2		
126	6	21	6	-8	33	23	33	18	3	-4
133	8	17	55	201	20	20	3	-4	5	-22
135	0	-3	20	15	6	-7	2	-5	5	-20
139	3	-8	24	13	19	-12	4	-20	2	24
146	1	1	4	7	3	0	1	1	4	-7
148	3	-12	15	-32	5	-16	18	-20	7	-10
149	5	11	2	-1	5	13	47	5	3	12
150	4	-10	6	-13	1	-21	2	-12	5	-3
152			17	4						
153	2	3	5	-6	1	17	9	6	5	-8
155	8	14	9	10	3	3	4	-13	1	-2
156	1	12	57	8	4	24	9	56	1	-3
157	11	11	6	158	3	25	9	-4	2	-3
158	4	11	6	2	4	14	0	-6	1	-3
160	1	3	13	-24	9	-9	5	-3	12	-9
164	1	-1	2	-4	4	18	1	-2	0	-7
165	4	-7	2	-4	13	-16	6	-18	1	-1

Table 31: Analytical methods used for the determination of chemical constituents in precipitation samples.

Constituents	Methods	Laboratory
SO ₄	1. Ion chromatography 2. Spectrophotometry 3. ICP-AES 4. Capillary electrophoresis 5. Turbidimetry	1, 3, 4, 5, 6, 7, 8, 10, 11, 12, 13, 14, 15, 16, 19, 20, 21, 22, 23, 24, 26, 27, 30, 31, 32, 33, 34, 35, 36, 38, 104, 107, 110, 114, 115, 116, 118, 124, 126, 135, 139, 146, 152, 153, 156, 157, 158, 160, 164, 165 18, 121, 109, 112, 113, 117, 120, 148 39 133
NO ₃	1. Ion chromatography 2. Spectrophotometry 3. Capillary electrophoresis	1, 3, 4, 5, 6, 7, 8, 10, 11, 12, 13, 14, 15, 16, 19, 20, 21, 22, 23, 24, 26, 27, 30, 31, 32, 33, 34, 35, 36, 38, 107, 110, 114, 115, 116, 118, 124, 126, 135, 136, 139, 146, 152, 153, 156, 157, 158, 164, 165 40, 104, 109, 112, 117, 120, 121, 133, 148, 160 39
NH ₄	1. Ion chromatography 2. Spectrophotometry	1, 5, 6, 7, 8, 12, 13, 15, 21, 22, 23, 30, 31, 35, 36, 107, 115, 124, 135, 136, 139, 153, 156, 158, 164, 165 3, 4, 10, 11, 14, 16, 18, 19, 20, 24, 26, 27, 32, 33, 34, 39, 40, 104, 109, 110, 112, 114, 116, 117, 118, 120, 121, 124, 126, 140, 133, 139, 146, 148, 152, , 157, 160,
H ⁺	Titration	6, 31, 124, 153, 165
Mg	1. Atomic absorption (AAS) 2. Ion chromatography 3. ICP-AES 4. ICP-MS	3, 4, 10, 16, 19, 22, 26, 33, 34, 38, 39, 40, 116, 133, 154, 156, 157, 160 1, 5, 6, 7, 8, 12, 13, 15, 20, 21, 23, 30, 31, 35, 36, 107, 113, 114, 126, 135, 146, 153, 158, 164, 165, 104, 109, 115, 117, 120, 11, 27, 112, 118, 121, 139, 148, 152 14
Na	1. F-AES 2. AAS 3. ICP-AES 4. Ion chromatography 5. ICP-MS	4, 32, 33, 38, 39, 116, 133, 156 3, 10, 16, 19, 26, 27, 34, 40, 124, 157, 160 11, 27, 104, 109, 110, 111, 112, 115, 117, 118, 120, 121, 139, 148, 152 1, 5, 6, 7, 8, 12, 13, 15, 20, 21, 22, 23, 30, 31, 35, 36, 107, 114, 126, 135, 146, 153, 158, 164, 165 14
Cl	1. Ion chromatography 2. Spectrophotometry 3. Potentiometric method 4. Capillary electrophoresis	1, 3, 4, 5, 6, 7, 8, 10, 11, 12, 13, 14, 15, 16, 19, 20, 21, 22, 23, 24, 26, 27, 30, 31, 32, 33, 34, 35, 36, 38, 104, 107, 110, 114, 115, 116, 118, 124, 126, 130, 131, 135, 139, 146, 152, 153, 156, 157, 158, 160, 164, 165 18, 117, 40, 120, 121, 148, 133, 109, 112 39
Ca	1. AAS 2. ICP-AES 3. Ion chromatography 4. AES 5. ICP-MS	3, 4, 10, 16, 19, 22, 26, 33, 34, 38, 39, 40, 116, 124, 133, 154, 156, 157, 160 11, 27, 104, 109, 110, 112, 115, 117, 118, 120, 121, 139, 148 1, 5, 6, 7, 8, 12, 13, 15, 20, 21, 23, 31, 35, 36, 107, 114, 126, 135, 153, 158, 164, 165 32 14
K	1. AAS 2. Ion chromatograph 3. AES 4. ICP-MS	3, 4, 10, 16, 19, 26, 34, 40, 124, 160 1, 5, 6, 7, 8, 12, 13, 15, 20, 21, 22, 23, 30, 31, 35, 36, 107,, 114, 126, 135, 146, 153, 158, 164, 165 11, 27, 32, 33, 39, 104, 109, 110, 112, 115, 116, 117, 118, 120, 121, 133, 139, 148, 152, 156, 157 14

Appendix 2

Figures – 23rd intercomparison





Figure 2: SO_2 in absorbing solution.

Figure 3: SO₂ in impregnated filter.





Figure 4: HNO_3 in impregnated filter.

Figure 5: NH_3 in impregnated filter.




Figure 6: Percent deviation from theoretical value for sulphate.




Figure 7: Percent deviation from theoretical value for nitrate.




Figure 8: Percent deviation from theoretical value for ammonium.




Figure 9: Percent deviation from theoretical value for pH.




Figure 10: Percent deviation from theoretical value for chloride.




Figure 11: Percent deviation from theoretical value for sodium.




Figure 12: Percent deviation from theoretical value for magnesium.




Figure 13: Percent deviation from theoretical value for calcium.




Figure 14: Percent deviation from theoretical value for potassium.




Figure 15: Percent deviation from theoretical value for conductivity.




Figure 16: The median compared to theoretical value.




Figure 17: Youden plot of SO₂-S in absorbing solution.




Figure 18: Youden plot of SO₂-S on impregnated filter.




Figure 19: Youden plot of HNO₃-N on impregnated filter.




Figure 20: Youden plot of NH₃-N on impregnated filter.




Figure 21: Youden plot of Ca in precipitation.




Figure 22: Youden plot of Cl in precipitation.




Figure 23: Youden plot of conductivity in precipitation.




Figure 24: Youden plot of K in precipitation.




Figure 25: Youden plot of Mg in precipitation.




Figure 26: Youden plot of Na in precipitation.




Figure 27: Youden plot of NH₄-N in precipitation.




Figure 28: Youden plot of strong acid calculated from pH.




Figure 29: Youden plot of NO_3-N in precipitation.




Figure 30: Youden plot of $SO_4\text{-}S$ in precipitation.




Figure 31: Youden plot of pH in precipitation.

Appendix 3

Tables – 24rd intercomparison

Table 32: Samples distributed for the twenty-fourth interlaboratory test.

A.	5 synthetic samples for determination of SO ₂ , consisting of 0.3% H ₂ O ₂ absorbing solution and containing different concentrations of sulphuric acid. One of the samples was an unidentified blank.
B.	5 KOH-impregnated Whatman 40 filters, comprising 1 blank and 4 filters to which different amounts of sulphuric acid and nitrate salt have been added.
C.	4 synthetic samples for determination of NO ₂ consisting of sodium nitrite diluted in water.
J.	6 Whatman 40 filters impregnated with 3% oxalic acid, comprising 2 blank and 4 filters to which different amounts of ammonium salt solution have been added.
G.	4 synthetic precipitation samples, containing SO ₄ ²⁻ , NO ₃ ⁻ , NH ₄ ⁺ , H ⁺ , Na ⁺ , Mg ²⁺ and Cl ⁻ , and Ca ²⁺ and K ⁺ .

Table 33a: EMEP laboratories participating in the twenty-fourth laboratory intercomparison. The numbers in front of the names are used in tables and figures.

Austria	(1)	Umweltbundesamt Zweigstelle Sud, Klagenfurt
Canada	(26)	The Ontario Ministry of the Environment, Laboratory Service Branch
Croatia	(35)	Meteorological and Hydrological Service of Croatia
Czech Republic	(3)	Czech Hydrometeorological Institute, Praha
Denmark	(4)	National Environmental Research Institute. Air Pollution Laboratory
Estonia	(38)	Estonian Environmental Research Centre, Tallinn
Finland	(5)	Finnish Meteorological Institute. Air Quality Department
France	(6)	SGS Multilab
Germany	(7)	IfE Leipzig GmbH, Umweltlabor
Germany	(8)	Umweltbundesamt, Messstelle Schauinsland
Hungary	(10)	Institute for Atmospheric Physics
Iceland	(11)	Idntæknistofnun Islands (Technological Inst. of Iceland)
Ireland	(12)	Met. Eirann, Dublin
Italy	(13)	C.N.R. Istituto Inquinamento Atmosferico
Italy	(30)	Join Research centre, Ispra
Latvia	(33)	Air Pollution Observation Laboratory
Lithuania	(32)	Atmospheric Pollution Research Laboratory, Institute of Physics, Vilnius
Netherlands	(14)	National Institute of Public Health and Environmental Protection (RIVM)
Norway	(15)	Norwegian Institute for Air Research (NILU)
Macedonia	(40)	Hydrometeorological Institute, Skopje
Poland	(16)	Institute of Meteorology and Water Management, Warsaw
Poland	(39)	Environmental Monitoring Laboratory, Institute of Environmental Protection
Portugal	(17)	Laboratorio Santo Andre
Russian Federation	(22)	Institute of Global Climate and Ecology
Serbia and Montenegro	(24)	Rep. Hydrometeorological Institute of Serbia
Slovakia	(31)	Slovak Hydrometeorological Institute
Slovenia	(36)	Hydrometeorological Institute of Slovenia
Spain	(19)	Centro Nacional de Sanidad Ambiental
Sweden	(20)	Swedish Environmental Research Institute (IVL), Gothenburg
Switzerland	(21)	Swiss Federal Laboratories for Materials Testing (EMPA)
United Kingdom	(23)	AEA Technology, National Environmental Technology Centre
United States of America	(27)	Illinois State Water Survey

Table 33b: Participating laboratories outside the EMEP network.

Germany	(102)	Ökologie-Zentrum Universität Kiel
Germany	(104)	Hessige Landwirtschaftliche
Finland	(107)	The Finnish Forest Institute
Germany	(109)	Institut für Bondenkunde und Waldernährung
Germany	(110)	Thüringer Landesanstalt für Landwirtschaft (TTL), Jena
Germany	(112)	Niedersächsische Forstliche Versuchsanstalt (NVF)
Germany	(115)	Bayerische Landesanstalt f. Wald- und Forstwirtschaft
Switzerland	(116)	Institute for Applied Plant Biology
Germany	(117)	Sächsische Landesanstalt für Forsten, Graupa
Germany	(118)	Forstliche Versuchs-und Forschungsanstalt
Germany	(120)	Landwirtschaftliche Untersuchungs- und Forschungsanstalt (LUFA)
Germany	(121)	Landesamt für Natur und Umwelt
Belgium	(124)	Laboratorium voor Bondenkunde, Gent
Germany	(125)	Bayerisches Landesamt für Umweltschutz, Augsburg
Italy	(126)	APPA Laboratorio Biologico Provinciale
Denmark	(139)	Danish Centre for Forest, Landscape and Planning, Hørsholm
Italy	(140)	C.N.R. Istituto di Ricerca sulle Acque
Austria	(147)	Institute of Pollution Research and Forest Chemistry
Netherlands	(148)	Alterra, Wageningen
Switzerland	(149)	WSL Zentrallabor, Birmensdorf
Belgium	(151)	Laboratoire de l'Unité des Eaux et Forêt (EFOR)
Norway	(152)	Norwegian Forest Research Institute, Ås
Slovenia	(153)	Slovenian Forestry Institute, Ljubljana
United Kingdom	(155)	Environmental Research Branch, Farnham
Japan	(158)	Acid Deposition and Oxidant Research Center (ADOCRC), Niigata
Ireland	(160)	Coillte Research Laboratory, Newtownmountkennedy
Slovenia	(161)	National Institute of Chemistry
Thailand	(163)	Environmental Researching and Training Center (ERTC)
Thailand	(164)	Pollution Control Department (PCD), Bangkok
Viet Nam	(165)	Institute of Meteorology and Hydrology, Ha Noi

Table 34: Analytical methods used at the participating laboratories for the determination of sulphur dioxide in absorbing solution (A).

Method	Laboratory
1. Ion chromatography	6, 15, 17, 21, 23, 36

Table 35: Analytical methods used at the participating laboratories for the determination of sulphur dioxide on impregnated filters (B).

Method	Laboratory
1. Spectrophotometry	16
2. Ion chromatography	3, 4, 5, 8, 11, 12, 13, 15, 20, 22, 31, 32, 33, 34, 36, 38, 116, 158, 164, 165
3. Capillary Ion Analysis	39

Table 36: Analytical methods used at the participating laboratories for determination of nitric acid on impregnated filters (B).

Method	Laboratory
1. Reduction to nitrite	16
2. Ion chromatography	3, 4, 5, 8, 11, 13, 15, 20, 22, 31, 32, 33, 34, 36, 116, 158, 64, 165
3. Capillary Ion Analysis	39

Table 37: Analytical methods for determination of ammonia on impregnated filters (J).

Method	Laboratory
1. Spectrophotometry	3, 4, 8, 10, 19, 20, 32, 33, 39, 116
3. Ion chromatography	5, 13, 15, 31, 36, 38, 158, 165

Table 38: Analytical results for sulphur dioxide in absorbing solution.

Method	Laboratory
Ion chromatography	19, 22
Spectrophotometry	3, 8, 10, 15, 16, 20, 23, 24, 31, 32, 33, 35, 38, 39

Table 39: Analytical results for sulphur dioxide in absorbing solution.

SO ₂ -S in absorbing solution				SO ₂ -S in absorbing solution			
Sample no.: A1		Sample no.: A3					
Theoretical value:	0.240	Theoretical value:	0.140				
Unit:		Unit:					
Run 1:		Run 1:					
Number of laboratories:	7	Number of laboratories:	7				
Arithmetic mean value:	0.229	Arithmetic mean value:	0.136				
Median:	0.241	Median:	0.136				
Standard deviation	0.036	Standard deviation	0.021				
Rel. st. deviation (%)	15.851	Rel. st. deviation (%)	15.117				
Run 2:		Run 2:					
Number of laboratories:	6	Number of laboratories:	7				
Arithmetic mean value:	0.241	Arithmetic mean value:	0.136				
Median:	0.244	Median:	0.136				
Standard deviation	0.015	Standard deviation	0.021				
Rel. st. deviation (%)	6.412	Rel. st. deviation (%)	15.117				
Results in decreasing order:				Results in decreasing order:			
17	0.263	36	0.231	17	0.170	36	0.134
23	0.248	15	0.218	21	0.141	15	0.130
21	0.247	19	0.153 (*)	23	0.139	19	0.100
6	0.241			6	0.136		
SO ₂ -S in absorbing solution				SO ₂ -S in absorbing solution			
Sample no.: A4		Sample no.: A5					
Theoretical value:	0.440	Theoretical value:	0.160				
Unit:		Unit:					
Run 1:		Run 1:					
Number of laboratories:	7	Number of laboratories:	7				
Arithmetic mean value:	0.416	Arithmetic mean value:	0.159				
Median:	0.420	Median:	0.158				
Standard deviation	0.036	Standard deviation	0.013				
Rel. st. deviation (%)	8.581	Rel. st. deviation (%)	8.283				
Run 2:		Run 2:					
Number of laboratories:	6	Number of laboratories:	7				
Arithmetic mean value:	0.429	Arithmetic mean value:	0.159				
Median:	0.428	Median:	0.158				
Standard deviation	0.013	Standard deviation	0.013				
Rel. st. deviation (%)	3.106	Rel. st. deviation (%)	8.283				
Results in decreasing order:				Results in decreasing order:			
21	0.443	15	0.418	17	0.183	36	0.151
23	0.442	17	0.413	23	0.166	19	0.147
6	0.437	19	0.340 (*)	21	0.162	15	0.145
36	0.420			6	0.158		

Table 40: The ratios of the theoretical values and the results found by the laboratories in the determination of sulphur dioxide in absorbing solutions.

Lab no.	Measured / Expected value Sample No.			
	A1	A3	A4	A5
6	1.00	0.97	0.99	0.99
15	0.91	0.93	0.95	0.90
17	1.09	1.21	0.94	1.14
19	0.64	0.71	0.77	0.92
21	1.03	1.01	1.00	1.01
23	1.03	0.99	1.00	1.04
36	0.96	0.96	0.95	0.94

Table 41: Analytical results for sulphur dioxide in impregnated filter.

SO2-S on impregnated filter Sample no.: B2 Theoretical value: 24.050 Unit: ug S/filter	SO2-S on impregnated filter Sample no.: B3 Theoretical value: 10.020 Unit: ug S/filter
Run 1:	Run 1:
Number of laboratories: 19 Arithmetic mean value: 22.522 Median: 23.190 Standard deviation 2.209 Rel. st. deviation (%) 9.808	Number of laboratories: 19 Arithmetic mean value: 9.301 Median: 9.500 Standard deviation 1.547 Rel. st. deviation (%) 16.636
Run 2:	Run 2:
Number of laboratories: 18 Arithmetic mean value: 22.815 Median: 23.245 Standard deviation 1.855 Rel. st. deviation (%) 8.131	Number of laboratories: 18 Arithmetic mean value: 9.541 Median: 9.553 Standard deviation 1.173 Rel. st. deviation (%) 12.292
Results in decreasing order:	Results in decreasing order:
8 25.700 20 22.727 31 24.751 36 22.360 15 24.700 116 21.600 3 24.440 32 21.600 33 24.188 38 21.270 5 24.013 165 20.490 4 23.740 39 20.400 16 23.650 164 18.550 22 23.300 11 17.250 (*) 158 23.190	33 11.348 15 9.420 8 11.000 20 9.180 16 10.950 165 9.020 3 10.760 4 8.920 5 10.573 38 8.770 158 10.090 32 8.350 36 9.860 116 7.900 22 9.700 164 6.800 31 9.605 11 4.980 (*) 39 9.500
SO2-S on impregnated filter Sample no.: B4 Theoretical value: 72.100 Unit: ug S/filter	SO2-S on impregnated filter Sample no.: B5 Theoretical value: 56.080 Unit: ug S/filter
Run 1:	Run 1:
Number of laboratories: 19 Arithmetic mean value: 69.999 Median: 71.400 Standard deviation 4.631 Rel. st. deviation (%) 6.616	Number of laboratories: 19 Arithmetic mean value: 54.330 Median: 54.640 Standard deviation 3.768 Rel. st. deviation (%) 6.935
Run 2:	Run 2:
Number of laboratories: 18 Arithmetic mean value: 70.563 Median: 71.621 Standard deviation 4.039 Rel. st. deviation (%) 5.723	Number of laboratories: 18 Arithmetic mean value: 54.751 Median: 54.980 Standard deviation 3.386 Rel. st. deviation (%) 6.184
Results in decreasing order:	Results in decreasing order:
33 76.393 36 71.160 22 74.900 158 70.850 15 74.200 4 70.780 5 74.043 20 68.920 16 73.120 38 65.970 3 72.110 165 65.840 116 72.000 39 63.000 32 71.990 11 61.620 31 71.843 164 59.850 (*) 8 71.400	22 61.100 116 54.500 33 58.655 32 53.960 31 57.680 20 53.020 15 57.400 38 52.470 5 57.333 36 51.960 16 57.230 165 50.600 158 56.350 39 49.900 8 55.700 11 47.700 3 55.320 164 46.750 (*) 4 54.640

Table 42: The ratios of the theoretical values and the results found by the laboratories in the determination of sulphur dioxide on impregnated filters. The reported results are corrected for blank value (B1).

Lab no.	Measured / expected value Sample no.				
	B 2	B 3	B 4	B 5	Average
3	1.02	1.07	1.00	0.99	1.02
4	0.99	0.89	0.98	0.97	0.96
5	1.00	1.06	1.03	1.02	1.03
8	1.07	1.10	0.99	0.99	1.04
11	0.72	0.50	0.85	0.85	0.73
15	1.03	0.94	1.03	1.02	1.00
16	0.98	1.09	1.01	1.02	1.03
20	0.94	0.92	0.96	0.95	0.94
22	0.97	0.97	1.04	1.09	1.02
31	1.03	0.96	1.00	1.03	1.00
32	0.90	0.83	1.00	0.96	0.92
33	1.01	1.13	1.06	1.05	1.06
36	0.93	0.98	0.99	0.93	0.96
38	0.88	0.88	0.91	0.94	0.90
39	0.85	0.95	0.87	0.89	0.89
116	0.90	0.79	1.00	0.97	0.91
158	0.96	1.01	0.98	1.00	0.99
164	0.77	0.68	0.83	0.83	0.78
165	0.85	0.90	0.91	0.90	0.89

Table 43: Analytical results for nitric acid on impregnated filter.

HNO₃-N on impregnated filter
 Sample no.: B2
 Theoretical value: 16.360
 Unit: ug N/filter

Run 1:

Number of laboratories: 19
 Arithmetic mean value: 15.757
 Median: 15.966
 Standard deviation 1.329
 Rel. st. deviation (%) 8.432

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 16.015
 Median: 16.009
 Standard deviation 0.723
 Rel. st. deviation (%) 4.514

Results in decreasing order:

33	17.350	4	15.880
15	16.900	32	15.780
8	16.900	31	15.696
158	16.560	38	15.600
36	16.490	16	15.500
22	16.350	116	15.400
3	16.290	165	14.930
11	16.230	39	14.400
20	16.053	164	11.100 (*)
5	15.966		

HNO₃-N on impregnated filter
 Sample no.: B3
 Theoretical value: 9.816
 Unit: ug N/filter

Run 1:

Number of laboratories: 19
 Arithmetic mean value: 10.038
 Median: 10.010
 Standard deviation 0.803
 Rel. st. deviation (%) 7.997

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 10.170
 Median: 10.100
 Standard deviation 0.573
 Rel. st. deviation (%) 5.633

Results in decreasing order:

8	11.200	3	10.000
15	11.100	16	10.000
32	10.990	20	9.920
5	10.736	11	9.860
4	10.360	165	9.830
33	10.304	31	9.619
22	10.250	39	9.600
38	10.200	116	8.900
36	10.190	164	7.650 (*)
158	10.010		

HNO₃-N on impregnated filter
 Sample no.: B4
 Theoretical value: 32.720
 Unit: ug N/filter

Run 1:

Number of laboratories: 19
 Arithmetic mean value: 32.669
 Median: 32.660
 Standard deviation 1.679
 Rel. st. deviation (%) 5.138

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 32.934
 Median: 32.763
 Standard deviation 1.254
 Rel. st. deviation (%) 3.806

Results in decreasing order:

22	35.350	3	32.500
15	34.800	8	32.200
33	34.605	36	32.190
5	34.456	31	32.124
158	33.540	165	31.700
38	33.400	32	31.440
4	33.360	116	31.400
11	33.120	39	31.100
20	32.867	164	27.900 (*)
16	32.660		

HNO₃-N on impregnated filter
 Sample no.: B5
 Theoretical value: 29.450
 Unit: ug N/filter

Run 1:

Number of laboratories: 19
 Arithmetic mean value: 28.921
 Median: 29.490
 Standard deviation 2.028
 Rel. st. deviation (%) 7.012

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 29.305
 Median: 29.540
 Standard deviation 1.175
 Rel. st. deviation (%) 4.009

Results in decreasing order:

33	31.220	31	28.998
15	31.000	165	28.990
5	30.716	20	28.693
22	30.150	11	28.210
158	30.140	32	28.120
38	29.700	116	28.000
8	29.600	3	27.140
36	29.590	164	22.000 (*)
16	29.490		

Table 3: The ratios of the theoretical values and the results found by the laboratories in the determination of nitric acid on impregnated filters. The reported results are corrected for blank value (B1).

Lab no.	Measured / expected value				Average	
	Sample no.					
	B 2	B 3	B 4	B 5		
3	1.00	1.02	0.99	0.92	0.98	
4	0.97	1.06	1.02	1.02	1.03	
5	0.98	1.09	1.05	1.04	1.06	
8	1.03	1.14	0.98	1.01	1.04	
11	0.99	1.00	1.01	0.96	0.99	
15	1.03	1.13	1.06	1.05	1.08	
16	0.95	1.02	1.00	1.00	1.01	
20	0.98	1.01	1.00	0.97	1.00	
22	1.00	1.04	1.08	1.02	1.05	
31	0.96	0.98	0.98	0.98	0.98	
32	0.96	1.12	0.96	0.95	1.01	
33	1.06	1.05	1.06	1.06	1.06	
36	1.01	1.04	0.98	1.00	1.01	
38	0.95	1.04	1.02	1.01	1.02	
39	0.88	0.98	0.95	0.94	0.96	
116	0.94	0.91	0.96	0.95	0.94	
158	1.01	1.02	1.03	1.02	1.02	
164	0.68	0.78	0.85	0.75	0.79	
165	0.91	1.00	0.97	0.98	0.98	

Table 44: Analytical results of nitrogen dioxide in absorbing solutions.

NO₂-N in absorbing solution
 Sample no.: C1
 Theoretical value: 0.084
 Unit: ug N/ml

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 0.086
 Median: 0.085
 Standard deviation 0.004
 Rel. st. deviation (%) 5.055

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 0.086
 Median: 0.085
 Standard deviation 0.004
 Rel. st. deviation (%) 5.055

Results in decreasing order:

32	0.093	10	0.085
3	0.091	31	0.084
33	0.090	8	0.084
35	0.090	16	0.083
20	0.089	38	0.082
22	0.089	19	0.082
24	0.089	23	0.080
12	0.088	15	0.079
4	0.086	39	0.079

NO₂-N in absorbing solution
 Sample no.: C2
 Theoretical value: 0.051
 Unit: ug N/ml

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 0.052
 Median: 0.052
 Standard deviation 0.004
 Rel. st. deviation (%) 8.211

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 0.053
 Median: 0.052
 Standard deviation 0.004
 Rel. st. deviation (%) 7.091

Results in decreasing order:

12	0.059	4	0.052
32	0.058	8	0.051
33	0.058	16	0.050
35	0.056	39	0.049
3	0.055	15	0.049
10	0.054	38	0.049
20	0.054	19	0.048
24	0.054	31	0.047
22	0.052	23	0.043 (*)

NO₂-N in absorbing solution
 Sample no.: C3
 Theoretical value: 0.108
 Unit: ug N/ml

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 0.108
 Median: 0.106
 Standard deviation 0.006
 Rel. st. deviation (%) 5.469

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 0.108
 Median: 0.106
 Standard deviation 0.006
 Rel. st. deviation (%) 5.469

Results in decreasing order:

32	0.119	19	0.105
35	0.118	38	0.105
12	0.114	31	0.105
33	0.114	24	0.105
22	0.112	16	0.104
20	0.111	39	0.103
8	0.110	3	0.100
10	0.108	15	0.100
4	0.108	23	0.100

NO₂-N in absorbing solution
 Sample no.: C4
 Theoretical value: 0.118
 Unit: ug N/ml

Run 1:

Number of laboratories: 17
 Arithmetic mean value: 0.112
 Median: 0.116
 Standard deviation 0.029
 Rel. st. deviation (%) 26.212

Run 2:

Number of laboratories: 16
 Arithmetic mean value: 0.119
 Median: 0.117
 Standard deviation 0.007
 Rel. st. deviation (%) 5.739

Results in decreasing order:

35	0.130	38	0.116
32	0.129	10	0.115
24	0.126	19	0.114
33	0.126	16	0.114
12	0.122	15	0.111
20	0.122	39	0.111
8	0.120	3	0.107
4	0.118	22	0.001 (*)
31	0.116		

Table 45: The ratios of the theoretical values and the results found by the laboratories in the determination of nitrogen dioxide in absorbing solutions.

Lab No.	Measured / expected value				Average	
	Sample No.					
	C1	C2	C3	C4		
3	1,08	1,08	0,93	0,90	1,00	
4	1,01	1,04	1,00	1,00	1,01	
8	1,00	1,01	1,02	1,02	1,01	
10	1,01	1,07	1,00	0,97	1,01	
12	1,04	1,17	1,06	1,03	1,07	
15	0,94	0,97	0,93	0,94	0,94	
16	0,98	0,99	0,96	0,97	0,98	
19	0,97	0,95	0,97	0,97	0,96	
20	1,05	1,06	1,03	1,04	1,05	
22	1,06	1,03	1,04	0,01	0,78	
23	0,95	0,85	0,93		0,91	
24	1,06	1,07	0,97	1,07	1,04	
31	1,00	0,93	0,97	0,98	0,97	
32	1,10	1,15	1,10	1,09	1,11	
33	1,07	1,15	1,06	1,07	1,08	
35	1,07	1,11	1,09	1,10	1,09	
38	0,97	0,97	0,97	0,98	0,97	
39	0,94	0,97	0,95	0,94	0,95	

Table 46: Analytical results for ammonia on impregnated filter.

NH3-N on impregnated filter
 Sample no.: J1
 Theoretical value: 10.030
 Unit: ug N/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 9.549
 Median: 9.540
 Standard deviation 1.759
 Rel. st. deviation (%) 18.424

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 9.302
 Median: 9.510
 Standard deviation 1.459
 Rel. st. deviation (%) 15.684

Results in decreasing order:

15	13.735	(*)	39	9.510
31	11.628		19	9.451
38	10.950		33	9.338
13	10.900		116	9.010
4	10.350		158	8.810
8	10.150		32	8.270
36	9.915		165	7.970
20	9.670		10	6.388
5	9.570		3	6.260

NH3-N on impregnated filter
 Sample no.: J2
 Theoretical value: 22.060
 Unit: ug N/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 20.992
 Median: 21.402
 Standard deviation 1.856
 Rel. st. deviation (%) 8.841

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 21.298
 Median: 21.484
 Standard deviation 1.367
 Rel. st. deviation (%) 6.419

Results in decreasing order:

38	23.650		20	21.320
31	22.788		4	21.150
13	22.310		32	20.960
15	22.135		5	20.880
33	22.018		39	20.410
36	22.015		165	20.310
8	21.850		158	19.410
116	21.720		3	17.660
19	21.484		10	15.790 (*)

NH3-N on impregnated filter
 Sample no.: J3
 Theoretical value: 32.080
 Unit: ug N/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 31.433
 Median: 31.985
 Standard deviation 2.554
 Rel. st. deviation (%) 8.127

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 31.849
 Median: 32.120
 Standard deviation 1.906
 Rel. st. deviation (%) 5.983

Results in decreasing order:

13	34.160		5	31.850
32	34.100		20	31.580
38	33.950		4	31.250
33	33.658		165	30.300
36	33.215		39	30.110
15	33.135		31	29.718
8	33.050		158	29.110
19	32.375		3	27.752
116	32.120		10	24.369 (*)

NH3-N on impregnated filter
 Sample no.: J4
 Theoretical value: 9.020
 Unit: ug N/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 9.442
 Median: 8.827
 Standard deviation 3.975
 Rel. st. deviation (%) 42.104

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 8.550
 Median: 8.824
 Standard deviation 1.267
 Rel. st. deviation (%) 14.814

Results in decreasing order:

165	24.590	(*)	19	8.824
38	10.050		20	8.740
13	9.990		116	8.560
36	9.915		33	8.479
31	9.588		32	7.920
4	9.350		158	7.800
39	9.010		8	7.750
15	8.875		10	6.530
5	8.830		3	5.147

Table 47: The ratios of the theoretical values and the results found by the laboratories in the determination of ammonia on impregnated filters. The reported results are corrected for an average blank value (J5 and J6).

Lab No.	Measured / expected value Sample no.				Average
	J 1	J 2	J 3	J 4	
3	0.62	0.80	0.87	0.57	0.72
4	1.03	0.96	0.97	1.04	1.00
5	0.95	0.95	0.99	0.98	0.97
8	1.01	0.99	1.03	0.86	0.97
10	0.64	0.72	0.76	0.72	0.71
13	1.09	1.01	1.06	1.11	1.07
15	1.37	1.00	1.03	0.98	1.10
19	0.94	0.97	1.01	0.98	0.98
20	0.96	0.97	0.98	0.97	0.97
31	1.16	1.03	0.93	1.06	1.05
32	0.82	0.95	1.06	0.88	0.93
33	0.93	1.00	1.05	0.94	0.98
36	0.99	1.00	1.04	1.10	1.03
38	1.09	1.07	1.06	1.11	1.08
39	0.95	0.93	0.94	1.00	0.95
116	0.90	0.98	1.00	0.95	0.96
158	0.88	0.88	0.91	0.86	0.88
165	0.80	0.92	0.94	2.73	1.35

Table 48: Analytical results for sulphate in precipitation samples.

Sulphate in precipitation			
Sample no.: G1		Sulphate in precipitation	
Theoretical value: 1.533		Sample no.: G2	
Unit: µg/l		Theoretical value: 1.766	
Run 1:			
Number of laboratories: 64		Run 1:	
Arithmetic mean value: 1.644		Number of laboratories: 63	
Median: 1.538		Arithmetic mean value: 1.905	
Standard deviation 0.573		Median: 1.757	
Rel. st. deviation (%) 34.879		Standard deviation 0.700	
Run 2:		Rel. st. deviation (%) 36.745	
Number of laboratories: 62		Run 2:	
Arithmetic mean value: 1.548		Number of laboratories: 61	
Median: 1.535		Arithmetic mean value: 1.786	
Standard deviation 0.196		Median: 1.757	
Rel. st. deviation (%) 12.693		Standard deviation 0.229	
Results in decreasing order:			
139 4.720 (*) 158 1.536		139 5.740 (*) 149 1.757	
104 4.530 (*) 27 1.535		104 5.320 (*) 27 1.756	
13 2.590 116 1.535		13 3.030 117 1.756	
10 2.184 5 1.525		10 2.456 5 1.755	
160 1.920 15 1.524		160 2.190 15 1.750	
14 1.760 6 1.520		7 2.015 102 1.750	
2 1.735 152 1.510		2 1.993 161 1.750	
7 1.734 26 1.505		14 1.990 32 1.741	
151 1.680 120 1.500		31 1.951 19 1.739	
107 1.660 12 1.500		151 1.930 26 1.733	
31 1.615 38 1.500		107 1.920 167 1.732	
8 1.610 166 1.500		22 1.873 152 1.730	
22 1.604 125 1.491		23 1.862 166 1.730	
153 1.604 140 1.480		147 1.840 11 1.726	
3 1.583 11 1.479		8 1.840 38 1.720	
30 1.580 118 1.470		110 1.820 12 1.720	
39 1.574 147 1.470		35 1.811 120 1.720	
24 1.570 167 1.469		30 1.810 140 1.710	
35 1.569 102 1.464		3 1.808 125 1.709	
121 1.560 36 1.460		165 1.804 36 1.694	
23 1.560 32 1.458		39 1.803 121 1.680	
21 1.554 19 1.457		115 1.790 20 1.667	
165 1.551 20 1.452		21 1.781 109 1.650	
115 1.550 109 1.410		24 1.780 112 1.620	
124 1.550 112 1.400		1 1.770 33 1.600	
16 1.548 33 1.400		4 1.768 118 1.570	
4 1.546 117 1.361		158 1.764 126 1.550	
149 1.542 126 1.320		163 1.763 148 1.500	
1 1.540 164 1.280		116 1.762 164 1.460	
110 1.540 148 1.280		124 1.760 17 1.416	
161 1.540 155 1.251		6 1.760 155 1.372	
163 1.539 17 1.240		16 1.757	
Sulphate in precipitation			
Sample no.: G3		Sulphate in precipitation	
Theoretical value: 0.983		Sample no.: G4	
Unit: µg/l		Theoretical value: 1.183	
Run 1:			
Number of laboratories: 64		Run 1:	
Arithmetic mean value: 1.024		Number of laboratories: 63	
Median: 0.978		Arithmetic mean value: 1.222	
Standard deviation 0.279		Median: 1.172	
Rel. st. deviation (%) 27.302		Standard deviation 0.334	
Run 2:		Rel. st. deviation (%) 27.362	
Number of laboratories: 62		Run 2:	
Arithmetic mean value: 0.983		Number of laboratories: 61	
Median: 0.977		Arithmetic mean value: 1.171	
Standard deviation 0.098		Median: 1.170	
Rel. st. deviation (%) 9.981		Standard deviation 0.093	
Results in decreasing order:			
104 3.000 (*) 1 0.978		104 3.630 (*) 6 1.170	
10 1.593 (*) 15 0.976		13 1.900 (*) 102 1.166	
13 1.400 158 0.972		7 1.405 167 1.164	
7 1.190 19 0.971		14 1.380 163 1.163	
14 1.150 124 0.970		160 1.380 38 1.160	
160 1.130 6 0.970		160 1.350 149 1.159	
112 1.120 166 0.970		31 1.294 26 1.159	
2 1.095 149 0.970		121 1.290 5 1.157	
22 1.052 26 0.967		107 1.280 19 1.157	
31 1.045 167 0.965		22 1.263 152 1.150	
107 1.040 11 0.965		23 1.243 12 1.150	
153 1.037 5 0.964		8 1.240 109 1.150	
102 1.036 16 0.962		151 1.240 16 1.148	
151 1.030 110 0.960		153 1.237 11 1.146	
8 1.030 116 0.957		115 1.220 140 1.140	
24 1.010 152 0.950		24 1.210 120 1.140	
115 1.010 12 0.950		35 1.209 112 1.140	
35 1.010 109 0.950		3 1.209 110 1.140	
39 1.009 140 0.940		21 1.196 125 1.125	
23 1.007 155 0.937		10 1.194 118 1.120	
3 1.003 125 0.935		30 1.190 36 1.114	
139 1.000 20 0.932		165 1.190 126 1.110	
21 0.995 38 0.930		4 1.189 20 1.109	
4 0.990 36 0.925		27 1.186 32 1.095	
27 0.990 147 0.920		39 1.181 33 1.060	
165 0.984 32 0.905		116 1.181 117 1.059	
121 0.981 126 0.900		1 1.180 124 1.050	
118 0.980 33 0.890		147 1.180 155 1.027	
120 0.980 117 0.889		161 1.180 164 0.990	
30 0.980 164 0.800		166 1.180 17 0.956	
161 0.980 17 0.800		158 1.173 148 0.860	
163 0.979 148 0.610		15 1.172	

Table 49: Analytical results for nitrate in precipitation samples.

Nitrate in precipitation			
Sample no.: G1			
Theoretical value:	0.497	Run 1:	Nitrate in precipitation
Unit: µg/l		Number of laboratories:	63
Run 1:	Number of laboratories:	63	Sample no.: G2
Number of laboratories:	63	Arithmetic mean value:	0.527
Arithmetic mean value:	0.527	Median:	0.484
Median:	0.484	Standard deviation	0.246
Standard deviation	0.246	Rel. st. deviation (%)	46.618
Run 2:	Number of laboratories:	61	Run 1:
Number of laboratories:	61	Number of laboratories:	62
Arithmetic mean value:	0.484	Arithmetic mean value:	0.471
Median:	0.484	Median:	0.430
Standard deviation	0.027	Standard deviation	0.224
Rel. st. deviation (%)	5.624	Rel. st. deviation (%)	47.588
Results in decreasing order:			
121 2.100 (*) 5 0.484	121 1.900 (*) 104 0.430		
167 1.580 (*) 21 0.482	167 1.410 (*) 6 0.430		
10 0.593 16 0.481	10 0.553 110 0.430		
2 0.549 6 0.480	13 0.480 117 0.430		
22 0.527 110 0.480	2 0.478 118 0.430		
32 0.525 26 0.479	32 0.472 24 0.430		
153 0.521 158 0.477	22 0.471 158 0.427		
13 0.520 7 0.476	8 0.468 7 0.425		
8 0.514 3 0.475	31 0.461 3 0.425		
11 0.514 155 0.474	11 0.461 15 0.422		
12 0.500 163 0.473	35 0.452 155 0.421		
151 0.500 39 0.471	12 0.450 163 0.421		
161 0.500 160 0.470	151 0.450 107 0.421		
112 0.500 117 0.470	161 0.450 120 0.420		
118 0.500 107 0.470	20 0.443 124 0.420		
23 0.499 124 0.470	165 0.443 14 0.420		
35 0.499 147 0.470	38 0.443 160 0.420		
31 0.494 15 0.469	1 0.442 19 0.418		
38 0.493 19 0.463	36 0.442 140 0.410		
148 0.492 109 0.460	148 0.441 33 0.410		
149 0.491 140 0.460	149 0.440 147 0.410		
24 0.490 14 0.460	112 0.440 125 0.406		
20 0.490 33 0.460	166 0.440 102 0.400		
104 0.490 120 0.460	5 0.439 152 0.400		
166 0.490 125 0.457	27 0.438 109 0.400		
36 0.489 116 0.455	4 0.436 116 0.398		
102 0.489 30 0.450	115 0.436 126 0.397		
27 0.488 152 0.450	23 0.433 17 0.393		
1 0.485 17 0.446	26 0.433 39 0.392		
4 0.485 126 0.438	21 0.431 30 0.380		
165 0.484 164 0.410	16 0.431 164 0.360		
115 0.484			
Results in decreasing order:			
121 2.720 (*) 155 0.637	121 3.140 (*) 6 0.740		
167 2.031 (*) 107 0.637	167 2.497 (*) 110 0.740		
10 0.776 102 0.636	2 0.856 158 0.740		
22 0.715 115 0.636	22 0.851 7 0.737		
2 0.715 165 0.635	13 0.809 163 0.736		
32 0.711 158 0.635	32 0.791 20 0.736		
11 0.695 3 0.634	12 0.790 155 0.735		
13 0.682 7 0.632	31 0.788 149 0.735		
151 0.680 147 0.630	8 0.773 15 0.731		
12 0.680 19 0.629	23 0.772 160 0.730		
104 0.670 163 0.628	153 0.771 120 0.730		
153 0.669 15 0.626	11 0.770 117 0.730		
8 0.668 26 0.626	166 0.770 112 0.730		
148 0.663 16 0.625	151 0.770 26 0.729		
23 0.662 39 0.623	161 0.770 19 0.728		
166 0.660 117 0.620	148 0.768 16 0.725		
31 0.659 33 0.620	104 0.760 116 0.715		
38 0.653 124 0.620	38 0.759 10 0.714		
35 0.651 118 0.620	27 0.756 14 0.710		
161 0.650 160 0.620	21 0.755 140 0.710		
36 0.649 120 0.610	35 0.753 124 0.710		
27 0.649 14 0.610	36 0.752 33 0.710		
4 0.647 140 0.610	4 0.752 17 0.703		
21 0.646 125 0.602	147 0.750 109 0.700		
1 0.641 116 0.600	165 0.750 152 0.700		
24 0.640 109 0.600	5 0.748 126 0.699		
112 0.640 17 0.600	1 0.747 125 0.697		
6 0.640 126 0.593	115 0.745 102 0.672		
110 0.640 152 0.590	107 0.742 30 0.670		
149 0.639 30 0.570	24 0.740 39 0.660		
5 0.639 164 0.560	118 0.740 164 0.650		
20 0.638	3 0.740		

Table 50: Analytical results for ammonium in precipitation samples.

Ammonium in precipitation			Ammonium in precipitation		
Sample no.: G1	Sample no.: G2	Theoretical value: 0.281	Theoretical value: 0.180	Unit: µg/l	Unit: µg/l
Run 1:					
Number of laboratories: 62	Number of laboratories: 60				
Arithmetic mean value: 0.285	Arithmetic mean value: 0.180				
Median: 0.285	Median: 0.183				
Standard deviation 0.034	Standard deviation 0.040				
Rel. st. deviation (%) 12.081	Rel. st. deviation (%) 21.915				
Run 2:					
Number of laboratories: 56	Number of laboratories: 58				
Arithmetic mean value: 0.284	Arithmetic mean value: 0.186				
Median: 0.285	Median: 0.185				
Standard deviation 0.015	Standard deviation 0.026				
Rel. st. deviation (%) 5.394	Rel. st. deviation (%) 13.795				
Results in decreasing order:					
2 0.378 (*) 167 0.285	2 0.255 4 0.182				
164 0.370 (*) 4 0.285	121 0.251 16 0.181				
121 0.365 (*) 36 0.285	147 0.250 163 0.181				
147 0.360 (*) 1 0.284	164 0.240 110 0.180				
32 0.320 21 0.284	32 0.234 107 0.180				
120 0.310 158 0.282	13 0.213 102 0.180				
39 0.303 10 0.282	39 0.212 33 0.180				
31 0.301 163 0.281	124 0.200 7 0.179				
24 0.300 30 0.280	24 0.200 5 0.177				
161 0.300 109 0.280	36 0.200 10 0.177				
112 0.300 14 0.280	112 0.200 22 0.177				
124 0.300 33 0.280	158 0.196 14 0.176				
17 0.298 166 0.280	1 0.195 118 0.175				
149 0.297 6 0.280	19 0.194 6 0.170				
22 0.295 152 0.280	31 0.194 104 0.170				
26 0.294 107 0.277	125 0.194 166 0.170				
13 0.294 118 0.275	20 0.191 155 0.170				
153 0.293 3 0.273	160 0.190 109 0.170				
125 0.292 16 0.273	17 0.190 152 0.170				
19 0.290 126 0.273	30 0.190 38 0.170				
117 0.290 5 0.272	3 0.190 8 0.167				
27 0.290 104 0.270	161 0.190 116 0.166				
15 0.290 23 0.270	12 0.190 126 0.164				
160 0.290 7 0.269	140 0.190 35 0.162				
110 0.290 38 0.264	15 0.190 148 0.157				
12 0.290 8 0.260	149 0.188 151 0.140				
140 0.290 116 0.248	27 0.188 120 0.130				
20 0.288 148 0.245	26 0.188 165 0.108				
35 0.287 151 0.230	21 0.186 23 0.026 (*)				
102 0.286 165 0.203 (*)	167 0.184 117 0.010 (*)				
155 0.285 115 0.127 (*)					
Ammonium in precipitation					
Sample no.: G3	Sample no.: G4				
Theoretical value: 0.401	Theoretical value: 0.481				
Unit: µg/l	Unit: µg/l				
Run 1:					
Number of laboratories: 62	Number of laboratories: 62				
Arithmetic mean value: 0.405	Arithmetic mean value: 0.480				
Median: 0.404	Median: 0.485				
Standard deviation 0.048	Standard deviation 0.076				
Rel. st. deviation (%) 11.861	Rel. st. deviation (%) 15.835				
Run 2:					
Number of laboratories: 56	Number of laboratories: 59				
Arithmetic mean value: 0.401	Arithmetic mean value: 0.487				
Median: 0.403	Median: 0.486				
Standard deviation 0.026	Standard deviation 0.044				
Rel. st. deviation (%) 6.585	Rel. st. deviation (%) 9.061				
Results in decreasing order:					
2 0.540 (*) 10 0.403	121 0.647 (*) 21 0.484				
121 0.535 (*) 4 0.403	2 0.624 4 0.481				
164 0.520 (*) 116 0.402	164 0.600 117 0.480				
147 0.510 (*) 126 0.402	22 0.590 160 0.480				
35 0.476 104 0.400	147 0.580 30 0.480				
32 0.464 120 0.400	32 0.549 10 0.480				
6 0.440 12 0.400	110 0.540 5 0.479				
161 0.430 33 0.400	39 0.522 165 0.478				
110 0.430 165 0.397	126 0.516 163 0.477				
39 0.423 5 0.397	35 0.516 116 0.476				
17 0.422 14 0.394	125 0.515 14 0.471				
36 0.420 163 0.394	149 0.507 12 0.470				
19 0.420 8 0.392	17 0.507 8 0.470				
149 0.420 15 0.391	36 0.503 7 0.468				
153 0.419 152 0.390	104 0.500 38 0.463				
125 0.417 7 0.388	140 0.500 15 0.463				
155 0.415 38 0.387	6 0.500 118 0.462				
13 0.415 31 0.387	152 0.500 166 0.460				
1 0.414 118 0.386	161 0.500 120 0.460				
102 0.412 109 0.380	31 0.499 24 0.460				
20 0.410 166 0.380	27 0.497 109 0.460				
140 0.410 22 0.377	19 0.497 102 0.454				
117 0.410 16 0.376	158 0.494 112 0.450				
30 0.410 24 0.370	167 0.494 16 0.450				
124 0.410 107 0.369	153 0.492 107 0.440				
158 0.409 160 0.360	124 0.490 148 0.424				
21 0.408 148 0.359	155 0.490 23 0.403				
27 0.408 3 0.333	33 0.490 3 0.388				
26 0.407 151 0.310	1 0.488 151 0.360				
167 0.406 112 0.280 (*)	20 0.486 115 0.326 (*)				
23 0.404 115 0.243 (*)	13 0.486 26 0.050 (*)				

Table 51: Analytical results for pH in precipitation samples.

pH in precipitation				pH in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 4.100				Theoretical value: 4.040			
Unit: pH-unit				Unit: pH-unit			
Run 1:		Run 1:					
Number of laboratories:	58	Number of laboratories:	57				
Arithmetic mean value:	4.206	Arithmetic mean value:	4.080				
Median:	4.140	Median:	4.070				
Standard deviation	0.323	Standard deviation	0.070				
Rel. st. deviation (%)	7.688	Rel. st. deviation (%)	1.724				
Run 2:		Run 2:					
Number of laboratories:	56	Number of laboratories:	56				
Arithmetic mean value:	4.152	Arithmetic mean value:	4.073				
Median:	4.140	Median:	4.070				
Standard deviation	0.090	Standard deviation	0.053				
Rel. st. deviation (%)	2.170	Rel. st. deviation (%)	1.291				
Results in decreasing order:				Results in decreasing order:			
2	6.360 (*)	6	4.140	117	4.430 (*)	148	4.070
121	5.050 (*)	27	4.140	164	4.200	149	4.070
160	4.480	124	4.140	19	4.184	12	4.070
38	4.470	153	4.130	2	4.170	121	4.070
164	4.400	163	4.130	22	4.170	139	4.060
148	4.300	149	4.130	112	4.160	163	4.060
19	4.232	139	4.130	166	4.150	5	4.060
166	4.220	8	4.130	14	4.140	6	4.060
22	4.220	155	4.130	38	4.130	24	4.060
112	4.200	12	4.130	126	4.120	110	4.050
14	4.190	32	4.130	26	4.120	152	4.050
26	4.190	24	4.120	140	4.120	21	4.050
140	4.190	13	4.120	36	4.110	39	4.040
116	4.190	20	4.120	158	4.110	16	4.040
117	4.190	5	4.120	160	4.110	23	4.040
126	4.190	21	4.110	116	4.100	151	4.040
17	4.180	120	4.110	17	4.100	120	4.030
158	4.170	39	4.110	118	4.090	10	4.030
36	4.170	165	4.100	155	4.090	11	4.020
15	4.150	151	4.090	109	4.090	165	4.020
35	4.150	16	4.090	4	4.080	30	4.010
107	4.150	33	4.060	15	4.080	3	4.006
110	4.150	11	4.050	32	4.080	33	4.000
109	4.150	1	4.050	8	4.080	147	4.000
118	4.150	102	4.043	27	4.080	13	4.000
30	4.150	10	4.040	35	4.080	1	3.980
152	4.150	23	4.040	107	4.080	102	3.972
3	4.140	7	4.040	124	4.080	7	3.970
4	4.140	147	4.000	20	4.070		
pH in precipitation				pH in precipitation			
Sample no.: G3		Sample no.: G4					
Theoretical value:	4.400	Theoretical value:	4.320				
Unit: pH-unit		Unit: pH-unit					
Run 1:		Run 1:					
Number of laboratories:	59	Number of laboratories:	59				
Arithmetic mean value:	4.457	Arithmetic mean value:	4.362				
Median:	4.440	Median:	4.360				
Standard deviation	0.203	Standard deviation	0.096				
Rel. st. deviation (%)	4.547	Rel. st. deviation (%)	2.208				
Run 2:		Run 2:					
Number of laboratories:	57	Number of laboratories:	56				
Arithmetic mean value:	4.440	Arithmetic mean value:	4.357				
Median:	4.440	Median:	4.360				
Standard deviation	0.077	Standard deviation	0.057				
Rel. st. deviation (%)	1.740	Rel. st. deviation (%)	1.310				
Results in decreasing order:				Results in decreasing order:			
38	5.820 (*)	148	4.430	38	4.750 (*)	8	4.360
112	4.760	32	4.430	112	4.650 (*)	109	4.350
104	4.620	2	4.430	22	4.490	163	4.350
166	4.620	149	4.430	19	4.486	152	4.350
19	4.581	110	4.420	140	4.480	165	4.350
1	4.500	5	4.420	14	4.470	5	4.340
22	4.500	152	4.420	1	4.440	30	4.340
164	4.500	24	4.420	158	4.430	32	4.340
26	4.500	12	4.420	116	4.420	24	4.340
158	4.490	21	4.410	166	4.420	13	4.330
14	4.490	33	4.410	36	4.410	139	4.330
36	4.480	139	4.410	26	4.400	20	4.330
117	4.480	165	4.410	155	4.400	21	4.330
140	4.470	20	4.410	164	4.400	3	4.330
30	4.470	6	4.400	118	4.380	6	4.330
109	4.470	10	4.400	2	4.380	10	4.320
17	4.460	16	4.400	104	4.380	160	4.320
27	4.460	39	4.390	17	4.370	126	4.320
155	4.460	13	4.380	15	4.370	16	4.320
118	4.460	23	4.370	121	4.370	39	4.310
35	4.460	3	4.365	124	4.370	110	4.300
107	4.450	120	4.360	35	4.370	120	4.290
124	4.450	151	4.350	4	4.370	33	4.290
126	4.450	121	4.340	117	4.370	147	4.280
4	4.440	11	4.340	107	4.360	23	4.270
8	4.440	102	4.333	153	4.360	102	4.262
153	4.440	147	4.300	12	4.360	11	4.260
15	4.440	7	4.280	27	4.360	7	4.190
163	4.440	160	4.040 (*)	149	4.360	151	4.010 (*)
116	4.440			148	4.360		

Table 52: Analytical results for strong acid calculated from pH.

Strong acid calculated from pH Sample no.: G1 Theoretical value: 80.000 Unit: μeq	Strong acid calculated from pH Sample no.: G2 Theoretical value: 92.000 Unit: μeq
Run 1: Number of laboratories: 12 Arithmettic mean value: 70.652 Median: 72.465 Standard deviation 17.726 Rel. st. deviation (%) 25.089	Run 1: Number of laboratories: 12 Arithmettic mean value: 85.655 Median: 82.240 Standard deviation 11.634 Rel. st. deviation (%) 13.583
Run 2: Number of laboratories: 11 Arithmettic mean value: 74.075 Median: 74.130 Standard deviation 13.819 Rel. st. deviation (%) 18.656	Run 2: Number of laboratories: 11 Arithmettic mean value: 83.169 Median: 81.300 Standard deviation 8.205 Rel. st. deviation (%) 9.865
Results in decreasing order: 124 103.000 118 70.800 14 82.000 35 70.790 165 79.400 158 67.800 6 78.000 126 66.000 120 77.600 164 45.300 155 74.130 160 33.000 (*)	Results in decreasing order: 124 113.000 (*) 118 81.300 165 95.500 155 81.280 120 93.300 160 78.000 14 92.000 158 77.100 6 88.000 126 76.000 35 83.180 164 69.200
Strong acid calculated from pH Sample no.: G3 Theoretical value: 40.000 Unit: μeq	Strong acid calculated from pH Sample no.: G4 Theoretical value: 48.000 Unit: μeq
Run 1: Number of laboratories: 13 Arithmettic mean value: 42.218 Median: 36.000 Standard deviation 17.356 Rel. st. deviation (%) 41.110	Run 1: Number of laboratories: 13 Arithmettic mean value: 46.059 Median: 45.000 Standard deviation 8.580 Rel. st. deviation (%) 18.628
Run 2: Number of laboratories: 12 Arithmettic mean value: 38.153 Median: 35.350 Standard deviation 9.709 Rel. st. deviation (%) 25.447	Run 2: Number of laboratories: 12 Arithmettic mean value: 43.981 Median: 43.830 Standard deviation 4.364 Rel. st. deviation (%) 9.923
Results in decreasing order: 160 91.000 (*) 118 34.700 124 64.000 35 34.670 14 44.000 155 34.670 120 43.700 164 33.100 165 38.900 158 32.100 6 38.000 104 24.000 126 36.000	Results in decreasing order: 124 71.000 (*) 35 42.660 120 51.300 104 42.000 126 50.000 118 41.700 160 48.000 155 39.810 14 46.000 164 38.900 165 45.000 158 37.400 6 45.000

Table 53: Analytical results for chloride in precipitation samples.

Chloride in precipitation
Sample no.: G1
Theoretical value: 0.145
Unit: µg/l

Run 1:
Number of laboratories: 57
Arithmetic mean value: 0.169
Median: 0.144
Standard deviation 0.121
Rel. st. deviation (%) 71.732

Run 2:
Number of laboratories: 54
Arithmetic mean value: 0.155
Median: 0.143
Standard deviation 0.049
Rel. st. deviation (%) 31.370

Results in decreasing order:
10 0.911 (*) 21 0.142
160 0.420 (*) 7 0.141
112 < 0.370
153 0.320 16 0.141
19 < 0.310
166 0.300 3 0.140
100 < 0.300
167 0.291 116 0.140
104 0.240 15 0.140
2 0.214 125 0.139

161 0.210 8 0.139
115 0.207 20 0.139
151 0.200 1 0.138
152 0.200 158 0.135
117 < 0.200
38 0.180 17 0.133
24 0.180 36 0.131
22 0.172 126 0.130
32 0.171 6 0.130
147 0.170 163 0.130
30 0.170 140 0.130
4 0.165 107 0.129
155 0.162 33 0.129
39 0.158 102 0.124
5 0.151 11 0.101
23 0.151 165 0.101
12 0.150 124 0.100
26 0.149 164 0.100
121 < 0.100
35 0.148 14 0.091
31 0.148 120 0.090
13 0.146 118 0.070
149 0.145 148 -0.100 (*)
27 0.144

Chloride in precipitation
Sample no.: G3
Theoretical value: 0.463
Unit: µg/l

Run 1:
Number of laboratories: 63
Arithmetic mean value: 0.460
Median: 0.449
Standard deviation 0.071
Rel. st. deviation (%) 15.537

Run 2:
Number of laboratories: 61
Arithmetic mean value: 0.451
Median: 0.448
Standard deviation 0.051
Rel. st. deviation (%) 11.219

Results in decreasing order:
10 0.804 (*) 16 0.448
160 0.660 (*) 22 0.446
153 0.601 4 0.443
112 0.570 36 0.443
167 0.562 155 0.440
166 0.560 158 0.437
104 0.530 163 0.436
2 0.522 15 0.434
39 0.496 147 0.430
115 0.493 118 0.430
38 0.490 124 0.430
151 0.490 6 0.430
161 0.490 32 0.428
23 0.487 125 0.428
13 0.485 107 0.424
35 0.480 126 0.420
31 0.474 140 0.420
149 0.471 120 0.420
30 0.470 165 0.419
148 0.470 33 0.416
27 0.469 1 0.414
5 0.468 11 0.410
20 0.467 109 0.410
26 0.467 17 0.410
3 0.465 152 0.400
8 0.463 117 0.390
12 0.460 116 0.385
24 0.460 110 0.370
7 0.455 164 0.370
19 0.454 14 0.341
21 0.450 121 0.330
102 0.449

Chloride in precipitation
Sample no.: G2
Theoretical value: 0.174
Unit: µg/l

Run 1:
Number of laboratories: 57
Arithmetic mean value: 0.218
Median: 0.179
Standard deviation 0.140
Rel. st. deviation (%) 64.143

Run 2:
Number of laboratories: 55
Arithmetic mean value: 0.195
Median: 0.178
Standard deviation 0.064
Rel. st. deviation (%) 32.630

Results in decreasing order:
10 1.049 (*) 26 0.178
117 0.610 (*) 32 0.177
160 0.440 35 0.176
112 < 0.370
124 0.340 149 0.172
166 0.320 16 0.172
167 0.316 3 0.170
19 < 0.310
110 < 0.300
161 0.280 102 0.170
104 0.260 21 0.169
22 0.233 7 0.168
115 0.223 15 0.167

151 0.210 17 0.166
38 0.210 8 0.165
2 0.208 20 0.165
30 0.200 158 0.162
152 0.200 125 0.162
148 0.200 155 0.162
165 0.191 33 0.161
39 0.191 140 0.160
1 0.190 6 0.160
120 0.190 36 0.158
24 0.190 163 0.156
4 0.189 107 0.156
13 0.188 118 0.140

116 0.182 126 0.140
12 0.180 11 0.126
31 0.180 164 0.120
147 0.180 14 0.111
5 0.179 121 < 0.100

Chloride in precipitation
Sample no.: G4
Theoretical value: 0.521
Unit: µg/l

Run 1:
Number of laboratories: 63
Arithmetic mean value: 0.519
Median: 0.505
Standard deviation 0.067
Rel. st. deviation (%) 12.917

Run 2:
Number of laboratories: 58
Arithmetic mean value: 0.510
Median: 0.503
Standard deviation 0.045
Rel. st. deviation (%) 8.879

Results in decreasing order:
160 0.780 (*) 16 0.504
167 0.673 (*) 4 0.502
153 0.664 (*) 152 0.500
10 0.655 (*) 6 0.500
23 0.613 15 0.495
166 0.600 163 0.493
161 0.600 158 0.493
2 0.597 155 0.492
104 0.580 147 0.490
30 0.560 117 0.490
112 0.560 124 0.490
38 0.560 148 0.490
151 0.560 32 0.489
35 0.558 125 0.486
24 0.550 107 0.485
31 0.549 118 0.480
26 0.542 140 0.480
13 0.542 11 0.477
115 0.540 19 0.472
102 0.539 1 0.471
5 0.537 33 0.471
27 0.531 120 0.470
149 0.530 126 0.470
39 0.526 109 0.460
3 0.525 17 0.453
20 0.523 110 0.450
8 0.523 116 0.445
7 0.521 22 0.435
12 0.520 164 0.420
21 0.510 14 0.396
165 0.506 121 0.360 (*)
36 0.505

Table 54: Analytical results for sodium in precipitation samples.

Sodium in precipitation			Sodium in precipitation		
Sample no.: G1	Theoretical value:	Unit: µg/l	Sample no.: G2	Theoretical value:	Unit: µg/l
Run 1:					
Number of laboratories:	63		Number of laboratories:	61	
Arithmetic mean value:	0.260		Arithmetic mean value:	0.360	
Median:	0.259		Median:	0.347	
Standard deviation	0.055		Standard deviation	0.097	
Rel. st. deviation (%)	21.085		Rel. st. deviation (%)	27.016	
Run 2:					
Number of laboratories:	61		Number of laboratories:	59	
Arithmetic mean value:	0.257		Arithmetic mean value:	0.345	
Median:	0.259		Median:	0.343	
Standard deviation	0.027		Standard deviation	0.050	
Rel. st. deviation (%)	10.633		Rel. st. deviation (%)	14.556	
Results in decreasing order:					
24 0.600 (*) 27	0.258		117 0.896 (*) 21	0.343	
22 0.341 31	0.255		23 0.708 (*) 155	0.342	
102 0.327 163	0.255		22 0.532 125	0.341	
11 0.314 26	0.254		24 0.500 149	0.341	
2 0.304 16	0.250		161 0.420 8	0.341	
19 0.299 110	0.250		115 0.417 110	0.340	
104 0.290 109	0.250		165 0.415 166	0.340	
116 0.290 12	0.250		2 0.390 118	0.340	
166 0.290 151	0.250		102 0.386 112	0.340	
4 0.282 6	0.250		116 0.382 151	0.340	
121 0.280 14	0.249		35 0.381 26	0.339	
161 0.270 107	0.247		121 0.370 11	0.335	
112 0.270 152	0.247		6 0.360 20	0.332	
15 0.270 20	0.246		13 0.357 14	0.331	
120 0.270 3	0.246		4 0.353 38	0.329	
13 0.269 149	0.246		7 0.353 107	0.329	
7 0.265 38	0.244		36 0.352 152	0.327	
125 0.264 32	0.243		5 0.351 16	0.320	
158 0.264 126	0.240		32 0.350 3	0.320	
23 0.264 147	0.240		30 0.350 160	0.320	
115 0.263 124	0.240		12 0.350 147	0.320	
165 0.262 117	0.237		158 0.350 124	0.320	
39 0.262 118	0.230		164 0.350 31	0.319	
8 0.262 139	0.228		120 0.350 33	0.305	
36 0.261 10	0.222		109 0.350 139	0.301	
5 0.261 33	0.221		104 0.350 10	0.292	
1 0.261 160	0.220		27 0.350 167	0.280	
21 0.261 148	0.200		1 0.348 126	0.260	
35 0.261 153	0.192		163 0.347 148	0.250	
30 0.260 167	0.189		39 0.347 17	0.167	
164 0.260 17	0.099 (*)		15 0.347 19 <	0.120	
155 0.259					
Sodium in precipitation					
Sample no.: G3			Sample no.: G4		
Theoretical value:	0.574		Theoretical value:	0.660	
Unit: µg/l			Unit: µg/l		
Run 1:					
Number of laboratories:	62		Number of laboratories:	63	
Arithmetic mean value:	0.571		Arithmetic mean value:	0.661	
Median:	0.579		Median:	0.660	
Standard deviation	0.059		Standard deviation	0.063	
Rel. st. deviation (%)	10.379		Rel. st. deviation (%)	9.557	
Run 2:					
Number of laboratories:	58		Number of laboratories:	59	
Arithmetic mean value:	0.570		Arithmetic mean value:	0.655	
Median:	0.579		Median:	0.660	
Standard deviation	0.037		Standard deviation	0.041	
Rel. st. deviation (%)	6.433		Rel. st. deviation (%)	6.293	
Results in decreasing order:					
112 0.840 (*) 115	0.579		24 0.900 (*) 161	0.660	
24 0.700 (*) 125	0.571		23 0.839 (*) 104	0.660	
2 0.645 151	0.570		112 0.800 (*) 151	0.660	
35 0.642 147	0.570		35 0.762 11	0.659	
22 0.631 109	0.570		22 0.756 163	0.658	
6 0.630 107	0.569		2 0.734 38	0.655	
3 0.617 26	0.568		121 0.700 26	0.654	
32 0.605 149	0.568		13 0.696 125	0.651	
5 0.604 155	0.565		31 0.696 155	0.651	
102 0.602 38	0.564		30 0.690 32	0.650	
30 0.600 161	0.560		166 0.690 15	0.650	
121 0.600 110	0.560		102 0.688 160	0.640	
4 0.594 11	0.560		115 0.685 107	0.638	
7 0.593 165	0.553		158 0.681 19	0.637	
36 0.592 16	0.550		6 0.680 124	0.630	
27 0.592 167	0.549		120 0.680 14	0.628	
120 0.590 148	0.540		165 0.680 152	0.622	
23 0.590 124	0.540		164 0.680 16	0.620	
13 0.590 152	0.536		110 0.680 147	0.620	
8 0.589 14	0.535		1 0.679 167	0.617	
20 0.588 31	0.533		8 0.679 153	0.611	
15 0.587 117	0.525		39 0.677 126	0.610	
158 0.586 153	0.516		36 0.675 148	0.600	
39 0.586 118	0.510		5 0.673 117	0.599	
1 0.585 10	0.502		7 0.673 118	0.590	
163 0.583 160	0.500		20 0.667 139	0.588	
164 0.580 33	0.492		21 0.667 10	0.588	
166 0.580 139	0.490		4 0.666 116	0.586	
12 0.580 116	0.482		149 0.662 33	0.563	
104 0.580 126	0.430 (*)		27 0.662 3	0.556	
21 0.579 17	0.401 (*)		12 0.660 17	0.465 (*)	
19 < 0.120			109 0.660		

Table 55: Analytical results for magnesium in precipitation samples.

Magnesium in precipitation			
Sample no.: G1			
Theoretical value:	0.085	Run 1:	Magnesium in precipitation
Unit: µg/l		Number of laboratories:	61
Run 1:			
Number of laboratories:	61	Arithmetic mean value:	0.081
Arithmetic mean value:	0.081	Median:	0.081
Median:	0.081	Standard deviation	0.015
Standard deviation	0.015	Rel. st. deviation (%)	18.720
Rel. st. deviation (%)	18.720	Run 1:	Magnesium in precipitation
Run 2:			
Number of laboratories:	57	Number of laboratories:	60
Arithmetic mean value:	0.080	Arithmetic mean value:	0.093
Median:	0.080	Median:	0.093
Standard deviation	0.009	Standard deviation	0.017
Rel. st. deviation (%)	11.577	Rel. st. deviation (%)	18.614
Rel. st. deviation (%)	11.577	Run 2:	Magnesium in precipitation
Results in decreasing order:			
126 0.140 (*) 30 0.080	126 0.150 (*) 116 0.093		
2 0.121 (*) 112 0.080	117 0.149 (*) 26 0.092		
24 0.120 (*) 161 0.080	2 0.132 (*) 166 0.092		
104 0.100 121 0.080	23 0.117 14 0.091		
35 < 0.1	35 < 0.100		
22 0.099 120 0.080	22 0.114 125 0.091		
167 0.094 109 0.080	24 0.110 11 0.091		
107 0.093 155 0.080	115 0.107 112 0.090		
163 0.090 166 0.080	12 0.103 155 0.090		
5 0.089 139 0.079	107 0.101 109 0.090		
8 0.089 23 0.079	161 0.100 152 0.090		
102 0.089 115 0.078	120 0.100 121 0.090		
12 0.088 1 0.078	30 0.100 104 0.090		
10 0.087 14 0.077	167 0.100 118 0.090		
7 0.086 3 0.077	102 0.099 3 0.089		
149 0.086 38 0.076	163 0.099 36 0.088		
151 0.085 125 0.076	19 0.097 139 0.088		
26 0.085 36 0.075	33 0.096 165 0.086		
116 0.085 148 0.074	10 0.096 20 0.085		
16 0.085 153 0.074	5 0.096 38 0.084		
21 0.084 117 0.074	7 0.096 4 0.082		
39 0.084 110 0.071	158 0.096 6 0.080		
27 0.084 6 0.070	149 0.095 124 0.080		
158 0.084 147 0.070	27 0.094 110 0.079		
19 0.084 124 0.070	8 0.094 148 0.078		
15 0.083 164 0.060	13 0.094 1 0.073		
11 0.083 160 0.060	39 0.094 147 0.070		
13 0.082 4 0.060	16 0.094 164 0.070		
118 0.082 31 0.059	21 0.093 160 0.070		
152 0.082 165 0.054	15 0.093 31 0.057 (*)		
33 0.082 17 0.031 (*)	151 0.093 17 0.036 (*)		
20 0.081			
Results in decreasing order:			
112 0.170 (*) 118 0.105	126 0.190 (*) 118 0.136		
2 0.150 (*) 11 0.105	22 0.188 (*) 11 0.134		
22 0.140 152 0.104	2 0.182 (*) 165 0.133		
126 0.140 23 0.104	112 0.180 (*) 1 0.133		
24 0.130 155 0.104	12 0.165 155 0.133		
3 0.126 20 0.103	107 0.153 33 0.132		
107 0.126 1 0.103	24 0.150 152 0.132		
30 0.120 13 0.103	23 0.150 161 0.130		
12 0.119 33 0.103	30 0.150 6 0.130		
102 0.115 139 0.102	167 0.145 104 0.130		
163 0.114 115 0.101	5 0.144 109 0.130		
5 0.113 125 0.101	163 0.144 110 0.130		
10 0.112 121 0.100	15 0.143 115 0.130		
8 0.111 161 0.100	102 0.143 139 0.130		
158 0.111 36 0.100	158 0.142 20 0.129		
149 0.111 6 0.100	10 0.141 153 0.127		
104 0.110 109 0.100	149 0.141 14 0.126		
120 0.110 14 0.099	120 0.140 38 0.125		
167 0.109 153 0.098	27 0.140 4 0.125		
35 0.109 38 0.097	19 0.140 148 0.124		
165 0.109 4 0.096	21 0.140 36 0.124		
7 0.109 117 0.096	7 0.140 116 0.122		
21 0.109 110 0.095	121 0.140 117 0.122		
166 0.108 148 0.093	8 0.138 147 0.120		
16 0.108 116 0.092	166 0.138 124 0.120		
19 0.108 164 0.090	26 0.138 160 0.120		
39 0.108 147 0.090	16 0.137 164 0.110		
27 0.108 124 0.090	125 0.137 3 0.109		
151 0.108 160 0.080	39 0.137 31 0.106		
15 0.107 31 0.067 (*)	35 < 0.100		
26 0.106 17 0.048 (*)	17 0.080 (*)		

Table 56: Analytical results for calcium in precipitation samples.

Calcium in precipitation			Calcium in precipitation		
Sample no.: G1	Sample no.: G2		Sample no.: G2	Theoretical value:	0.211
Theoretical value:	0.172	Unit: µg/l	Unit: µg/l		
Run 1:					
Number of laboratories:	62		Number of laboratories:	60	
Arithmetic mean value:	0.188		Arithmetic mean value:	0.223	
Median:	0.177		Median:	0.211	
Standard deviation	0.054		Standard deviation	0.071	
Rel. st. deviation (%)	28.701		Rel. st. deviation (%)	31.766	
Run 2:					
Number of laboratories:	57		Number of laboratories:	58	
Arithmetic mean value:	0.178		Arithmetic mean value:	0.212	
Median:	0.175		Median:	0.210	
Standard deviation	0.029		Standard deviation	0.033	
Rel. st. deviation (%)	16.142		Rel. st. deviation (%)	15.465	
Results in decreasing order:					
4 0.380 (*) 33 0.176			116 0.650 (*) 33 0.211		
116 0.379 (*) 2 0.175			117 0.438 (*) 118 0.210		
20 0.352 (*) 153 0.174			104 0.360 160 0.210		
165 0.309 (*) 13 0.173			23 0.294 147 0.210		
104 0.290 125 0.172			22 0.260 166 0.210		
148 0.240 8 0.172			165 0.248 6 0.210		
22 0.221 14 0.171			109 0.240 11 0.209		
126 0.220 36 0.171			163 0.240 7 0.208		
11 0.218 110 0.170			151 0.234 8 0.208		
102 0.213 118 0.170			2 0.231 115 0.207		
167 0.211 166 0.170			107 0.230 3 0.205		
160 0.210 120 0.170			121 0.230 158 0.204		
163 0.207 161 0.170			20 0.226 1 0.201		
151 0.202 7 0.168			102 0.224 4 0.200		
10 0.200 23 0.167			26 0.222 12 0.200		
107 0.196 158 0.165			21 0.221 24 0.200		
19 0.191 1 0.165			120 0.220 110 0.200		
147 0.190 38 0.162			30 0.220 17 0.197		
109 0.190 12 0.160			16 0.220 38 0.193		
152 0.190 6 0.160			112 0.220 155 0.192		
3 0.185 117 0.158			152 0.220 36 0.191		
115 0.185 155 0.156			10 0.220 15 0.190		
5 0.184 139 0.153			161 0.220 126 0.190		
26 0.184 124 0.150			5 0.219 139 0.188		
16 0.180 15 0.150			27 0.217 149 0.182		
112 0.180 17 0.144			167 0.216 148 0.180		
30 0.180 31 0.139			14 0.216 124 0.180		
121 0.180 164 0.130			39 0.214 164 0.170		
27 0.179 149 0.129			125 0.214 31 0.123		
21 0.179 24 0.100			13 0.211 32 0.120		
39 0.178 32 0.080 (*)					
Calcium in precipitation					
Sample no.: G3	Sample no.: G4		Sample no.: G4	Theoretical value:	0.287
Theoretical value:	0.249	Unit: µg/l	Unit: µg/l		
Run 1:					
Number of laboratories:	61		Number of laboratories:	62	
Arithmetic mean value:	0.250		Arithmetic mean value:	0.290	
Median:	0.244		Median:	0.287	
Standard deviation	0.056		Standard deviation	0.059	
Rel. st. deviation (%)	22.344		Rel. st. deviation (%)	20.212	
Run 2:					
Number of laboratories:	59		Number of laboratories:	58	
Arithmetic mean value:	0.247		Arithmetic mean value:	0.283	
Median:	0.244		Median:	0.285	
Standard deviation	0.036		Standard deviation	0.036	
Rel. st. deviation (%)	14.374		Rel. st. deviation (%)	12.864	
Results in decreasing order:					
112 0.550 (*) 7 0.244			112 0.560 (*) 26 0.287		
104 0.350 33 0.242			104 0.420 (*) 7 0.283		
165 0.341 158 0.242			160 0.410 (*) 1 0.283		
22 0.332 1 0.241			22 0.393 33 0.283		
107 0.304 115 0.241			116 0.381 27 0.281		
163 0.290 32 0.240			107 0.333 6 0.280		
116 0.282 110 0.240			165 0.331 118 0.280		
166 0.280 161 0.240			109 0.330 125 0.280		
109 0.280 12 0.240			167 0.330 11 0.277		
102 0.277 36 0.238			23 0.329 14 0.275		
121 0.270 20 0.236			32 0.320 20 0.275		
30 0.270 117 0.234			151 0.315 117 0.272		
147 0.270 11 0.233			163 0.310 110 0.270		
2 0.269 160 0.230			102 0.309 126 0.270		
3 0.269 4 0.230			2 0.306 38 0.265		
151 0.263 118 0.230			152 0.300 155 0.265		
6 0.260 155 0.229			120 0.300 8 0.263		
120 0.260 38 0.225			30 0.300 36 0.261		
152 0.260 167 0.225			158 0.295 15 0.260		
10 0.259 15 0.220			5 0.295 124 0.260		
21 0.258 124 0.220			39 0.292 139 0.253		
5 0.253 139 0.217			21 0.292 148 0.250		
39 0.252 148 0.210			10 0.292 147 0.250		
125 0.251 149 0.204			115 0.291 4 0.250		
13 0.250 24 0.200			161 0.290 164 0.240		
8 0.250 164 0.200			12 0.290 3 0.236		
27 0.248 126 0.200			121 0.290 149 0.231		
26 0.247 17 0.161			166 0.290 31 0.207		
16 0.246 31 0.143			13 0.289 24 0.200		
14 0.244 153 0.098 (*)			16 0.288 17 0.184		
23 0.244			19 0.287 153 0.127 (*)		

Table 57: Analytical results for potassium in precipitation samples.

Potassium in precipitation			
Sample no.: G1			
Theoretical value: 0.280			
Unit: µg/l			
Run 1:			
Number of laboratories:	62	Number of laboratories:	61
Arithmetic mean value:	0.287	Arithmetic mean value:	0.313
Median:	0.271	Median:	0.303
Standard deviation	0.143	Standard deviation	0.133
Rel. st. deviation (%)	49.939	Rel. st. deviation (%)	42.440
Run 2:			
Number of laboratories:	61	Number of laboratories:	59
Arithmetic mean value:	0.269	Arithmetic mean value:	0.303
Median:	0.270	Median:	0.303
Standard deviation	0.036	Standard deviation	0.044
Rel. st. deviation (%)	13.449	Rel. st. deviation (%)	14.602
Results in decreasing order:			
164	1.360 (*)	107	0.270
109	0.410	4	0.270
2	0.345	151	0.270
161	0.340	115	0.269
166	0.310	152	0.268
160	0.310	155	0.268
120	0.300	165	0.262
21	0.296	12	0.260
10	0.295	158	0.260
14	0.294	6	0.260
22	0.293	124	0.260
5	0.293	126	0.260
19	0.291	16	0.255
112	0.290	36	0.253
110	0.290	38	0.252
13	0.287	118	0.250
26	0.285	30	0.250
27	0.282	147	0.250
7	0.282	121	0.250
116	0.282	117	0.247
104	0.280	23	0.247
39	0.278	32	0.245
125	0.278	20	0.242
33	0.277	11	0.235
153	0.277	102	0.226
163	0.276	139	0.223
1	0.275	17	0.220
8	0.275	167	0.216
3	0.273	35	0.213
15	0.273	24	0.200
		149 <	0.200
31	0.271	148	0.150
Run 2:			
Number of laboratories:	59	Number of laboratories:	57
Arithmetic mean value:	0.269	Arithmetic mean value:	0.171
Median:	0.180	Median:	0.170
Standard deviation	0.027	Standard deviation	0.024
Rel. st. deviation (%)	14.735	Rel. st. deviation (%)	14.110
Results in decreasing order:			
109	0.400 (*)	6	0.180
149	0.264	164	0.180
2	0.259	12	0.180
161	0.240	120	0.180
31	0.234	4	0.180
11	0.230	158	0.179
112	0.210	36	0.177
110	0.210	3	0.176
13	0.204	32	0.175
10	0.204	8	0.175
115	0.200	117	0.174
21	0.200	20	0.174
125	0.199	30	0.170
5	0.199	118	0.170
27	0.197	126	0.170
26	0.195	152	0.166
7	0.194	153	0.166
14	0.192	116	0.164
155	0.192	38	0.163
39	0.192	167	0.161
163	0.189	124	0.160
15	0.187	104	0.160
165	0.185	139	0.154
107	0.184	35	0.153
1	0.184	102	0.152
23	0.183	166	0.150
16	0.183	147	0.140
22	0.182	121	0.130
33	0.181	148	0.110
151	0.180	24	0.100 (*)
160	0.180	17	0.092 (*)
Potassium in precipitation			
Sample no.: G2			
Theoretical value: 0.316			
Unit: µg/l			
Run 1:			
Number of laboratories:	61	Number of laboratories:	60
Arithmetic mean value:	0.313	Arithmetic mean value:	0.174
Median:	0.303	Median:	0.170
Standard deviation	0.133	Standard deviation	0.047
Rel. st. deviation (%)	42.440	Rel. st. deviation (%)	26.839
Run 2:			
Number of laboratories:	59	Number of laboratories:	57
Arithmetic mean value:	0.303	Arithmetic mean value:	0.171
Median:	0.303	Median:	0.170
Standard deviation	0.044	Standard deviation	0.024
Rel. st. deviation (%)	14.602	Rel. st. deviation (%)	14.110
Results in decreasing order:			
109	0.410 (*)	4	0.170
164	0.270 (*)	104	0.170
2	0.237	120	0.170
126	0.230	12	0.170
161	0.230	151	0.170
149 <	0.200	31	0.199
		22	0.195
		3	0.191
160	0.190	36	0.167
112	0.190	117	0.166
5	0.188	16	0.163
10	0.187	8	0.163
115	0.186	116	0.163
19	0.184	35	0.162
23	0.184	118	0.160
163	0.183	6	0.160
21	0.183	32	0.160
125	0.182	20	0.157
27	0.181	167	0.155
110	0.180	166	0.150
26	0.178	30	0.150
39	0.178	124	0.150
7	0.177	152	0.149
13	0.177	38	0.148
15	0.177	139	0.143
155	0.176	102	0.140
107	0.173	147	0.130
14	0.172	11	0.101
		121 <	0.100
33	0.172	24	0.100
		17 <	0.077
1	0.170	148	0.000 (*)

Table 58: Analytical results for conductivity in precipitation samples.

Conductivity in precipitation
 Sample no.: G1
 Theoretical value: 41.773
 Unit: $\mu\text{S}/\text{cm}$

Run 1:
 Number of laboratories: 57
 Arithmetic mean value: 40.562
 Median: 40.800
 Standard deviation 6.068
 Rel. st. deviation (%) 14.959

Run 2:
 Number of laboratories: 53
 Arithmetic mean value: 40.802
 Median: 40.800
 Standard deviation 2.240
 Rel. st. deviation (%) 5.490

Results in decreasing order:
 121 73.600 (*) 7 40.700
 13 47.500 1 40.500
 120 45.000 2 40.500
 155 44.000 140 40.300
 139 43.900 6 40.200
 30 43.200 151 40.100
 5 43.100 109 40.100
 24 42.750 27 40.100
 3 42.700 148 40.000
 20 42.500 112 39.700
 22 42.300 39 39.600
 12 42.300 165 39.500
 33 42.300 17 39.400
 14 42.200 166 39.400
 19 42.100 11 39.400
 16 42.000 126 39.300
 10 41.900 164 39.200
 21 41.790 32 39.020
 15 41.700 110 38.900
 163 41.640 107 38.700
 147 41.600 8 38.600
 4 41.500 118 38.000
 102 41.300 36 37.600
 23 41.200 152 35.100
 117 41.100 153 33.360
 149 41.070 116 26.680 (*)
 35 41.000 160 26.000 (*)
 158 40.800 38 23.200 (*)
 124 40.800

Conductivity in precipitation
 Sample no.: G3
 Theoretical value: 28.049
 Unit: $\mu\text{S}/\text{cm}$

Run 1:
 Number of laboratories: 58
 Arithmetic mean value: 28.098
 Median: 27.880
 Standard deviation 3.253
 Rel. st. deviation (%) 11.578

Run 2:
 Number of laboratories: 56
 Arithmetic mean value: 27.605
 Median: 27.830
 Standard deviation 1.688
 Rel. st. deviation (%) 6.114

Results in decreasing order:
 160 47.000 (*) 149 27.860
 148 36.800 (*) 109 27.800
 13 33.000 21 27.800
 155 30.800 140 27.600
 6 30.200 16 27.600
 3 29.800 117 27.600
 14 29.100 11 27.500
 22 29.000 158 27.500
 5 29.000 126 27.500
 120 29.000 2 27.400
 139 28.900 38 27.300
 147 28.900 107 26.900
 121 28.800 151 26.900
 30 28.800 110 26.900
 15 28.700 35 26.800
 24 28.530 152 26.700
 4 28.500 32 26.700
 20 28.470 39 26.500
 19 28.400 27 26.500
 124 28.300 17 26.400
 12 28.100 153 25.890
 1 28.100 164 25.800
 10 28.100 8 25.700
 23 28.100 36 25.500
 165 28.000 166 25.100
 163 28.000 118 25.000
 33 28.000 104 24.000
 7 28.000 116 23.540
 102 27.900 112 23.100

Conductivity in precipitation
 Sample no.: G2
 Theoretical value: 46.810
 Unit: $\mu\text{S}/\text{cm}$

Run 1:
 Number of laboratories: 56
 Arithmetic mean value: 44.743
 Median: 45.600
 Standard deviation 4.249
 Rel. st. deviation (%) 9.496

Run 2:
 Number of laboratories: 54
 Arithmetic mean value: 45.408
 Median: 45.600
 Standard deviation 2.087
 Rel. st. deviation (%) 4.597

Results in decreasing order:
 120 49.000 16 45.600
 155 48.900 124 45.500
 139 48.600 109 45.500
 13 48.500 23 45.400
 5 48.000 35 45.300
 30 47.900 140 45.300
 22 47.600 158 44.900
 24 47.400 151 44.800
 1 47.300 110 44.600
 6 47.100 27 44.600
 19 46.900 11 44.400
 10 46.900 166 44.400
 165 46.800 126 44.300
 3 46.800 118 44.000
 163 46.700 112 43.600
 14 46.600 164 43.600
 4 46.500 17 43.300
 102 46.500 32 43.130
 15 46.500 2 43.100
 12 46.400 8 42.800
 147 46.400 39 42.800
 20 46.350 121 42.800
 33 46.200 36 42.300
 21 46.120 160 41.000
 149 45.950 152 40.900
 7 45.900 117 39.100
 107 45.600 116 33.530 (*)
 148 45.600 38 20.000 (*)

Conductivity in precipitation
 Sample no.: G4
 Theoretical value: 33.321
 Unit: $\mu\text{S}/\text{cm}$

Run 1:
 Number of laboratories: 58
 Arithmetic mean value: 32.324
 Median: 32.820
 Standard deviation 2.327
 Rel. st. deviation (%) 7.200

Run 2:
 Number of laboratories: 54
 Arithmetic mean value: 32.598
 Median: 32.865
 Standard deviation 1.658
 Rel. st. deviation (%) 5.087

Results in decreasing order:
 148 37.100 (*) 2 32.800
 13 36.500 7 32.800
 155 36.300 166 32.700
 120 35.000 109 32.600
 6 34.700 158 32.500
 5 34.500 102 32.400
 139 34.200 126 32.400
 160 34.000 16 32.400
 3 34.000 27 32.200
 30 33.900 151 32.200
 24 33.790 11 32.200
 19 33.600 35 32.100
 10 33.600 107 31.800
 15 33.600 110 31.800
 14 33.600 153 31.160
 22 33.400 32 31.100
 1 33.400 118 31.000
 147 33.300 17 31.000
 124 33.300 164 30.700
 33 33.200 39 30.700
 163 33.200 140 30.600
 4 33.200 8 30.500
 121 33.200 36 29.700
 12 33.200 104 29.000
 23 33.000 152 28.500
 165 32.900 149 28.190
 117 32.900 116 27.120 (*)
 21 32.890 112 26.900 (*)
 20 32.840 38 23.400 (*)

Table 59: Ratio of the measured to the calculated conductivity in synthetic precipitation samples (G1-G4).

Lab.No.	Measured / calculated value				Remarks
	G1	G2	G3	G4	
1	0.90	0.92	1.12	1.15	
2	2.48	1.06	0.93	0.97	
3	1.09	0.95	1.03	1.06	
4	1.04	1.06	1.07	1.06	
5	1.07	1.06	1.06	1.06	
6	1.03	1.05	1.07	1.06	
7	0.87	0.86	0.84	0.82	
8	0.96	0.97	0.95	0.95	
10	0.82	0.88	0.87	1.01	
11	0.91	0.95	0.98	0.98	NH ₄ ⁺ missing
12	1.07	1.05	1.03	1.04	
13	1.04	0.86	1.06	0.99	
14	1.13	1.14	1.14	1.17	
15	1.08	1.07	1.08	1.08	
16	0.99	0.98	1.00	0.99	
17	1.14	1.09	1.10	-0.02	
19					Cl ⁻ , Ca ²⁺ , K ⁺ , and Na ⁺ values < LOD
20	1.05	1.05	1.04	1.01	
21	1.02	1.00	1.00	1.00	
22	1.17	1.20	1.11	1.11	
23	0.90	0.95	0.96	0.91	
24	1.04	1.04	1.04	1.03	
26					Cond missing
27	1.02	1.02	1.01	1.01	
30	1.12	0.98	1.12	1.05	
31					pH missing
32	1.00	0.99	1.00	0.97	
33	0.98	0.95	1.05	0.99	
35	1.08	1.06	1.03	1.05	
36	1.02	1.02	1.00	1.00	
38	0.91	0.50	1.93	1.04	
39	0.96	0.92	0.93	0.92	
102	0.91	0.90	0.91	0.92	
104					PH-values missing
107	0.99	1.03	1.01	0.99	
109	1.06	1.07	1.08	1.02	
110	1.02	0.98	0.99	0.94	
112	1.13	1.15	1.04	1.06	
115					pH and Cond. values missing
116	0.73	0.77	0.90	0.92	
117	1.17	1.32	1.12	1.08	
118	1.00	1.05	0.97	1.02	
120	1.10	1.05	1.00	1.03	
121	2.90	0.83	0.70	0.75	
124	1.05	1.04	1.09	1.10	
125					
126	1.11	1.11	1.08	0.98	
139					NO ³⁻ , Cl ⁻ and NH ₄ ⁺ -values missing
140					
147	0.85	0.93	0.91	0.95	
148	1.37	1.08	1.49	1.27	
149	1.05	1.03	1.03	0.90	
151	0.93	0.95	0.91	0.64	
152	0.91	0.90	1.00	0.90	
153	0.82		0.96	0.98	

Table 59, cont.

Lab.No.	Measured / calculated value				Remarks
	G1	G2	G3	G4	
155	1.15	1.20	1.19	1.23	
158	1.09	1.07	1.09	1.10	
160	0.93	0.92	1.01	0.97	
161					pH-values missing
163	1.05	1.03	1.05	1.04	
164	1.38	1.19	1.07	1.04	
165	0.95	0.97	1.01	1.02	
166	1.12	1.13	1.11	1.09	
167					pH-values missing

Table 60: Ratio of equivalent concentration of anions to the equivalent concentration of cation measured in synthetic precipitation samples.

Lab.No.	Measured value/Calculated				Average
	G1	G2	G3	G4	
1	0.95	0.94	1.06	1.08	1.01
2	2.26	1.22	0.99	1.05	1.38
3	1.09	0.98	0.99	1.15	1.05
4	1.00	1.08	1.05	1.07	1.05
5	1.03	1.03	1.00	1.00	1.01
6	1.07	1.04	0.95	0.99	1.02
7	1.01	1.00	1.00	0.99	1.00
8	1.11	1.13	1.08	1.09	1.10
10	1.40	1.52	1.51	1.04	1.37
11	1.04	1.06	1.26	1.26	1.16
12	1.05	1.03	1.02	1.03	1.03
13	1.55	1.46	1.22	1.35	1.39
14	1.23	1.25	1.15	1.20	1.21
15	1.07	1.05	1.03	1.05	1.05
16	1.01	1.01	1.01	1.01	1.01
17	1.06	1.02	1.08	0.00	0.79
19					
20	0.94	1.01	0.99	0.98	0.98
21	1.02	1.02	1.01	1.01	1.01
22	1.16	1.18	1.09	1.03	1.12
23	0.96	1.02	1.01	0.98	0.99
24	0.96	1.00	1.03	1.01	1.00
26	1.10	1.10	1.06	1.37	1.16
27	1.06	1.05	1.05	1.04	1.05
30	1.09	0.95	0.99	0.98	1.00
31					
32	1.13	1.15	1.06	1.04	1.10
33	0.89	0.88	0.98	0.93	0.92
35	1.28	1.28	1.13	1.24	1.23
36	1.08	1.08	1.03	1.04	1.06
38	1.59	1.17	1.54	1.34	1.41
39	1.02	0.99	0.99	0.93	0.98
102	0.87	0.88	0.95	0.92	0.91
104					
107	1.12	1.12	1.05	1.07	1.09
109	0.94	0.93	0.96	0.94	0.94
110	1.05	1.02	0.98	0.93	1.00
112	1.03	1.06	1.09	1.02	1.05
115					
116	1.03	0.92	1.02	1.05	1.01
117	1.02	1.33	1.01	1.01	1.09
118	1.06	0.99	1.08	1.05	1.05
120	0.96	0.98	0.94	0.95	0.96
121	3.58	1.68	1.96	2.17	2.35
124	1.07	1.12	1.06	0.99	1.06
125					
126	0.95	1.00	1.02	0.91	0.97
139					
140					
147	0.83	0.96	0.84	0.96	0.90
148	1.14	1.07	0.92	1.00	1.03
149	1.13	1.05	1.01	1.06	1.06
151	1.10	1.11	1.07	0.81	1.02
152	1.06	1.01	0.98	0.99	1.01

Table 60, cont.

Lab.No.	Measured value/Calculated				Average
	G1	G2	G3	G4	
153	1.17	0.00	1.22	1.19	0.90
155	0.92	0.91	1.02	0.99	0.96
158	1.11	1.10	1.05	1.05	1.08
160	1.91	1.38	0.83	1.10	1.30
161					
163	1.03	1.01	1.00	1.00	1.01
164	0.91	0.85	0.87	0.86	0.87
165	1.02	1.01	0.97	1.01	1.00
166	1.18	1.21	1.20	1.12	1.18

Table 61: The ratio of the median values to the theoretical values for all parameters and samples.

Parameter	Sample no.	Median/Expected
SO ₄ -S	G1	1.00
	G2	0.99
	G3	1.00
	G4	0.99
NO ₃ -N	G1	1.01
	G2	1.00
	G3	1.00
	G4	1.00
NH ₄ -N	G1	1.02
	G2	1.01
	G3	1.00
	G4	1.00
pH	G1	1.01
	G2	1.01
	G3	1.01
	G4	1.01
H	G1	0.91
	G2	0.93
	G3	0.91
	G4	0.91
Mg	G1	0.94
	G2	1.00
	G3	0.97
	G4	0.98
Na	G1	1.01
	G2	1.02
	G3	1.01
	G4	1.00
Cl	G1	0.99
	G2	1.03
	G3	0.97
	G4	0.97
Ca	G1	1.02
	G2	1.00
	G3	0.98
	G4	0.99
K	G1	0.96
	G2	0.96
	G3	0.93
	G4	0.95
Cond	G1	0.97
	G2	0.97
	G3	0.99
	G4	0.98

Table 62: Relative random and systematic errors obtained by the different laboratories in the analysis of each parameter in the precipitation samples.

Lab. no.	SO ₄ ²⁻		NO ₃ ⁻		NH ₄ ⁺		Mg ²⁺		H ⁺ calc	
	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %
1	0	0	1	1	1	3	6	-6	15	1
2	3	14	5	13	8	35	3	38	48	-24
3	1	2	0	-1	12	-11	18	-6	9	1
4	0	0	0	1	1	1	5	-13	3	-10
5	1	-1	0	1	1	-1	1	4	2	-5
6	0	-1	0	0	6	3	3	-10	5	-5
7	1	16	0	-1	1	-4	1	1	3	21
8	1	5	0	6	1	-4	2	2	3	-8
10	20	46	12	21	1	0	1	2	7	1
11	1	-3	2	6			1	-3	4	10
12	0	-2	2	6	3	1	9	10	3	-8
13	25	65	2	8	3	4	2	-3	8	0
14	2	15	1	-4	1	-2	4	-8	7	-23
15	0	-1	0	-2	3	0	2	-1	3	-11
16	1	-1	1	-1	4	-5	1	0	1	0
17	5	-19	0	-7	2	6	2	-55	5	-14
19	2	-2	1	-2	1	4	2	0	8	-28
20	1	-6	1	1	1	2	2	-6	4	-4
21	0	1	1	1	1	1	1	0	1	-3
22	1	6	5	11	16	2	13	25	10	-27
23	2	3	2	4	19	-13	12	3	8	6
24	1	2	1	0	6	0	9	18	2	-5
26	1	-2	1	-1	54	2	1	-1	5	-18
27	1	0	1	2	1	3	1	0	3	-10
30	1	2	3	-10	1	1	6	8	9	-6
31	4	7	2	5	4	5	6	-33		
32	2	-6	2	8	3	17			4	-7
33	2	-9	1	-4	1	0	4	-4	6	8
35	1	2	1	3	11	6	0	1	2	-11
36	0	-5	0	2	2	6	4	-9	5	-17
38	1	0	0	3	1	-5	2	-10	18	-53
39	1	2	5	-5	2	8	1	-1	2	0
102	4	-1	6	-3	5	1	1	5	5	13
104	46	199	2	3	4	-2	9	-1	6	-17
107	3	8	1	-1	5	-5	4	10	3	-10
109	3	-5	2	-6	2	-5	2	-6		
110	3	-1	0	0	7	6	2	-13	7	-4
112	8	-6	2	1	17	-2	26	18	5	-35
115	1	2	0	1	1	-46	9	-7		
116	1	0	1	-5	4	-3	7	-8	7	-17
117	5	-8	1	-2	22	1	28	-11	31	-17
118	6	-5	3	0	2	-3	0	-3	3	-12
120	1	-3	1	-3	10	-3	5	1	4	4
121	6	1	66	324	12	33	3	-4	48	-9
124	4	-1	2	-2	1	4	2	-16	3	-10
125	0	-4	2	-5	3	4	3	-5		
126	4	-11	1	-7	6	-1	10	50	10	-15
139	118	233					2	-6	3	-5
140	0	-4	1	-4	1	3			6	-23
147	4	-2	2	-1	5	27	3	-18	10	14
148	4	-22	1	3	4	-12	2	-14	17	-9
149	1	-1	1	1	2	5	1	2	3	-8
151	3	7	2	4	10	-21	1	0	32	5
152	0	-2	1	-6	4	-2	2	-4	5	-5
153	1	4	1	6	1	4	1	-10	1	-7
155	10	-16	1	-1	3	2	1	-4	3	-11
158	0	0	0	0	2	3	2	2	4	-18
160	8	21	1	-2	6	1	3	-23	62	-11
161	1	0	1	3	2	6	6	-6		
163	1	0	0	-1	1	-1	1	5	2	-7
164	4	-16	2	-13	7	31	4	-23	20	-29
165	1	1	1	1	9	-11	12	-6	4	-1
166	1	-2	1	3	2	-5	2	-1	7	-28
167	1	-2	55	218	1	1	3	6		

Table 62, cont.

Lab. no.	Na ⁺		Cl ⁻		Ca ²⁺		K ⁺		Cond.	
	Random error %	Systematic error %								
1	1	2	8	-9	1	-3	1	-4	2	0
2	2	13	5	20	3	8	2	27	3	-3
3	13	-3	1	0	13	1	5	-3	2	2
4	2	4	5	-1	43	-6	3	-5	1	-1
5	2	3	1	2	1	3	1	4	0	3
6	5	5	2	-6	4	-2	1	-7	4	2
7	1	3	1	-1	0	-2	1	0	1	-2
8	1	2	1	-1	4	-1	2	-5	2	-8
10	4	-13	93	170	4	4	1	5	0	0
11	7	0	1	-14	11	-3	19	-10	2	-5
12	1	1	1	1	3	-4	2	-6	1	0
13	2	4	3	5	0	0	2	1	4	11
14	3	-4	9	-28	3	-1	3	0	1	1
15	2	2	3	-5	1	-11	1	-3	1	0
16	3	-5	2	-3	2	2	6	-8	2	-2
17	4	-38	8	-10	16	-25	7	-35	2	-6
19	6	2	5	-9	3	4	1	3	0	1
20	2	0	2	-1	34	1	3	-8	1	0
21	0	1	1	-2	1	3	2	5	1	-1
22	12	20	18	2	10	29	5	0	1	2
23	32	21	30	18	16	8	7	-6	2	-1
24	20	44	5	7	14	-26	6	-36	1	1
26	1	-1	2	1	2	2	1	1		
27	1	2	1	1	2	1	0	1	1	-4
30	2	4	4	8	2	5	2	-12	1	2
31	7	-2	3	3	13	-37	12	2		
32	4	0	8	-4	22	-22	5	-11	3	-7
33	6	-13	5	-10	2	-1	2	-3	1	0
35	9	12	4	3			11	-22	1	-3
36	1	3	1	-5	4	-7	6	-9	2	-10
38	1	-2	2	11	2	-9	0	-12	28	-38
39	1	2	3	5	0	2	0	-1	3	-6
102	4	8	5	-3	5	11	4	-20	1	-1
104	3	2	5	23	9	54	6	-5	0	-11
107	2	-2	3	-8	6	15	1	-4	2	-4
109	1	0	1	-18	4	13	17	83	2	-3
110	3	0	3	-25	3	-4	2	5	2	-5
112	24	17	9	22	52	61	2	4	5	-11
115	7	4	5	12	4	0	9	3		
116	12	-4	11	-12	72	65	5	-5	12	-26
117	55	-7	64	-10	43	-6	4	-9	8	-1
118	6	-10	5	-11	3	-2	2	-9	2	-8
120	1	3	9	-15	3	4	6	-1	2	5
121	1	6	3	-45	3	6	14	-26	39	1
124	2	-5	26	-10	1	-12	3	-13	2	-1
125	2	0	4	-7	2	0	2	0		
126	12	-14	5	-12	17	-8	13	-5	2	-5
139	5	-12			3	-12	4	-20	1	4
140			4	-9					2	-4
147	3	-4	7	-4	10	4	4	-17	1	0
148	5	-13	34	-4	19	-15	8	-44	11	3
149	1	0	1	1	5	-19	5	23	5	-2
151	1	0	4	12	3	11	1	-4	1	-4
152	3	-5	15	1	1	5	3	-10	6	-14
153	1	-13	5	44	29	-66	4	-4	7	-6
155	1	-1	6	-5	1	-8	2	-3	1	7
158	1	3	2	-6	3	-3	3	-7	1	-2
160	5	-6	10	81	25	8	7	7	38	-7
161	9	2	10	22	3	0	2	22		
163	1	1	2	-7	3	14	2	0	0	0
164	1	2	7	-23	1	-20	185	208	1	-7
165	9	3	8	-9	17	29	6	-6	2	-1
166	3	4	10	37	6	0	12	-11	3	-6
167	4	-11	7	44	12	10	7	-20		

Appendix 4

Figures – 24rd intercomparison





Figure 32: SO_2 in absorbing solution.

Figure 33: SO₂ in impregnated filter.






Figure 34: HNO₃ in impregnated filter.

Figure 35: NO_2 in absorbing solution.

Figure 36: NH_3 in impregnated filter.




Figure 37: Percent deviation from theoretical value for sulphate.




Figure 38: Percent deviation from theoretical value for nitrate.




Figure 39: Percent deviation from theoretical value for ammonium.




Figure 40: Percent deviation from theoretical value for pH.




Figure 41: Percent deviation from theoretical value for chloride.




Figure 42: Percent deviation from theoretical value for sodium.




Figure 43: Percent deviation from theoretical value for magnesium.




Figure 44: Percent deviation from theoretical value for calcium.




Figure 45: Percent deviation from theoretical value for potassium.




Figure 46: Percent deviation from theoretical value for conductivity.




Figure 47: The median compared to theoretical value.





Figure 48: Youden plot of SO₂-S in absorbing solution.






Figure 50: Youden plot of HNO₃-N on impregnated filter.




Figure 51: Youden plot of $\text{NO}_2\text{-N}$ in absorbing solution.




Figure 52: Youden plot of NH₃-N on impregnated filter.




Figure 53: Youden plot of Ca in precipitation.




Figure 54: Youden plot of Cl in precipitation.




Figure 55: Youden plot of conductivity in precipitation.




Figure 56: Youden plot of K in precipitation.




Figure 57: Youden plot of Mg in precipitation.




Figure 58: Youden plot of Na in precipitation.




Figure 59: Youden plot of NH₄-N in precipitation.




Figure 60: Youden plot of strong acid calculated from pH.




Figure 61: Youden plot of $\text{NO}_3\text{-N}$ in precipitation.




Figure 62: Youden plot of $SO_4\text{-}S$ in precipitation.




Figure 63: Youden plot of pH in precipitation.