


The twenty-seventh and twenty-eighth intercomparison of analytical methods within EMEP

Hilde Th. Uggerud and Anne-Gunn Hjellbrekke



NILU : EMEP/CCC-Report 5/2011
REFERENCE : O-7726
DATE : AUGUST 2011

**EMEP Co-operative Programme for Monitoring and Evaluation
of the Long-range Transmission of Air Pollutants
in Europe**

**The twenty-seventh and twenty-eighth
intercomparison of
analytical methods within EMEP**

Hilde Th. Uggerud and Anne-Gunn Hjellbrekke



NILU – Norwegian Institute for Air Research
PO Box 100, NO-2027 Kjeller, Norway

Contents

	Page
1. Introduction.....	5
2. Organisation of the intercomparison	5
2.1 The twenty-seventh intercomparison	5
2.2 The twenty-eighth intercomparison.....	6
3. Data handling	6
3.1 Data analysis.....	6
3.2 Youden plot	6
3.3 Estimating random and systematic errors from laboratory comparisons.....	8
3.3.1 Estimating random errors	8
3.3.2 Estimating systematic errors.....	10
4. Results – 27th intercomparison.....	10
4.1 Sulphur dioxide in absorbing solution (A-samples).....	10
4.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples) ...	10
4.3 Nitrogen dioxide in absorbing solution (C-samples).....	11
4.4 Ammonia on impregnated filters (J-samples)	11
4.5 Precipitation (G-samples)	11
5. Results – 28th intercomparison.....	12
5.1 Sulphur dioxide in absorbing solution (A-samples).....	12
5.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples) ...	12
5.3 Nitrogen dioxide in absorbing solution (C-samples).....	13
5.4 Ammonia on impregnated filters (J-samples)	13
5.5 Precipitation (G-samples)	13
6. Summary.....	14
7. References	14
Appendix 1 Tables – 27th intercomparison	17
Appendix 2 Figures – 27th intercomparison	45
Appendix 3 Tables – 28th intercomparison	63
Appendix 4 Figures – 28th intercomparison	91

The twenty-seventh and twenty-eighth intercomparison of analytical methods within EMEP

1. Introduction

41 different laboratories in European countries are performing chemical analysis of air and precipitation samples within EMEP (Co-operative Programme for Monitoring and Evaluation of Long-range Transmission of Air Pollutants in Europe). Since the measurement programme is based on individual national networks, the participating laboratories apply different sampling and analytical methods. Most of the methods used are described in the manual for sampling and chemical analysis (EMEP/CCC, 1995).

In order to improve the data comparability and to get a picture of the different laboratories' performance, interlaboratory comparisons are organised by the Chemical Co-ordinating Centre (CCC) at the Norwegian Institute for Air Research (NILU). So far twenty-six intercomparisons have been arranged (Hanssen, 1988, 1990; Hanssen et al., 1983; Hanssen and Ladegård, 1984, 1985, 1987; Hanssen and Skjelmoen, 1992, 1994, 1995, 1996, 1997, 2001; Thrane, 1978, 1980a, 1980b, 1981; Uggerud et al., 2001, 2002, 2003, 2004; Hjellbrekke et al., 2005; Uggerud and Hjellbrekke, 2007, 2008, 2009).

Since 2000 the laboratory intercomparisons within EMEP have also been open for participation of laboratories from other networks.

This report gives the results of the twenty-seventh and twenty-eighth inter-laboratory tests.

2. Organisation of the intercomparison

2.1 The twenty-seventh intercomparison

The samples for the twenty-seventh intercomparison (see Table 2) were prepared and distributed to 84 laboratories in July 2009.

Most of the laboratories had returned their results to the CCC within one month after the deadline given as 15 November 2009. A total of 71 laboratories have returned their results. This includes 36 EMEP-laboratories.

Expected values were released on the CCC home page in December 2009. In accordance with the decision of the Steering Body of EMEP, the results are presented in such a way that the different laboratories are identified. Tables 3a and 3b give the names of the participating laboratories together with the numbers used when presenting the results in tables and figures.

Information received on the analytical methods used is given in Tables 4–9.

2.2 The twenty-eighth intercomparison

The samples for the twenty-eighth intercomparison (see Table 29) were prepared and distributed to 72 laboratories in July 2010.

Most of the laboratories had returned their results to the CCC within one month after the deadline given as 12 November 2010. A total of 61 laboratories have returned their results. This includes 33 EMEP-laboratories.

Expected values were released on the CCC home page in January 2011. In accordance with the decision of the Steering Body of EMEP, the results are presented in such a way that the different laboratories are identified. Tables 3a and 3b give the names of the participating laboratories together with the numbers used when presenting the results in tables and figures.

Information received on the analytical methods used is given in Tables 31–36.

3. Data handling

The data reported for the twenty-seventh intercomparison from the participants are presented in Tables 13–28 and Figures 2–17. An overview of all results is presented in Tables 10 and 11.

The data reported for the twenty-eighth intercomparison from the participants are presented in Tables 40–55 and Figures 18–33. An overview of all results is presented in Tables 37 and 38.

3.1 Data analysis

The reported values are presented in the tables in decreasing order together with the number of the laboratory. The expected (theoretical) value, the number of results, the arithmetic mean value, the median, the standard deviation and the relative standard deviation in percent are also given. After the first statistical run with all results included, the calculation was repeated with the outliers excluded. The outliers (unused) are defined as the results more than two standard deviations from the mean value in the first run.

Reported results expressed as percent deviation from expected value are presented in Table 10 and Table 11 for EMEP 27th and in Table 37 and Table 38 for EMEP 28th.

3.2 Youden plot

The Youden plot is a graphical method to analyse inter-laboratory data where the samples are ordered in pairs with similar concentrations. One plot is made for each pair of samples and gives results for all participating laboratories. The plots visualize both systematic and random errors.

The plot is drawn as a scatter plot where each point represents a pair of concentrations for one laboratory. The expected values for the two samples are drawn as solid blue lines. The arithmetic average of the measured values

excluding outliers is drawn as dotted lines. The solid lines divide the plot in four quadrants and a 45° reference line going through the intercept of the solid lines may be added.

If errors are due to random factors, the points will be evenly distributed around the mean value and situated in all four quadrants.

If systematic errors dominate, the results will be close to 45° reference line, and be situated in the upper right quadrant (overestimation) or lower left quadrant (underestimation).

Drawing a line from a given point perpendicular on the 45° reference line gives two line segments, one from the point to the intercept on the reference line (a), and one continuing from the intercept to the point representing the expected values (b). The lengths of these line segments are measures of the random and systematic errors respectively.

Ellipses with radii corresponding to the data quality objectives (DQO, Table 1) are added in each plot. The data points are colour coded depending on the magnitude of errors as given in Table 1.




Figure 1: Youden plot showing concentrations for a pair of samples (green), expected values (blue lines), average of measured values (dotted lines) and random and systematic errors (red lines)

In Figures 2–33 the reported data are presented in Youden plots.

Table 1: Youden plot parameters.


Radii	Components
10%	SO ₂ in abs.sol, NO ₂ in abs.sol.
20%	SO ₂ , HNO ₃ and NH ₃ in impregnated filter
Radii = DQO	Components
10% accuracy or better	SO ₄ ²⁻ , NO ₃ ⁻
15% accuracy or better	NH ₄ ⁺ , Cl ⁻ , Ca ²⁺ , K ⁺ , Mg ²⁺ , Na ⁺ , cond, H+ (from pH)
0.1 units	pH
Criteria	Colour
Within 0.5*DQO	Blue
Within DQO	Green
Within 2*DQO	Orange
> 2*DQO	Red

3.3 Estimating random and systematic errors from laboratory comparisons

Table 12 and Table 39 presents relative random and relative systematic errors obtained by the different laboratories in the analysis of each parameter in the precipitation samples. The calculation method and assumptions used are given in Chapter 3.3.1 and Chapter 3.3.2.

3.3.1 Estimating random errors

Systematic errors or bias in the laboratory analyses give a constant shift in the results from the expected ones at a particular concentration level. It is assumed that laboratories taking part in comparisons will obtain results near the expected ones when this bias is removed, and that the differences between expected and obtained results more often will be close to zero than not. A triangular distribution, based upon this assumption, can be used to quantify the random errors in the laboratory results (Eurachem/CITAC, 2000).



The triangle distribution is symmetric with a baseline 2a. The height in the triangle will be 1/a when the triangle area equals 1. The standard uncertainty is given by

$$u(x) = \frac{a}{\sqrt{6}} \quad (1)$$

and more than 95 % of the data will be within $\pm 2 \cdot u(x)$. The distance from -a to a (i.e. 2a) is called the range. When applied on the laboratory comparison results, the range equals the distance between the largest and smallest of the four differences between expected and found concentrations. As long as the bias can be assumed to be constant for the samples in the comparison of a specific component, it cannot have an effect on the distance corresponding to 2a. The bias may be dependent upon the concentrations, but can be considered approximate constant for the concentrations used here in the comparison of the main components in precipitation, since the differences between the concentrations are small.

L and T represent the laboratories' and the expected concentrations respectively, and D is the difference. The difference for the lowest concentration is

$$D_1 = L_1 - T_1 \quad (2)$$

and the differences are D_1, D_2, D_3, D_4 in increasing order.

The range is $D_4 - D_1$ and the standard uncertainty for the differences $u(D)$ becomes

$$u(D) = \frac{D_4 - D_1}{\sqrt{6}}. \quad (3)$$

The average expected concentration T for the four samples is given by

$$T = \frac{T_1 + T_2 + T_3 + T_4}{4} \quad (4)$$

The relative standard uncertainty, RSD, for 4 samples is given by $\frac{u(T)}{T}$, or

$$RSD = \frac{2 \cdot (D_4 - D_1) \cdot 100}{\sqrt{6} \cdot (T_1 + T_2 + T_3 + T_4)} \%, \quad (5)$$

and 95 per cent of the laboratory results in this comparison are expected to be within $\pm 2 \cdot RSD$.

If the data quality objectives (DQO) likewise are looked upon as 95 percentiles, then 95 per cent of the laboratory analytical results should not be more than 10 or 15 per cent from the correct values (10 per cent for S and N containing components and 15 per cent for other components).

Correspondingly, the values $2 \cdot RSD$ should therefore be less than 10 or 15 per cent in order to comply with the DQO.

3.3.2 Estimating systematic errors

An estimation of bias in single measurements requires a long data series, and four samples as we normally have in laboratory comparison, are merely able to give an indication of the bias or a very coarse estimate.

Coarse estimates have been performed here in the cases where the four samples had similar concentrations and where all four laboratory results were either higher or lower than the expected concentrations. The median of the differences D_i , as defined above, was taken as a measure of the bias, B , in these cases.

$$B = \text{median}[D_i] \quad (6)$$

A relative bias, RB , was also calculated based upon the average expected concentration T , as defined in (4).

$$RB = \frac{4 \cdot \text{median}[D_i]}{T_1 + T_2 + T_3 + T_4} \cdot 100\% \quad (7)$$

4. Results – 27th intercomparison

4.1 Sulphur dioxide in absorbing solution (A-samples)

Four samples and one blank solution were distributed to the laboratories that use the hydrogen peroxide absorption solution method. For those laboratories that reported a blank value this has been subtracted from the reported results.

7 laboratories have reported values for SO_2 in absorbing solution. 71% of the reported values are within 10% of expected value. One laboratory reports 3 values above 30% of expected value.

The results are presented in Table 13 and Figure 2.

4.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples)

Five impregnated filter samples (including one blank) for determination of sulphur dioxide were analysed by 21 laboratories. The value reported for the blank filter was subtracted from the other values before the data were used.

This year the amount of sulphur on the distributed filters was a factor of 10 higher than usual. The amount of sulphur on the distributed filters corresponds to air concentrations between 5.6–28.8 $\mu\text{g S m}^{-3}$ when 25 m^3 is sampled.

In addition to sulphate, nitrate was added to the same impregnated filters for determination of $\text{HNO}_3\text{-N}$. The value reported for the blank filter was subtracted from the other values before using the data.

The amount of nitrogen on the distributed filters corresponds to air concentrations between 0.52 $\mu\text{g N m}^{-3}$ –1.18 $\mu\text{g N m}^{-3}$ when 25 m^3 sampling volume is used.

80% of the reported results for sulphur dioxide and 75% of the reported results for nitric acid are within 10% of expected value.

The results are presented in Tables 14 and 15 and Figures 3 and 4.

4.3 Nitrogen dioxide in absorbing solution (C-samples)

The four samples distributed were made to represent both absorption solutions and extracts from iodide-impregnated glass filters. The samples contain known amounts of sodium nitrite diluted in water. In order to assure sample stability and to give the laboratories the opportunity to use the matrix they use in their daily routine, the distributed samples were to be diluted 1:10. The results should be reported as the diluted concentrations.

The 10 times diluted samples correspond to air concentrations between $2.60\text{-}5.15 \mu\text{g NO}_2\text{-N m}^{-3}$, when 70 ml absorbing solution and 1.4 m^3 are used. If 4 ml extraction solution and 0.7 m^3 sampling volume are used, the samples correspond to air concentrations between $0.30\text{-}0.59 \mu\text{g NO}_2\text{-N m}^{-3}$.

15 laboratories have reported results. 38% of the reported results are within 10% of expected value. Four laboratories report results that deviates more than 20% from expected value, all of which above the expected value.

The results are presented in Table 16 and Figure 5.

4.4 Ammonia on impregnated filters (J-samples)

Six impregnated filters inclusive two unidentified blank filters were sent to 22 laboratories. All laboratories have reported their analytical results. The two blank values reported by each laboratory were averaged and subtracted from the other values reported before the data were used. The results are shown in Table 17 and Figure 6.

The amount of nitrogen on the filters correspond to air concentrations between $0.40\text{-}1.20 \mu\text{g N m}^{-3}$, if 25 m^3 sampling volume is used.

31% of the reported values are within 10% of expected value. Four laboratories report values outside DQO. These values are distributed both above and below expected value.

4.5 Precipitation (G-samples)

Four precipitation samples were distributed and 2872 single results from 71 laboratories were reported. 79 results were identified as outliers. This is less than 3% of the data, which is a decrease compared to earlier rounds.

Results for sulphate and sodium are overall in good agreement with expected value. For both parameters, only 6 laboratories report results outside DQO. Standard deviation when outliers are excluded varies between 7-10%, which is comparable to earlier rounds.

The reported results for nitrate are mostly in good agreement with expected value. Only 5 laboratories report values outside DQO, which is an improvement compared to last year. In addition, the results for ammonium has improved compared to emep25 and emep26.

Determination of calcium and chloride seems to be troublesome for many laboratories. 35% of the laboratories report Ca-values outside DQO. Even so, this is an improvement compared to last year. More than 32 % of the laboratories report Cl-values outside DQO, which is an increase compared to earlier year's results.

The results are presented in Tables 18-28 and in Figures 7-17.

5. Results – 28th intercomparison

5.1 Sulphur dioxide in absorbing solution (A-samples)

Four samples and one blank solution were distributed to the laboratories that use the hydrogen peroxide absorption solution method. For those laboratories that reported a blank value this has been subtracted from the reported results.

8 laboratories have reported values for SO₂ in absorbing solution. 57% of the reported values are within 10% of expected value. One laboratory reports 3 values above 30% of expected value.

The results are presented in Table 40 and Figure 18.

5.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples)

Five impregnated filter samples (including one blank) for determination of sulphur dioxide were distributed to 26 laboratories. 17 laboratories reported results. The value reported for the blank filter was subtracted from the other values before the data were used.

The amount of sulphur on the distributed filters corresponds to air concentrations between 0.72-2.56 µg S m⁻³ when 25 m³ is sampled.

In addition to sulphate, nitrate was added to the same impregnated filters for determination of HNO₃-N. The value reported for the blank filter was subtracted from the other values before using the data.

The amount of nitrogen on the distributed filters corresponds to air concentrations between 0.59 µg N m⁻³–1.05 µg N m⁻³ when 25 m³ sampling volume is used.

84% of the reported results for sulphur dioxide are within 10% of expected value. This is an increase compared to last year. 68 % of the reported results for nitric acid are within 10% of expected value. This is a decrease compared to last year.

The results are presented in Tables 41 and 42 and Figures 19 and 20.

5.3 Nitrogen dioxide in absorbing solution (C-samples)

The four samples distributed were made to represent both absorption solutions and extracts from iodide-impregnated glass filters. The samples contain known amounts of sodium nitrite diluted in water. In order to assure sample stability and to give the laboratories the opportunity to use the matrix they use in their daily routine, the distributed samples were to be diluted 1:10. The results should be reported as the diluted concentrations.

The 10 times diluted samples correspond to air concentrations between 2.9-4.45 $\mu\text{g NO}_2\text{-N m}^{-3}$, when 70 ml absorbing solution and 1.4 m^3 are used. If 4 ml extraction solution and 0.7 m^3 sampling volume are used, the samples correspond to air concentrations between 0.33-0.62 $\mu\text{g NO}_2\text{-N m}^{-3}$.

C-samples were distributed to 24 laboratories. 16 laboratories have reported results. 80% of the reported results are within 10% of expected value. This is much better than last year's result. Two laboratories report results that deviates more than 20% from expected value. All of these deviating results are above expected value.

The results are presented in Table 43 and Figure 21.

5.4 Ammonia on impregnated filters (J-samples)

Six impregnated filters inclusive two unidentified blank filters were sent to 27 laboratories. 17 laboratories have reported their analytical results. The two blank values reported by each laboratory were averaged and subtracted from the other values reported before the data were used. The results are shown in Table 44 and Figure 22.

The amount of nitrogen on the filters correspond to air concentrations between 0.44-1.44 $\mu\text{g N m}^{-3}$, if 25 m^3 sampling volume is used.

62% of the reported values are within 10% of expected value. Four laboratories report values outside DQO.

5.5 Precipitation (G-samples)

Four precipitation samples were distributed and 2861 single results from 61 laboratories were reported. 120 results were identified as outliers. This is 4,2% of the data, which is an increase compared to last year, but comparable to earlier rounds.

Results for nitrate, sodium, potassium and ammonia are overall in good agreement with expected value. For all of these parameters, only 6 laboratories or less, report results outside DQO. Standard deviation when outliers are excluded varies mostly between 4-10%, which is comparable to earlier rounds.

The reported results for chloride show an improvement. 15% of the reported results are outside DQO, compared to 32% in emep27.

Determination of calcium seems to be troublesome for many laboratories. 28% of the laboratories report Ca-values outside DQO. Even so, this is an improvement compared to emep26 and emep27.

The results are presented in Tables 45-55 and in Figures 23-33.

6. Summary

A total of 71 laboratories participated in the twenty-seventh intercomparison. 36 of these laboratories are within the EMEP network. For the twenty-eight intercomparison 61 laboratories participated, out of which 33 are within EMEP.

As in earlier intercomparisons, outliers are defined as values that deviates more than two standard deviations from the mean value. Outliers occur for all samples and parameters. Out of a total of 3200 single results reported for emep27, 101 are defined as outliers. This is close to 3% of the reported data, which is very good.

For emep28, 139 out of a total of 3156 results are defined as outliers. This is 4.4% of the reported data, which is an increase compared to the two last rounds.

7. References

- Eurachem/CITAC (2000) Quantifying uncertainty in analytical measurements.
2nd ed. URL: <http://www.measurementuncertainty.org/mu/QUAM2000-1.pdf>.
- EMEP/CCC (1995) Manual for sampling and chemical analysis. Kjeller,
Norwegian Institute for Air Research (EMEP/CCC Report 1/95) (Last rev.
2002). URL: <http://tarantula.nilu.no/projects/ccc/manual/index.html>
[2011-08-30].
- Hanssen, J.E. (1988) The tenth intercomparison of analytical methods within
EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-
Report 3/88).
- Hanssen, J.E. (1990) The eleventh intercomparison of analytical methods within
EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-
Report 1/90).
- Hanssen, J.E., Ladegård, N.E. (1984) The seventh intercomparison of analytical
methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research
(EMEP/CCC-Report 1/84).
- Hanssen, J.E., Ladegård, N.E. (1985) The eighth intercomparison of analytical
methods within EMEP. Lillestrøm, Norwegian Institute for Air Research
(EMEP/CCC-Report 4/85).
- Hanssen, J.E., Ladegård, N.E. (1987) The ninth intercomparison of analytical
methods within EMEP. Lillestrøm, Norwegian Institute for Air Research
(EMEP/CCC-Report 4/87).

- Hanssen, J.E., Ladegård, N.E., Thrane, K.E. (1983) The sixth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/83).
- Hanssen, J.E., Skjelmoen, J.E. (1992) The twelfth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 7/92).
- Hanssen, J.E., Skjelmoen, J.E. (1994) The thirteenth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/94).
- Hanssen, J.E., Skjelmoen, J.E. (1995) The fourteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 3/95).
- Hanssen, J.E., Skjelmoen, J.E. (1996) The fifteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 2/96).
- Hanssen, J.E., Skjelmoen, J.E. (1997) The sixteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 2/97).
- Hanssen, J.E., Skjelmoen, J.E. (2001) The seventeenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 10/2001).
- Hjellbrekke, A.-G., Uggerud, H., Hanssen, J.E., Schaug, J. (2005) The twenty-second intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 8/2005).
- Thrane, K.E. (1978) Report on the first intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/78).
- Thrane, K.E. (1980a) Report on the second and third intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/80).
- Thrane, K.E. (1980b) Report on the fourth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 5/80).
- Thrane, K.E. (1981) The fifth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/81).
- Uggerud, H.T., Hanssen, J.E., Skjelmoen, J.E. (2001) The eighteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 11/2001).

Uggerud, H.T., Hanssen, J.E., Schaug, J., Skjelmoen, J.E. (2002) The nineteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 1/2002).

Uggerud, H.T., Hanssen, J.E., Schaug, J., Skjelmoen, J.E. (2003) The twentieth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 8/2003).

Uggerud, H.T., Hanssen, J.E., Schaug, J., Hjellbrekke, A.-G. (2004) The twenty-first intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 6/2004).

Uggerud, H., Hjellbrekke, A.-G. (2007) The twenty-third and twenty-fourth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 5/2007).

Uggerud, H., Hjellbrekke, A.-G. (2008) The twenty-fifth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 7/2008).

Uggerud, H., Hjellbrekke, A.-G. (2009) The twenty-sixth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 6/2009).

Appendix 1

Tables – 27th intercomparison

Table 2: Samples distributed for the twenty-seventh interlaboratory test.

A.	5 synthetic samples for determination of SO ₂ , consisting of 0.3% H ₂ O ₂ absorbing solution and containing different concentrations of sulphuric acid. One of the samples was an unidentified blank.
B.	5 KOH-impregnated Whatman 40 filters, comprising 1 blank and 4 filters to which different amounts of sulphuric acid and nitrate salt have been added.
C.	4 synthetic samples for determination of NO ₂ consisting of sodium nitrite diluted in water.
J.	6 Whatman 40 filters impregnated with 3% oxalic acid, comprising 2 blank and 4 filters to which different amounts of ammonium salt solution have been added.
G.	4 synthetic precipitation samples, containing SO ₄ ²⁻ , NO ₃ ⁻ , NH ₄ ⁺ , H ⁺ , Na ⁺ , Mg ²⁺ and Cl ⁻ , and Ca ²⁺ and K ⁺ .

Table 3a: EMEP laboratories participating in the twenty-seventh laboratory intercomparison. The numbers in front of the names are used in tables and figures.

Armenia	(45)	Dept of Environm. Health and Research
Austria	(1)	Umweltbundesamt Zweigstelle Sud, Klagenfurt
Belgia	(2)	Flemish Environment Agency, Antwerpen
Czech Republic	(3)	Czech Hydrometeorological Institute, Praha
Denmark	(4)	National Environmental Research Institute. Air Pollution Laboratory
Estonia	(38)	Estonian Environmental Research Centre, Tallinn
Finland	(5)	Finnish Meteorological Institute. Air Quality Department
France	(41)	Micro Pollutants Technology
Georgia	(43)	Centre for Monitoring and Prognostication, Tbilisi
Germany	(7)	IfE Leipzig GmbH, Umweltlabor
Germany	(8)	Umweltbundesamt, Messstelle Schauinsland
Hungary	(10)	Institute for Atmospheric Physics
Iceland	(11)	Innovation Center Iceland
Ireland	(12)	Met. Eirann, Dublin
Italy	(13)	C.N.R. Istituto Inquinamento Atmosferico
Italy	(30)	Joint Research Centre, Ispra
Kazakhstan	(44)	Hydrometeorological Monitoring, Astana
Latvia	(33)	Air Pollution Observation Laboratory
Lithuania	(32)	Environmental Physics and Chemistry Laboratory
Netherlands	(14)	National Institute of Public Health and Environmental Protection (RIVM)
Norway	(15)	Norwegian Institute for Air Research (NILU)
Macedonia	(40)	Hydrometeorological Institute, Skopje
Moldova	(42)	State Hydrometeorological Service, Chisinau
Poland	(16)	Institute of Meteorology and Water Management, Warsaw
Poland	(39)	Environmental Monitoring Laboratory, Institute of Environmental Protection
Poland	(46)	Diabla Gora station
Portugal	(17)	Laboratorio Santo Andre
Russian Federation	(22)	Institute of Global Climate and Ecology
Serbia	(24)	Rep. Hydrometeorological Institute of Serbia
Slovakia	(31)	Slovak Hydrometeorological Institute
Slovenia	(36)	Hydrometeorological Institute of Slovenia
Spain	(19)	Centro Nacional de Sanidad Ambiental
Sweden	(20)	Swedish Environmental Research Institute (IVL), Gothenburg
Switzerland	(21)	Swiss Federal Laboratories for Materials Testing (EMPA)
United Kingdom	(23)	AEA Technology, National Environmental Technology Centre
United Kingdom	(167)	CEH Edinburgh

Table 3b: Participating laboratories outside the EMEP network.

Canada	(26)	Meteorological Service of Canada, Toronto
United States of America	(27)	Illinois State Water Survey
Germany	(104)	Hessige Landwirtschaftliche
Germany	(108)	Institut f. Bondenkunde und Standortlehre, Dredsten
Germany	(109)	Bügen-Institute, dep. of Soil Science of Temporal and Boreal Ecosystems
Germany	(110)	Thüringer Landesanstalt für Landwirtschaft (TTL), Jena
Germany	(112)	Niedersächsische Forstliche Versuchsanstalt (NVF)
Germany	(113)	Landesforstanstalt Eberswalde, abt. Waldökologie
Italy	(114)	C.N.R. Istituto Italiano di Idrobiologia
Germany	(115)	Bayerische Landesanstalt f. Wald- und Forstwirtschaft
Switzerland	(116)	Institute for Applied Plant Biology
Germany	(118)	Forstliche Versuchs-und Forschungsanstalt
Germany	(120)	Landwirtschaftliche Untersuchungs- und Forschungsanstalt (LUFA)
Germany	(121)	Landeslabor Schleswig-Holstein
Belgium	(124)	Laboratorium voor Bondenkunde, Gent
Germany	(125)	Bayerisches Landesamt für Umweltschutz, Augsburg
Italy	(126)	APPA Laboratorio Biologico Provinciale
Italy	(140)	C.N.R. Istituto di Ricerca sulle Acque
Luxembourg	(146)	Cellule de Recherche en environment et Biotechnologies Public Research Center-Gabriel Lippman
Spain	(150)	Fundación Centro de Estudios ambientals del mediterrain
Norway	(152)	Norwegian Forest Research Institute, Ås
Slovenia	(153)	Slovenian Forestry Institute, Ljubljana
United Kingdom	(155)	Environmental Research Branch, Forest Research
Hungary	(157)	Ecological Laboratory of Forest research Institute
Japan	(158)	Acid Deposition and Oxidant Research Center (ADOCRC), Niigata
Ireland	(160)	Collite, Wicklow
Thailand	(163)	Environmental Researching and Training Center (ERTC)
Thailand	(164)	Pollution Control Department (PCD)
Viet Nam	(165)	Institute of Meteorology and Hydrology
Poland	(166)	Forest Research Institute, Laboratory of Forest Habitat Chemistry
France	(172)	UMR SAS INRA
Russian Federation	(175)	Primorsky Environmental Pollution Monitoring center
Russian Federation	(176)	Main Geophysical Observatory
Russian Federation	(179)	Murmansk Environmental Monitoring Centre
Germany	(180)	vTI-Institut für Agrarrelevante Klimaforschung

Table 4: Analytical methods used at the participating laboratories for the determination of sulphur dioxide in absorbing solution (A).

Method	Laboratory
1. Ion chromatography	15, 17, 19, 23, 36, 45
2. Spectrophotometry	173

Table 5: Analytical methods used at the participating laboratories for the determination of sulphur dioxide on impregnated filters (B).

Method	Laboratory
1. Ion chromatography	3, 4, 5, 8, 15, 22, 31 33, 36, 38, 41, 45, 116 158, 163, 165, 172
2. Capillary Electrophoresis	39
3. Spectrophotometry	16, 11

Table 6: Analytical methods used at the participating laboratories for determination of nitric acid on impregnated filters (B).

Method	Laboratory
1. Ion chromatography	3, 4, 5, 8, 15, 20, 31 33, 36, 41, 45, 116 158, 163, 165, 172
2. Capillary Electrophoresis	39
3. Spectrophotometry	16

Table 7: Analytical method for determination of ammonia on impregnated filters (J).

Method	Laboratory
1. Spectrophotometry	3, 4, 10, 15, 16, 19, 20, 32, 33, 39, 45, 116, 172
3. Ion chromatography	5, 8, 13, 31, 36, 38, 41, 158, 165

Table 8: Analytical method used for NO₂ in absorbing solution (C).

Method	Laboratory
1. Spectrophotometry	3, 8, 10, 15, 16, 20, 23, 31, 32, 33, 36, 39, 45, 173
2. Ion chromatography	19

Table 9: Analytical methods used for the determination of chemical constituents in precipitation samples.

Lab no	Network	SO ₄ ²⁺	NH ₄ ⁺	NO ₃ ⁻	Na ⁺	Mg ²⁺	Cl ⁻	Ca ²⁺	K ⁺
1	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
2	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
3	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
4	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AES	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AES
5	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
7	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
8	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
10	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
11	EMEP	ICP-AES	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Spectrophotometry	ICP-AES	ICP-AES
12	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
13	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
14	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-MS	ICP-MS	Ion chromatography	ICP-MS	ICP-MS
15	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
16	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
17	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
19	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
20	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography			Ion chromatography		
21	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
22	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
23	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
24	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
26	Other	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
27	Other	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
30	EMEP	Ion chromatography	Ion chromatography	Spectrophotometry	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
31	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
32	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS		Ion chromatography	Flame-AAS	Flame-AAS
33	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-MS	ICP-MS	Ion chromatography	Ca methods, ICP-MS	ICP-MS
36	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
38	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
39	EMEP	CE	Spectrophotometry	CE	Flame-AES	ICP-AES	CE	ICP-AES	Flame-AES
40	EMEP		Spectrophotometry	Spectrophotometry	Flame-AAS	Flame-AAS	Spectrophotometry	Flame-AAS	Flame-AAS
41	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
42	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
43	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Na methods, Flame photometry		Ion chromatography	titration	Flame photometry
44	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
45	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-MS	ICP-MS	Ion chromatography	ICP-MS	ICP-MS
46	EMEP								
104	ICP-Forest	Ion chromatography	CFA	CFA	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
108	EMEP	ICP-AES	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Potensiometric method	ICP-AES	ICP-AES
109	EMEP	ICP-AES	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Photometry Hg(SCN)2	ICP-AES	ICP-AES

Table 9, cont.

Lab no	Network	SO_4^{2-}	NH_4^+	NO_3^-	Na^+	Mg^{2+}	Cl^-	Ca^{2+}	K^+
110	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
112	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	Ion chromatography
113	Other			NO ₃ methods	ICP-AES			ICP-AES	
114	ICP-Forest	Ion chromatography	Spectrophotometry	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
115	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
116	Other	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AES	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
118	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
120	ICP-Forest	ICP-AES	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Spectrophotometry	ICP-AES	ICP-AES
121		Spectrophotometry	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Spectrophotometry	ICP-AES	
124	Other	Ion chromatography	Ion chromatography	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
125	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
126	ICP-Forest	Ion chromatography	Spectrophotometry	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
140	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
146	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
150	Other	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
152	ICP-Forest	Ion chromatography	SFA	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
153	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
155	ICP-Forest	Ion chromatography	Spectrophotometry, Colorimetry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
157	ICP-Forest	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AES	Flame-AES	Ion chromatography	Flame-AES	Flame-AES
158	EANET	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
160	EMEP	Ion chromatography	Colorimetry FIA	Colorimetry FIA	ICPMS	ICPMS	Ion chromatography	ICPMS	ICPMS
163	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
164	EMEP								
165	EANET	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
166	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
167	EMEP	Ion chromatography	FIA	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
172	Other	Ion chromatography	Spectrophotometry	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
175	WMO-GAW	Nephelometry	Spectrophotometry	Spectrophotometry	Flame-AAS	Flame-AAS	Titration	Flame-AAS	Flame-AAS
176	WMO-GAW	Ion chromatography	Ion chromatography	Ion chromatography					
179	WMO-GAW	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
180	Other	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography

Table 10: Reported results for precipitation samples expressed as % deviation from expected value.

Lab no	SO ₄ ²⁻				NH ₄ ⁺				NO ₃ ⁻				Na ⁺				Mg ²⁺				Cl ⁻				Ca ²⁺				K ⁺				pH				Cond							
	% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				pH-units from expected value				% deviation from expected value											
	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4																								
1	1	-1	0	0	7	4	2	2	2	0	0	0	1	0	1	-1	0	-2	-1	-1	-5	-8	-6	-6	-3	-9	-7	-7	-4	-5	-6	-4	0.09	0.09	0.06	0.08	-15	-11	-4	-10				
2	4	5	6	8	13	12	12	12	2	4	4	4	-8	-6	-4	-4	5	10	8	6	13	11	12	8	-15	-11	-4	-3	-2	-1	-1	-1	-0.19	-0.245	-0.27	-0.3	-3	-3	-1	-3				
3	2	1	2	1	22	15	6	9	1	0	0	0	0	0	-2	-4	-1	-1	-1	-1	-2	-2	0	-1	2	-6	-7	-5	7	9	12	13	0.009	0.021	0.002	0.009	1	2	2	0				
4	-2	-2	0	-2	-2	-4	-4	-2	-1	-1	-1	-1	-1	2	-5	1	-7	-6	-8	-8	1	0	-3	8	0	1	-8	-5	-7	0	-4	0.04	0.02	0.01	0.01	-4	-3	2	0					
5	0	-1	-1	-2	-1	-1	-2	-1	2	2	1	1	3	5	4	3	2	3	3	1	4	3	2	2	4	4	3	4	5	4	3	-1	0.03	0.03	0.02	0.02	-2	-2	0	-1				
7	-3	-1	-4	-2	8	3	0	2	-5	-4	-3	-1	6	9	4	7	5	5	2	3	21	18	11	6	56	11	18	15	3	7	2	1	0.09	0.09	0.08	0.09	-10	-6	0	-5				
8	-1	-3	-1	-2	2	6	7	7	-1	-2	2	-1	-2	0	12	-1	-2	-1	-2	-3	-3	-3	-3	-5	49	-1	-6	-7	-5	-3	-1	0.06	0.06	0.06	0.06	-11	-13	-9	-10					
10	6	6	9	7	-11	-12	-11	-9	2	5	3	4	-3	-6	-6	-6	2	6	4	3	-1	-3	-7	-2	13	4	9	9	6	3	6	3	0.08	0.08	0.05	0.08	0	-1	1	-1				
11	-7	-7	-1	-2	-23	-17	-12	-10	7	8	8	8	-10	-11	-8	-4	1	2	1	2	10	18	10	7	3	0	-1	16	-6	-18	7	-0.1	-0.13	-0.15	-0.12	-8	-7	-4	-4					
12	-4	-5	-2	-3	-10	-4	0	2	-4	-5	-5	-5	1	4	3	3	-4	-19	-12	-6	19	13	11	7	-2	-7	-5	-3	-5	-7	-6	-4	1	-1	1	-2								
13	2	1	-1	0	-2	-5	0	-2	-1	0	2	2	0	1	3	1	-1	1	1	4	-3	-7	-6	-5	-3	-1	-3	-2	0	-4	-1	-7	0.13	0.13	0.1	0.13	-7	-8	0	-7				
14	-3	-3	-2	-2	-10	-12	-13	-10	-7	-8	-7	-7	-5	-5	-3	-6	-9	-11	-7	-12	-29	-32	-28	-28	-2	-7	-5	-3	-3	2	-4	5	0.15	0.15	0.19	0.09	8	4	5	3				
15	-3	2	3	-1	12	7	8	9	-4	3	0	1	5	7	5	5	-3	-16	-3	-3	-17	-6	0	1	8	-7	1	1	4	3	6	3	0.05	0.05	0.04	0.05	1	-6	4	1				
16	0	0	1	0	2	4	-2	-2	-5	-4	-3	-3	5	-1	-6	-7	-6	-4	-10	-8	-2	-2	-2	-3	24	16	4	4	-15	-12	-19	-16	0.04	0.02	0	0.01	-6	-3	1	-1				
17	-1	-2	-3	-4	27	12	20	19	4	1	-1	-1	-47	-34	-28	-26	-23	-35	-26	-23	-38	-7	-9	-10	-8	1	10	8	28	-20	-25	-37	0.06	0.09	0.05	0.09	1	1	2	2				
19	-12	-12	-5	-6	-11	-1	-2	-2	-7	-6	-5	-5	0	-1	-3	-3	-1	1	1	1	-1	-2	-2	0	0	21	10	19	5	2	2	2	3	0.166	0.131	0.125	0.127	1	-1	2	1			
20	-8	-12	-2	-6	1	1	3	3	2	0	-1	-1	-2	-2	-2	-3	-1	-2	-2	-2	-2	-2	-2	-2	-2	-2	0.05	0.06	0.05	0.06	-2	-2	3	-1										
21	1	0	2	1	-5	1	1	2	0	0	0	0	1	0	-1	-1	0	-2	-2	-1	5	2	2	1	16	4	3	2	3	2	1	-1	0.02	0.03	0.02	0.03	-2	-3	-1	-3				
22	4	1	6	-4	11	15	9	4	-2	-3	1	1	-15	-7	-16	-13	-23	-27	-22	-26	-17	-21	-26	-18	40	4	4	-4	-25	-10	-18	-12	-16	0.07	0.09	0.09	0.15	-8	-8	-6	-8			
23	3	-4	-2	-3	4	4	7	4	7	6	6	6	-3	-1	5	0	-3	2	-3	0	31	-4	-3	-2	1	1	6	4	-4	-8	-3	-6	0	0	0	0	-15	-13	-7	-9				
24	-11	-8	-9	-8	-41	-34	-43	-20	-18	-15	-29	-9	-23	-45	10	8	7	13	52	48	28	12	351	250	179	153	-37	-41	-47	-41	0.13	0.28	0.22	0.28	-7	-6	0	-5						
26	-2	-3	-1	-2	-3	-3	-2	-2	3	2	0	-1	-2	0	-2	-3	-1	-1	-1	-2	1	0	-1	0	7	2	1	0	-1	-1	-1	-1	0.02	0.06	0.01	0.03								
27	-1	-2	-1	-2	0	-1	-1	0	-1	-2	-2	-3	3	3	3	2	3	1	2	2	-1	-1	-1	0	7	2	2	2	-1	0	-1	2	0.04	0.06	0.02	0.04	-5	-6	-2	-4				
30	3	1	1	-1	5	2	-1	1	-2	-3	-2	-2	7	10	7	6	2	1	4	2	-23	31	-2	2	2	5	1	0	-5	-5	-6	-9	-0.02	-0.04	-0.06	-0.01	-13	-13	-12	-15				
31	-4	-3	-3	-2	37	49	29	20	0	1	1	2	-5	-2	-3	-2	-8	-12	-12	-10	-5	-4	-4	-3	-13	-13	-13	-12	-12	-17	-11	-15	0.06	0.06	0.06	0.06	0	-1	2	0				
32	-7	-7	2	-5	12	4	4	4	1	-2	2	2	-2	-2	-3	-3	-3	-7	-1	-4	-8	-22	-7	6	1	-12	-2	-6	-4	0.05	0.03	0	0.02	-3	-3	-1	-4							
33	2	-1	2	-1	-4	-3	-7	-3	-4	-5	0	-3	-2	-2	-3	-3	-3	-16	-13	-11	9	-1	7	-2	-12	-14	-16	-12	-5	-7	-6	-12	0.11	0.09	0.09	0.11	-2	-3	3	-1				
36	-4	-5	-2	-3	4	4	4	4	-4	-4	-4	-4	-2	-2	-3	-3	-3	0	1	-3	-1	-1	1	2	2	-5	-3	-1	-3	-6	-7	-10	-14	0.06	0.06	0.09	0.12	-10	-12	-10	-13			
38	3	7	-1	0	-10	-9	-5	-5	-7	-11	-3	-3	-20	-18	-15	-16	-68	-76	-71	-76	-17	-20	-19	-16	-78	-77	-5	-7	-6	-12	0.27	0	0.03	-0.06	1	22	9	10						
39	10	14	9	1	-3	-6	-7	-10	1	1	1	5	4	10	17	6	20	-4	-1	-1	-2	-1	2	4	0	5	0	-1	0	1	5	-1	-12	-2	-3	-1	-2	-15	-5	-23				
40	-3	-6	-7	-10	1	1	5	4	10	17	6	20	-4	-1	-1	-2	-1	-3	0	2	-15	91	70	59	54	35	31	14	5	4	6	10	0.05	-0.02	0.08	0.12	-1	10	76	45				
41	8	9	6	10	11	3	3	0	-5	-3	-2	1	1	4	0	-1	1	8	3	0	-1	-5	-4	-2	-10	-24	-25	-15	-2	3	-8	-6	0.05	0.5	-0.04	0.04	-14	-20	155	17				
42	-19	-48	-20	-16	-3	-5	-4	-6	-39	-13	-23	-26	-2	-3	-6	-6	-1	-3	0	2	-15	91	70	59	54	35	31	14	5	4	6	10	0.05	-0.02	0.08	0.12	-1	10	76	45				
43	168	169	163	179	49	40	39	83	-10	-11	-4	0	-1	-1	-2	-2	-1	0	2	704	-10	-17	-1	-3	9	10	22	11	-18	-14	-13	-12	-30	-26	-35	-41	0.07	0.06	0.05	0.06	-7	-9	-5	-5
44	36	37	37	34	-13	-33	-28	-30	47	43	46	46	-15	-10	-9	-11	-37	-37	-66	-68	9	10</td																						

Table 10, cont.

Lab no	SO ₄ ²⁻				NH ₄ ⁺				NO ₃ ⁻				Na ⁺				Mg ²⁺				Cl ⁻				Ca ²⁺				K ⁺				pH				Cond				
	% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				pH-units from expected value				% deviation from expected value								
	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4					
109	-7	-7	0	-1	-14	-9	-8	0	0	0	-1	-23	-31	-33	-33	10	45	17	13	-19	27	7	23	15	-14	-17	-12	-12	0.16	0.09	0.05	0.12	-18	-9	3	-5					
110	4	4	22	-1	42	-9	-13	-12	13	0	18	5	5	1	-2	-2	1	2	2	-2	7	10	2	1	-1	0	3	0	-4	0.16	0.09	0.05	0.12	-5	-5	-4	-5				
112	2	1	3	1	-13	-9	-8	-6	3	2	0	0	0	1	0	0	-3	-4	-3	-3	1	19	14	11	11	7	6	3	-9	-8	-6	-4	0.13	0.11	0.1	0.11	0.13	0.11	0.1	0.11	
113	-4	-4	-2	-2	-3	-4	-5	-5	-4	0	0	-1	-2	1	0	1	10	-4	-3	-3	-7	-6	-7	-5	-2	14	12	61	-10	-12	-12	-4	0.4	0.25	0.15	0.14	-38	-34	-24	-25	
114	-1	-4	5	0	-22	-29	-13	-22	2	-2	-3	-5	3	-2	-1	-2	-9	-8	-9	-10	71	60	48	31	20	6	-2	-3	3	-11	-5	-7	0.01	-0.01	-0.01	2.26	-10	-10	-10	36	
115	-2	0	-1	0	37	33	31	32	-6	-5	-5	-3	0	-15	-7	5	0	-2	0	23	-3	1	-4	-5	8	29	1	49	5	-5	14	48	0.12	0.11	0.11	0.1	-7	0	4	3	
116	-2	0	2	-1	-5	-5	-6	-5	-16	0	0	-3	-13	-12	-15	-14	-5	-4	-7	-6	-12	-15	-15	-14	8	0	-5	-8	-14	-12	-18	-12	-0.01	0.01	-0.03	0	-7	-9	-7	-7	
118	-7	-6	-5	-7	18	-3	-5	1	-4	-2	-1	2	1	2	-1	1	-4	-4	-5	-6	-12	-10	-4	-16	0	-2	-7	-7	9	3	6	10	0.09	0.04	0.03	0.04	0	-3	0	0	
120	-9	-7	-3	-9	4	1	0	0	-1	-1	-1	0	1	4	3	1	3	-4	-3	-3	-33	-31	-31	-31	8	7	1	6	0	0.05	0.04	0.15	0.11	3	3	11	6				
121	0	6	-2	-2	-3	2	0	-1	-4	3	-3	-3	19	20	15	16	-10	-4	-3	-11	24	37	26	18	-2	0	6	1	-1	-4	-2	-12	-4	0.05	0.03	0.05	0.06	-10	-6	-5	-5
124	0	-1	-1	-1	5	1	5	6	-2	-4	-4	-4	24	15	11	7	-3	-7	-6	-7	68	-1	-3	1	8	10	3	4	93	4	14	0	-10	-7	-4	-3					
125	-5	-5	-3	-4	-8	-7	-6	-6	-1	-2	-3	-4	-6	-2	-3	-3	-10	-4	-13	-11	-7	-6	-4	-2	18	21	12	6	-5	-7	-6	-12	0.19	0.12	0.13	0.14	-29	-22	-12	-23	
140	-3	-4	-1	-1	-11	-8	-7	-5	-4	-5	-5	-6	-5	0	1	10	-4	-3	5	-7	-10	-7	-5	-31	-21	-11	-8	4	-2	6	3	0.13	0.06	0.01	0.11	-29	-22	-12	-23		
146	-3	-4	-3	-3	5	2	2	2	-5	-5	-4	-5	-2	-1	-1	-1	15	27	-1	7	-13	-14	-12	-11	214	71	9	20	-9	-9	-14	-19	0.18	0.1	0.05	0.08	-4	-3	1	-5	
150	1	1	2	0	10	4	1	1	-4	-2	-3	-3	8	-3	-2	-3	-10	-5	-4	-4	2	15	2	3	-25	-28	-21	-24	-1	-13	-6	-12	0.64	0.49	0.42	0.48	-15	-8	-11	-10	
152	-14	-15	-13	-15	-5	7	-2	4	-20	-19	-17	-17	-3	-4	-3	-4	-6	-8	-8	-8	4	-6	-26	-34	21	14	10	5	-4	-5	-6	-8	0.08	0.06	0.06	0.06	-8	-8	-1	-4	
153	-13	-17	-12	-30	-3	2	-5	-5	-23	-27	-28	-47	-6	-2	0	1	10	8	26	21	-7	-15	-15	-37	57	29	17	15	9	-7	24	47	0.11	0.11	0.11	0.12	-4	-2	0	-3	
155	-7	-8	-7	-7	-1	-10	16	2	-8	-7	-4	-2	2	3	2	1	4	2	3	3	-17	-16	-16	-13	29	13	8	7	10	2	-1	0	0.04	0.04	0.04	0.05	-1	-1	1	-1	
157	-18	-13	-11	-14	4	2	-5	-5	-4	0	-5	-3	-2	17	77	41	-10	-4	-13	-11	45	-6	-4	1	155	114	54	34	131	76	18	54	0.24	0.06	-0.02	-0.04	-14	-6	-5	-5	
158	2	1	0	1	11	9	9	8	4	3	4	6	2	4	6	7	3	2	6	7	-6	-6	-5	-6	-5	-1	4	6	0	0	1	1	0.07	0.06	0.09	0.1	-6	-8	-5	-9	
160	-4	-4	-5	-6	-3	2	0	-5	3	0	2	-1	22	20	15	16	23	358	337	344	-71	312	221	179	140	86	13	65	69	-2	-2	-3	-5	0.12	0.12	0.11	0.1	-2	-2	-1	-3
163	0	0	0	0	2	2	1	2	-2	-1	-1	-1	3	5	4	3	6	6	6	8	6	-7	-6	-5	-4	13	16	19	15	-2	-2	-3	-5	0.12	0.12	0.11	0.1	-2	-2	-1	-3
164	-12	-10	-9	-10	4	2	4	5	-16	-11	-7	-7	-9	-5	-5	-5	-23	-28	-22	-19	-33	-25	-19	-19	-31	-29	-27	-26	-14	-12	-12	-12	0.04	0.02	0.02	0	-5	-4	-10	-12	
165	2	2	2	0	0	2	-2	3	3	1	0	0	-55	-48	-37	-31	-21	7	-2	6	-48	-42	-33	-24	154	101	25	18	8	-11	-24	18	0.05	0.02	-0.05	-0.02	-8	-8	3	-3	
166	2	2	1	1	0	2	5	6	5	2	0	0	-2	4	0	2	3	5	4	2	-1	-1	-3	-2	4	4	3	3	0.04	0.01	0.04	0.05	3	3	6	4					
167	-1	3	9	8	2	-4	-2	-5	-3	-2	-2	-2	-47	-34	-27	-24	3	-4	-2	-1	66	52	31	29	-14	-22	0	-11	-36	-25	-19	-51	-3	-2	0	-1					
172	1	3	2	3	32	21	19	16	3	4	1	1	-10	-9	-8	-6	-45	-50	-49	-4	-12	-7	-7	-15	-31	-22	-24	-6	-9	-14	-18	-0.23	-0.24	-0.31	-0.27	-3	-2	0	-1		
175	1	-4	4	8	28	-4	-16	-10	0	-2	1	-1	-13	-5	-3	-3	20	19	20	21	9	-1	-48	-34	90	33	40	22	18	13	176	12	0.07	0.05	0.05	0.04	-7	-3	-5	-5	
176	177	176	176	171	-10	23	12	19	308	304	300	298	-27	-24	-23	-22	71	-10	-11	-8	-9	-15	-19	-20	1	0	-1	-1	0	-3	-4	-5	0.08	0.03	0.03	0.03	-3	-3	0	-5	
179	-6	-2	0	11	0	-1	-1	-2	-1	97732	-1	0	-18	-14	-9	-2	26	18	17	13	-9	-15	-19	-20	1	0	-1	-1	0	-3	-4	-5	0.08	0.03	0.03	0.03	-3	-3	0	-5	
180	1	1	2	0	-1	-3	-1	-1	3	3	2	1	3	0	2	1	5	0	4	2	-5	-5	-3	-4	25	1	11	5	2	8	11	13	-3	-2	0	-1					




Table 11: Reported results for filter samples expressed as % deviation from expected value.

Air and aerosols																						
	Absorbing solution				Impregnated filter					Impregnated filter					Absorbing solution				Impregnated filter			
	SO ₂ -S		% deviation from expected value		SO ₂ -S		% deviation from expected value		HNO ₃ -N			% deviation from expected value		NO ₂ -N		% deviation from expected value		NH ₃ -N		% deviation from expected value		
	A1	A3	A4	A5	B1	B2	B4	B5	B1	B2	B4	B5	B1	C1	C2	C3	C4	J1	J2	J3	J6	
3					-10	-13	-5	-6	-1	-9	-1	-3	-17	1	-8	-8	4	-4	-5	0		
4					-2	-2	1	1	-12	-14	-4	-7		1	-4	-5	-3					
5					2	2	2	2	1	1	2	4		-4	-4	-6	-7					
8					2	4	6	2	2	-5	5	2	-2	0	1	1	-3	2	0	-2		
10													-17	-10	-9	-9	-16	-17	-20	-17		
11					-2	0	1	1					-8	-4	-5	-3		-1	-4	-4	1	
12																						
13																						
15	17	15	11	10	-20	10	-10	-4	4	7	-1	2	15	7	9	7	6	5	1	14		
16					-1	-1	3	1	-2	-5	-2	-3	-2	-2	-1	-1	-3	-7	-11	-2		
17	4	1	1	1																		
19	7	-3	-10	-6																		
20									95	-95	-95	-95	208	207	218	219	3	1	5	-1		
22					8	9	10	8														
23	-5	-3	-1	0					-8				-26	-15	-15	-12						
31									-4				-6	-3	-8	-6	-2	-13	-16	-5		
32													2	8	5	8	8	5	-1	2		
33					1	8	5	2	-1	1	1	-2					-1	-1	-1	-13		
36	2	3	7	3	-5	-4	-2	-3	0	2	-2	0	-8	-4	-6	-6	0	-5	-5	-1		
38					-53	-78	-52	-40									5	2	3	5		
39					0	-3	-7	-5	-12	-20	-26	-31	-12	-13	-13	-6	-6	-10	-10	-6		
41					2	1	2	-4	-2	1	-8	-7					-28	-22	-23	-22		
45	3	3	5	4	32	20	15	21	40	21	19	21	60	70	67	71	-8	-11	17	29		
116					-3	6	0	-22	-21	-21	-7	-22					38	23	22	26		
158					-2	-1	-1	-1	4	2	7	4					-12	-19	-21	-5		
163					0	1	3	3	0	-6	3	2										
164					-3	-5	4	4	-5	-12	-1	-2										
165					13	2	22	2	-4	-4	-9	-8					8	9	4	-7		
172					1	0	-3	2	-3	3	-1	-1					13	15	6	8		
173	42	45	39	0									-4	-4	-5	-3						

 Results between 10 and 20% or between -10 and -20% from expected value
 more than $\pm 20\%$ from expected value

Table 12: Relative random and systematic errors obtained by the different laboratories in the analysis of each parameter in the precipitation samples.

Lab. no.	SO ₄ ²⁻		NO ³⁻		NH ₄ ⁺		Mg ²⁺		H ⁺ calc	
	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %
1	1	0	1	0	1	3	1	-1	7	-17
2	1	6	1	4	3	12	0	7	22	71
3	0	1	1	0	3	13	0	-1	3	-1
4	1	-2	0	-1	1	-3	2	-8	4	-3
5	1	-1	0	1	1	-1	1	2	3	-5
7	1	-2	1	-3	2	3	2	3	6	-18
8	1	-2	2	-1	4	6	1	-2	4	-12
10	1	7	1	4	2	-12	1	3	7	-16
11	3	-5	2	8	0	-14	1	2	11	30
12	2	-4	1	-5	4	-2	4	-9		
13	1	0	1	1	2	-2	2	1	10	-25
14	1	-3	1	-8	3	-12	3	-10	11	-29
15	2	1	2	0	2	8	4	-4	4	-10
16	0	0	0	-4	2	-1	2	-8	4	-3
17	1	-2	2	0	6	20	3	-25	8	-14
19	5	-10	0	-5	4	-1	1	0	10	-27
20	5	-8	1	0	2	2			5	-11
21	1	1	0	0	2	1	1	-1	3	-5
22	3	3	2	0	3	9	5	-24	6	-19
23	3	-2	1	6	2	5	2	-2	6	-3
24	3	-8	2	-15	9	-29	3	9	17	-33
26	1	-2	1	1	0	-2	1	-1	7	-5
27	1	-1	1	-2	0	0	1	2	6	-8
30	2	1	1	-2	2	1	1	3	5	8
31	1	-3	1	1	8	30	1	-10	4	-12
32	4	-6	2	1	2	5			5	-6
33	1	0	2	-3	2	-4	3	-11	7	-21
36	2	-3	1	-4	1	4	1	0	3	-16
38	4	2	3	-4	1	-7	15	-72	28	-2
39	7	9	1	6	5	-3	0	-1		
40			2	3	5	-7	2	-3	19	58
41	3	9	2	-2	3	3	2	2	42	-9
42	19	-19	8	-27	2	-4	2	0		
43	36	176	4	-6	34	39			10	-11
44	7	36	9	45	14	-30	19	-54	5	-13
45	1	2	1	1	88	94	2	-15	16	40
46										
104	2	-6	2	-11	2	0	2	-4	8	-17
108	2	-2	2	-13	1	-3	1	-12	31	-50
109	4	-5	1	0			8	14	13	-21
110	7	5	8	8	17	-11	1	2		
112	0	2	1	1	0	-9	1	-3	8	-23
113							81	-8	23	-40
114	1	-3	1	-1	2	-4	7	-3	9	-24
115	3	0	3	-3	7	-20	2	-9	30	0
116	1	-1	1	-5	8	32	11	0	7	-22
118	1	-1	5	-2	2	-6	2	-6	3	2
120	2	-7	3	-1	7	0	1	-5	8	-9
121	3	-8	0	-1	1	0	3	-3	2	-14
124	4	-1	3	-3	1	-1	4	-7	3	-9
125	0	-1	2	-4	3	4	2	-5		
126	2	-4	2	-2	1	-7	4	-12	11	-26
140	2	-2	2	-5	0	-7	6	1	13	-17
146	1	-3	1	-5	1	2	8	13	15	-20
150	0	1	1	-3	2	2	7	-3	24	-71
152	4	-15	2	-18	4	2	1	-8	5	-13
153	7	-18	17	-28	3	-4	7	18	7	-22
155	2	-7	2	-5	9	-5	1	3	3	-9
157	6	-14	2	-3	4	-2	4	-12	25	-7
158	1	1	2	4	2	9	2	4	3	-16
160	0	-4	1	1	3	-1	1	1	277	
163	0	0	0	-1	1	1	1	7	8	-22
164	3	-10	2	-9	2	4	4	-20	4	-4
165	1	2	1	0	3	1	14	2	9	-1
166	1	2	2	1	3	4	1	3	3	-6
167	3	5	0	-2	3	-3	2	-1		
172	1	2	1	2	2	21			18	78
175	5	2	1	0	14	-9	5	20	5	-11
176	35	178	52	301	13	17			7	-7
179	7	-1	0	1	1	-1	9	14		
180	0	1	1	2	1	-1	3	3		

Table 12, cont.

Lab. no.	Na+		Cl-		K+		Cond.	
	Random error %	Systematic error %						
1	1	0	2	-7	1	-5	6	-11
2	0	-5	1	11	1	-1	1	-3
3	2	-1	1	-1	1	10	1	1
4	3	0			3	-5	2	-2
5	1	4	0	3	3	3	1	-1
7	2	6	3	13	3	3	4	-6
8	6	-1	2	-3	3	-4	4	-11
10	2	-6	3	-3	2	5	1	-1
11	2	-8	3	10	15	-1	2	-6
12	1	3	2	11	2	-6	1	0
13	1	1	2	-6	2	-3	4	-7
14	2	-4	7	-28	3	0	3	4
15	1	6	6	-2	1	4	4	1
16	5	-4	1	-2	2	-15	3	-2
17	3	-31	10	-12	5	-25	0	2
19	2	-2			0	2	1	1
20			1	-1			2	-2
21	1	-1	1	2	2	2	1	-2
22	5	-14	6	-21	4	-12	3	-8
23	3	0	11	-3	2	-5	5	-12
24	33	-16	10	34	6	-44	3	-6
26	1	-1	1	0	0	-1		
27	0	3	0	-1	1	-1	2	-5
30	1	7	17	0	0	-6	3	-14
31	1	-3	0	-4	4	-13	1	0
32	1	-2	4	-5	5	-4	1	-3
33	1	-2	4	3	1	-7	2	-1
36	1	-2	1	2	1	-9	3	-11
38	2	-16	4	-18	1	-7	11	8
39	2	-5	2	1	2	0		
40	8	11	14	75	2	-4	20	-11
41	2	1	2	-3	4	-4	57	-1
42	2	-4	37	75	1	6		
43			224	-5			24	26
44	3	-10	7	12	3	-32	3	-6
45	3	-8	1	-1	4	-18	1	-1
46							4	-2
104	2	11	3	23	4	19	14	24
108	2	-2	40	75	6	-12	5	-10
109	10	-31			4	-14	9	-8
110	3	-1			3	0	1	-5
112	0	0	6	15	3	-7		
113			25	-94			10	-32
114	1	0	1	-6	4	-11	4	-1
115	2	-1	5	50	6	-5	18	-9
116	8	-4	3	-3	17	9	5	1
118	3	-13	4	-14	3	-15	3	-7
120	1	1	8	-9	3	7	2	0
121	1	2	7	-30			2	5
124	2	17	5	26	6	-7	3	-6
125	4	12	22	1	46	9		
126	1	-4	1	-4	1	-7	4	-6
140	2	-2	1	-7	3	4	10	-24
146	0	-1	2	-12	2	-12	2	-4
150	4	-2	5	3	6	-7	4	-9
152	1	-3	20	-16	0	-6	4	-6
153	2	-1	18	-14	17	16	2	-3
155	0	2	2	-15	5	1	1	-1
157	34	34	16	-1	58	63	5	-6
158	3	5	2	-5	1	1	3	-7
160	2	17			36	56		
163	1	4	0	-5	1	-2	1	-2
164	1	-6	3	-23	3	-12	2	-7
165	1	-41	2	-35	14	-1	5	-6
166	2	1	1	-2			0	3
167	3	-30	7	41	11	-33		
172	0	-8	3	-9	2	-11	1	-2
175	3	-4	23	-23	63	18	2	-4
176	3	-23	27	-10	3	-48	2	-4
179	4	-11	8	-16	1	-4		
180	1	1	1	-4	3	9		

Table 13: Analytical results for sulphur dioxide in absorbing solution.

SO ₂ -S in absorbing solution Sample no.: A1 Theoretical value: 0.120 Unit: µg S/ml	SO ₂ -S in absorbing solution Sample no.: A2 Theoretical value: 0.200 Unit: µg S/ml
Run 1:	Run 1:
Number of laboratories: 7 Arithmetic mean value: 0.132 Median: 0.125 Standard deviation 0.019 Rel. st. deviation (%) 14.196	Number of laboratories: 7 Arithmetic mean value: 0.218 Median: 0.207 Standard deviation 0.034 Rel. st. deviation (%) 15.679
Run 2:	Run 2:
Number of laboratories: 6 Arithmetic mean value: 0.125 Median: 0.124 Standard deviation 0.009 Rel. st. deviation (%) 6.950	Number of laboratories: 6 Arithmetic mean value: 0.206 Median: 0.204 Standard deviation 0.013 Rel. st. deviation (%) 6.447
Results in decreasing order:	Results in decreasing order:
173 0.170 (*) 45 0.123 15 0.140 36 0.122 19 0.128 23 0.114 17 0.125	173 0.290 (*) 17 0.201 15 0.230 23 0.195 36 0.207 19 0.194 45 0.207
SO ₂ -S in absorbing solution Sample no.: A4 Theoretical value: 0.280 Unit: µg S/ml	SO ₂ -S in absorbing solution Sample no.: A5 Theoretical value: 0.400 Unit: µg S/ml
Run 1:	Run 1:
Number of laboratories: 7 Arithmetic mean value: 0.301 Median: 0.296 Standard deviation 0.044 Rel. st. deviation (%) 14.538	Number of laboratories: 7 Arithmetic mean value: 0.408 Median: 0.405 Standard deviation 0.019 Rel. st. deviation (%) 4.749
Run 2:	Run 2:
Number of laboratories: 6 Arithmetic mean value: 0.286 Median: 0.289 Standard deviation 0.021 Rel. st. deviation (%) 7.283	Number of laboratories: 7 Arithmetic mean value: 0.408 Median: 0.405 Standard deviation 0.019 Rel. st. deviation (%) 4.749
Results in decreasing order:	Results in decreasing order:
173 0.390 (*) 17 0.282 15 0.310 23 0.277 36 0.299 19 0.251 45 0.296	15 0.440 23 0.401 45 0.418 173 0.400 36 0.414 19 0.377 17 0.405

Table 14: Analytical results for sulphur dioxide on impregnated filter.

SO₂-S on impregnated filter
 Sample no.: B1
 Theoretical value: 200.400
 Unit: ug S/filter

Run 1:

Number of laboratories: 21
 Arithmetic mean value: 195.901
 Median: 198.800
 Standard deviation 30.308
 Rel. st. deviation (%) 15.471

Run 2:

Number of laboratories: 19
 Arithmetic mean value: 197.625
 Median: 198.800
 Standard deviation 13.609
 Rel. st. deviation (%) 6.886

Results in decreasing order:

45	264.507	(*)	158	196.010
165	227.000		11	195.951
22	217.280		4	195.670
41	205.130		164	195.310
5	204.620		116	194.290
8	204.000		36	190.710
172	201.690		31	183.736
33	201.466		3	180.520
163	200.810		15	161.190
39	200.700		38	94.540 (*)
16	198.800			

SO₂-S on impregnated filter
 Sample no.: B4
 Theoretical value: 560.800
 Unit: ug S/filter

Run 1:

Number of laboratories: 20
 Arithmetic mean value: 559.015
 Median: 568.990
 Standard deviation 78.637
 Rel. st. deviation (%) 14.067

Run 2:

Number of laboratories: 19
 Arithmetic mean value: 574.156
 Median: 571.280
 Standard deviation 41.081
 Rel. st. deviation (%) 7.155

Results in decreasing order:

165	684.000		4	566.700
45	644.857		11	564.251
22	615.960		116	561.890
8	595.000		158	554.890
33	586.249		36	549.310
164	581.330		172	543.810
16	576.800		3	534.920
163	575.810		39	522.700
5	572.210		15	506.990
41	571.280		38	271.340 (*)

SO₂-S on impregnated filter
 Sample no.: B3
 Theoretical value: 140.300
 Unit: ug S/filter

Run 1:

Number of laboratories: 20
 Arithmetic mean value: 137.129
 Median: 141.065
 Standard deviation 26.853
 Rel. st. deviation (%) 19.582

Run 2:

Number of laboratories: 19
 Arithmetic mean value: 142.744
 Median: 141.680
 Standard deviation 9.771
 Rel. st. deviation (%) 6.845

Results in decreasing order:

45	168.601		11	140.451
15	153.990		172	139.670
22	153.320		16	139.000
33	151.230		158	138.740
116	148.090		4	137.750
8	146.000		39	135.700
5	143.120		36	134.510
165	143.000		164	133.260
163	141.810		3	122.220
41	141.680		38	30.440 (*)

SO₂-S on impregnated filter
 Sample no.: B5
 Theoretical value: 721.000
 Unit: ug S/filter

Run 1:

Number of laboratories: 20
 Arithmetic mean value: 706.965
 Median: 728.226
 Standard deviation 85.947
 Rel. st. deviation (%) 12.157

Run 2:

Number of laboratories: 19
 Arithmetic mean value: 721.550
 Median: 730.551
 Standard deviation 57.500
 Rel. st. deviation (%) 7.969

Results in decreasing order:

45	871.469		4	725.900
22	780.350		16	725.700
164	748.130		158	716.110
163	740.810		36	697.610
165	734.000		41	694.780
8	734.000		15	691.990
5	733.960		39	682.700
33	733.145		3	674.220
172	732.840		116	561.190
11	730.551		38	429.840 (*)

Table 15: Analytical results for nitric acid on impregnated filter.

HNO ₃ -N on impregnated filter Sample no.: B1 Theoretical value: 16.360 Unit: ug N/filter	HNO ₃ -N on impregnated filter Sample no.: B3 Theoretical value: 13.090 Unit: ug N/filter
Run 1:	Run 1:
Number of laboratories: 19 Arithmetic mean value: 15.408 Median: 16.100 Standard deviation 4.024 Rel. st. deviation (%) 26.118	Number of laboratories: 18 Arithmetic mean value: 11.983 Median: 12.550 Standard deviation 3.110 Rel. st. deviation (%) 25.953
Run 2:	Run 2:
Number of laboratories: 18 Arithmetic mean value: 16.221 Median: 16.130 Standard deviation 1.964 Rel. st. deviation (%) 12.109	Number of laboratories: 17 Arithmetic mean value: 12.651 Median: 12.600 Standard deviation 1.324 Rel. st. deviation (%) 10.464
Results in decreasing order:	Results in decreasing order:
45 22.914 41 16.092 15 17.090 172 15.860 158 17.080 31 15.780 8 16.700 165 15.700 5 16.460 164 15.490 163 16.359 4 14.430 36 16.290 39 14.400 33 16.206 116 12.860 3 16.160 20 0.778 (*) 16 16.100	45 15.783 8 12.500 15 13.990 16 12.470 172 13.460 163 12.359 36 13.390 3 11.860 158 13.300 164 11.530 5 13.270 4 11.270 33 13.226 39 10.500 41 13.192 116 10.360 165 12.600 20 0.634 (*)
HNO ₃ -N on impregnated filter Sample no.: B4 Theoretical value: 24.540 Unit: ug N/filter	HNO ₃ -N on impregnated filter Sample no.: B5 Theoretical value: 29.450 Unit: ug N/filter
Run 1:	Run 1:
Number of laboratories: 18 Arithmetic mean value: 22.892 Median: 24.210 Standard deviation 5.828 Rel. st. deviation (%) 25.457	Number of laboratories: 18 Arithmetic mean value: 27.063 Median: 28.878 Standard deviation 7.121 Rel. st. deviation (%) 26.313
Run 2:	Run 2:
Number of laboratories: 17 Arithmetic mean value: 24.167 Median: 24.220 Standard deviation 2.242 Rel. st. deviation (%) 9.277	Number of laboratories: 17 Arithmetic mean value: 28.568 Median: 28.966 Standard deviation 3.246 Rel. st. deviation (%) 11.361
Results in decreasing order:	Results in decreasing order:
45 29.305 3 24.200 158 26.230 36 24.040 8 25.700 16 24.000 163 25.359 4 23.440 5 25.120 116 22.760 33 24.669 41 22.652 172 24.350 165 22.400 15 24.290 39 18.100 164 24.220 20 1.228 (*)	45 35.533 164 28.790 5 30.650 16 28.550 158 30.530 3 28.450 163 30.159 4 27.510 8 30.100 41 27.302 15 29.990 165 27.000 36 29.550 116 22.960 172 29.220 39 20.400 33 28.966 20 1.470 (*)

Table 16: Analytical results for nitrogen dioxide in absorbing solution.

NO₂-N in absorbing solution
 Sample no.: C1
 Theoretical value: 0.052
 Unit: ug N/ml

Run 1:

Number of laboratories:	16
Arithmetical mean value:	0.060
Median:	0.051
Standard deviation	0.029
Rel. st. deviation (%)	49.031

Run 2:

Number of laboratories:	15
Arithmetical mean value:	0.053
Median:	0.050
Standard deviation	0.012
Rel. st. deviation (%)	22.716

Results in decreasing order:

20	0.160 (*)	173	0.050
45	0.083	31	0.049
19	0.076	12	0.048
15	0.060	36	0.048
33	0.053	39	0.046
32	0.053	10	0.043
8	0.051	3	0.043
16	0.051	23	0.039

NO₂-N in absorbing solution
 Sample no.: C3
 Theoretical value: 0.079
 Unit: ug N/ml

Run 1:

Number of laboratories:	16
Arithmetical mean value:	0.093
Median:	0.076
Standard deviation	0.046
Rel. st. deviation (%)	48.970

Run 2:

Number of laboratories:	15
Arithmetical mean value:	0.083
Median:	0.075
Standard deviation	0.018
Rel. st. deviation (%)	22.253

Results in decreasing order:

20	0.251 (*)	12	0.075
45	0.132	173	0.075
19	0.119	36	0.074
33	0.086	31	0.073
15	0.086	3	0.073
32	0.083	10	0.071
8	0.080	39	0.069
16	0.078	23	0.067

NO₂-N in absorbing solution
 Sample no.: C2
 Theoretical value: 0.103
 Unit: ug N/ml

Run 1:

Number of laboratories:	16
Arithmetical mean value:	0.123
Median:	0.102
Standard deviation	0.059
Rel. st. deviation (%)	48.217

Run 2:

Number of laboratories:	15
Arithmetical mean value:	0.109
Median:	0.101
Standard deviation	0.024
Rel. st. deviation (%)	22.236

Results in decreasing order:

20	0.326 (*)	16	0.101
45	0.175	31	0.100
19	0.157	12	0.099
32	0.111	36	0.099
15	0.110	173	0.099
33	0.105	10	0.094
3	0.104	39	0.090
8	0.103	23	0.088

NO₂-N in absorbing solution
 Sample no.: C4
 Theoretical value: 0.086
 Unit: ug N/ml

Run 1:

Number of laboratories:	16
Arithmetical mean value:	0.103
Median:	0.084
Standard deviation	0.050
Rel. st. deviation (%)	48.302

Run 2:

Number of laboratories:	15
Arithmetical mean value:	0.091
Median:	0.083
Standard deviation	0.020
Rel. st. deviation (%)	22.095

Results in decreasing order:

20	0.274 (*)	12	0.083
45	0.147	173	0.083
19	0.130	36	0.081
33	0.095	39	0.081
32	0.093	31	0.081
15	0.092	3	0.079
8	0.087	10	0.078
16	0.085	23	0.076

Table 17: Analytical results for ammonia on impregnated filter. The reported results are corrected for an average blank value (J2 and J4).

NH3-N on impregnated filter Sample no.: J1 Theoretical value: 16.040 Unit: ug N/filter	NH3-N on impregnated filter Sample no.: J3 Theoretical value: 24.060 Unit: ug N/filter
Run 1:	Run 1:
Number of laboratories: 22 Arithmetic mean value: 16.111 Median: 15.964 Standard deviation 1.990 Rel. st. deviation (%) 12.353	Number of laboratories: 22 Arithmetic mean value: 23.292 Median: 23.135 Standard deviation 2.417 Rel. st. deviation (%) 10.377
Run 2:	Run 2:
Number of laboratories: 20 Arithmetic mean value: 16.041 Median: 15.964 Standard deviation 1.167 Rel. st. deviation (%) 7.274	Number of laboratories: 21 Arithmetic mean value: 22.997 Median: 23.120 Standard deviation 2.032 Rel. st. deviation (%) 8.834
Results in decreasing order:	Results in decreasing order:
116 22.110 (*) 33 15.908 172 18.100 13 15.800 19 17.731 31 15.639 32 17.250 8 15.550 165 17.250 16 15.480 15 17.027 5 15.460 38 16.795 39 15.065 3 16.720 45 14.697 20 16.460 158 14.160 4 16.160 10 13.545 36 16.020 41 11.505 (*)	116 29.480 (*) 4 23.120 165 26.150 5 23.110 15 25.328 13 23.000 32 25.310 36 22.960 172 25.300 16 22.340 38 24.575 39 21.665 8 24.450 45 21.500 20 24.380 31 21.000 33 23.902 10 20.037 19 23.589 158 19.390 3 23.150 41 18.686
NH3-N on impregnated filter Sample no.: J5 Theoretical value: 30.080 Unit: ug N/filter	NH3-N on impregnated filter Sample no.: J6 Theoretical value: 10.030 Unit: ug N/filter
Run 1:	Run 1:
Number of laboratories: 22 Arithmetic mean value: 29.204 Median: 29.280 Standard deviation 3.343 Rel. st. deviation (%) 11.447	Number of laboratories: 22 Arithmetic mean value: 10.010 Median: 9.865 Standard deviation 1.193 Rel. st. deviation (%) 11.915
Run 2:	Run 2:
Number of laboratories: 21 Arithmetic mean value: 28.840 Median: 28.800 Standard deviation 2.946 Rel. st. deviation (%) 10.215	Number of laboratories: 20 Arithmetic mean value: 9.730 Median: 9.840 Standard deviation 0.812 Rel. st. deviation (%) 8.346
Results in decreasing order:	Results in decreasing order:
116 36.840 (*) 13 28.800 45 35.106 3 28.720 172 32.000 4 28.690 20 31.460 36 28.690 165 31.350 5 28.380 38 31.085 39 27.065 19 30.677 16 26.890 15 30.477 31 25.360 8 30.050 10 24.190 33 29.767 158 23.900 32 29.760 41 23.224	45 12.985 (*) 8 9.850 116 12.640 (*) 16 9.830 15 11.428 4 9.720 172 10.800 158 9.520 38 10.545 31 9.511 32 10.250 39 9.455 19 10.215 5 9.360 13 10.100 165 9.350 3 10.020 33 8.693 36 9.890 10 8.325 20 9.880 41 7.848

Table 18: Analytical results for sulphate in precipitations samples.

Sulphate in precipitation				Sulphate in precipitation			
Sample no.: G1	Theoretical value:	1.141	Sample no.: G2	Theoretical value:	1.183	Unit: µg/l	Unit: µg/l
Run 1:							
Number of laboratories:	68		Number of laboratories:	68			
Arithmetic mean value:	1.179		Arithmetic mean value:	1.218			
Median:	1.130		Median:	1.166			
Standard deviation	0.349		Standard deviation	0.367			
Rel. st. deviation (%)	29.569		Rel. st. deviation (%)	30.179			
Run 2:							
Number of laboratories:	66		Number of laboratories:	66			
Arithmetic mean value:	1.121		Arithmetic mean value:	1.157			
Median:	1.130		Median:	1.161			
Standard deviation	0.085		Standard deviation	0.110			
Rel. st. deviation (%)	7.551		Rel. st. deviation (%)	9.467			
Results in decreasing order:							
176	3.160 (*)	115	1.130	176	3.260 (*)	17	1.162
43	3.059 (*)	167	1.130	43	3.180 (*)	4	1.161
44	1.557	116	1.123	44	1.617	27	1.159
39	1.254	118	1.120	39	1.352	179	1.159
41	1.232	4	1.116	41	1.288	108	1.150
10	1.210	26	1.113	38	1.260	8	1.150
2	1.191	15	1.110	10	1.259	14	1.150
22	1.190	14	1.110	124	1.250	26	1.149
110	1.190	146	1.110	2	1.241	31	1.147
30	1.177	140	1.110	110	1.230	146	1.141
45	1.176	7	1.105	167	1.214	115	1.140
23	1.173	108	1.100	172	1.213	114	1.140
38	1.170	160	1.100	45	1.210	140	1.140
166	1.166	114	1.100	15	1.210	160	1.140
165	1.162	36	1.099	166	1.209	175	1.140
3	1.161	31	1.098	165	1.205	23	1.133
158	1.161	12	1.090	112	1.200	36	1.126
33	1.160	126	1.085	22	1.195	12	1.120
112	1.160	179	1.077	3	1.193	126	1.118
13	1.160	104	1.070	158	1.192	120	1.110
172	1.158	11	1.060	30	1.192	104	1.110
21	1.153	109	1.060	180	1.192	11	1.100
1	1.150	120	1.060	150	1.190	109	1.100
150	1.150	155	1.060	13	1.190	32	1.100
175	1.150	32	1.057	21	1.184	121	1.100
180	1.148	20	1.045	16	1.182	24	1.087
5	1.146	121	1.040	118	1.180	155	1.083
163	1.140	24	1.019	116	1.180	164	1.070
124	1.140	19	1.004	163	1.180	20	1.042
125	1.138	164	1.000	7	1.173	19	1.040
16	1.136	153	0.990	5	1.171	157	1.030
27	1.131	152	0.980	33	1.170	152	1.003
17	1.131	157	0.930	1	1.170	153	0.980
8	1.130	42	0.924	125	1.170	42	0.610
Sulphate in precipitation							
Sample no.: G3			Sample no.: G4				
Theoretical value:	0.707		Theoretical value:	0.818			
Unit: µg/l			Unit: µg/l				
Run 1:				Run 1:			
Number of laboratories:	68		Number of laboratories:	68			
Arithmetic mean value:	0.742		Arithmetic mean value:	0.848			
Median:	0.704		Median:	0.810			
Standard deviation	0.211		Standard deviation	0.254			
Rel. st. deviation (%)	28.399		Rel. st. deviation (%)	29.903			
Run 2:				Run 2:			
Number of laboratories:	66		Number of laboratories:	66			
Arithmetic mean value:	0.706		Arithmetic mean value:	0.806			
Median:	0.703		Median:	0.810			
Standard deviation	0.052		Standard deviation	0.062			
Rel. st. deviation (%)	7.417		Rel. st. deviation (%)	7.690			
Results in decreasing order:							
176	1.950 (*)	5	0.703	43	2.283 (*)	33	0.810
43	1.860 (*)	13	0.703	176	2.220 (*)	108	0.810
44	0.968	125	0.702	44	1.093	7	0.805
110	0.863	140	0.700	179	0.911	26	0.804
167	0.771	38	0.700	41	0.903	4	0.803
39	0.770	26	0.700	167	0.887	31	0.802
10	0.769	27	0.700	175	0.883	27	0.802
22	0.752	8	0.699	2	0.880	5	0.800
2	0.750	11	0.699	10	0.875	114	0.800
41	0.748	116	0.699	172	0.845	8	0.800
115	0.740	36	0.696	45	0.836	11	0.800
175	0.733	20	0.695	112	0.830	14	0.800
112	0.730	12	0.690	166	0.828	124	0.800
15	0.730	114	0.690	3	0.828	36	0.797
45	0.722	124	0.690	39	0.828	23	0.795
165	0.721	14	0.690	21	0.826	146	0.790
118	0.720	23	0.690	158	0.823	12	0.790
33	0.720	126	0.688	150	0.820	17	0.788
172	0.720	31	0.687	115	0.820	126	0.788
32	0.720	146	0.684	38	0.820	22	0.785
21	0.720	121	0.684	165	0.819	104	0.781
150	0.720	17	0.684	180	0.819	32	0.780
180	0.719	7	0.676	16	0.818	160	0.770
3	0.719	19	0.674	163	0.817	20	0.768
166	0.714	160	0.670	13	0.816	19	0.767
30	0.713	120	0.670	116	0.815	120	0.760
16	0.712	104	0.670	1	0.814	155	0.758
108	0.710	155	0.656	110	0.813	24	0.756
109	0.710	24	0.645	30	0.812	121	0.741
179	0.709	164	0.640	125	0.812	164	0.740
1	0.707	157	0.630	109	0.810	157	0.700
163	0.707	153	0.620	118	0.810	152	0.693
158	0.704	152	0.613	140	0.810	42	0.686
4	0.704	42	0.566	15	0.810	153	0.570

Table 19: Analytical results for nitrate in precipitations samples.

Nitrate in precipitation				Nitrate in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 0.311				Theoretical value: 0.369			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:	69	Number of laboratories:	69	Number of laboratories:	69	Number of laboratories:	69
Arithmetic mean value:	0.318	Arithmetic mean value:	5.605	Arithmetic mean value:	0.364	Arithmetic mean value:	0.364
Median:	0.307	Median:	0.364	Median:	0.364	Median:	0.364
Standard deviation	0.120	Standard deviation	43.414	Standard deviation	0.140	Standard deviation	0.140
Rel. st. deviation (%)	37.850	Rel. st. deviation (%)	774.503	Rel. st. deviation (%)	36.923	Rel. st. deviation (%)	36.923
Run 2:				Run 2:			
Number of laboratories:	68	Number of laboratories:	68	Number of laboratories:	68	Number of laboratories:	68
Arithmetic mean value:	0.304	Arithmetic mean value:	0.379	Arithmetic mean value:	0.360	Arithmetic mean value:	0.360
Median:	0.307	Median:	0.364	Median:	0.364	Median:	0.364
Standard deviation	0.031	Standard deviation	0.140	Standard deviation	0.140	Standard deviation	0.140
Rel. st. deviation (%)	10.187	Rel. st. deviation (%)	36.923	Rel. st. deviation (%)	36.923	Rel. st. deviation (%)	36.923
Results in decreasing order:							
176	1.270 (*)	4	0.307	179	361.000 (*)	121	0.364
44	0.456	30	0.306	176	1.490	126	0.363
110	0.352	163	0.306	44	0.528	8	0.363
23	0.334	22	0.306	11	0.400	115	0.363
11	0.332	125	0.305	23	0.392	120	0.362
39	0.329	167	0.302	39	0.389	167	0.362
166	0.328	114	0.300	10	0.389	27	0.362
17	0.324	12	0.300	172	0.384	175	0.361
158	0.323	15	0.300	2	0.382	32	0.360
26	0.320	33	0.300	158	0.381	150	0.360
172	0.320	157	0.300	124	0.380	41	0.359
112	0.320	124	0.300	15	0.380	22	0.358
160	0.320	150	0.300	180	0.379	30	0.357
180	0.319	140	0.299	166	0.378	16	0.354
165	0.319	36	0.299	26	0.376	125	0.354
1	0.318	120	0.298	5	0.376	7	0.353
2	0.318	16	0.297	112	0.375	36	0.353
20	0.317	146	0.296	40	0.374	140	0.352
10	0.317	41	0.296	31	0.374	116	0.351
115	0.316	7	0.294	17	0.372	146	0.351
5	0.316	116	0.292	165	0.371	12	0.350
32	0.314	38	0.290	118	0.370	33	0.350
3	0.314	14	0.290	160	0.370	19	0.347
40	0.313	19	0.290	114	0.370	155	0.343
45	0.312	155	0.286	109	0.370	14	0.340
175	0.312	43	0.279	157	0.370	104	0.332
31	0.311	104	0.267	3	0.370	38	0.330
109	0.310	118	0.260	1	0.369	164	0.330
21	0.310	164	0.260	13	0.369	43	0.330
8	0.309	108	0.260	45	0.369	42	0.323
179	0.309	24	0.254	110	0.368	108	0.320
13	0.309	152	0.250	21	0.368	24	0.315
27	0.308	153	0.240	20	0.368	152	0.300
126	0.307	42	0.189	4	0.366	153	0.270
121	0.307			163	0.364		
Nitrate in precipitation							
Sample no.: G3				Sample no.: G4			
Theoretical value:	0.432	Theoretical value:	0.495	Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:	69	Number of laboratories:	69	Number of laboratories:	69	Number of laboratories:	69
Arithmetic mean value:	0.445	Arithmetic mean value:	0.508	Arithmetic mean value:	0.490	Arithmetic mean value:	0.490
Median:	0.429	Median:	0.485	Median:	0.485	Median:	0.485
Standard deviation	0.161	Standard deviation	0.185	Standard deviation	0.185	Standard deviation	0.185
Rel. st. deviation (%)	36.262	Rel. st. deviation (%)	36.401	Rel. st. deviation (%)	36.401	Rel. st. deviation (%)	36.401
Run 2:				Run 2:			
Number of laboratories:	68	Number of laboratories:	68	Number of laboratories:	68	Number of laboratories:	68
Arithmetic mean value:	0.426	Arithmetic mean value:	0.486	Arithmetic mean value:	0.486	Arithmetic mean value:	0.486
Median:	0.428	Median:	0.490	Median:	0.490	Median:	0.490
Standard deviation	0.038	Standard deviation	0.048	Standard deviation	0.048	Standard deviation	0.048
Rel. st. deviation (%)	8.910	Rel. st. deviation (%)	9.854	Rel. st. deviation (%)	9.854	Rel. st. deviation (%)	9.854
Results in decreasing order:							
176	1.730 (*)	120	0.428	176	1.970 (*)	26	0.490
44	0.632	121	0.428	44	0.725	8	0.489
110	0.510	179	0.428	11	0.536	17	0.489
11	0.468	163	0.427	158	0.524	4	0.488
39	0.459	17	0.426	23	0.524	7	0.488
23	0.457	30	0.424	110	0.521	20	0.488
40	0.455	167	0.422	39	0.521	155	0.487
158	0.451	41	0.422	2	0.516	167	0.484
2	0.448	27	0.422	10	0.515	30	0.483
10	0.443	7	0.421	40	0.515	27	0.482
8	0.442	124	0.420	120	0.507	38	0.480
180	0.442	126	0.420	13	0.507	124	0.480
160	0.440	150	0.420	32	0.504	150	0.480
13	0.440	38	0.420	31	0.504	157	0.480
32	0.440	16	0.419	41	0.502	33	0.480
31	0.438	115	0.417	5	0.500	116	0.480
175	0.438	43	0.415	180	0.500	118	0.480
172	0.438	125	0.414	15	0.500	16	0.479
45	0.437	155	0.414	22	0.500	36	0.475
22	0.437	146	0.413	172	0.500	125	0.473
5	0.434	36	0.413	45	0.499	126	0.473
165	0.434	116	0.412	165	0.497	19	0.472
112	0.434	19	0.411	179	0.497	146	0.472
1	0.433	157	0.410	1	0.496	12	0.470
3	0.432	12	0.410	166	0.495	140	0.470
166	0.432	140	0.404	43	0.495	115	0.468
33	0.430	164	0.400	21	0.494	14	0.460
15	0.430	14	0.400	112	0.494	164	0.460
109	0.430	104	0.386	3	0.493	108	0.440
118	0.430	24	0.372	121	0.493	104	0.440
21	0.430	108	0.360	163	0.492	24	0.426
114	0.430	152	0.360	160	0.490	152	0.410
26	0.430	42	0.335	109	0.490	42	0.366
20	0.429	153	0.310	114	0.490	153	0.260
4	0.429			175	0.490		

Table 20: Analytical results for ammonium in precipitations sample.

Ammonium in precipitation				Ammonium in precipitation			
Sample no.:	G1	Sample no.:	G2	Theoretical value:	0.134	Theoretical value:	0.187
Unit:	µg/l	Unit:	µg/l				
Run 1:							
Number of laboratories:	68	Number of laboratories:	69	Arithmetic mean value:	0.189	Median:	0.188
Arithmetic mean value:	0.138	Arithmetic mean value:	0.189	Standard deviation	0.028	Standard deviation	0.028
Median:	0.134	Median:	0.188	Rel. st. deviation (%)	14.810	Rel. st. deviation (%)	14.570
Standard deviation	0.020	Standard deviation	0.028				
Rel. st. deviation (%)	14.810	Rel. st. deviation (%)	14.570				
Run 2:							
Number of laboratories:	63	Number of laboratories:	62	Arithmetic mean value:	0.187	Median:	0.188
Arithmetic mean value:	0.135	Arithmetic mean value:	0.187	Standard deviation	0.014	Standard deviation	0.014
Median:	0.134	Median:	0.188	Rel. st. deviation (%)	10.827	Rel. st. deviation (%)	7.662
Standard deviation	0.015	Standard deviation	0.014				
Rel. st. deviation (%)	10.827	Rel. st. deviation (%)	7.662				
Results in decreasing order:							
43	0.199 (*)	179	0.134	45	0.293 (*)	121	0.188
110	0.190 (*)	27	0.134	31	0.278 (*)	19	0.186
31	0.184 (*)	166	0.134	43	0.261 (*)	27	0.186
116	0.183 (*)	5	0.133	116	0.248 (*)	104	0.186
172	0.177	180	0.133	176	0.230	179	0.186
175	0.171	155	0.132	172	0.226	5	0.184
17	0.170	4	0.131	3	0.215	120	0.182
3	0.164	13	0.131	22	0.215	33	0.182
120	0.158	108	0.130	17	0.210	26	0.182
22	0.152	114	0.130	2	0.209	180	0.182
15	0.150	124	0.130	158	0.204	114	0.180
32	0.150	153	0.130	152	0.200	4	0.180
109 <	0.150			20	0.188		
158	0.149	26	0.130	15	0.200	12	0.180
22	0.149	40	0.130	8	0.198	108	0.180
41	0.149	160	0.130	32	0.195	167	0.179
150	0.147	42	0.130	23	0.195	175	0.179
7	0.145	33	0.128	1	0.194	42	0.178
39	0.145	21	0.127	16	0.194	13	0.178
1	0.144	118	0.127	150	0.194	118	0.177
30	0.141	152	0.127	36	0.194	40	0.175
146	0.141	126	0.123	41	0.193	126	0.174
125	0.141	12	0.120	7	0.193	39	0.174
164	0.140	38	0.120	166	0.191	140	0.172
104	0.140	14	0.120	165	0.191	110	0.170
36	0.140	176	0.120	146	0.191	112	0.170
157	0.140	10	0.119	124	0.190	38	0.170
23	0.140	19	0.119	30	0.190	155	0.168
121	0.139	140	0.119	153	0.190	14	0.165
167	0.137	44	0.116	157	0.190	10	0.164
163	0.137	112	0.116	160	0.190	109	0.160
8	0.137	45	0.109	163	0.190	11	0.156
16	0.137	115	0.105	164	0.190	115	0.132 (*)
20	0.136	11	0.103	125	0.189	44	0.125 (*)
165	0.134	24	0.079 (*)	21	0.188	24	0.124 (*)
				20	0.188		
Ammonium in precipitation							
Sample no.:	G3	Sample no.:	G4	Theoretical value:	0.241	Theoretical value:	0.294
Unit:	µg/l	Unit:	µg/l				
Run 1:							
Number of laboratories:	69	Number of laboratories:	69	Arithmetic mean value:	0.303	Median:	0.293
Arithmetic mean value:	0.244	Arithmetic mean value:	0.303	Standard deviation	0.066	Standard deviation	0.066
Median:	0.239	Median:	0.293	Rel. st. deviation (%)	18.659	Rel. st. deviation (%)	21.694
Standard deviation	0.046	Standard deviation	0.028				
Rel. st. deviation (%)	18.659	Rel. st. deviation (%)	9.487				
Run 2:							
Number of laboratories:	66	Number of laboratories:	67	Arithmetic mean value:	0.294	Median:	0.293
Arithmetic mean value:	0.240	Arithmetic mean value:	0.294	Standard deviation	0.028	Standard deviation	0.028
Median:	0.239	Median:	0.293	Rel. st. deviation (%)	9.621	Rel. st. deviation (%)	9.487
Standard deviation	0.023	Standard deviation	0.028				
Rel. st. deviation (%)	9.621	Rel. st. deviation (%)	9.487				
Results in decreasing order:							
45	0.537 (*)	27	0.239	45	0.730 (*)	41	0.293
43	0.336 (*)	179	0.239	43	0.538 (*)	5	0.292
116	0.316	180	0.239	116	0.387	180	0.292
31	0.310	5	0.237	31	0.353	124	0.290
17	0.290	152	0.237	17	0.350	108	0.290
172	0.286	19	0.236	176	0.350	26	0.289
2	0.271	167	0.235	172	0.342	13	0.288
176	0.270	26	0.235	2	0.328	16	0.288
22	0.263	165	0.235	3	0.321	179	0.288
158	0.262	16	0.235	15	0.320	4	0.287
15	0.260	42	0.232	158	0.318	104	0.287
23	0.258	4	0.231	8	0.316	33	0.284
8	0.257	38	0.230	125	0.312	114	0.280
3	0.256	108	0.230	166	0.312	153	0.280
166	0.253	114	0.230	164	0.310	167	0.280
125	0.252	153	0.230	23	0.307	38	0.280
32	0.250	120	0.230	36	0.307	157	0.280
36	0.250	157	0.230	152	0.307	160	0.280
164	0.250	126	0.226	32	0.305	39	0.279
41	0.249	118	0.226	22	0.305	140	0.279
20	0.248	140	0.225	20	0.304	42	0.278
1	0.246	33	0.224	165	0.302	118	0.278
146	0.245	40	0.223	155	0.301	126	0.277
21	0.244	112	0.222	7	0.301	112	0.275
163	0.243	109	0.220	19	0.300	109	0.270
150	0.243	10	0.214	12	0.300	10	0.267
39	0.241	11	0.211	1	0.299	14	0.266
104	0.241	14	0.210	163	0.299	40	0.265
7	0.241	110	0.210	21	0.299	11	0.265
121	0.241	115	0.210	146	0.299	175	0.264
124	0.240	175	0.202	120	0.297	110	0.260
160	0.240	155	0.202	30	0.296	24	0.234
13	0.240	44	0.173	150	0.296	115	0.229
12	0.240	24	0.138 (*)	121	0.295	44	0.205
30	0.239			27	0.293		

Table 21: Analytical results for pH in precipitations samples.

pH in precipitation					
Sample no.: G1					
Theoretical value:	4.270	Theoretical value:	4.220		
Unit: pH-unit		Unit: pH-unit			
Run 1:					
Number of laboratories:	62	Number of laboratories:	62		
Arithmetic mean value:	4.357	Arithmetic mean value:	4.283		
Median:	4.330	Median:	4.280		
Standard deviation	0.144	Standard deviation	0.125		
Rel. st. deviation (%)	3.301	Rel. st. deviation (%)	2.911		
Run 2:					
Number of laboratories:	58	Number of laboratories:	57		
Arithmetic mean value:	4.337	Arithmetic mean value:	4.283		
Median:	4.330	Median:	4.280		
Standard deviation	0.078	Standard deviation	0.075		
Rel. st. deviation (%)	1.795	Rel. st. deviation (%)	1.748		
Results in decreasing order:					
108	4.980 (*) 31	4.330	41	4.720 (*) 140	4.280
150	4.910 (*) 17	4.330	150	4.710 (*) 152	4.280
113	4.670 (*) 36	4.330	108	4.510 157	4.280
38	4.540 124	4.320	24	4.500 23	4.270
157	4.510 15	4.320	113	4.470 15	4.270
126	4.460 20	4.320	14	4.370 175	4.270
146	4.450 32	4.320	19	4.351 46	4.270
19	4.436 43	4.320	13	4.350 121	4.260
109	4.430 41	4.320	126	4.340 155	4.260
14	4.420 165	4.320	114	4.340 120	4.260
24	4.400 46	4.320	163	4.340 124	4.250
140	4.400 121	4.320	153	4.330 32	4.250
112	4.400 166	4.310	116	4.330 5	4.250
114	4.400 16	4.310	112	4.330 21	4.250
13	4.400 27	4.310	146	4.320 176	4.250
116	4.390 155	4.310	17	4.310 3	4.241
163	4.390 4	4.310	33	4.310 16	4.240
33	4.380 164	4.310	22	4.310 164	4.240
153	4.380 5	4.300	109	4.310 165	4.240
104	4.370 21	4.290	7	4.310 4	4.240
1	4.360 23	4.290	1	4.310 166	4.230
120	4.360 26	4.290	104	4.300 118	4.230
7	4.360 115	4.280	10	4.300 38	4.220
10	4.350 3	4.279	158	4.280 115	4.210
152	4.350 118	4.260	8	4.280 43	4.200
176	4.350 30	4.250	20	4.280 30	4.180
44	4.340 45	4.180	26	4.280 11	4.090
175	4.340 11	4.170	27	4.280 45	4.050
22	4.340 40	4.150	31	4.280 40	4.010 (*)
158	4.340 2	4.080	36	4.280 172	3.980 (*)
8	4.330 172	4.040 (*)	44	4.280 2	3.975 (*)
pH in precipitation					
Sample no.: G3					
Theoretical value:	4.570	Theoretical value:	4.520		
Unit: pH-unit		Unit: pH-unit			
Run 1:					
Number of laboratories:	62	Number of laboratories:	62		
Arithmetic mean value:	4.607	Arithmetic mean value:	4.609		
Median:	4.620	Median:	4.580		
Standard deviation	0.115	Standard deviation	0.303		
Rel. st. deviation (%)	2.487	Rel. st. deviation (%)	6.570		
Run 2:					
Number of laboratories:	57	Number of laboratories:	61		
Arithmetic mean value:	4.623	Arithmetic mean value:	4.573		
Median:	4.620	Median:	4.580		
Standard deviation	0.065	Standard deviation	0.116		
Rel. st. deviation (%)	1.407	Rel. st. deviation (%)	2.529		
Results in decreasing order:					
150	4.990 (*) 109	4.620	115	6.780 (*) 124	4.580
24	4.790 104	4.610	150	5.000 152	4.580
108	4.780 166	4.610	24	4.800 8	4.580
14	4.760 155	4.610	108	4.780 166	4.570
121	4.720 15	4.610	22	4.670 155	4.570
113	4.720 120	4.600	113	4.660 15	4.570
126	4.700 38	4.600	126	4.660 41	4.560
19	4.695 46	4.600	13	4.650 175	4.560
153	4.680 176	4.600	19	4.647 27	4.560
116	4.680 21	4.590	109	4.640 120	4.560
163	4.680 27	4.590	43	4.640 21	4.550
112	4.670 5	4.590	153	4.640 46	4.550
114	4.670 164	4.590	36	4.640 26	4.550
13	4.670 26	4.580	121	4.630 176	4.550
36	4.660 140	4.580	114	4.630 5	4.540
33	4.660 23	4.580	112	4.630 32	4.540
158	4.660 4	4.580	33	4.630 4	4.530
22	4.660 3	4.572	140	4.630 16	4.530
7	4.650 16	4.570	158	4.620 3	4.529
43	4.650 32	4.570	116	4.620 23	4.520
1	4.630 115	4.560	163	4.620 164	4.520
8	4.630 157	4.550	7	4.610 118	4.520
152	4.630 118	4.540	17	4.610 30	4.510
31	4.630 41	4.530	14	4.610 165	4.500
124	4.620 165	4.520	10	4.600 157	4.480
146	4.620 30	4.510	146	4.600 38	4.460
175	4.620 11	4.420	1	4.600 11	4.400
10	4.620 45	4.346 (*)	104	4.590 45	4.342
17	4.620 2	4.300 (*)	20	4.580 40	4.290
20	4.620 40	4.260 (*)	31	4.580 172	4.250
44	4.620 172	4.260 (*)	44	4.580 2	4.220
pH in precipitation					
Sample no.: G4					
Theoretical value:	4.520	Theoretical value:	4.520		
Unit: pH-unit		Unit: pH-unit			
Run 1:					
Number of laboratories:	62	Number of laboratories:	62		
Arithmetic mean value:	4.609	Arithmetic mean value:	4.609		
Median:	4.580	Median:	4.580		
Standard deviation	0.303	Standard deviation	0.303		
Rel. st. deviation (%)	6.570	Rel. st. deviation (%)	6.570		
Run 2:					
Number of laboratories:	61	Number of laboratories:	61		
Arithmetic mean value:	4.573	Arithmetic mean value:	4.573		
Median:	4.580	Median:	4.580		
Standard deviation	0.116	Standard deviation	0.116		
Rel. st. deviation (%)	2.529	Rel. st. deviation (%)	2.529		
Results in decreasing order:					
150	6.780 (*) 124	4.580	115	6.780 (*) 124	4.580
24	5.000 152	4.580	150	5.000 152	4.580
24	4.800 8	4.580	24	4.800 8	4.580
108	4.780 166	4.570	108	4.780 166	4.570
22	4.670 155	4.570	22	4.670 155	4.570
113	4.660 15	4.570	113	4.660 15	4.570
126	4.660 41	4.560	126	4.660 41	4.560
19	4.650 175	4.560	13	4.650 175	4.560
153	4.647 27	4.560	19	4.647 27	4.560
116	4.640 120	4.560	109	4.640 120	4.560
163	4.640 21	4.550	43	4.640 21	4.550
153	4.640 46	4.550	153	4.640 46	4.550
36	4.640 26	4.550	36	4.640 26	4.550
121	4.630 176	4.550	121	4.630 176	4.550
114	4.630 5	4.540	114	4.630 5	4.540
112	4.630 32	4.540	112	4.630 32	4.540
33	4.630 4	4.530	33	4.630 4	4.530
158	4.630 118	4.530	140	4.630 16	4.530
22	4.620 3	4.529	158	4.620 3	4.529
7	4.620 23	4.520	116	4.620 23	4.520
43	4.620 164	4.520	116	4.620 23	4.520
1	4.620 164	4.520	163	4.620 164	4.520
8	4.620 118	4.520	7	4.610 118	4.520
152	4.620 30	4.510	17	4.610 30	4.510
31	4.620 165	4.510	14	4.610 165	4.500
124	4.620 41	4.510	10	4.600 157	4.480
146	4.620 30	4.510	146	4.600 38	4.460
175	4.620 11	4.420	1	4.600 11	4.400
10	4.620 45	4.346 (*)	104	4.590 45	4.342
17	4.620 2	4.300 (*)	20	4.580 40	4.290
20	4.620 40	4.260 (*)	31	4.580 172	4.250
44	4.620 172	4.260 (*)	44	4.580 2	4.220

Table 22: Analytical results for strong acid calculated from pH.

Strong acid calculated from pH				Strong acid calculated from pH			
Sample no.: G1		Sample no.: G2		Theoretical value:	53.000	Theoretical value:	60.000
Unit:		Unit:					
Run 1:							
Number of laboratories:	62	Number of laboratories:	62	Arithmetic mean value:	45.999	Arithmetic mean value:	54.211
Median:	46.770	Median:	52.480	Standard deviation	12.586	Standard deviation	15.372
Rel. st. deviation (%)	27.361	Rel. st. deviation (%)	28.356				
Run 2:							
Number of laboratories:	58	Number of laboratories:	56	Arithmetic mean value:	45.775	Arithmetic mean value:	52.229
Median:	46.770	Median:	52.480	Standard deviation	8.223	Standard deviation	7.843
Rel. st. deviation (%)	17.964	Rel. st. deviation (%)	15.016				
Results in decreasing order:							
172 91.000 (*) 31 46.770		2 105.930 (*) 140 52.480		172 105.000 (*) 152 52.480		40 97.720 (*) 157 52.480	
2 83.180 (*) 175 46.000		40 89.130 (*) 158 52.480		45 81.280 8 52.480		30 66.070 20 52.480	
40 70.790 44 45.710		30 63.000 26 52.480		43 61.660 27 52.480		118 60.260 104 50.120	
11 67.610 158 45.710		118 59.000 10 50.120		115 56.230 166 58.880		26 55.900 33 48.980	
45 66.070 22 45.710		166 58.000 109 48.980		164 58.000 17 48.980		21 51.290 7 43.650	
30 56.230 10 44.670		4 57.540 1 48.980		165 57.540 7 48.980		23 51.290 104 42.660	
118 55.000 152 44.670		165 57.410 22 48.980		165 57.410 146 47.860		5 50.120 153 42.000	
3 52.600 176 44.670		32 56.230 153 47.000		165 56.230 112 46.770		164 49.000 33 41.690	
115 52.480 120 43.650		121 54.950 163 45.710		165 56.230 116 46.770		166 48.980 163 40.740	
26 51.290 1 43.650		120 54.950 163 45.710		165 56.230 114 45.710		155 48.980 116 40.740	
21 51.290 7 43.650		163 54.950 126 45.710		165 56.230 114 45.710		155 48.980 24 39.810	
23 51.290 104 42.660		126 54.000 19 44.570		165 56.230 114 45.710		27 48.980 13 39.810	
5 50.120 153 42.000		175 54.000 14 42.660		165 56.230 114 45.710		16 48.980 13 39.810	
164 49.000 33 41.690		23 53.700 113 33.880		165 56.230 114 45.710		4 48.980 140 39.810	
166 48.980 163 40.740		15 53.700 24 31.620		165 56.230 114 45.710		15 47.860 109 37.150	
155 48.980 116 40.740		31 52.480 108 30.900		165 56.230 114 45.710		124 47.860 19 36.640	
27 48.980 24 39.810		36 52.480 150 19.500	(*)	165 56.230 114 45.710		121 47.860 146 35.480	
16 48.980 13 39.810		44 52.480 41 19.000	(*)			20 47.860 126 34.670	
4 48.980 140 39.810						165 47.860 157 30.900	
46 48.000 112 39.810						32 47.860 38 28.840	
41 48.000 114 39.810						36 46.770 113 21.380	
43 48.000 14 38.020						17 46.770 150 12.300	(*)
155 47.860 109 37.150						8 46.770 108 10.470	(*)
Strong acid calculated from pH							
Sample no.: G3		Strong acid calculated from pH		Sample no.: G4		Strong acid calculated from pH	
Theoretical value:	27.000	Theoretical value:	30.000	Theoretical value:	30.000	Theoretical value:	30.000
Unit:		Unit:		Unit:		Unit:	
Run 1:							
Number of laboratories:	62	Number of laboratories:	62	Arithmetic mean value:	25.640	Arithmetic mean value:	27.261
Median:	23.995	Median:	26.300	Standard deviation	7.914	Standard deviation	8.861
Rel. st. deviation (%)	30.866	Rel. st. deviation (%)	32.505				
Run 2:							
Number of laboratories:	58	Number of laboratories:	57	Arithmetic mean value:	23.871	Arithmetic mean value:	25.912
Median:	23.990	Median:	26.300	Standard deviation	4.063	Standard deviation	4.487
Rel. st. deviation (%)	17.019	Rel. st. deviation (%)	17.315				
Results in decreasing order:							
172 55.000 (*) 109 23.990		2 60.260 (*) 124 26.300		172 56.000 (*) 152 26.300		40 51.290 (*) 8 26.300	
40 54.950 (*) 124 23.990		45 45.500 (*) 104 25.700		45 45.500 (*) 104 25.700		11 38.020 17 23.990	
2 50.120 (*) 146 23.990		11 39.810 1 25.120		165 33.110 146 25.120		30 30.900 20 23.990	
45 45.080 (*) 10 23.990		38 34.670 10 25.120		165 31.620 7 24.550		41 30.000 1 23.440	
11 38.020 17 23.990		165 30.900 17 24.550		165 31.620 7 24.550		118 29.000 152 23.440	
30 30.900 20 23.990		30 30.900 17 24.550		157 33.110 14 24.550		157 28.180 8 23.440	
165 30.200 44 23.990		23 30.200 14 24.550		164 30.000 116 23.990		115 27.540 31 23.440	
41 30.000 1 23.440		164 30.000 116 23.990		118 30.000 163 23.990		32 26.920 7 23.390	
118 29.000 152 23.440		3 29.580 158 23.990		16 29.510 112 23.440		16 26.790 36 21.880	
157 28.180 8 23.440		16 29.510 112 23.440		16 29.510 114 23.440		140 26.300 33 21.880	
115 27.540 31 23.440		32 28.840 121 23.440		16 29.510 114 23.440		4 26.300 158 21.880	
32 26.920 7 23.390		5 28.840 140 23.440		16 28.840 140 23.440		23 26.300 22 21.880	
16 26.790 36 21.880		26 28.180 33 23.440		176 28.180 43 23.000		26 26.300 13 21.880	
140 26.300 33 21.880		21 28.180 153 23.000		176 28.180 43 23.000		5 25.700 112 21.880	
4 26.300 158 21.880		46 28.000 109 22.910		175 28.000 36 22.910		27 25.700 153 21.880	
23 26.300 22 21.880		41 28.000 19 22.540		175 28.000 19 22.540		21 25.700 163 21.880	
26 26.300 13 21.880		27 27.540 13 22.390		176 27.540 13 22.390		120 25.120 116 21.880	
164 26.000 114 21.880		120 27.540 113 21.880		176 27.540 113 21.880		176 25.120 19 20.180	
5 25.700 112 21.880		176 28.180 43 23.000		176 28.180 43 23.000		38 25.120 126 19.950	
27 25.700 153 21.000		175 28.000 109 22.910		175 28.000 109 22.910		46 25.000 113 19.050	
21 25.700 163 20.890		175 28.000 36 22.910		175 28.000 36 22.910		155 24.550 121 19.050	
120 25.120 116 20.890		41 28.000 19 22.540		175 28.000 19 22.540		166 24.550 14 17.380	
176 25.120 19 20.180		27 27.540 13 22.390		175 26.920 122 21.380		15 24.550 108 16.600	
38 25.120 126 19.950		120 27.540 113 21.880		175 26.920 108 16.600		104 24.550 24 16.220	
46 25.000 113 19.050		176 28.180 43 23.000		176 26.920 108 16.600		104 24.550 150 10.230	
155 24.550 121 19.050		175 28.000 109 22.910		175 28.000 109 22.910		175 24.000 150 4.0.170	(*)

Table 23: Analytical results for chloride in precipitations samples.

Chloride in precipitation
Sample no.: G1
Theoretical value: 0.193
Unit: µg/l

Run 1:
Number of laboratories: 64
Arithmetic mean value: 0.221
Median: 0.190
Standard deviation 0.176
Rel. st. deviation (%) 79.667

Run 2:
Number of laboratories: 63
Arithmetic mean value: 0.200
Median: 0.189
Standard deviation 0.051
Rel. st. deviation (%) 25.469

Results in decreasing order:
43 1.552 (*) 20 0.189
40 0.352 16 0.189
176 0.330 8 0.188
115 0.330 116 0.187
125 0.324 13 0.187
167 0.320 45 0.186
19 < 0.310
110 < 0.300
24 0.293 172 0.185
157 0.290 1 0.194
104 0.254 31 0.184
23 0.253 180 0.183
109 < 0.250
124 0.240 158 0.182
7 0.234 163 0.180
12 0.230 114 0.180
2 0.218 126 0.180
11 0.213 140 0.180
44 0.211 153 0.180
108 0.210 32 0.180
175 0.210 179 0.175
33 0.210 118 0.170
21 0.202 120 0.170
5 0.201 146 0.167
152 0.200 42 0.163
4 < 0.200
150 0.196 22 0.161
39 0.195 15 0.160
26 0.195 155 0.160
112 0.195 338 0.160
36 0.192 30 0.149
10 0.192 14 0.137
27 0.191 121 0.130
166 0.191 164 0.130
41 0.191 17 0.119
3 0.190 165 0.100

Chloride in precipitation
Sample no.: G2
Theoretical value: 0.212
Unit: µg/l

Run 1:
Number of laboratories: 65
Arithmetic mean value: 0.218
Median: 0.206
Standard deviation 0.052
Rel. st. deviation (%) 23.702

Run 2:
Number of laboratories: 61
Arithmetic mean value: 0.209
Median: 0.205
Standard deviation 0.036
Rel. st. deviation (%) 17.097

Results in decreasing order:
42 0.406 (*) 8 0.206
40 0.376 (*) 10 0.205
115 0.340 (*) 31 0.204
108 0.340 23 0.203
167 0.322 180 0.202
24 0.314 41 0.201
19 < 0.310
110 < 0.300
124 0.290 126 0.200
30 0.277 114 0.200
104 0.270 157 0.200
112 0.252 152 0.200
11 0.251 15 0.200
7 0.250 158 0.199
109 < 0.250
150 0.244 163 0.199
12 0.240 17 0.198
2 0.235 13 0.197
44 0.234 1 0.194
5 0.218 120 0.190
39 0.216 140 0.190
21 0.216 176 0.190
36 0.215 43 0.190
4 0.214 172 0.187
116 0.214 146 0.183
45 0.206

Chloride in precipitation
Sample no.: G3
Theoretical value: 0.270
Unit: µg/l

Run 1:
Number of laboratories: 65
Arithmetic mean value: 0.272
Median: 0.261
Standard deviation 0.065
Rel. st. deviation (%) 23.945

Run 2:
Number of laboratories: 61
Arithmetic mean value: 0.264
Median: 0.260
Standard deviation 0.040
Rel. st. deviation (%) 15.205

Results in decreasing order:
108 0.540 (*) 157 0.260
40 0.490 (*) 31 0.260
42 0.459 (*) 32 0.260
115 0.400 39 0.260
167 0.355 116 0.260
24 0.345 120 0.260
124 0.340 126 0.260
104 0.332 41 0.258
44 0.330 158 0.257
19 < 0.310
112 0.309 163 0.257
2 0.302 13 0.255
12 0.300 1 0.253
110 < 0.300
7 0.299 10 0.251
11 0.298 172 0.250
33 0.290 140 0.250
36 0.276 114 < 0.250
21 0.275 109 < 0.250
5 0.275 176 0.240
150 0.275 146 0.238
3 0.271 153 0.230
20 0.271 118 0.230
45 0.270 155 0.228
15 0.270 43 0.224
4 0.269 38 0.220
27 0.268 164 0.220
26 0.267 179 0.220
30 0.265 22 0.201
16 0.264 152 0.200
23 0.263 14 0.194
166 0.262 121 0.185
125 0.262 121 0.182
8 0.261 165 0.182
180 0.261 175 0.140 (*)

Chloride in precipitation
Sample no.: G4
Theoretical value: 0.347
Unit: µg/l

Run 1:
Number of laboratories: 68
Arithmetic mean value: 0.340
Median: 0.340
Standard deviation 0.074
Rel. st. deviation (%) 21.721

Run 2:
Number of laboratories: 64
Arithmetic mean value: 0.333
Median: 0.339
Standard deviation 0.045
Rel. st. deviation (%) 13.649

Results in decreasing order:
108 0.600 (*) 166 0.339
40 0.592 (*) 10 0.339
42 0.550 (*) 31 0.338
115 0.455 4 0.337
167 0.446 16 0.335
124 0.410 163 0.333
104 0.392 180 0.333
24 0.389 13 0.330
44 0.385 114 0.330
112 0.384 116 0.330
2 0.376 140 0.330
11 0.372 8 0.330
12 0.370 1 0.327
110 0.370 158 0.326
7 0.368 172 0.322
150 0.359 176 0.320
5 0.355 32 0.320
30 0.354 17 0.311
19 < 0.310
125 0.353 146 0.309
15 0.350 155 0.302
157 0.350 118 0.300
21 0.349 120 0.290
39 0.348 22 0.285
20 0.348 109 0.280
26 0.346 164 0.280
27 0.346 179 0.278
45 0.346 165 0.264
3 0.345 14 0.250
43 0.343 121 0.241
41 0.341 152 0.230
23 0.341 175 0.230
126 0.340 153 0.220
33 0.340 160 0.100 (*)

Table 24: Analytical results for sodium in precipitations samples.

Sodium in precipitation				Sodium in precipitation			
Sample no.: G1		Sample no.: G2		Theoretical value:		Theoretical value:	
Unit: µg/l		Unit: µg/l		0.286		0.317	
Run 1:							
Number of laboratories:	67	Number of laboratories:	67	Arithmetic mean value:	0.277	Arithmetic mean value:	0.309
Median:	0.281	Median:	0.281	Standard deviation	0.040	Standard deviation	0.038
Rel. st. deviation (%)	14.561	Rel. st. deviation (%)	12.333				
Run 2:							
Number of laboratories:	63	Number of laboratories:	63	Arithmetic mean value:	0.282	Arithmetic mean value:	0.315
Median:	0.281	Median:	0.281	Standard deviation	0.026	Standard deviation	0.027
Rel. st. deviation (%)	9.375	Rel. st. deviation (%)	8.489				
Results in decreasing order:							
43 < 1.000		43 < 0.100					
24 0.370 (*) 8	0.281	160 0.380 19	0.313				
125 0.356 33	0.280	124 0.380 31	0.312				
160 0.350 108	0.280	40 0.372 115	0.311				
124 0.340 114	0.280	157 0.370 126	0.310				
104 0.330 32	0.280	125 0.364 33	0.310				
40 0.314 157	0.280	104 0.360 153	0.310				
150 0.308 42	0.279	30 0.348 36	0.310				
30 0.307 36	0.279	7 0.344 32	0.310				
7 0.304 10	0.278	15 0.340 150	0.308				
16 0.300 23	0.277	5 0.334 42	0.308				
110 0.300 152	0.276	163 0.334 152	0.305				
15 0.300 45	0.273	158 0.331 39	0.302				
180 0.296 31	0.273	41 0.330 14	0.301				
5 0.296 14	0.272	121 0.330 164	0.300				
163 0.296 126	0.270	12 0.330 175	0.300				
27 0.294 140	0.270	166 0.329 140	0.300				
115 0.294 153	0.270	27 0.326 2	0.298				
155 0.293 39	0.269	155 0.326 10	0.298				
158 0.291 2	0.264	120 0.324 22	0.296				
120 0.290 164	0.260	4 0.323 24	0.290				
121 0.290 11	0.257	13 0.321 172	0.289				
12 0.290 172	0.256	108 0.320 44	0.286				
1 0.289 118	0.250	114 0.320 45	0.284				
41 0.289 175	0.250	110 0.319 11	0.282				
21 0.288 44	0.243	112 0.319 118	0.280				
13 0.287 22	0.242	3 0.318 179	0.272				
3 0.286 179	0.235	180 0.318 116	0.269				
19 0.286 38	0.230	1 0.317 38	0.260				
112 0.286 109	0.220	21 0.317 176	0.240				
116 0.286 176	0.210	26 0.316 109	0.220	(*)			
4 0.282 17	0.151 (*)	8 0.316 167	0.210	(*)			
26 0.281 167	0.151 (*)	16 0.315 17	0.210	(*)			
146 0.281 165	0.128 (*)	146 0.314 165	0.164	(*)			
166 0.281		23 0.313					
Sodium in precipitation							
Sample no.: G3		Sample no.: G4		Theoretical value:		Theoretical value:	
Theoretical value:		0.390		0.476		0.476	
Unit: µg/l		Unit: µg/l					
Run 1:							
Number of laboratories:	67	Number of laboratories:	67	Arithmetic mean value:	0.382	Arithmetic mean value:	0.462
Median:	0.384	Median:	0.467	Standard deviation	0.055	Standard deviation	0.057
Rel. st. deviation (%)	14.448	Rel. st. deviation (%)	12.292				
Run 2:							
Number of laboratories:	64	Number of laboratories:	63	Arithmetic mean value:	0.381	Arithmetic mean value:	0.467
Median:	0.384	Median:	0.468	Standard deviation	0.034	Standard deviation	0.038
Rel. st. deviation (%)	8.875	Rel. st. deviation (%)	8.147				
Results in decreasing order:							
43 < 1.000		43 1.000					
157 0.690 (*) 150	0.383	157 0.670 (*) 110	0.466				
124 0.450 33	0.380	40 0.570 179	0.465				
160 0.450 36	0.380	124 0.550 26	0.464				
8 0.436 108	0.380	160 0.550 150	0.464				
125 0.434 126	0.380	158 0.511 19	0.462				
104 0.430 175	0.380	7 0.510 36	0.461				
30 0.416 32	0.380	125 0.508 32	0.460				
40 0.414 19	0.379	30 0.503 33	0.460				
158 0.414 152	0.379	15 0.500 175	0.460				
15 0.410 14	0.379	104 0.500 108	0.460				
23 0.408 31	0.378	116 0.498 126	0.460				
163 0.406 2	0.371	163 0.491 3	0.457				
5 0.405 4	0.370	12 0.490 11	0.456				
7 0.404 164	0.370	5 0.489 152	0.456				
13 0.403 39	0.368	27 0.487 2	0.455				
12 0.400 42	0.368	166 0.486 164	0.450				
121 0.400 10	0.366	155 0.483 42	0.448				
27 0.400 16	0.365	13 0.482 14	0.448				
180 0.397 116	0.362	153 0.480 172	0.447				
155 0.396 45	0.361	121 0.480 10	0.446				
1 0.393 11	0.359	4 0.480 16	0.443				
112 0.391 172	0.358	114 0.480 39	0.441				
41 0.391 44	0.356	140 0.480 45	0.433				
166 0.390 179	0.354	120 0.479 44	0.422				
140 0.390 118	0.330	180 0.479 22	0.414				
153 0.390 38	0.330	23 0.478 118	0.410				
114 0.390 22	0.327	112 0.476 38	0.400				
115 0.388 176	0.300	1 0.473 176	0.370				
146 0.387 24	0.300	21 0.472 167	0.361				
120 0.387 167	0.286	41 0.472 17	0.354				
21 0.386 17	0.281	146 0.470 165	0.328	(*)			
26 0.384 109	0.260	8 0.469 109	0.320	(*)			
3 0.384 165	0.245	31 0.468 24	0.260	(*)			
110 0.384		115 0.467					

Table 25: Analytical results for magnesium in precipitations samples.

Magnesium in precipitation				Magnesium in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 0.155				Theoretical value: 0.083			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:	66	Number of laboratories:	65	Arithmetic mean value:	0.088	Arithmetic mean value:	0.088
Arithmetic mean value:	0.150	Median:	0.154	Median:	0.081	Standard deviation	0.048
Median:	0.154	Standard deviation	0.022	Rel. st. deviation (%)	54.175	Rel. st. deviation (%)	16.222
Standard deviation	0.022	Rel. st. deviation (%)	14.641	Run 2:			
Rel. st. deviation (%)	14.641	Run 2:		Number of laboratories:	63	Number of laboratories:	63
Number of laboratories:	62	Arithmetic mean value:	0.153	Arithmetic mean value:	0.080	Median:	0.081
Arithmetic mean value:	0.153	Median:	0.154	Standard deviation	0.013	Standard deviation	0.080
Median:	0.154	Standard deviation	0.014	Rel. st. deviation (%)	16.222	Rel. st. deviation (%)	9.091
Standard deviation	0.014	Rel. st. deviation (%)	9.091	Results in decreasing order:		Results in decreasing order:	
Rel. st. deviation (%)	9.091	179	0.195 (*) 42	0.154	160	0.380 (*) 116	0.081
7	0.162 118	160	0.190 13	0.153	113	0.304 (*) 1	0.081
155	0.161 16	175	0.186 26	0.153	109	0.120 42	0.081
166	0.160 113	146	0.178 8	0.152	146	0.105 118	0.080
121	0.160 152	24	0.170 125	0.151	175	0.099 120	0.080
127	0.159 4	140	0.170 15	0.150	179	0.098 121	0.080
167	0.159 31	109	0.170 33	0.150	2	0.091 124	0.080
158	0.159 14	153	0.170 23	0.150	153	0.090 140	0.080
30	0.158 115	150	0.170 104	0.150	24	0.090 126	0.080
10	0.158 108	114	0.170 112	0.150	41	0.090 157	0.080
5	0.158 124	163	0.164 120	0.149	165	0.089 16	0.080
110	0.157 126	2	0.163 40	0.149	163	0.088 167	0.080
41	0.157 157	180	0.163 12	0.149	10	0.088 114	0.080
11	0.157 45	7	0.162 118	0.147	7	0.087 112	0.080
21	0.155 165	155	0.161 16	0.146	166	0.087 150	0.079
1	0.155 164	166	0.160 113	0.145	155	0.085 104	0.079
116	0.155 22	121	0.160 152	0.145	23	0.085 4	0.078
36	0.155 17	27	0.158 108	0.140	5	0.085 125	0.077
3	0.154 44	11	0.158 124	0.140	11	0.085 115	0.076
19	0.154 172	158	0.159 14	0.141	158	0.085 152	0.076
39	0.154 38	30	0.158 115	0.141	110	0.085 14	0.074
		10	0.158 108	0.140	27	0.084 31	0.073
		5	0.158 124	0.140	30	0.084 33	0.070
		110	0.157 126	0.140	19	0.084 15	0.070
		41	0.157 157	0.140	36	0.084 108	0.070
		11	0.157 45	0.134	13	0.084 45	0.067
		21	0.155 165	0.122	180	0.083 12	0.067
		1	0.155 164	0.120	26	0.082 22	0.061
		116	0.155 22	0.119	40	0.082 164	0.060
		36	0.155 17	0.119	39	0.082 17	0.054
		3	0.154 44	0.097 (*)	3	0.082 44	0.052
		19	0.154 172	0.085 (*)	8	0.082 38	0.020
		39	0.154 38	0.050 (*)	21	0.081	
Results in decreasing order:		Run 1:		Results in decreasing order:		Run 1:	
160	0.450 (*) 112	Number of laboratories:	66	160	0.550 (*) 39	Number of laboratories:	66
153	0.130 114	Arithmetic mean value:	0.103	116	0.153 26	Arithmetic mean value:	0.125
175	0.124 140	Median:	0.101	153	0.150 40	Median:	0.122
109	0.120 121	Standard deviation	0.046	175	0.150 121	Standard deviation	0.057
179	0.120 124	Rel. st. deviation (%)	44.762	179	0.140 114	Rel. st. deviation (%)	45.461
2	0.111 36	Run 2:		179	0.140 114	Run 2:	
163	0.111 23	Number of laboratories:	65	24	0.140 8	Number of laboratories:	65
24	0.110 15	Arithmetic mean value:	0.098	109	0.140 112	Arithmetic mean value:	0.118
158	0.109 150	Median:	0.100	146	0.133 15	Median:	0.122
30	0.107 104	Standard deviation	0.016	158	0.133 150	Standard deviation	0.020
10	0.107 120	Rel. st. deviation (%)	16.481	2	0.132 120	Rel. st. deviation (%)	17.090
180	0.107 40	Results in decreasing order:		163	0.132 118	Results in decreasing order:	
166	0.107 125	180	0.107 40	165	0.131 113	160	0.550 (*) 39
155	0.106 118	166	0.107 120	140	0.130 12	116	0.153 26
41	0.106 14	13	0.106 118	109	0.129 125	115	0.150 40
5	0.106 152	10	0.106 14	10	0.128 104	126	0.127 152
7	0.105 4	7	0.106 152	7	0.128 4	126	0.127 152
110	0.105 115	155	0.105 4	166	0.127 115	126	0.127 152
27	0.105 113	11	0.105 115	11	0.127 115	126	0.127 152
13	0.104 16	30	0.104 16	30	0.127 31	126	0.127 152
11	0.104 12	42	0.104 12	42	0.126 124	126	0.126 124
19	0.104 31	27	0.104 31	27	0.126 157	126	0.126 157
116	0.103 157	5	0.103 157	5	0.126 126	126	0.126 126
42	0.103 108	180	0.103 108	180	0.126 108	126	0.126 108
1	0.102 33	41	0.102 33	41	0.124 33	126	0.124 33
3	0.102 126	1	0.102 126	1	0.123 14	126	0.123 14
146	0.102 45	19	0.102 45	19	0.123 45	126	0.123 45
39	0.102 22	21	0.102 22	21	0.123 45	126	0.123 45
26	0.102 164	167	0.102 164	167	0.123 17	126	0.123 17
165	0.101 17	23	0.101 17	23	0.123 22	126	0.123 22
8	0.101 172	36	0.101 172	36	0.123 172	126	0.123 172
167	0.101 44	3	0.101 44	3	0.122 44	126	0.122 44
21	0.101 38	38	0.101 38	38	0.122 38	126	0.122 38

Table 26: Analytical results for calcium in precipitations samples.

Calcium in precipitation				Calcium in precipitation			
Sample no.: G1	Theoretical value:	0.102	Unit: µg/l	Sample no.: G2	Theoretical value:	0.140	Unit: µg/l
Run 1:				Run 1:			
Number of laboratories:	66			Number of laboratories:	66		
Arithmetic mean value:	0.131			Arithmetic mean value:	0.159		
Median:	0.110			Median:	0.143		
Standard deviation	0.071			Standard deviation	0.066		
Rel. st. deviation (%)	54.378			Rel. st. deviation (%)	41.332		
Run 2:				Run 2:			
Number of laboratories:	63			Number of laboratories:	63		
Arithmetic mean value:	0.119			Arithmetic mean value:	0.147		
Median:	0.109			Median:	0.142		
Standard deviation	0.039			Standard deviation	0.031		
Rel. st. deviation (%)	32.882			Rel. st. deviation (%)	21.067		
Results in decreasing order:				Results in decreasing order:			
43 < 0.500				43 < 0.500			
24 0.460 (*) 118 0.110				24 0.490 (*) 26 0.143			
160 0.420 (*) 26 0.109				160 0.450 (*) 23 0.142			
146 0.320 (*) 27 0.109				157 0.300 (*) 180 0.142			
157 0.260 39 0.107				165 0.281 17 0.141			
165 0.259 166 0.106				146 0.239 124 0.140			
40 0.231 5 0.106				108 0.230 4 0.140			
108 0.200 11 0.105				42 0.190 11 0.140			
175 0.194 30 0.104				175 0.186 40 0.140			
153 0.160 3 0.104				113 0.184 39 0.140			
7 0.159 179 0.103				153 0.180 179 0.140			
42 0.157 23 0.103				116 0.180 118 0.140			
8 0.152 120 0.102				104 0.171 158 0.139			
113 0.148 124 0.100				126 0.170 13 0.139			
22 0.143 114 0.100				163 0.163 8 0.138			
155 0.132 12 0.100				16 0.163 120 0.137			
109 0.130 14 0.100				152 0.160 36 0.136			
180 0.128 1 0.099				114 0.160 3 0.132			
16 0.126 13 0.099				155 0.158 32 0.130			
19 0.123 158 0.097				7 0.156 15 0.130			
152 0.123 36 0.097				19 0.154 12 0.130			
115 0.122 17 0.094				125 0.154 14 0.130			
104 0.120 41 0.092				109 0.150 1 0.128			
126 0.120 33 0.090				121 0.150 2 0.125			
21 0.118 45 0.090				112 0.150 31 0.122			
163 0.115 31 0.089				115 0.149 44 0.120			
10 0.115 167 0.088				30 0.147 33 0.120			
112 0.113 2 0.087				22 0.146 45 0.120			
110 0.112 172 0.087				21 0.146 140 0.110			
121 0.110 44 0.084				10 0.145 167 0.109			
125 0.110 32 0.080				5 0.145 41 0.107			
15 0.110 150 0.076				166 0.145 150 0.101			
4 0.110 140 0.070				27 0.143 164 0.100			
116 0.110 164 0.070				110 0.143 172 0.096			
38 < 0.040				38 < 0.040			
Calcium in precipitation				Calcium in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value:	0.179			Theoretical value:	0.217		
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:	67			Number of laboratories:	67		
Arithmetic mean value:	0.193			Arithmetic mean value:	0.230		
Median:	0.183			Median:	0.221		
Standard deviation	0.063			Standard deviation	0.067		
Rel. st. deviation (%)	32.420			Rel. st. deviation (%)	28.955		
Run 2:				Run 2:			
Number of laboratories:	64			Number of laboratories:	64		
Arithmetic mean value:	0.186			Arithmetic mean value:	0.223		
Median:	0.182			Median:	0.221		
Standard deviation	0.027			Standard deviation	0.034		
Rel. st. deviation (%)	14.272			Rel. st. deviation (%)	15.028		
Results in decreasing order:				Results in decreasing order:			
160 0.500 (*) 27 0.182				24 0.550 (*) 32 0.220			
24 0.500 (*) 30 0.181				160 0.520 (*) 15 0.220			
43 < 0.500				43 < 0.500			
157 0.275 26 0.180				114 0.350 124 0.220			
175 0.251 116 0.180				116 0.324 30 0.218			
108 0.250 110 0.180				108 0.300 26 0.217			
42 0.234 121 0.180				157 0.290 39 0.216			
165 0.224 15 0.180				175 0.265 110 0.214			
113 0.223 4 0.180				146 0.261 179 0.214			
109 0.220 167 0.179				165 0.257 13 0.213			
163 0.213 39 0.178				7 0.250 2 0.211			
19 0.213 36 0.178				109 0.250 36 0.211			
7 0.212 179 0.178				163 0.250 11 0.211			
153 0.210 11 0.177				153 0.250 14 0.210			
104 0.207 115 0.176				42 0.247 12 0.210			
114 0.200 13 0.173				113 0.244 115 0.210			
126 0.200 2 0.171				10 0.237 3 0.207			
180 0.199 22 0.171				17 0.234 8 0.204			
17 0.197 12 0.170				104 0.232 1 0.202			
152 0.197 14 0.170				155 0.232 120 0.201			
146 0.196 118 0.170				121 0.230 118 0.200			
10 0.196 1 0.167				126 0.230 4 0.200			
155 0.193 120 0.167				40 0.229 140 0.200			
23 0.190 3 0.167				158 0.229 45 0.194			
112 0.190 140 0.160				19 0.228 167 0.193			
32 0.190 45 0.159				180 0.228 31 0.192			
8 0.190 31 0.156				152 0.227 44 0.190			
124 0.190 44 0.156				5 0.226 33 0.190			
158 0.187 33 0.150				16 0.226 41 0.184			
16 0.187 150 0.141				23 0.226 150 0.166			
125 0.185 172 0.140				125 0.225 172 0.164			
21 0.184 41 0.134				166 0.223 22 0.163			
166 0.184 164 0.130				112 0.223 164 0.160			
5 0.184 38 0.040 (*)				27 0.222 38 0.050 (*)			
40 0.183				21 0.221			

Table 27: Analytical results for potassium in precipitations samples.

Potassium in precipitation
Sample no.: G1
Theoretical value: 0.221
Unit: $\mu\text{g/l}$

Run 1:
Number of laboratories: 66
Arithmetic mean value: 0.222
Median: 0.216
Standard deviation 0.056
Rel. st. deviation (%) 25.397

Run 2:
Number of laboratories: 63
Arithmetic mean value: 0.212
Median: 0.215
Standard deviation 0.026
Rel. st. deviation (%) 12.058

Results in decreasing order:
 $43 < 1.000$

157	0.510	(*)	41	0.216
125	0.426	(*)	14	0.215
160	0.410	(*)	23	0.212
166	<	0.410		
104	0.262		152	0.212
175	0.260		1	0.212
11	0.256		40	0.211
155	0.242		12	0.210
153	0.240		126	0.210
120	0.240		33	0.210
165	0.238		38	0.210
3	0.236		4	0.210
10	0.234		30	0.209
5	0.232		172	0.207
42	0.232		36	0.207
116	0.231		8	0.205
140	0.230		146	0.202
15	0.230		112	0.201
115	0.228		108	0.200
21	0.228		114	0.200
7	0.227		22	0.198
19	0.226		32	0.195
180	0.225		31	0.194
13	0.221		118	0.190
158	0.220		45	0.190
121	0.220		109	0.190
110	0.220		124	0.190
179	0.220		64	0.190
150	0.219		16	0.188
39	0.219		17	0.159
26	0.218		44	0.155
27	0.218		167	0.141
2	0.217		176	0.140
163	0.217		24	0.140

Potassium in precipitation
Sample no.: G3
Theoretical value: 0.170
Unit: $\mu\text{g/l}$

Run 1:
Number of laboratories: 65
Arithmetic mean value: 0.168
Median: 0.163
Standard deviation 0.047
Rel. st. deviation (%) 27.843

Run 2:
Number of laboratories: 63
Arithmetic mean value: 0.161
Median: 0.163
Standard deviation 0.023
Rel. st. deviation (%) 14.238

Results in decreasing order:
 $43 < 1.000$

175	0.470	(*)	40	0.163
166	<	0.410		
160	0.280	(*)	115	0.161
153	0.210		112	0.160
104	0.201		38	0.160
157	0.200		33	0.160
125	0.194		126	0.160
116	0.193		32	0.160
3	0.190		12	0.160
180	0.188		150	0.159
42	0.181		152	0.159
15	0.180		1	0.159
120	0.180		30	0.159
10	0.180		41	0.157
140	0.180		36	0.153
5	0.175		31	0.153
7	0.174		24	0.150
19	0.174		64	0.150
21	0.172		114	0.150
158	0.172		109	0.150
39	0.171		121	< 0.150
108	0.170		22	0.149
4	0.170		172	0.147
110	0.170		146	0.146
2	0.169		11	0.140
155	0.169		167	0.138
27	0.169		16	0.138
13	0.169		45	0.137
26	0.168		165	0.130
23	0.166		17	0.127
163	0.165		44	0.110
8	0.165		24	0.090
14	0.164		176	0.080
179	0.163			

Potassium in precipitation
Sample no.: G2
Theoretical value: 0.204
Unit: $\mu\text{g/l}$

Run 1:
Number of laboratories: 65
Arithmetic mean value: 0.195
Median: 0.194
Standard deviation 0.031
Rel. st. deviation (%) 15.707

Run 2:
Number of laboratories: 62
Arithmetic mean value: 0.195
Median: 0.194
Standard deviation 0.018
Rel. st. deviation (%) 9.302

Results in decreasing order:

157	0.410	(*)	152	0.194
166	<	0.410		
157	0.360	(*)	152	0.194
104	0.241		1	0.194
160	0.230		30	0.193
175	0.230		11	0.192
173	0.222		4	0.190
180	0.221		12	0.190
7	0.229		38	0.190
42	0.213		33	0.190
125	0.212		126	0.190
41	0.211		153	0.190
5	0.211		23	0.189
10	0.211		36	0.189
120	0.210		112	0.188
110	0.210		146	0.186
15	0.210		172	0.185
21	0.209		115	0.182
14	0.209		165	0.182
19	0.208		114	0.180
155	0.208		108	0.180
158	0.205		164	0.180
39	0.204		118	0.180
27	0.203		16	0.180
26	0.202		150	0.177
2	0.201		109	0.170
140	0.200		31	0.169
32	0.200		22	0.167
124	0.200		17	0.164
163	0.200		45	0.160
179	0.197		167	0.153
13	0.195		44	0.150
121	<	0.150		
116	0.194		121	< 0.150
8	0.194		176	0.110 (*)
116	0.194			

Potassium in precipitation
Sample no.: G4
Theoretical value: 0.136
Unit: $\mu\text{g/l}$

Run 1:
Number of laboratories: 65
Arithmetic mean value: 0.132
Median: 0.130
Standard deviation 0.029
Rel. st. deviation (%) 21.701

Run 2:
Number of laboratories: 59
Arithmetic mean value: 0.129
Median: 0.130
Standard deviation 0.016
Rel. st. deviation (%) 12.778

Results in decreasing order:

166	<	0.410	110	0.130
160	0.230	(*)	110	0.130
157	0.210	(*)	112	0.130
116	0.201	(*)	114	0.130
153	0.200	(*)	124	0.130
165	0.160		179	0.129
104	0.160		163	0.129
3	0.154		41	0.128
180	0.153		23	0.128
175	0.152		13	0.127
120	0.150		115	0.127
42	0.150		152	0.125
121	<	0.150		
11	0.146		30	0.124
14	0.143		150	0.120
39	0.143		118	0.120
15	0.140		109	0.120
19	0.140		164	0.120
10	0.140		33	0.120
140	0.140		38	0.120
27	0.139		126	0.120
7	0.137		36	0.117
158	0.137		31	0.116
155	0.136		16	0.114
125	0.136		22	0.114
21	0.135		172	0.111
2	0.135		108	0.110
40	0.134		146	0.110
5	0.134		45	0.108
8	0.134		17	0.086
26	0.134		44	0.080
1	0.131		24	0.080
12	0.130		167	0.067 (*)
4	0.130		176	0.050 (*)
32	0.130			

Table 28: Analytical results for conductivity in precipitations samples.

Conductivity in precipitation				Conductivity in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 29.000				Theoretical value: 32.000			
Unit: $\mu\text{S}/\text{cm}$				Unit: $\mu\text{S}/\text{cm}$			
Run 1:				Run 1:			
Number of laboratories:	62	Number of laboratories:	62	Number of laboratories:	62	Number of laboratories:	58
Arithmetic mean value:	27.510	Arithmetic mean value:	30.409	Arithmetic mean value:	30.600	Arithmetic mean value:	30.000
Median:	27.635	Median:	30.800	Median:	31.000	Median:	31.000
Standard deviation	2.794	Standard deviation	2.436	Standard deviation	1.584	Standard deviation	1.584
Rel. st. deviation (%)	10.156	Rel. st. deviation (%)	8.010	Rel. st. deviation (%)	5.178	Rel. st. deviation (%)	5.178
Run 2:				Run 2:			
Number of laboratories:	58	Number of laboratories:	58	Number of laboratories:	58	Number of laboratories:	58
Arithmetic mean value:	27.455	Arithmetic mean value:	30.600	Arithmetic mean value:	31.000	Arithmetic mean value:	30.000
Median:	27.635	Median:	30.800	Median:	31.000	Median:	31.000
Standard deviation	1.578	Standard deviation	1.584	Standard deviation	1.584	Standard deviation	1.584
Rel. st. deviation (%)	5.746	Rel. st. deviation (%)	5.178	Rel. st. deviation (%)	5.178	Rel. st. deviation (%)	5.178
Results in decreasing order:				Results in decreasing order:			
104	38.000 (*)	110	27.600	38	39.100 (*)	164	30.600
40	36.700 (*)	27	27.500	43	35.200	110	30.500
14	31.300	158	27.200	104	34.000	46	30.480
121	29.900	16	27.200	14	33.200	27	30.200
166	29.800	118	27.100	121	33.100	7	30.170
17	29.340	114	27.100	166	32.900	24	30.000
15	29.200	13	27.000	3	32.800	157	30.000
19	29.200	24	27.000	17	32.250	124	30.000
38	29.200	175	27.000	116	31.900	15	30.000
12	29.200	44	27.000	19	31.800	11	29.900
3	29.180	116	26.980	31	31.730	126	29.800
31	29.000	11	26.800	12	31.700	165	29.600
120	29.000	22	26.600	155	31.700	152	29.500
10	28.900	165	26.600	10	31.650	150	29.400
45	28.800	152	26.600	45	31.600	13	29.400
155	28.700	7	26.220	172	31.400	158	29.400
43	28.600	36	26.200	5	31.400	22	29.300
33	28.500	115	26.100	114	31.400	109	29.200
5	28.500	124	26.000	163	31.400	44	29.000
21	28.480	126	26.000	20	31.250	118	29.000
163	28.400	8	25.800	153	31.220	115	28.800
20	28.300	30	25.300	16	31.200	108	28.800
2	28.200	157	25.000	33	31.200	1	28.630
172	28.100	41	24.920	21	31.180	36	28.100
176	28.000	1	24.770	32	31.130	23	27.800
32	28.000	150	24.700	4	31.100	30	27.800
153	27.970	23	24.600	2	31.000	8	27.700
46	27.960	108	24.400	120	31.000	40	27.300
146	27.900	109	23.900	146	31.000	41	25.490 (*)
4	27.900	140	20.500 (*)	175	31.000	140	24.900 (*)
164	27.670	113	18.000 (*)	176	31.000	113	21.100 (*)
Conductivity in precipitation				Conductivity in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value: 19.000				Theoretical value: 22.000			
Unit: $\mu\text{S}/\text{cm}$				Unit: $\mu\text{S}/\text{cm}$			
Run 1:				Run 1:			
Number of laboratories:	62	Number of laboratories:	62	Number of laboratories:	62	Number of laboratories:	59
Arithmetic mean value:	19.566	Arithmetic mean value:	21.521	Arithmetic mean value:	21.521	Arithmetic mean value:	21.283
Median:	19.000	Median:	21.390	Median:	21.390	Median:	21.380
Standard deviation	4.536	Standard deviation	2.351	Standard deviation	2.351	Standard deviation	1.485
Rel. st. deviation (%)	23.184	Rel. st. deviation (%)	10.923	Rel. st. deviation (%)	6.975	Rel. st. deviation (%)	6.975
Run 2:				Run 2:			
Number of laboratories:	59	Number of laboratories:	59	Number of laboratories:	59	Number of laboratories:	59
Arithmetic mean value:	18.664	Arithmetic mean value:	21.283	Arithmetic mean value:	21.283	Arithmetic mean value:	21.283
Median:	18.940	Median:	21.380	Median:	21.380	Median:	21.380
Standard deviation	1.120	Standard deviation	1.485	Standard deviation	1.485	Standard deviation	1.485
Rel. st. deviation (%)	6.003	Rel. st. deviation (%)	6.975	Rel. st. deviation (%)	6.975	Rel. st. deviation (%)	6.975
Results in decreasing order:				Results in decreasing order:			
41	48.400 (*)	176	19.000	43	32.000 (*)	153	21.380
43	33.500 (*)	7	18.940	115	30.000 (*)	126	21.300
104	30.000 (*)	153	18.940	41	25.780	11	21.200
121	21.000	21	18.900	104	25.000	32	21.200
38	20.800	163	18.900	38	24.300	152	21.100
46	20.260	2	18.900	121	23.400	27	21.100
166	20.100	152	18.800	166	22.800	124	21.000
14	20.000	32	18.800	14	22.700	146	21.000
116	19.800	27	18.600	116	22.550	157	21.000
15	19.700	110	18.300	17	22.460	175	21.000
33	19.600	11	18.300	15	22.300	44	21.000
20	19.550	1	18.300	19	22.200	176	21.000
165	19.500	126	18.230	46	22.160	24	21.000
114	19.500	124	18.000	3	22.070	109	20.900
109	19.500	40	18.000	31	22.000	7	20.810
17	19.440	44	18.000	120	22.000	110	20.800
3	19.440	157	18.000	114	22.000	118	20.500
31	19.430	175	18.000	4	21.900	13	20.400
19	19.400	158	18.000	20	21.870	22	20.200
4	19.300	22	17.900	10	21.800	108	20.200
12	19.200	108	17.700	155	21.800	23	20.100
155	19.200	23	17.600	33	21.800	158	20.000
10	19.150	118	17.600	5	21.800	150	19.840
45	19.100	8	17.200	16	21.700	1	19.770
16	19.100	36	17.100	172	21.700	8	19.700
146	19.100	115	17.100	45	21.600	164	19.380
24	19.000	164	17.070	12	21.500	36	19.200
5	19.000	150	16.930	21	21.450	30	18.800
120	19.000	140	16.800	163	21.400	140	17.000
13	19.000	30	16.700	2	21.400	40	17.000
172	19.000	113	14.400	165	21.400	113	16.600 (*)

Appendix 2

Figures – 27th intercomparison




Figure 2: Youden plot of SO₂-S in absorbing solution.




Figure 3: Youden plot of SO₂-S on impregnated filter.




Figure 4: Youden plot of HNO₃-N on impregnated filter.




Figure 5: Youden plot of NO₂-N in absorbing solution.




Figure 6: Youden plot of NH₃-N on impregnated filter.




Figure 7: Youden plot of SO_4 -S in precipitation.




Figure 8: Youden plot of NO_3 -N in precipitation.




Figure 9: Youden plot of $\text{NH}_4\text{-N}$ in precipitation.




Figure 10: Youden plot of pH in precipitation.




Figure 11: Youden plot of strong acid in precipitation.




Figure 12: Youden plot of Cl in precipitation.




Figure 13: Youden plot of Na in precipitation.




Figure 14: Youden plot of Mg in precipitation.




Figure 15: Youden plot of Ca in precipitation.




Figure 16: Youden plot of K in precipitation.




Figure 17: Youden plot of conductivity in precipitation.

Appendix 3

Tables – 28th intercomparison

Table 29: Samples distributed for the twenty-eighth interlaboratory test.

A.	5 synthetic samples for determination of SO ₂ , consisting of 0.3% H ₂ O ₂ absorbing solution and containing different concentrations of sulphuric acid. One of the samples was an unidentified blank.
B.	5 KOH-impregnated Whatman 40 filters, comprising 1 blank and 4 filters to which different amounts of sulphuric acid and nitrate salt have been added.
C.	4 synthetic samples for determination of NO ₂ consisting of sodium nitrite diluted in water.
J.	6 Whatman 40 filters impregnated with 3% oxalic acid, comprising 2 blank and 4 filters to which different amounts of ammonium salt solution have been added.
G.	4 synthetic precipitation samples, containing SO ₄ ²⁻ , NO ₃ ⁻ , NH ₄ ⁺ , H ⁺ , Na ⁺ , Mg ²⁺ and Cl ⁻ , and Ca ²⁺ and K ⁺ .

Table 30a: EMEP laboratories participating in the twenty-eighth laboratory intercomparison. The numbers in front of the names are used in tables and figures.

Armenia	(45)	Dept of Environm. Health and Research
Belgia	(2)	Flemish Environment Agency, Antwerpen
Croatia	(35)	Meteorological and Hydrological Service of Croatia
Czech Republic	(3)	Czech Hydrometeorological Institute, Praha
Denmark	(4)	National Environmental Research Institute. Air Pollution Laboratory
Estonia	(38)	Estonian Environmental Research Centre, Tallinn
Finland	(5)	Finnish Meteorological Institute. Air Quality Department
France	(41)	Micro Pollutants Technology
Georgia	(43)	Centre for Monitoring and Prognostication, Tbilisi
Germany	(7)	IfE Leipzig GmbH, Umweltlabor
Germany	(8)	Umweltbundesamt, Messtelle Schauinsland
Hungary	(10)	Institute for Atmospheric Physics
Ireland	(12)	Met. Eirann, Dublin
Italy	(13)	C.N.R. Istituto Inquinamento Atmosferico
Italy	(30)	Joint Research Centre, Ispra
Latvia	(33)	Air Pollution Observation Laboratory
Lithuania	(32)	Environmental Physics and Chemistry Laboratory
Netherlands	(14)	National Institute of Public Health and Environmental Protection (RIVM)
Norway	(15)	Norwegian Institute for Air Research (NILU)
Macedonia	(40)	Hydrometeorological Institute, Skopje
Moldova	(42)	State Hydrometeorological Service, Chisinau
Poland	(16)	Institute of Meteorology and Water Management, Warsaw
Poland	(46)	Diabla Gora station
Portugal	(17)	Laboratorio Santo Andre
Russian Federation	(22)	Institute of Global Climate and Ecology
Serbia	(24)	Rep. Hydrometeorological Institute of Serbia
Slovakia	(31)	Slovak Hydrometeorological Institute
Slovenia	(36)	Hydrometeorological Institute of Slovenia
Spain	(19)	Centro Nacional de Sanidad Ambiental
Sweden	(20)	Swedish Environmental Research Institute (IVL), Gothenburg
Switzerland	(21)	Swiss Federal Laboratories for Materials Testing (EMPA)
United Kingdom	(23)	AEA Technology, National Environmental Technology Centre
United Kingdom	(167)	CEH Edinburgh

Table 30b: Participating laboratories outside the EMEP network.

Canada	(26)	Meteorological Service of Canada, Toronto
United States of America	(27)	Illinois State Water Survey
Germany	(104)	Hessige Landwirtschaftliche
Finland	(107)	The Finnish Forest Institute
Germany	(110)	Thüringer Landesanstalt für Landwirtschaft (TTL), Jena
Germany	(112)	Niedersächsische Forstliche Versuchsanstalt (N VF)
Germany	(115)	Bayerische Landesanstalt f. Wald- und Forstwirtschaft
Switzerland	(116)	Institute for Applied Plant Biology
Germany	(118)	Forstliche Versuchs-und Forschungsanstalt
Germany	(120)	Landwirtschaftliche Untersuchungs- und Forschungsanstalt (LUFA)
Germany	(121)	Landeslabor Schleswig-Holstein
Belgium	(124)	Laboratorium voor Bondemkunde, Gent
Germany	(125)	Bayerisches Landesamt für Umweltschutz, Augsburg
Italy	(126)	APPA Laboratorio Biologico Provinciale
Spain	(150)	Fundación Centro de Estudios ambientales del mediterráneo
Slovenia	(153)	Slovenian Forestry Institute, Ljubljana
United Kingdom	(155)	Environmental Research Branch, Forest Research
Hungary	(157)	Ecological Laboratory of Forest research Institute
Japan	(158)	Acid Deposition and Oxidant Research Center (ADOCRC), Niigata
Slovenia	(161)	National Institute of Chemistry, Ljubljana
Thailand	(163)	Environmental Researching and Training Center (ERTC)
Thailand	(164)	Pollution Control Department (PCD)
Poland	(166)	Forest Research Institute, Laboratory of Forest Habitat Chemistry
France	(172)	UMR SAS INRA
Russian Federation	(176)	Main Geophysical Observatory
Russian Federation	(178)	Environmental Chemistry and EANET Monitoring Laboratory. Limnological Institute RAS/SB, Irkutsk
Russian Federation	(179)	Murmansk Environmental Monitoring Centre
Germany	(180)	vTI-Institut für Agrarrelevante Klimaforschung

Table 31: Analytical methods used at the participating laboratories for the determination of sulphur dioxide in absorbing solution (A).

Method	Laboratory
3. Ion chromatography	15, 17, 19, 23, 36, 159, 173
4. Spectrophotometry	178

Table 32: Analytical methods used at the participating laboratories for the determination of sulphur dioxide on impregnated filters (B).

Method	Laboratory
4. Ion chromatography	3, 4, 5, 8, 15, 20, 22, 31, 32, 33, 36, 38, 41, 158, 163, 178
5. Spectrophotometry	16

Table 33: Analytical methods used at the participating laboratories for determination of nitric acid on impregnated filters (B).

Method	Laboratory
1. Ion chromatography	3, 4, 5, 8, 15, 20, 22, 31, 32, 33, 38, 41, 158, 163, 178
2. Spectrophotometry	16

Table 34: Analytical method for determination of ammonia on impregnated filters (J).

Method	Laboratory
1. Spectrophotometry	3, 4, 8, 10, 15, 19, 32, 33
3. Ion chromatography	5, 13, 31, 36, 38, 41, 158, 178, 180

Table 35: Analytical method used for NO₂ in absorbing solution (C).

Method	Laboratory
1. Spectrophotometry	3, 8, 10, 12, 15, 16, 23, 24, 31, 32, 33, 35, 36, 42, 173
2. Ion chromatography	19

Table 36: Analytical methods used for the determination of chemical constituents in precipitation samples

Table 36, cont.

Lab no	Network	SO_4^{2-}	NH_4^+	NO_3^-	Na^+	Mg^{2+}	Cl^-	Ca^{2+}	K^+
120	ICP-Forest	ICP-AES	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Spectrophotometry	ICP-AES	ICP-AES
121	EMEP	Spectrophotometry	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Potensiometric method	ICP-AES	ICP-AES
124	Other	Ion chromatography	Ion chromatography	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
125	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
126		Ion chromatography	Spectrophotometry	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
150	Other	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
153	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
155	ICP-Forest	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
157	ICP-Forest	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
158	EANET	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
161	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
163	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
164	EANET								
166	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
167	EMEP	Ion chromatography	Ammonia Flow Injection Analysis	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
167	EMEP	Ion chromatography	Ammonia Flow Injection Analysis	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
172	EMEP		Spectrophotometry						
176	EMEP	Spectrophotometry	Spectrophotometry	Spectrophotometry	Flame-AES	Flame-AAS	Potensiometric method	Flame-AAS	Flame-AES
178	EANET	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
179	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Ion chromatography	Ion chromatography	Flame-AAS	Flame-AAS
180	Other	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography

Table 37: Reported results for precipitation samples expressed as % deviation from expected value.

Lab no	Precipitation																pH				Cond				pH-units from expected value				% deviation from expected value											
	SO ₄ ²⁻				NH ₄ ⁺				NO ₃ ⁻				Na ⁺				Mg ²⁺				Cl ⁻				Ca ²⁺				K ⁺				pH-units from expected value				% deviation from expected value			
	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4								
2	9	11	9	8	20	16	15	14	-3	-2	-3	-1	-7	-6	-10	-7	6	5	7	3	6	11	6	4	-39	-20	-14	-17	-1	-1	-3	-5	-0.31	-0.25	-0.32	-0.32	-90	-90	-90	-90
3	1	1	0	0	18	11	9	8	-2	-1	-2	-3	0	3	-3	-5	-7	-4	-6	-5	-1	7	0	2	1	10	1	3	0	-3	-6	-3	0.058	0.064	0.031	0.031	-2	-3	9	1
4	-4	-3	-4	-3	-1	-3	-2	-2	-4	-3	-3	-3	-7	-3	-7	-5	-6	-6	-6	-5	10	10	6	8	-1	-2	-4	-4	-7	-8	-9	-8	0.04	0.04	0.04	0.04	-1	-4	0	-1
5	1	1	1	0	-2	-3	-5	-2	0	2	0	0	9	9	7	4	0	0	-2	-1	3	5	2	0	1	0	-1	-1	15	16	10	6	0.02	0.02	0.02	0.02	-1	-4	0	-1
7	7	7	4	4	-3	-3	-2	0	2	3	2	2	-4	-4	-6	-7	6	4	2	0	11	11	5	2	5	5	2	-3	0	0	1	3	0.11	0.08	0.07	0.07	-21	-15	-6	-7
8	-3	-2	-3	-4	0	0	-2	-2	0	-1	-1	-6	-4	-1	-4	-3	0	0	-1	-4	0	-2	-2	2	0	-2	-1	-3	-3	-4	-3	0.06	0.06	0.05	0.05	-4	-5	6	4	
10	-1	4	-1	-1	-6	-6	-6	-6	-6	-3	-1	-1	-5	-4	-8	-8	-2	-2	-3	-2	-10	-19	-7	-16	6	-1	6	3	1	-3	-5	0.11	0.1	0.11	0.11	8	-1	2	-1	
12	-1	-2	-1	-2	0	3	1	2	0	-2	-3	-2	-2	-6	-12	-6	-6	-9	-2	-5	-5	-6	-69	-41	-38	-36	-7	-8	-9	-8	0.04	0.03	0.03	0.03	1	-2	-3	-1		
13	6	4	1	4	13	12	9	9	6	-2	-3	0	16	15	9	11	-3	-4	3	4	17	19	12	10	15	-2	-1	8	3	5	2	-2	0.24	0.36	0.31	0.31	-27	-21	-10	-16
14	-4	-5	-6	-9	5	-2	-2	-2	-6	25	-7	-4	3	1	12	5	-2	-1	4	1	-5	-21	-15	-10	4	5	9	8	-2	-5	-2	-4	0.12	0.11	0.14	0.14	5	2	7	5
15	1	4	14	6	3	1	0	1	-2	-1	-3	0	4	4	2	0	-1	-4	-6	-6	-23	-16	-5	-4	-10	-7	-10	-8	0	-3	1	5	0.07	0.08	0.07	0.07	3	1	6	4
16	-1	-2	-2	-2	4	-7	-5	-6	-3	-4	-4	-3	1	1	0	-1	-2	-1	2	6	6	1	-1	10	3	-1	3	14	1	-2	-1	0.03	0.02	0.03	0.03	0	-2	1	0	
17	0	-2	-5	-5	15	8	7	3	10	-5	17	11	27	34	25	10	47	16	3	-1	119	34	26	36	23	26	43	16	-0.37	-0.13	-0.42	-0.42	-16	-9	0	-13				
19	-6	-6	-3	-5	6	7	5	5	-10	-7	-7	-6	11	15	9	7	3	8	7	4	0	4	18	4	3	-2	1	-4	-8	0.21	0.27	0.22	0.22	-3	-5	-2	-1			
20	3	1	1	-1	0	3	1	-1	0	-2	-3	-2	2	5	2	0	-3	-11	-5	3	4	4	3	0	3	1	2	-2	0.1	0.03	0.08	0.08	0	-1	4	3				
21	5	5	4	4	4	1	2	3	2	2	2	3	6	6	4	2	3	0	2	2	3	6	1	2	2	0	-1	0	0.05	0.05	0.04	0.04	-2	-4	0	-1				
22	-5	-7	-10	-10	36	26	20	16	-9	-7	-5	-4	21	-4	6	-4	1	-11	-10	-4	8	-4	-3	-12	44	-7	-20	-17	2	-7	-9	-13	0.07	0.05	0.05	0.05	-13	-17	-12	-12
23	-3	-3	-3	-3	7	6	8	8	3	3	3	3	-3	1	-1	0	4	4	3	-12	-5	-6	-5	14	11	6	12	-3	-2	0	-5	0	0.02	0	0	0	-20	-17	-11	-13
24	-1	0	2	-3	-9	-8	-1	1	-6	-6	-5	-6	-7	-6	-8	-8	-12	-14	-12	-10	-16	-10	-1	-6	-2	-5	-2	3	-11	-11	-12	-14	-0.44	-0.41	-0.52	-0.52	-4	-13	0	0
26	-2	-2	-2	-2	2	0	0	0	0	-1	-1	-2	-1	-3	-4	-4	-2	-2	1	7	2	1	1	5	3	1	-8	-8	-5	-7	0.04	0.05	0.02	0.02	-2	-3	2	1		
27	-1	-2	-1	-2	1	1	1	0	-1	-1	-2	4	2	0	0	-1	0	0	1	3	0	0	4	3	2	1	0	-2	-3	-5	0.04	0.05	0.02	0.02	-2	-3	2	1		
30	-7	-7	-9	-9	6	-2	1	5	-6	-4	-3	-2	2	-1	-1	-4	10	8	7	4	4	4	-1	-9	-15	18	20	18	-17	-16	-20	-22	0.38	0.06	0.07	0.07	1	-3	3	0
31	-1	-1	-3	-2	15	8	6	3	-4	-2	-3	-2	0	1	-3	-3	-2	-2	-8	-10	-4	-4	-3	-4	3	1	-5	-7	-3	-3	-16	-8	0.06	0.06	0.06	0.06	-1	-2	3	1
32	0	1	1	1	3	-8	-9	-8	-4	-2	-1	-1	-7	-2	-2	0	-3	-2	-2	-1	6	0	-1	4	-22	-11	3	-2	-5	-4	-2	0.05	0.03	0.01	0.01	-11	-10	-4	-4	
33	4	4	1	3	-12	-11	-5	-11	19	20	24	15	11	10	7	6	-9	-7	-3	-3	10	9	8	7	3	-7	-6	0	0	-3	-7	-8	0.04	0.05	0.04	0.04	1	-1	5	3
35	3	2	-5	2	6	0	-2	2	4	-3	-3	-1	6	0	1	-3	-8	-16	-10	-8	-18	-9	-9	-6	-23	-19	-15	-15	3	3	1	1	0.05	0.05	0.05	0.05	0	-2	0	-6
36	-1	-2	-2	-2	-1	-4	-2	1	-5	-4	-5	-4	2	4	3	1	-1	1	3	-19	-15	-16	-13	16	7	-1	3	2	3	5	7	0.1	0.1	0.1	0.1	-7	-8	-4	-8	
38	2	3	0	1	0	-2	-3	-1	-6	0	-3	-2	-11	-11	-11	-15	10	-4	17	-10	-16	-7	-9	-13	88	50	72	52	-22	-20	-20	-22	0.13	0.05	0	0	-1	-24	17	55
40	4	-3	-3	-6	-6	15	13	5	1	-11	-12	-10	4	-1	-2	-6	28	25	21	20	5	1	-5	-13	11	29	24	7	6	15	9	10	0.2	0.08	0.08	0.08	-28	-30	-26	-26
41	-3	-3	-6	-6	16	-16	-17	-17	-1	-1	-2	-6	28	25	21	20	5	1	-5	-13	11	29	24	7	6	15	9	10	0.2	0.08	0.08	0.08	-28	-30	-26	-26				
42	-29	-34	-10	-36	6	-12	-4	-1	143	4	222	-20	26	27	21	27	28	23	12	27	574	449	343	85	124	60	55	68	39	19	18	26	0.06	0.04	-0.02	-0.02	-16	-18	-15	-12
43	12	8	12	9	80	-56	-94	-66	9	9	2	-2	8	7	2	0	3	5	5	3	-6	-1	-2	-2	28	25	10	13	4	3	-3	-3	-0.1	-0.07	-0.12	-0.12	-21	-6	162	-8
45	20	22	19	19	3	-2	-2	-1	-1	-1	-1	-1	8	7	2	0	3	5	5	3	-6	-1	-2	-2	122	8	4	2	2	2	-1	0	-1	-3	1	0				
46	104	200	201	198	0	-2	-3	2	-13	-14	-14	-14	-12	-11	-13	-12	-12	-13	-14	-12	-3	9	-1	0	-10	-12	-16	-15	-2	-11	-16	-10	2.89	1.68	1.14	1.14	-26	-7	-5	0
107	195	198	210	-5	-3	-2	0	-6	-4	1	-5	-6	-6	-5	-4	-3	-3	-4	-12	-6	-3	-4	-12	-36	-38	-22	-22	-17	3	9	12	18	-0.12	-0.09	-0.23	-0.23	-17	-13	-7	-4
110	6	-1	6	-2	-25	-25	-21	-25	45	10	28	11	-2	2	-4	0	-3</td																							

Table 37, cont.

 SO₄²⁻ and NO₃⁻ between \pm 10 and 20%
 SO₄²⁻ and NO₃⁻ more than + 20%

NH₄⁺, NO₃⁻, Cl⁻, Na⁺, Mg²⁺, Ca²⁺, K⁺ and cond: between \pm 15-25%

 pH: between \pm 0,1-0,2 pH-units
 pH: more than \pm 0,2 pH- units

Table 38: Reported results for filter samples expressed as % deviation from expected value.

Air and aerosols																							
	Absorbing solution					Impregnated filter					Impregnated filter					Absorbing solution				Impregnated filter			
	SO ₂ -S				% deviation form expected value	SO ₂ -S				% deviation form expected value	HNO ₃ -N				% deviation form expected value	NO ₂ -N				NH ₃ -N			
	A1	A3	A4	A5	B1	B2	B3	B4	B5	B1	B2	B3	B4	B5	C1	C2	C3	C4	J1	J2	J3	J4	
3					7	7	3	5		1	4	4	4	7	-10	-7	-10	-4	1	0	-4	-1	
4					-17	-13	-6	-2		-20	-9	-10		-5					1	-4	-3	-1	
5					1	1	2	2		-9	0	1	3						0	-3	-3	-8	
8					-1	-1	0	1		-2	-1	0	3		-2	-1	-1	-1	-11	3	-3	-3	
10															-16	-9	-9	-8	-22	-20	-26	-28	
12															-3	1	-3	0					
13																			20	10	7	10	
15	-9	-8	1	14	7	1	0	3		-5	-2	-1	1		5	-4	12	-6	19	12	11	7	
16					5	-3	1	0		-8	-4	-1	1		0	-4	0	-3					
17	-7	-3	-9	-5																			
19	-6	8	19	25												38	41	47	48	-20	-3	-8	-24
20						-1	-1	4	6	-14	-14	-10		-9									
22						-9	-10	-5	-2	-3	1	5	11										
23	3	1	1	5												-2	-1	0	-1				
24																-5	-6	-3	-5				
31						-14	-15	-12	-11	-6	-4	-3	0		-2	-4	-4	-5	-5	-8	2	-7	
32						2	4	2	1	-7	7	8	9		3	5	0	5	0	13	1	5	
33						5	5	5	7	-5	-6	-4	0		9	7	9	8	0	-10	8	3	
35															-3	-4	-4	-3					
36	14	14	14		7	22	18	5	3	-13	-5	-5	-2		0	-4	-4	-4	-10	0	-5	-2	
38						-7	-2	2	2									5	0	5	5		
41						-3	-1	0	1	-23	-14	-14	-9					25	17	10	26		
42															176	253	312	243					
158						13	15	4	9	1	2	0	6					-7	-10	-14	-2		
159						19	42	59	73														
163						-3	-1	-1	1	-3	-1	2	3										
173	-31	-8	0	2											-3	-5	-4	-7	-21	-30	-43		
178	-5	-5	-5	-6		29	-8	2	3	12	5	15	7					-1	-15	-1	-6		
180																							



Results between 10 and 20% or between -10 and -20% from expected value
more than $\pm 20\%$ from expected value

Table 39: Relative random and systematic errors obtained by the different laboratories in the analysis of each parameter in the precipitation.

Lab. no.	SO_4^{2-}		NO^{3-}		NH_4^+		Mg^{2+}		H^+ calc	
	Random error %	Systematic error %								
2	2	9	1	-2	2	16	1	5	19	92
3	0	0	1	-2	1	10	2	-6	5	-10
4	1	-3	0	-3	1	-3	1	-6	2	-8
5	1	1	1	0	2	-3	1	0	1	-5
7	2	6	1	2	1	-2	3	2	7	-17
8	0	-3	3	-1	1	-1	0	0	3	-11
10					2	-6	0	-3	26	-26
12	0	-1	1	-2	1	2	1	-8	2	-7
13	2	4	3	-1	1	10	4	-3	14	-46
14	2	-6	13	-5	3	-2	2	0	4	-24
15	4	5	1	-1	1	1	2	-4	4	-14
16	0	-2	0	-4	4	-6	1	0	1	-6
17	2	-3	3	8			9	23	99	82
19	2	-5	0	-7	1	6	1	6	12	-41
20	2	1	1	-2	2	1	2	-4	8	-10
21	1	5	1	2	1	2	1	2	3	-10
22	1	-8	1	-6	2	24	4	-6	4	-11
23	1	-3	1	3	2	7	1	3	2	0
24	2	-1	1	-5	3	-4	2	-11	23	186
26	0	-2	1	0	1	0	1	-3		
27	1	-1	1	-1	0	1	0	0	4	-8
30	0	-8	1	-4	3	3	3	6	27	-13
31	1	-1	0	-3	3	7	4	-5	3	-14
32	0	1	1	-2	5	-9	4	-4	5	-7
33	1	3	4	19	3	-9	4	-4	3	-8
35	3	2	1	-3	3	1	1	-10	2	-10
36	0	-2	1	-5	2	-2	2	1	5	-21
38	1	1	2	-3	1	-2	11	5	13	-9
40			7	-9	4	8			50	396
41	1	-4	4	-17	2	7	7	21	13	-17
42	12	-33	103	56	6	-3	10	23	13	-3
43	2	10	5	4	23	-72			29	-68
45	3	20	1	-1	2	-2	0	4		
46									2	-6
104	28	204	3	-14	2	-1	2	-13	24	-94
107	17	206	3	-5	1	-2			6	4
110	3	2	10	21	6	-24				
112	0	1	1	2	1	3	3	-3	6	-26
115	1	1	1	1	4	-8	1	-9	6	-25
116	3	0	1	-6	2	22	2	3	1	-18
118	1	3	1	-2	2	-6	1	-5	3	0
120	1	-5	1	-8	2	15	3	-13	3	-8
121	2	-8	0	-1	2	-4	3	-9	13	-11
124	4	-3	3	-4	4	-10	1	-1	4	-10
125	1	0	0	-1	1	-5	5	-15		
126	3	-10	2	-8	14	-4	4	-37	20	1
150	1	-2	2	-6	2	7	3	4	11	-22
153	1	-3	1	-4	1	-3	2	-6	5	-32
155	3	-7	1	0	3	-6	1	0	4	-17
157	3	3	6	11	3	-4	21	-11		
158	1	2	1	0	2	4	1	0	3	-3
161	2	12	2	4						
163	0	-1	0	-2	0	2	2	8	4	-12
164	4	4	11	22	35	1	1	-3	3	-6
166	1	4	2	-1	4	-1	1	-1	5	-11
167	9	15	18	1	2	-8	20	-12		
172					0	8			24	177
176	7	-17	5	3	14	6	8	-29	7	-6
178	1	-3	7	27	3	-5	4	14	2	5
179	101	179	1	5	2	-3	15	14		
180	3	1	1	0	0	1	1	4		

Table 39, cont.

Lab. no.	Na+		Cl-		K+		Cond.	
	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %
2	2	-7	2	6	1	-2	15	-90
3	4	-1	3	1	2	-3	4	-1
4	2	-5	2	8	1	-8		
5	1	7	2	2	6	13	2	-1
7	3	-4	3	7	1	1	7	-12
8	1	-3	1	-2	1	-4	5	0
10	3	-6			3	-1		
12	2	-1	3	-6	1	-8	1	-2
13	3	13	2	13	3	2	9	-20
14	4	5	6	-15	2	-3	1	5
15	1	2	4	-10	3	0	1	4
16	1	0	3	3	6	0	1	0
17	9	11	14	8	12	28	7	-11
19	3	9			3	-3	2	-2
20	2	2	1	3	2	1	2	1
21	1	4	1	2	3	6	2	-1
22	9	1	9	-3	5	-8	5	-13
23	1	0	2	-7	1	-3	5	-16
24	3	-6	4	-9	1	-12	7	-2
26	2	-2	2	2	2	-7		
27	2	1	1	0	1	-2	2	-1
30	3	-1	7	1	1	-18	2	1
31	2	-1	4	-3	5	-5	2	1
32	3	-1	3	0	2	-3	4	-8
33	1	8	1	8	3	-5	2	2
35	3	1	2	-9	1	2	2	-1
36	1	2	2	-15	1	5	3	-7
38	6	-10	5	-10	3	-21	32	6
40			3	9			2	-16
41	3	-1	9	-2	5	8	6	-27
42	8	22	114	375	10	22	4	-14
43			243	18			62	-7
45	2	4	1	-2	3	0	2	-1
46							5	-4
104	4	-11	4	0	5	-10	11	-6
107			6	-5			1	-2
110	2	-1			4	12	6	-11
112	1	6	8	13	4	4	3	-4
115	3	5	3	44	3	5	3	-9
116	3	4	3	11	4	5	7	0
118	6	-15	3	-8	1	-13	2	-7
120	1	6	23	33	4	3	2	0
121	6	-6	8	5			4	2
124	3	5	9	8	10	-9	3	-4
125	3	-9	34	5	2	-8		
126	50	-13	3	-6	3	-13	12	-2
150	6	4	7	-4	5	20	2	-4
153	1	-1	3	-8	3	-9	3	-8
155	1	4	2	-5	4	-1	2	0
157	8	11	15	38	11	0		
158	1	4	5	4	1	4	2	-2
161			3	5				
163	1	4	2	-5	1	1	2	-2
164	1	4	6	1	2	1	1	-7
166	2	-7	2	-2	2	-7	2	-2
167	43	-3	26	16	13	-24		
172							3	0
176	16	6	20	15	56	58	7	-4
178	4	22	1	3	8	-54	2	1
179	12	1	7	-18	12	-12		
180	2	3	4	0	3	-17		

Table 40: Analytical results for sulphur dioxide in absorbing solution.

SO ₂ -S in absorbing solution Sample no.: A1 Theoretical value: 0.160 Unit: µg S/ml	SO ₂ -S in absorbing solution Sample no.: A3 Theoretical value: 0.240 Unit: µg S/ml
Run 1:	Run 1:
Number of laboratories: 8 Arithmetic mean value: 0.155 Median: 0.151 Standard deviation 0.025 Rel. st. deviation (%) 15.885	Number of laboratories: 8 Arithmetic mean value: 0.252 Median: 0.237 Standard deviation 0.040 Rel. st. deviation (%) 15.996
Run 2:	Run 2:
Number of laboratories: 8 Arithmetic mean value: 0.155 Median: 0.151 Standard deviation 0.025 Rel. st. deviation (%) 15.885	Number of laboratories: 7 Arithmetic mean value: 0.240 Median: 0.232 Standard deviation 0.021 Rel. st. deviation (%) 8.633
Results in decreasing order:	Results in decreasing order:
159 0.190 19 0.150 36 0.183 17 0.149 23 0.164 15 0.146 178 0.151 173 0.110	159 0.340 (*) 17 0.232 36 0.274 178 0.228 19 0.260 15 0.220 23 0.243 173 0.220
SO ₂ -S in absorbing solution Sample no.: A4 Theoretical value: 0.320 Unit: µg S/ml	SO ₂ -S in absorbing solution Sample no.: A5 Theoretical value: 0.440 Unit: µg S/ml
Run 1:	Run 1:
Number of laboratories: 8 Arithmetic mean value: 0.352 Median: 0.324 Standard deviation 0.070 Rel. st. deviation (%) 19.931	Number of laboratories: 8 Arithmetic mean value: 0.503 Median: 0.465 Standard deviation 0.113 Rel. st. deviation (%) 22.480
Run 2:	Run 2:
Number of laboratories: 7 Arithmetic mean value: 0.330 Median: 0.324 Standard deviation 0.032 Rel. st. deviation (%) 9.615	Number of laboratories: 7 Arithmetic mean value: 0.466 Median: 0.461 Standard deviation 0.048 Rel. st. deviation (%) 10.328
Results in decreasing order:	Results in decreasing order:
159 0.510 (*) 23 0.324 19 0.380 173 0.320 36 0.364 178 0.305 15 0.324 17 0.290	159 0.760 (*) 23 0.461 19 0.550 173 0.450 15 0.503 17 0.416 36 0.469 178 0.414

Table 41 Analytical results for sulphur dioxide on impregnated filter.

SO₂-S on impregnated filter
 Sample no.: B1
 Theoretical value: 18.000
 Unit: ug S/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 18.216
 Median: 18.030
 Standard deviation 2.168
 Rel. st. deviation (%) 11.904

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 17.923
 Median: 17.900
 Standard deviation 1.832
 Rel. st. deviation (%) 10.219

Results in decreasing order:

178	23.196	(*)	8	17.900
36	21.920		20	17.800
158	20.370		163	17.540
15	19.300		41	17.470
3	19.290		38	16.700
33	18.927		22	16.400
16	18.900		31	15.436
32	18.270		31	15.436
5	18.160		4	14.880

SO₂-S on impregnated filter
 Sample no.: B4
 Theoretical value: 48.100
 Unit: ug S/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 47.896
 Median: 48.785
 Standard deviation 2.529
 Rel. st. deviation (%) 5.280

Run 2:

Number of laboratories: 16
 Arithmetic mean value: 48.616
 Median: 49.010
 Standard deviation 1.508
 Rel. st. deviation (%) 3.101

Results in decreasing order:

33	50.623		16	48.570
36	50.290		41	48.110
158	50.060		15	48.000
20	49.800		8	47.900
3	49.400		163	47.640
32	49.270		22	45.650
178	49.249		4	45.280
5	49.020		31	42.137 (*)
38	49.000		31	42.137 (*)

SO₂-S on impregnated filter
 Sample no.: B2
 Theoretical value: 22.000
 Unit: ug S/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 21.760
 Median: 21.750
 Standard deviation 2.014
 Rel. st. deviation (%) 9.255

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 21.509
 Median: 21.720
 Standard deviation 1.763
 Rel. st. deviation (%) 8.197

Results in decreasing order:

36	26.020	(*)	163	21.720
158	25.230		8	21.700
3	23.550		38	21.500
33	23.126		16	21.360
32	22.890		178	20.303
15	22.300		22	19.700
5	22.130		4	19.080
20	21.800		31	18.743
41	21.780		31	18.743

SO₂-S on impregnated filter
 Sample no.: B5
 Theoretical value: 64.100
 Unit: ug S/filter

Run 1:

Number of laboratories: 18
 Arithmetic mean value: 64.648
 Median: 65.070
 Standard deviation 3.333
 Rel. st. deviation (%) 5.156

Run 2:

Number of laboratories: 16
 Arithmetic mean value: 65.613
 Median: 65.200
 Standard deviation 1.911
 Rel. st. deviation (%) 2.913

Results in decreasing order:

158	69.620		163	65.040
33	68.360		41	64.920
20	67.800		32	64.800
3	67.220		8	64.500
178	66.191		16	63.840
15	65.900		22	62.840
36	65.840		4	62.540
5	65.300		31	56.929 (*)
38	65.100		31	56.929 (*)

Table 42: Analytical results for nitric acid on impregnated filter.

HNO₃-N on impregnated filter
 Sample no.: B1
 Theoretical value: 14.720
 Unit: ug N/filter

Run 1:

Number of laboratories: 17
 Arithmetic mean value: 13.769
 Median: 13.887
 Standard deviation 1.210
 Rel. st. deviation (%) 8.788

Run 2:

Number of laboratories: 15
 Arithmetic mean value: 13.748
 Median: 13.887
 Standard deviation 0.840
 Rel. st. deviation (%) 6.107

Results in decreasing order:

178	16.527	(*)	31	13.887
3	14.870		32	13.660
158	14.810		16	13.530
8	14.400		5	13.450
22	14.350		36	12.880
163	14.215		20	12.600
15	14.000		4	11.720
33	13.960		41	11.330 (*)
31	13.887			

HNO₃-N on impregnated filter
 Sample no.: B4
 Theoretical value: 19.630
 Unit: ug N/filter

Run 1:

Number of laboratories: 17
 Arithmetic mean value: 19.437
 Median: 19.470
 Standard deviation 1.377
 Rel. st. deviation (%) 7.085

Run 2:

Number of laboratories: 16
 Arithmetic mean value: 19.239
 Median: 19.435
 Standard deviation 1.144
 Rel. st. deviation (%) 5.946

Results in decreasing order:

178	22.613	(*)	15	19.400
32	21.150		31	18.964
22	20.630		31	18.964
3	20.490		33	18.766
163	19.945		36	18.600
5	19.900		4	17.660
158	19.710		20	17.600
8	19.600		41	16.970
16	19.470			

HNO₃-N on impregnated filter
 Sample no.: B2
 Theoretical value: 18.000
 Unit: ug N/filter

Run 1:

Number of laboratories: 17
 Arithmetic mean value: 17.525
 Median: 17.700
 Standard deviation 1.083
 Rel. st. deviation (%) 6.178

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 17.525
 Median: 17.700
 Standard deviation 1.083
 Rel. st. deviation (%) 6.178

Results in decreasing order:

32	19.240		31	17.316
178	18.972		31	17.316
3	18.730		16	17.280
158	18.340		36	17.070
22	18.130		33	16.938
5	17.930		4	16.440
163	17.905		41	15.420
8	17.800		20	15.400
15	17.700			

HNO₃-N on impregnated filter
 Sample no.: B5
 Theoretical value: 26.180
 Unit: ug N/filter

Run 1:

Number of laboratories: 17
 Arithmetic mean value: 26.596
 Median: 26.500
 Standard deviation 1.506
 Rel. st. deviation (%) 5.664

Run 2:

Number of laboratories: 17
 Arithmetic mean value: 26.596
 Median: 26.500
 Standard deviation 1.506
 Rel. st. deviation (%) 5.664

Results in decreasing order:

22	29.190		16	26.400
32	28.500		31	26.195
3	28.090		31	26.195
178	28.047		33	26.160
158	27.640		36	25.740
163	27.055		4	25.000
8	27.000		20	23.800
5	26.920		41	23.700
15	26.500			

Table 43: Analytical results for nitrogen dioxide in absorbing solution.

NO₂-N in absorbing solution
 Sample no.: C1
 Theoretical value: 0.058
 Unit: ug N/ml

Run 1:

Number of laboratories:	16
Arithmetical mean value:	0.065
Median:	0.057
Standard deviation	0.026
Rel. st. deviation (%)	40.624

Run 2:

Number of laboratories:	15
Arithmetical mean value:	0.058
Median:	0.057
Standard deviation	0.007
Rel. st. deviation (%)	11.790

Results in decreasing order:

42	0.160 (*)	23	0.057
19	0.080	31	0.057
33	0.063	12	0.056
15	0.061	35	0.056
32	0.060	173	0.056
16	0.058	24	0.055
36	0.058	3	0.052
8	0.057	10	0.049

NO₂-N in absorbing solution
 Sample no.: C3
 Theoretical value: 0.068
 Unit: ug N/ml

Run 1:

Number of laboratories:	16
Arithmetical mean value:	0.082
Median:	0.067
Standard deviation	0.054
Rel. st. deviation (%)	65.071

Run 2:

Number of laboratories:	15
Arithmetical mean value:	0.069
Median:	0.066
Standard deviation	0.009
Rel. st. deviation (%)	13.615

Results in decreasing order:

42	0.280 (*)	12	0.066
19	0.100	24	0.066
15	0.076	31	0.065
33	0.074	35	0.065
16	0.068	36	0.065
23	0.068	173	0.065
32	0.068	10	0.062
8	0.067	3	0.061

NO₂-N in absorbing solution
 Sample no.: C2
 Theoretical value: 0.085
 Unit: ug N/ml

Run 1:

Number of laboratories:	16
Arithmetical mean value:	0.099
Median:	0.082
Standard deviation	0.055
Rel. st. deviation (%)	55.213

Run 2:

Number of laboratories:	15
Arithmetical mean value:	0.085
Median:	0.082
Standard deviation	0.010
Rel. st. deviation (%)	11.975

Results in decreasing order:

42	0.300 (*)	16	0.082
19	0.120	31	0.082
33	0.091	35	0.082
32	0.089	36	0.082
12	0.086	173	0.081
8	0.084	24	0.080
23	0.084	3	0.079
15	0.082	10	0.077

NO₂-N in absorbing solution
 Sample no.: C4
 Theoretical value: 0.108
 Unit: ug N/ml

Run 1:

Number of laboratories:	16
Arithmetical mean value:	0.125
Median:	0.105
Standard deviation	0.067
Rel. st. deviation (%)	53.299

Run 2:

Number of laboratories:	15
Arithmetical mean value:	0.109
Median:	0.105
Standard deviation	0.015
Rel. st. deviation (%)	13.630

Results in decreasing order:

42	0.370 (*)	35	0.105
19	0.160	3	0.104
33	0.117	36	0.104
32	0.113	24	0.103
12	0.108	31	0.103
8	0.107	15	0.101
23	0.106	173	0.100
16	0.105	10	0.099

Table 44: Analytical results for ammonia on impregnated filter. The reported results are corrected for an average blank value (J1 and J4).

NH3-N on impregnated filter Sample no.: J2 Theoretical value: 14.020 Unit: ug N/filter	NH3-N on impregnated filter Sample no.: J3 Theoretical value: 20.050 Unit: ug N/filter
Run 1:	Run 1:
Number of laboratories: 17 Arithmetic mean value: 13.802 Median: 13.990 Standard deviation 1.907 Rel. st. deviation (%) 13.819	Number of laboratories: 17 Arithmetic mean value: 19.471 Median: 19.395 Standard deviation 2.456 Rel. st. deviation (%) 12.615
Run 2:	Run 2:
Number of laboratories: 17 Arithmetic mean value: 13.802 Median: 13.990 Standard deviation 1.907 Rel. st. deviation (%) 13.819	Number of laboratories: 16 Arithmetic mean value: 19.814 Median: 19.700 Standard deviation 2.075 Rel. st. deviation (%) 10.474
Results in decreasing order:	Results in decreasing order:
41 17.493 180 13.870 13 16.800 31 13.309 15 16.680 158 13.000 38 14.705 36 12.650 4 14.140 8 12.500 3 14.104 19 11.270 33 14.078 178 11.035 5 14.065 10 10.951 32 13.990	41 23.463 19 19.380 32 22.640 4 19.160 15 22.530 31 18.409 13 22.000 33 18.128 8 20.600 158 17.990 36 20.150 180 17.080 3 20.015 10 16.082 38 20.005 178 13.989 (*) 5 19.395
NH3-N on impregnated filter Sample no.: J5 Theoretical value: 36.090 Unit: ug N/filter	NH3-N on impregnated filter Sample no.: J6 Theoretical value: 11.030 Unit: ug N/filter
Run 1:	Run 1:
Number of laboratories: 17 Arithmetic mean value: 34.681 Median: 35.040 Standard deviation 4.857 Rel. st. deviation (%) 14.005	Number of laboratories: 16 Arithmetic mean value: 10.852 Median: 10.830 Standard deviation 1.387 Rel. st. deviation (%) 12.777
Run 2:	Run 2:
Number of laboratories: 16 Arithmetic mean value: 35.552 Median: 35.450 Standard deviation 3.376 Rel. st. deviation (%) 9.496	Number of laboratories: 14 Arithmetic mean value: 10.838 Median: 10.830 Standard deviation 0.917 Rel. st. deviation (%) 8.463
Results in decreasing order:	Results in decreasing order:
15 40.150 5 34.905 41 39.543 8 34.900 33 38.828 3 34.605 13 38.500 36 34.410 38 37.905 19 33.210 31 36.719 158 31.090 32 36.450 10 26.722 180 35.860 178 20.741 (*) 4 35.040	41 13.943 (*) 36 10.790 13 12.100 158 10.760 15 11.840 8 10.700 38 11.605 180 10.410 32 11.540 31 10.309 33 11.377 5 10.145 3 10.895 19 8.390 4 10.870 10 7.963 (*)

Table 45: Analytical results for sulphate in precipitations samples.

Sulphate in precipitation				Sulphate in precipitation			
Sample no.: G1		Sample no.: G2					
Theoretical value:		1.069		Theoretical value:		1.087	
Unit: µg/l		Unit: µg/l					
Run 1:		Run 1:		Run 2:		Run 2:	
Number of laboratories:	58	Number of laboratories:	58	Number of laboratories:	55	Number of laboratories:	55
Arithmetic mean value:	1.140	Arithmetic mean value:	1.201	Arithmetic mean value:	1.084	Arithmetic mean value:	1.084
Median:	1.070	Median:	1.087	Median:	1.080	Median:	1.080
Standard deviation	0.397	Standard deviation	0.513	Standard deviation	0.081	Standard deviation	0.081
Rel. st. deviation (%)	34.845	Rel. st. deviation (%)	42.685	Rel. st. deviation (%)	7.453	Rel. st. deviation (%)	7.453
Run 2:		Results in decreasing order:		Run 1:		Results in decreasing order:	
Number of laboratories:	56	104 3.210 (*) 164 1.070	179 3.532 (*) 24 1.083	Number of laboratories:	58	104 3.270 (*) 31 1.080	
Arithmetic mean value:	1.067	107 3.156 (*) 17 1.066	107 3.243 (*) 110 1.080	Arithmetic mean value:	1.201	107 3.243 (*) 110 1.080	
Median:	1.068	45 1.286 150 1.064	45 1.329 116 1.080	Median:	1.087	45 1.329 116 1.080	
Standard deviation	0.397	161 1.200 12 1.060	167 1.242 163 1.080	Standard deviation	0.513	167 1.242 163 1.080	
Rel. st. deviation (%)	34.845	167 1.195 27 1.060	161 1.210 124 1.079	Rel. st. deviation (%)	42.685	161 1.210 124 1.079	
Run 2:		179 1.193 125 1.060	179 1.173 163 1.060	Run 2:		179 1.173 163 1.060	
Number of laboratories:	56	179 1.173 163 1.060	179 1.173 163 1.060	Number of laboratories:	55	179 1.173 163 1.060	
Arithmetic mean value:	1.067	179 1.173 163 1.060	179 1.173 163 1.060	Arithmetic mean value:	1.084	179 1.173 163 1.060	
Median:	1.068	179 1.173 163 1.060	179 1.173 163 1.060	Median:	1.087	179 1.173 163 1.060	
Standard deviation	0.081	179 1.173 163 1.060	179 1.173 163 1.060	Standard deviation	0.081	179 1.173 163 1.060	
Rel. st. deviation (%)	7.548	179 1.173 163 1.060	179 1.173 163 1.060	Rel. st. deviation (%)	7.453	179 1.173 163 1.060	
Results in decreasing order:		179 1.173 163 1.060	179 1.173 163 1.060	Results in decreasing order:		179 1.173 163 1.060	
104 3.210 (*) 164 1.070		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
107 3.156 (*) 17 1.066		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
45 1.286 150 1.064		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
161 1.200 12 1.060		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
167 1.195 27 1.060		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
43 1.193 125 1.060		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
2 1.160 31 1.059		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
7 1.143 16 1.058		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
13 1.130 24 1.056		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
110 1.130 10 1.053		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
21 1.122 36 1.053		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
166 1.117 26 1.048		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
33 1.110 178 1.047		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
124 1.110 8 1.040		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
20 1.100 23 1.038		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
35 1.100 41 1.036		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
118 1.100 153 1.033		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
180 1.100 14 1.030		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
38 1.090 4 1.029		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
5 1.083 120 1.017		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
112 1.080 22 1.011		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
116 1.080 19 1.010		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
15 1.077 30 0.990		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
3 1.075 155 0.988		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
32 1.074 121 0.983		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
158 1.071 126 0.950		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
115 1.070 176 0.810		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
157 1.070 42 0.760		179 1.173 163 1.060		179 1.173 163 1.060		179 1.173 163 1.060	
Run 1:		179 1.173 163 1.060		Run 1:		179 1.173 163 1.060	
Number of laboratories:	58	179 1.173 163 1.060		Number of laboratories:	57	179 1.173 163 1.060	
Arithmetic mean value:	0.857	179 1.173 163 1.060		Arithmetic mean value:	0.955	179 1.173 163 1.060	
Median:	0.767	179 1.173 163 1.060		Median:	0.860	179 1.173 163 1.060	
Standard deviation	0.376	179 1.173 163 1.060		Standard deviation	0.405	179 1.173 163 1.060	
Rel. st. deviation (%)	43.840	179 1.173 163 1.060		Rel. st. deviation (%)	42.392	179 1.173 163 1.060	
Run 2:		179 1.173 163 1.060		Run 2:		179 1.173 163 1.060	
Number of laboratories:	55	179 1.173 163 1.060		Number of laboratories:	54	179 1.173 163 1.060	
Arithmetic mean value:	0.771	179 1.173 163 1.060		Arithmetic mean value:	0.862	179 1.173 163 1.060	
Median:	0.764	179 1.173 163 1.060		Median:	0.859	179 1.173 163 1.060	
Standard deviation	0.057	179 1.173 163 1.060		Standard deviation	0.069	179 1.173 163 1.060	
Rel. st. deviation (%)	7.343	179 1.173 163 1.060		Rel. st. deviation (%)	8.049	179 1.173 163 1.060	
Results in decreasing order:		179 1.173 163 1.060		Results in decreasing order:		179 1.173 163 1.060	
107 2.522 (*) 10 0.764		179 1.173 163 1.060		107 2.710 (*) 110 0.860		179 1.173 163 1.060	
179 2.473 (*) 27 0.764		179 1.173 163 1.060		104 2.610 (*) 16 0.859		179 1.173 163 1.060	
104 2.300 (*) 12 0.760		179 1.173 163 1.060		179 2.578 (*) 31 0.859		179 1.173 163 1.060	
167 0.965 163 0.758		179 1.173 163 1.060		45 1.044 26 0.857		179 1.173 163 1.060	
45 0.942 180 0.758		179 1.173 163 1.060		161 0.960 27 0.856		179 1.173 163 1.060	
15 0.875 36 0.757		179 1.173 163 1.060		164 0.960 24 0.853		179 1.173 163 1.060	
161 0.870 16 0.756		179 1.173 163 1.060		161 0.956 150 0.852		179 1.173 163 1.060	
43 0.866 150 0.756		179 1.173 163 1.060		157 0.950 167 0.851		179 1.173 163 1.060	
2 0.840 26 0.754		179 1.173 163 1.060		2 0.948 4 0.848		179 1.173 163 1.060	
110 0.820 19 0.750		179 1.173 163 1.060		15 0.928 23 0.848		179 1.173 163 1.060	
164 0.810 23 0.749		179 1.173 163 1.060		166 0.923 153 0.847		179 1.173 163 1.060	
21 0.804 31 0.749		179 1.173 163 1.060		13 0.910 178 0.845		179 1.173 163 1.060	
7 0.800 153 0.747		179 1.173 163 1.060		21 0.910 8 0.842		179 1.173 163 1.060	
157 0.800 8 0.745		179 1.173 163 1.060		7 0.909 120 0.839		179 1.173 163 1.060	
166 0.799 4 0.742		179 1.173 163 1.060		33 0.900 116 0.837		179 1.173 163 1.060	
116 0.793 155 0.742		179 1.173 163 1.060		158 0.897 180 0.833		179 1.173 163 1.060	
158 0.792 178 0.741		179 1.173 163 1.060		35 0.895 17 0.830		179 1.173 163 1.060	
115 0.791 35 0.729		179 1.173 163 1.060		118 0.890 19 0.830		179 1.173 163 1.060	
118 0.790 17 0.728		179 1.173 163 1.060		115 0.886 124 0.827		179 1.173 163 1.060	
24 0.783 41 0.726		179 1.173 163 1.060		32 0.880 41 0.826		179 1.173 163 1.060	
13 0.780 120 0.724		179 1.173 163 1.060		38 0.880 155 0.824		179 1.173 163 1.060	
20 0.780 121 0.722		179 1.173 163 1.060		112 0.880 126 0.810		179 1.173 163 1.060	
33 0.780 14 0.720		179 1.173 163 1.060		5 0.875 121 0.807		179 1.173 163 1.060	
112 0.780 124 0.709		179 1.173 163 1.060		3 0.874 14 0.800		179 1.173 163 1.060	
125 0.779 30 0.700		179 1.173 163 1.060		125 0.874 30 0.800		179 1.173 163 1.060	
32 0.778 42 0.694		179 1.173 163 1.060		163 0.862 176 0.784		179 1.173 163 1.060	
5 0.776 22 0.691		179 1.173 163 1.060		20 0.870 22 0.784		179 1.173 163 1.060	
3 0.771 126 0.690		179 1.173 163 1.060		12 0.860 42 0.560		179 1.173 163 1.060	
38 0.770 176 0.650		179 1.173 163 1.060		36 0.860		179 1.173 163 1.060	

Table 46: Analytical results for nitrate in precipitations samples.

Nitrate in precipitation				Nitrate in precipitation			
Sample no.: G1		Sample no.: G2		Theoretical value: 0.310		Theoretical value: 0.408	
Unit: µg/l		Unit: µg/l					
Run 1:		Run 1:					
Number of laboratories:	59	Number of laboratories:	59				
Arithmetic mean value:	0.317	Arithmetic mean value:	0.408				
Median:	0.308	Median:	0.402				
Standard deviation	0.065	Standard deviation	0.033				
Rel. st. deviation (%)	20.434	Rel. st. deviation (%)	8.181				
Run 2:		Run 2:					
Number of laboratories:	57	Number of laboratories:	54				
Arithmetic mean value:	0.307	Arithmetic mean value:	0.402				
Median:	0.305	Median:	0.401				
Standard deviation	0.022	Standard deviation	0.020				
Rel. st. deviation (%)	7.286	Rel. st. deviation (%)	5.051				
Results in decreasing order:							
42	0.755 (*)	3	0.305	178	0.521 (*)	125	0.401
110	0.450 (*)	15	0.305	14	0.510 (*)	12	0.400
178	0.386	158	0.305	33	0.490 (*)	13	0.400
33	0.370	155	0.302	164	0.490 (*)	20	0.400
17	0.358	163	0.302	110	0.450	32	0.400
164	0.350	2	0.301	157	0.450	118	0.400
43	0.337	16	0.300	43	0.443	163	0.400
13	0.330	31	0.299	17	0.442	2	0.398
179	0.327	124	0.299	179	0.429	31	0.398
112	0.320	150	0.299	42	0.423	166	0.397
161	0.320	35	0.298	7	0.420	4	0.396
176	0.320	4	0.297	112	0.420	10	0.396
23	0.318	32	0.297	161	0.420	35	0.395
115	0.316	36	0.296	176	0.420	16	0.393
7	0.315	153	0.294	23	0.419	36	0.391
21	0.315	10	0.291	21	0.416	30	0.390
26	0.315	116	0.291	5	0.415	107	0.390
180	0.315	14	0.290	115	0.415	153	0.390
167	0.314	24	0.290	167	0.412	150	0.387
166	0.313	30	0.290	38	0.410	24	0.384
8	0.311	38	0.290	155	0.409	116	0.384
5	0.310	107	0.290	180	0.409	19	0.380
12	0.310	120	0.286	26	0.408	22	0.379
20	0.310	22	0.283	158	0.408	120	0.376
118	0.310	19	0.280	121	0.407	124	0.374
157	0.310	40	0.277	27	0.405	126	0.362
27	0.308	126	0.271	3	0.404	40	0.359
45	0.308	104	0.270	15	0.403	104	0.350
121	0.308	41	0.260	8	0.402	41	0.341 (*)
125	0.308			45	0.402		
Nitrate in precipitation							
Sample no.: G3		Sample no.: G4		Theoretical value: 0.421		Theoretical value: 0.512	
Unit: µg/l		Unit: µg/l					
Run 1:		Run 1:					
Number of laboratories:	59	Number of laboratories:	58				
Arithmetic mean value:	0.437	Arithmetic mean value:	0.508				
Median:	0.415	Median:	0.504				
Standard deviation	0.127	Standard deviation	0.042				
Rel. st. deviation (%)	29.115	Rel. st. deviation (%)	8.246				
Run 2:		Run 2:					
Number of laboratories:	58	Number of laboratories:	53				
Arithmetic mean value:	0.421	Arithmetic mean value:	0.507				
Median:	0.414	Median:	0.504				
Standard deviation	0.037	Standard deviation	0.024				
Rel. st. deviation (%)	8.838	Rel. st. deviation (%)	4.822				
Results in decreasing order:							
42	1.357 (*)	163	0.413	164	0.660 (*)	27	0.503
110	0.540	3	0.412	178	0.656 (*)	45	0.503
178	0.530	12	0.410	33	0.590	12	0.500
33	0.520	13	0.410	110	0.570	20	0.500
164	0.520	15	0.410	157	0.570	30	0.500
167	0.512	20	0.410	179	0.543	31	0.500
157	0.470	30	0.410	161	0.540	38	0.500
17	0.450	31	0.410	40	0.535	43	0.500
179	0.441	38	0.410	176	0.530	118	0.500
161	0.440	118	0.410	17	0.529	3	0.499
23	0.432	166	0.410	21	0.525	16	0.498
21	0.430	2	0.408	23	0.525	4	0.497
43	0.430	4	0.408	7	0.523	22	0.494
112	0.430	35	0.407	166	0.519	124	0.494
7	0.428	16	0.406	158	0.515	14	0.490
107	0.427	24	0.402	5	0.513	126	0.490
5	0.423	153	0.402	115	0.512	153	0.490
115	0.423	22	0.400	155	0.511	36	0.489
155	0.422	36	0.400	13	0.510	107	0.488
158	0.422	116	0.396	15	0.510	116	0.486
180	0.421	150	0.396	112	0.510	150	0.482
26	0.419	14	0.390	180	0.509	8	0.481
121	0.418	19	0.390	2	0.507	120	0.481
27	0.417	126	0.390	121	0.507	19	0.480
45	0.417	176	0.390	32	0.506	24	0.480
124	0.416	120	0.388	125	0.506	104	0.440
8	0.415	40	0.381	26	0.505	167	0.424 (*)
10	0.415	104	0.360	35	0.505	41	0.423 (*)
32	0.415	41	0.348	163	0.504	42	0.410 (*)
125	0.415						

Table 47: Analytical results for ammonium in precipitations sample.

Ammonium in precipitation				Ammonium in precipitation			
Sample no.: G1		Sample no.: G2		Theoretical value:		Theoretical value:	
Unit: µg/l		Unit: µg/l					
Run 1:		Run 1:					
Number of laboratories:	58	Number of laboratories:	58				
Arithmetic mean value:	0.162	Arithmetic mean value:	0.216				
Median:	0.163	Median:	0.210				
Standard deviation	0.026	Standard deviation	0.032				
Rel. st. deviation (%)	16.126	Rel. st. deviation (%)	14.626				
Run 2:		Run 2:					
Number of laboratories:	56	Number of laboratories:	56				
Arithmetic mean value:	0.163	Arithmetic mean value:	0.215				
Median:	0.163	Median:	0.210				
Standard deviation	0.016	Standard deviation	0.019				
Rel. st. deviation (%)	9.799	Rel. st. deviation (%)	9.053				
Results in decreasing order:				Results in decreasing order:			
22	0.218 (*)	166	0.162	164	0.360 (*)	14	0.210
116	0.199	179	0.162	22	0.274	30	0.210
12	0.192	27	0.161	116	0.267	38	0.210
120	0.191	8	0.160	120	0.252	104	0.210
176	0.190	12	0.160	2	0.249	157	0.210
3	0.188	20	0.160	40	0.241	176	0.210
31	0.184	38	0.160	13	0.240	45	0.209
40	0.184	104	0.160	150	0.239	4	0.208
41	0.184	112	0.160	3	0.237	5	0.208
172	0.181	157	0.160	41	0.237	7	0.208
13	0.180	4	0.158	172	0.233	107	0.208
150	0.178	36	0.158	31	0.232	153	0.206
158	0.177	5	0.157	19	0.230	36	0.205
23	0.172	7	0.156	23	0.228	179	0.204
19	0.170	155	0.155	158	0.226	121	0.203
30	0.170	107	0.152	12	0.220	125	0.203
164	0.170	118	0.152	20	0.220	10	0.202
35	0.169	10	0.151	112	0.220	155	0.202
42	0.169	121	0.151	126	0.219	118	0.201
14	0.168	125	0.149	180	0.219	16	0.199
16	0.166	153	0.149	163	0.218	115	0.198
21	0.166	24	0.145	21	0.217	32	0.197
15	0.165	167	0.144	15	0.216	24	0.196
45	0.165	33	0.141	27	0.216	167	0.194
163	0.165	178	0.139	26	0.215	33	0.191
32	0.164	124	0.135	8	0.213	124	0.190
26	0.163	126	0.124	35	0.213	42	0.189
115	0.163	110	0.120	166	0.211	110	0.160
180	0.163	43	0.016 (*)	178	0.211	43	0.093 (*)
Ammonium in precipitation				Ammonium in precipitation			
Sample no.: G3		Sample no.: G4		Theoretical value:		Theoretical value:	
Theoretical value:		0.267		0.294		0.294	
Unit: µg/l		Unit: µg/l					
Run 1:		Run 1:					
Number of laboratories:	58	Number of laboratories:	58				
Arithmetic mean value:	0.261	Arithmetic mean value:	0.293				
Median:	0.261	Median:	0.294				
Standard deviation	0.038	Standard deviation	0.035				
Rel. st. deviation (%)	14.555	Rel. st. deviation (%)	11.982				
Run 2:		Run 2:					
Number of laboratories:	57	Number of laboratories:	55				
Arithmetic mean value:	0.265	Arithmetic mean value:	0.296				
Median:	0.261	Median:	0.294				
Standard deviation	0.019	Standard deviation	0.020				
Rel. st. deviation (%)	7.296	Rel. st. deviation (%)	6.740				
Results in decreasing order:				Results in decreasing order:			
22	0.320	14	0.261	176	0.370 (*)	26	0.294
116	0.318	35	0.261	116	0.345	107	0.293
2	0.308	36	0.261	22	0.342	166	0.293
120	0.295	45	0.261	126	0.340	42	0.292
3	0.291	38	0.260	2	0.336	45	0.291
13	0.290	104	0.260	120	0.335	20	0.290
23	0.288	153	0.260	13	0.320	38	0.290
172	0.287	164	0.260	3	0.318	8	0.289
31	0.284	176	0.260	23	0.318	179	0.289
40	0.281	179	0.258	172	0.313	5	0.288
19	0.280	42	0.257	19	0.310	153	0.288
41	0.279	125	0.256	30	0.310	4	0.287
150	0.278	178	0.256	150	0.310	14	0.287
112	0.275	33	0.254	41	0.306	121	0.285
158	0.275	5	0.253	21	0.302	124	0.283
21	0.272	16	0.253	31	0.302	157	0.280
163	0.271	155	0.252	158	0.301	167	0.280
12	0.270	118	0.251	12	0.300	178	0.280
20	0.270	10	0.250	35	0.300	125	0.278
30	0.270	157	0.250	104	0.300	16	0.277
180	0.270	121	0.249	112	0.300	10	0.276
27	0.269	166	0.248	180	0.298	118	0.275
26	0.268	115	0.245	40	0.297	115	0.272
15	0.267	32	0.244	163	0.297	155	0.272
24	0.263	126	0.242	15	0.296	32	0.271
7	0.262	167	0.240	24	0.296	33	0.261
8	0.262	124	0.231	36	0.296	164	0.240
107	0.262	110	0.210	27	0.295	110	0.220
4	0.261	43	0.016 (*)	7	0.294	43	0.101 (*)

Table 48: Analytical results for pH in precipitations samples.

pH in precipitation			pH in precipitation		
Sample no.: G1	Theoretical value:	4.300	Sample no.: G2	Theoretical value:	4.270
Unit: pH-unit			Unit: pH-unit		
Run 1:					
Number of laboratories:	54		Number of laboratories:	54	
Arithmetic mean value:	4.391		Arithmetic mean value:	4.326	
Median:	4.354		Median:	4.320	
Standard deviation	0.445		Standard deviation	0.274	
Rel. st. deviation (%)	10.128		Rel. st. deviation (%)	6.333	
Run 2:					
Number of laboratories:	53		Number of laboratories:	52	
Arithmetic mean value:	4.336		Arithmetic mean value:	4.308	
Median:	4.350		Median:	4.320	
Standard deviation	0.192		Standard deviation	0.127	
Rel. st. deviation (%)	4.430		Rel. st. deviation (%)	2.954	
Results in decreasing order:					
104	7.290 (*)	21	4.350	104	5.950 (*)
43	4.970	32	4.350	13	4.630
30	4.680	35	4.350	43	4.590
13	4.540	4	4.340	19	4.540
19	4.510	12	4.340	153	4.410
41	4.500	27	4.340	112	4.400
153	4.490	33	4.340	115	4.390
121	4.460	120	4.340	14	4.380
38	4.430	16	4.330	10	4.370
112	4.430	46	4.330	36	4.370
115	4.430	5	4.320	150	4.370
150	4.430	107	4.320	7	4.350
14	4.420	118	4.320	15	4.350
7	4.410	124	4.320	41	4.350
10	4.410	126	4.320	126	4.350
20	4.400	158	4.320	155	4.350
36	4.400	164	4.320	3	4.334
155	4.380	23	4.300	8	4.330
15	4.370	178	4.290	30	4.330
22	4.370	176	4.270	31	4.330
116	4.370	45	4.200	116	4.330
163	4.370	110	4.180	121	4.330
166	4.370	2	3.990	163	4.330
8	4.360	17	3.930	166	4.330
31	4.360	172	3.880	21	4.320
42	4.360	24	3.860	22	4.320
3	4.358	40	3.690	27	4.320
Results in decreasing order:					
pH in precipitation			pH in precipitation		
Sample no.: G3	Theoretical value:	4.520	Sample no.: G4	Theoretical value:	4.480
Unit: pH-unit			Unit: pH-unit		
Run 1:					
Number of laboratories:	54		Number of laboratories:	52	
Arithmetic mean value:	4.569		Arithmetic mean value:	4.499	
Median:	4.556		Median:	4.510	
Standard deviation	0.325		Standard deviation	0.211	
Rel. st. deviation (%)	7.119		Rel. st. deviation (%)	4.698	
Run 2:					
Number of laboratories:	51		Number of laboratories:	48	
Arithmetic mean value:	4.532		Arithmetic mean value:	4.520	
Median:	4.551		Median:	4.515	
Standard deviation	0.154		Standard deviation	0.094	
Rel. st. deviation (%)	3.408		Rel. st. deviation (%)	2.078	
Results in decreasing order:					
43	6.160 (*)	3	4.551	104	5.320 (*)
104	5.660 (*)	12	4.550	13	4.790
13	4.830	16	4.550	19	4.730
153	4.760	121	4.550	153	4.680
19	4.740	150	4.550	43	4.630
14	4.660	166	4.550	112	4.610
115	4.660	176	4.550	14	4.590
112	4.650	5	4.540	116	4.590
116	4.650	27	4.540	150	4.590
10	4.630	46	4.540	36	4.580
36	4.620	32	4.530	115	4.580
124	4.620	158	4.530	124	4.570
20	4.600	23	4.520	155	4.560
41	4.600	38	4.520	7	4.550
155	4.600	164	4.520	31	4.550
7	4.590	107	4.510	41	4.550
15	4.590	118	4.510	15	4.540
30	4.590	42	4.500	22	4.530
31	4.580	178	4.490	163	4.530
8	4.570	126	4.480	8	4.520
22	4.570	45	4.400	16	4.520
35	4.570	110	4.290	21	4.520
163	4.570	2	4.200	35	4.520
4	4.560	17	4.100	164	4.520
21	4.560	172	4.060	176	4.520
33	4.560	24	4.000	4	4.510
120	4.560	40	3.780 (*)	40	3.620 (*)

Table 49: Analytical results for strong acid calculated from pH.

Strong acid calculated from pH				Strong acid calculated from pH			
Sample no.: G1	Theoretical value:	50.000	Unit:	Sample no.: G2	Theoretical value:	53.000	Unit:
Run 1:							
Number of laboratories:	52	Number of laboratories:	52	Arithmetic mean value:	50.077	Arithmetic mean value:	53.788
Median:	43.752	Median:	43.752	Standard deviation	32.567	Standard deviation	33.537
Rel. st. deviation (%)	65.033	Rel. st. deviation (%)	62.351				
Run 2:							
Number of laboratories:	48	Number of laboratories:	49	Arithmetic mean value:	41.927	Arithmetic mean value:	46.678
Median:	43.652	Median:	43.652	Standard deviation	13.202	Standard deviation	12.133
Rel. st. deviation (%)	31.488	Rel. st. deviation (%)	25.993				
Results in decreasing order:							
40	204.174 (*)	8	43.652	40	239.883 (*)	38	47.863
24	138.038 (*)	31	43.652	24	138.038 (*)	120	47.863
172	131.826 (*)	42	43.652	172	131.826 (*)	8	46.774
17	117.490 (*)	15	42.658	2	95.499	30	46.774
2	102.329	22	42.658	17	72.444	31	46.774
176	53.703	116	42.658	107	57.544	116	46.774
178	51.286	163	42.658	178	54.954	121	46.774
23	50.119	166	42.658	118	52.481	163	46.774
5	47.863	155	41.687	5	51.286	166	46.774
107	47.863	20	39.811	16	51.286	3	46.345
118	47.863	36	39.811	23	51.286	7	44.668
124	47.863	7	38.905	124	51.286	15	44.668
126	47.863	10	38.905	12	50.119	41	44.668
158	47.863	14	38.019	20	50.119	126	44.668
164	47.863	38	37.154	32	50.119	155	44.668
16	46.774	112	37.154	46	50.119	10	42.658
46	46.774	115	37.154	158	50.119	36	42.658
4	45.709	150	37.154	164	50.119	150	42.658
12	45.709	121	34.674	176	50.119	14	41.687
27	45.709	153	32.359	4	48.978	115	40.738
33	45.709	41	31.623	42	48.978	112	39.811
120	45.709	19	30.903	21	47.863	153	38.905
21	44.668	13	28.840	22	47.863	19	28.840
32	44.668	30	20.893	27	47.863	43	25.704
35	44.668	43	10.715	33	47.863	13	23.442
3	43.853	104	0.051	35	47.863	104	1.122
Strong acid calculated from pH							
Sample no.: G3				Strong acid calculated from pH			
Theoretical value:	30.000			Sample no.: G4			
Unit:				Theoretical value:	33.000		
Run 1:							
Number of laboratories:	52	Number of laboratories:	50	Arithmetic mean value:	32.647	Arithmetic mean value:	36.312
Median:	27.542	Median:	30.551	Standard deviation	25.281	Standard deviation	29.594
Rel. st. deviation (%)	77.437	Rel. st. deviation (%)	81.500				
Run 2:							
Number of laboratories:	49	Number of laboratories:	47	Arithmetic mean value:	27.441	Arithmetic mean value:	29.928
Median:	27.542	Median:	30.200	Standard deviation	11.175	Standard deviation	8.171
Rel. st. deviation (%)	40.722	Rel. st. deviation (%)	27.304				
Results in decreasing order:							
40	165.959 (*)	21	27.542	40	208.930 (*)	8	30.200
24	100.000 (*)	33	27.542	172	102.329 (*)	16	30.200
172	87.096 (*)	120	27.542	24	97.724 (*)	21	30.200
17	79.433	8	26.915	2	67.608	35	30.200
2	63.096	22	26.915	126	44.668	164	30.200
126	33.113	35	26.915	42	39.811	176	30.200
178	32.359	163	26.915	178	36.308	22	29.512
42	31.623	31	26.303	107	35.481	163	29.512
107	30.903	7	25.704	118	33.884	15	28.840
118	30.903	15	25.704	23	33.113	7	28.184
23	30.200	30	25.704	158	33.113	31	28.184
38	30.200	20	25.119	5	31.623	41	28.184
164	30.200	41	25.119	20	31.623	155	27.542
32	29.512	155	25.119	30	31.623	124	26.915
158	29.512	36	23.988	38	31.623	36	26.303
5	28.840	124	23.988	46	31.623	115	26.303
27	28.840	10	23.442	3	30.974	14	25.704
46	28.840	112	22.387	4	30.903	116	25.704
12	28.184	116	22.387	12	30.903	150	25.704
16	28.184	14	21.878	27	30.903	112	24.547
121	28.184	115	21.878	32	30.903	43	23.442
150	28.184	19	18.197	33	30.903	153	20.893
166	28.184	153	17.378	120	30.903	19	18.621
176	28.184	13	14.791	121	30.903	13	16.218
3	28.119	104	2.188	166	30.903	104	4.786
4	27.542	43	0.692				

Table 50: Analytical results for chloride in precipitations samples.

Chloride in precipitation

Sample no.: G1
Theoretical value: 0.154
Unit: µg/l

Run 1:
Number of laboratories: 58
Arithmetic mean value: 0.181
Median: 0.158
Standard deviation 0.122
Rel. st. deviation (%) 67.398

Run 2:
Number of laboratories: 57
Arithmetic mean value: 0.166
Median: 0.157
Standard deviation 0.043
Rel. st. deviation (%) 26.037

Results in decreasing order:
42 1.038 (*) 178 0.157
125 0.342 26 0.155
110 0.340 3 0.153
19 < 0.310
115 0.260 32 0.153
157 0.260 27 0.152
17 0.227 104 0.150
43 0.202 120 0.150
167 0.187 166 0.150
116 0.181 8 0.148
124 0.181 31 0.148
13 0.180 158 0.148
121 0.180 14 0.147
176 0.180 45 0.145
40 0.179 155 0.145
150 0.173 107 0.144
7 0.171 163 0.143
4 0.170 12 0.140
33 0.170 118 0.140
161 0.170 126 0.140
22 0.167 153 0.140
2 0.163 164 0.140
16 0.163 10 0.138
41 0.162 23 0.136
20 0.160 179 0.132
30 0.160 38 0.130
112 0.160 24 0.129
180 0.159 35 0.127
5 0.158 36 0.125
21 0.158 15 0.118

Chloride in precipitation

Sample no.: G3
Theoretical value: 0.232
Unit: µg/l

Run 1:
Number of laboratories: 57
Arithmetic mean value: 0.250
Median: 0.232
Standard deviation 0.108
Rel. st. deviation (%) 43.464

Run 2:
Number of laboratories: 56
Arithmetic mean value: 0.236
Median: 0.232
Standard deviation 0.028
Rel. st. deviation (%) 11.688

Results in decreasing order:
42 1.028 (*) 3 0.231
167 0.334 30 0.230
115 0.323 104 0.230
120 0.310 24 0.229
19 < 0.310
110 < 0.300
112 0.280 180 0.229
157 0.280 45 0.228
43 0.262 166 0.228
13 0.260 8 0.227
116 0.253 22 0.226
33 0.250 31 0.225
2 0.247 107 0.222
4 0.247 41 0.221
7 0.244 12 0.220
40 0.243 15 0.220
125 0.241 126 0.220
158 0.241 150 0.220
20 0.240 163 0.220
161 0.240 155 0.219
164 0.240 23 0.217
178 0.240 10 0.216
124 0.239 35 0.212
17 0.238 153 0.212
121 0.237 38 0.210
5 0.236 118 0.210
26 0.236 176 0.210
16 0.235 14 0.197
21 0.235 36 0.194
27 0.233 179 0.180
32 0.232

Chloride in precipitation

Sample no.: G2
Theoretical value: 0.193
Unit: µg/l

Run 1:
Number of laboratories: 57
Arithmetic mean value: 0.241
Median: 0.200
Standard deviation 0.207
Rel. st. deviation (%) 85.992

Run 2:
Number of laboratories: 55
Arithmetic mean value: 0.203
Median: 0.200
Standard deviation 0.030
Rel. st. deviation (%) 14.824

Results in decreasing order:
43 1.511 (*) 30 0.200
42 1.060 (*) 31 0.200
157 0.320 178 0.200
19 < 0.310
110 < 0.300
115 0.283 27 0.199
176 0.280 180 0.197
120 0.260 41 0.195
124 0.234 8 0.193
116 0.233 45 0.192
167 0.229 12 0.190
17 0.224 126 0.190
2 0.214 164 0.190
7 0.214 107 0.188
4 0.213 163 0.188
40 0.212 22 0.186
121 0.211 150 0.186
33 0.210 23 0.184
104 0.210 153 0.183
112 0.210 155 0.183
161 0.210 38 0.180
125 0.208 118 0.180
3 0.206 35 0.176
26 0.206 24 0.174
16 0.205 36 0.164
32 0.205 179 0.163
21 0.204 15 0.162
5 0.203 10 0.157
158 0.203 14 0.152
20 0.200

Chloride in precipitation

Sample no.: G4
Theoretical value: 0.309
Unit: µg/l

Run 1:
Number of laboratories: 57
Arithmetic mean value: 0.314
Median: 0.306
Standard deviation 0.049
Rel. st. deviation (%) 15.714

Run 2:
Number of laboratories: 53
Arithmetic mean value: 0.303
Median: 0.302
Standard deviation 0.024
Rel. st. deviation (%) 7.861

Results in decreasing order:
42 0.571 (*) 43 0.303
120 0.430 (*) 8 0.302
110 0.420 (*) 45 0.302
115 0.413 (*) 124 0.302
157 0.370 15 0.297
112 0.350 31 0.296
176 0.350 23 0.295
13 0.340 166 0.295
4 0.333 121 0.294
116 0.333 163 0.294
33 0.330 180 0.292
158 0.330 150 0.291
164 0.330 12 0.290
40 0.328 24 0.290
2 0.321 35 0.290
3 0.316 126 0.290
7 0.315 155 0.290
178 0.315 153 0.284
21 0.314 30 0.280
125 0.314 118 0.280
26 0.313 14 0.277
19 < 0.310
5 0.310 107 0.273
20 0.310 22 0.271
104 0.310 38 0.270
161 0.310 41 0.270
27 0.309 36 0.269
32 0.307 167 0.267
16 0.306 179 0.247
17 0.306

Table 51: Analytical results for sodium in precipitations samples.

Sodium in precipitation				Sodium in precipitation			
Sample no.: G1		Sample no.: G2		Theoretical value:		0.304	
Unit: µg/l		Unit: µg/l					
Run 1:							
Number of laboratories:	55	Number of laboratories:	55	Arithmetic mean value:	0.310	Median:	0.310
Arithmetic mean value:	0.233	Arithmetic mean value:	0.230	Standard deviation	0.026	Standard deviation	0.026
Median:	0.230	Median:	0.230	Rel. st. deviation (%)	8.291	Rel. st. deviation (%)	8.291
Standard deviation	0.028	Standard deviation	0.021	Rel. st. deviation (%)	11.961	Rel. st. deviation (%)	9.014
Run 2:							
Number of laboratories:	54	Number of laboratories:	53	Arithmetic mean value:	0.307	Median:	0.308
Arithmetic mean value:	0.231	Arithmetic mean value:	0.230	Standard deviation	0.021	Standard deviation	0.021
Median:	0.230	Median:	0.230	Rel. st. deviation (%)	6.840	Rel. st. deviation (%)	9.014
Standard deviation	0.021	Standard deviation	0.021	Rel. st. deviation (%)	11.961	Rel. st. deviation (%)	9.014
Results in decreasing order:							
43 < 0.672		43 < 0.672		178 < 0.391 (*)	23	0.308	
176 0.370 (*) 30	0.230	178 0.385 (*) 31	0.308	42 0.329	13	0.306	
42 0.285 36	0.229	42 0.350 14	0.306	19 0.350 16	19	0.306	
178 0.280 180	0.229	19 0.350 16	0.306	33 0.334 35	33	0.305	
22 0.273 16	0.227	33 0.334 35	0.305	150 0.334 153	150	0.304	
157 0.270 3	0.226	150 0.334 153	0.304	5 0.332 41	5	0.302	
13 0.260 31	0.224	5 0.332 41	0.302	112 0.330 8	112	0.301	
116 0.252 41	0.223	112 0.330 8	0.301	157 0.330 30	157	0.300	
19 0.250 26	0.222	157 0.330 30	0.300	115 0.326 26	115	0.298	
33 0.250 12	0.220	115 0.326 26	0.298	124 0.326 32	124	0.297	
150 0.250 110	0.220	124 0.326 32	0.297	158 0.326 4	158	0.295	
17 0.247 23	0.218	158 0.326 4	0.295	163 0.325 7	163	0.293	
5 0.246 7	0.217	163 0.325 7	0.293	45 0.324 167	45	0.293	
124 0.245 8	0.217	45 0.324 167	0.293	120 0.324 10	120	0.291	
45 0.242 153	0.215	120 0.324 10	0.291	155 0.322 22	155	0.291	
120 0.241 10	0.214	155 0.322 22	0.291	21 0.321 166	21	0.290	
112 0.240 167	0.214	21 0.321 166	0.290	163 0.321 17	163	0.288	
155 0.240 2	0.210	163 0.321 17	0.288	20 0.320 2	20	0.286	
163 0.240 4	0.210	20 0.320 2	0.286	116 0.320 24	116	0.286	
164 0.240 24	0.210	116 0.320 24	0.286	164 0.320 121	164	0.283	
115 0.239 32	0.210	164 0.320 121	0.283	176 0.320 125	176	0.278	
21 0.238 166	0.206	176 0.320 125	0.278	15 0.316 179	15	0.272	
35 0.238 121	0.205	15 0.316 179	0.272	36 0.315 104	36	0.271	
158 0.238 38	0.200	36 0.315 104	0.271	3 0.312 38	3	0.270	
179 0.235 118	0.200	3 0.312 38	0.270	12 0.310 118	12	0.260	
15 0.233 125	0.200	12 0.310 118	0.260	27 0.310 126	27	0.260	
27 0.233 104	0.199	27 0.310 126	0.260	110 0.310 110	110	0.260	
14 0.231 126	0.190	110 0.310 110	0.260				
20 0.230							
Sodium in precipitation				Sodium in precipitation			
Sample no.: G3		Sample no.: G4		Theoretical value:		0.469	
Theoretical value:		0.293		Unit: µg/l		0.469	
Run 1:							
Number of laboratories:	55	Number of laboratories:	55	Arithmetic mean value:	0.701	Median:	0.468
Arithmetic mean value:	0.300	Arithmetic mean value:	0.300	Standard deviation	61.156	Standard deviation	61.156
Median:	0.299	Median:	0.299	Rel. st. deviation (%)	702.826	Rel. st. deviation (%)	702.826
Standard deviation	0.031	Standard deviation	0.025	Rel. st. deviation (%)	10.422	Rel. st. deviation (%)	8.471
Run 2:							
Number of laboratories:	54	Number of laboratories:	54	Arithmetic mean value:	0.455	Median:	0.467
Arithmetic mean value:	0.297	Arithmetic mean value:	0.297	Standard deviation	0.072	Standard deviation	0.072
Median:	0.299	Median:	0.299	Rel. st. deviation (%)	15.747	Rel. st. deviation (%)	15.747
Standard deviation	0.025	Standard deviation	0.025	Rel. st. deviation (%)	10.422	Rel. st. deviation (%)	8.471
Results in decreasing order:							
43 < 0.672		43 < 0.672		179 < 0.465			
167 0.437 (*) 45	0.299	179 < 0.465		42 0.594 27			
42 0.354 35	0.297	42 0.594 27		178 0.553 16			
179 0.354 150	0.297	178 0.553 16		17 0.521 153			
178 0.351 121	0.295	17 0.521 153		13 0.520 12			
157 0.350 16	0.293	13 0.520 12		19 0.500 157			
17 0.342 27	0.292	19 0.500 157		33 0.495 8			
14 0.327 153	0.292	33 0.495 8		14 0.494 35			
13 0.320 12	0.290	14 0.494 35		31 0.454 31			
19 0.320 30	0.290	31 0.454 31		5 0.490 26			
33 0.314 23	0.289	5 0.490 26		112 0.490 150			
5 0.313 26	0.289	112 0.490 150		120 0.490 30			
22 0.310 41	0.287	120 0.490 30		176 0.490 3			
112 0.310 31	0.285	176 0.490 3		158 0.484 22			
176 0.310 3	0.284	158 0.484 22		164 0.480 166			
115 0.309 8	0.282	164 0.480 166		163 0.479 4			
180 0.308 110	0.280	163 0.479 4		21 0.477 41			
124 0.307 7	0.276	21 0.477 41		155 0.476 2			
158 0.307 4	0.273	155 0.476 2		36 0.475 7			
163 0.307 24	0.271	36 0.475 7		23 0.473 24			
21 0.306 10	0.270	23 0.473 24		180 0.473 10			
116 0.306 2	0.264	180 0.473 10		15 0.471 121			
120 0.304 125	0.264	15 0.471 121		115 0.471 125			
155 0.302 166	0.263	115 0.471 125		20 0.470 104			
36 0.301 38	0.260	20 0.470 104		45 0.470 38			
20 0.300 104	0.256	45 0.470 38		110 0.470 118			
32 0.300 126	0.250	110 0.470 118		116 0.469 167			
164 0.300 118	0.240	116 0.469 167		124 0.469 126			
15 0.299		124 0.469 126		32 0.468 0.040			

Table 52: Analytical results for magnesium in precipitations samples.

Magnesium in precipitation				Magnesium in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 0.145				Theoretical value: 0.083			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:	54			Number of laboratories:	54		
Arithmetic mean value:	0.145			Arithmetic mean value:	0.082		
Median:	0.144			Median:	0.081		
Standard deviation	0.018			Standard deviation	0.010		
Rel. st. deviation (%)	12.232			Rel. st. deviation (%)	12.132		
Run 2:				Run 2:			
Number of laboratories:	48			Number of laboratories:	49		
Arithmetic mean value:	0.142			Arithmetic mean value:	0.080		
Median:	0.143			Median:	0.081		
Standard deviation	0.010			Standard deviation	0.006		
Rel. st. deviation (%)	7.160			Rel. st. deviation (%)	7.776		
Results in decreasing order:							
179	0.195 (*)	155	0.144	43 < 0.145			
157	0.190 (*)	166	0.144	17	0.111 (*)	31	0.081
42	0.186 (*)	10	0.142	41	0.104 (*)	118	0.081
41	0.185 (*)	14	0.142	178	0.103 (*)	3	0.080
17	0.184 (*)	31	0.142	42	0.102 (*)	13	0.080
30	0.160	124	0.142	179	0.098	15	0.080
38	0.160	13	0.140	19	0.090	26	0.080
150	0.155	20	0.140	180	0.090	110	0.080
7	0.154	110	0.140	150	0.089	112	0.080
163	0.154	153	0.140	2	0.087	164	0.080
178	0.154	164	0.140	45	0.087	4	0.078
2	0.153	26	0.139	163	0.087	33	0.077
19	0.150	118	0.139	7	0.086	121	0.076
45	0.150	4	0.136	23	0.086	153	0.075
112	0.150	12	0.136	116	0.084	20	0.074
21	0.149	3	0.135	5	0.083	22	0.074
180	0.149	115	0.134	8	0.083	115	0.074
16	0.147	35	0.133	21	0.083	12	0.073
22	0.146	33	0.132	27	0.083	120	0.073
36	0.146	121	0.130	124	0.083	167	0.073
116	0.146	104	0.128	155	0.083	104	0.072
43 < 0.145				14	0.082	24	0.071
5	0.145	24	0.127	16	0.082	35	0.070
8	0.145	120	0.127	36	0.082	125	0.070
23	0.145	167	0.126	158	0.082	157	0.070
27	0.145	125	0.118	166	0.082	176	0.070
158	0.145	176	0.110	10	0.081	126	0.050 (*)
Magnesium in precipitation				Magnesium in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value:	0.103			Theoretical value:	0.134		
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:	54			Number of laboratories:	54		
Arithmetic mean value:	0.102			Arithmetic mean value:	0.131		
Median:	0.101			Median:	0.131		
Standard deviation	0.012			Standard deviation	0.013		
Rel. st. deviation (%)	11.712			Rel. st. deviation (%)	10.268		
Run 2:				Run 2:			
Number of laboratories:	51			Number of laboratories:	49		
Arithmetic mean value:	0.103			Arithmetic mean value:	0.131		
Median:	0.101			Median:	0.131		
Standard deviation	0.009			Standard deviation	0.008		
Rel. st. deviation (%)	8.700			Rel. st. deviation (%)	6.220		
Results in decreasing order:							
43 < 0.145				Results in decreasing order:			
17	0.129 (*)	16	0.101	42	0.170 (*)	124	0.131
41	0.125	26	0.101	41	0.161 (*)	166	0.131
167	0.122	10	0.100	178	0.151	20	0.130
38	0.120	13	0.100	17	0.148	33	0.130
179	0.120	33	0.100	43 < 0.145			
178	0.119	110	0.100	163	0.143	110	0.130
42	0.115	112	0.100	116	0.142	112	0.130
163	0.112	164	0.100	13	0.140	164	0.130
2	0.110	20	0.098	19	0.140	22	0.128
19	0.110	118	0.098	30	0.140	118	0.128
30	0.110	153	0.098	179	0.140	3	0.127
116	0.110	3	0.097	2	0.138	4	0.127
180	0.109	4	0.097	23	0.138	12	0.126
45	0.108	15	0.097	36	0.138	15	0.126
14	0.107	121	0.096	45	0.138	153	0.124
23	0.107	31	0.095	180	0.138	35	0.123
150	0.107	22	0.093	21	0.137	31	0.121
7	0.105	35	0.093	14	0.136	115	0.121
21	0.105	115	0.093	16	0.136	24	0.120
155	0.105	12	0.091	150	0.136	38	0.120
27	0.104	24	0.091	155	0.135	157	0.120
36	0.104	157	0.090	158	0.135	120	0.119
158	0.104	104	0.089	7	0.134	121	0.119
8	0.103	120	0.088	27	0.134	104	0.118
124	0.103	125	0.088	5	0.133	125	0.114
166	0.102	176	0.070 (*)	8	0.133	176	0.100 (*)
5	0.101	126	0.060 (*)	10	0.131	167	0.095 (*)
				26	0.131	126	0.090 (*)

Table 53: Analytical results for calcium in precipitations samples.

Calcium in precipitation				Calcium in precipitation			
Sample no.: G1		Sample no.: G2					
Theoretical value:		0.096		Theoretical value:		0.153	
Unit: $\mu\text{g/l}$		Unit: $\mu\text{g/l}$					
Run 1:		Run 1:		Run 1:		Run 1:	
Number of laboratories:	54	Number of laboratories:	54	Number of laboratories:	54	Number of laboratories:	54
Arithmetic mean value:	0.104	Arithmetic mean value:	0.158	Arithmetic mean value:	0.157	Arithmetic mean value:	0.158
Median:	0.100	Median:	0.157	Median:	0.157	Median:	0.156
Standard deviation	0.033	Standard deviation	0.034	Standard deviation	0.034	Standard deviation	0.034
Rel. st. deviation (%)	32.139	Rel. st. deviation (%)	21.260	Rel. st. deviation (%)	21.260	Rel. st. deviation (%)	21.260
Run 2:		Run 2:		Run 2:		Run 2:	
Number of laboratories:	49	Number of laboratories:	49	Number of laboratories:	49	Number of laboratories:	49
Arithmetic mean value:	0.101	Arithmetic mean value:	0.156	Arithmetic mean value:	0.156	Arithmetic mean value:	0.156
Median:	0.100	Median:	0.156	Median:	0.156	Median:	0.156
Standard deviation	0.018	Standard deviation	0.022	Standard deviation	0.022	Standard deviation	0.022
Rel. st. deviation (%)	17.708	Rel. st. deviation (%)	14.149	Rel. st. deviation (%)	14.149	Rel. st. deviation (%)	14.149
Results in decreasing order:				Results in decreasing order:			
43 < 0.299		43 < 0.299		43 < 0.299		43 < 0.299	
42 0.215 (*) 150	0.100	42 0.259 (*) 125	0.156	42 0.245 (*) 158	0.155	42 0.245 (*) 158	0.155
17 0.210 (*) 157	0.100	38 0.230 (*) 31	0.154	38 0.230 (*) 31	0.154	38 0.230 (*) 31	0.154
38 0.180 (*) 164	0.100	178 0.220 155	0.154	178 0.220 155	0.154	178 0.220 155	0.154
178 0.150 31	0.099	17 0.205 166	0.154	17 0.205 166	0.154	17 0.205 166	0.154
116 0.144 33	0.099	41 0.197 5	0.153	41 0.197 5	0.153	41 0.197 5	0.153
22 0.138 8	0.098	45 0.192 8	0.153	45 0.192 8	0.153	45 0.192 8	0.153
124 0.135 21	0.098	19 0.180 21	0.153	19 0.180 21	0.153	19 0.180 21	0.153
45 0.123 125	0.098	30 0.180 4	0.150	30 0.180 4	0.150	30 0.180 4	0.150
176 0.120 158	0.098	157 0.180 13	0.150	157 0.180 13	0.150	157 0.180 13	0.150
180 0.120 166	0.098	116 0.176 24	0.146	116 0.176 24	0.146	116 0.176 24	0.146
121 0.113 3	0.097	163 0.172 33	0.143	163 0.172 33	0.143	163 0.172 33	0.143
36 0.111 5	0.097	23 0.170 15	0.142	23 0.170 15	0.142	23 0.170 15	0.142
163 0.111 4	0.095	126 0.170 22	0.142	126 0.170 22	0.142	126 0.170 22	0.142
13 0.110 155	0.095	124 0.169 118	0.140	124 0.169 118	0.140	124 0.169 118	0.140
23 0.110 24	0.094	3 0.168 164	0.140	3 0.168 164	0.140	3 0.168 164	0.140
30 0.110 126	0.090	180 0.168 176	0.140	180 0.168 176	0.140	180 0.168 176	0.140
115 0.109 15	0.086	36 0.164 104	0.135	36 0.164 104	0.135	36 0.164 104	0.135
41 0.107 104	0.086	10 0.162 120	0.129	10 0.162 120	0.129	10 0.162 120	0.129
16 0.106 120	0.083	7 0.160 35	0.124	7 0.160 35	0.124	7 0.160 35	0.124
179 0.102 10	0.081	14 0.160 153	0.124	14 0.160 153	0.124	14 0.160 153	0.124
7 0.101 118	0.080	112 0.160 2	0.123	112 0.160 2	0.123	112 0.160 2	0.123
26 0.101 153	0.076	150 0.159 32	0.120	150 0.159 32	0.120	150 0.159 32	0.120
14 0.100 35	0.074	16 0.157 110	0.120	16 0.157 110	0.120	16 0.157 110	0.120
19 0.100 110	0.060	26 0.157 179	0.110	26 0.157 179	0.110	26 0.157 179	0.110
27 0.100 2	0.059	27 0.157 12	0.090	27 0.157 12	0.090	27 0.157 12	0.090
32 0.100 12	0.030 (*)	115 0.157 167	0.065 (*)	115 0.157 167	0.065 (*)	115 0.157 167	0.065 (*)
112 0.100 167	0.012 (*)						
Calcium in precipitation		Calcium in precipitation		Calcium in precipitation		Calcium in precipitation	
Sample no.: G3		Sample no.: G4		Sample no.: G4		Sample no.: G4	
Theoretical value:	0.192	Theoretical value:	0.204	Theoretical value:	0.204	Theoretical value:	0.204
Unit: $\mu\text{g/l}$		Unit: $\mu\text{g/l}$		Unit: $\mu\text{g/l}$		Unit: $\mu\text{g/l}$	
Run 1:		Run 1:		Run 1:		Run 1:	
Number of laboratories:	54	Number of laboratories:	54	Number of laboratories:	54	Number of laboratories:	54
Arithmetic mean value:	0.194	Arithmetic mean value:	0.206	Arithmetic mean value:	0.206	Arithmetic mean value:	0.206
Median:	0.190	Median:	0.205	Median:	0.205	Median:	0.205
Standard deviation	0.038	Standard deviation	0.038	Standard deviation	0.038	Standard deviation	0.038
Rel. st. deviation (%)	19.703	Rel. st. deviation (%)	18.258	Rel. st. deviation (%)	18.258	Rel. st. deviation (%)	18.258
Run 2:		Run 2:		Run 2:		Run 2:	
Number of laboratories:	51	Number of laboratories:	50	Number of laboratories:	50	Number of laboratories:	50
Arithmetic mean value:	0.187	Arithmetic mean value:	0.204	Arithmetic mean value:	0.204	Arithmetic mean value:	0.204
Median:	0.190	Median:	0.205	Median:	0.205	Median:	0.205
Standard deviation	0.026	Standard deviation	0.025	Standard deviation	0.025	Standard deviation	0.025
Rel. st. deviation (%)	14.039	Rel. st. deviation (%)	12.296	Rel. st. deviation (%)	12.296	Rel. st. deviation (%)	12.296
Results in decreasing order:				Results in decreasing order:			
38 0.330 (*) 10	0.190	42 0.343 (*) 158	0.205	42 0.343 (*) 158	0.205	42 0.343 (*) 158	0.205
116 0.301 (*) 13	0.190	38 0.310 (*) 125	0.203	38 0.310 (*) 125	0.203	38 0.310 (*) 125	0.203
43 < 0.299		43 < 0.299		43 < 0.299		43 < 0.299	
42 0.297 (*) 112	0.190	178 0.279 5	0.202	178 0.279 5	0.202	178 0.279 5	0.202
178 0.262 164	0.190	17 0.277 8	0.202	17 0.277 8	0.202	17 0.277 8	0.202
17 0.242 8	0.189	116 0.245 21	0.202	116 0.245 21	0.202	116 0.245 21	0.202
41 0.239 150	0.189	30 0.240 112	0.200	30 0.240 112	0.200	30 0.240 112	0.200
30 0.230 24	0.188	45 0.231 121	0.200	45 0.231 121	0.200	45 0.231 121	0.200
163 0.220 155	0.187	157 0.230 7	0.198	157 0.230 7	0.198	157 0.230 7	0.198
45 0.211 166	0.187	163 0.230 166	0.198	163 0.230 166	0.198	163 0.230 166	0.198
14 0.210 4	0.185	23 0.228 4	0.195	23 0.228 4	0.195	23 0.228 4	0.195
157 0.210 153	0.185	13 0.220 155	0.193	13 0.220 155	0.193	13 0.220 155	0.193
23 0.204 31	0.183	14 0.220 150	0.192	14 0.220 150	0.192	14 0.220 150	0.192
19 0.200 33	0.181	41 0.219 164	0.190	41 0.219 164	0.190	41 0.219 164	0.190
180 0.199 179	0.178	180 0.217 31	0.189	180 0.217 31	0.189	180 0.217 31	0.189
26 0.197 15	0.173	10 0.216 15	0.187	10 0.216 15	0.187	10 0.216 15	0.187
7 0.196 32	0.170	179 0.214 153	0.181	179 0.214 153	0.181	179 0.214 153	0.181
27 0.195 118	0.170	3 0.211 118	0.180	3 0.211 118	0.180	3 0.211 118	0.180
115 0.194 2	0.166	24 0.211 35	0.173	24 0.211 35	0.173	24 0.211 35	0.173
3 0.193 35	0.164	36 0.211 104	0.173	36 0.211 104	0.173	36 0.211 104	0.173
158 0.193 104	0.161	16 0.210 120	0.171	16 0.210 120	0.171	16 0.210 120	0.171
121 0.192 120	0.155	19 0.210 110	0.170	19 0.210 110	0.170	19 0.210 110	0.170
124 0.192 22	0.154	32 0.210 176	0.170	32 0.210 176	0.170	32 0.210 176	0.170
16 0.191 110	0.150	115 0.210 2	0.169	115 0.210 2	0.169	115 0.210 2	0.169
21 0.191 126	0.150	26 0.207 22	0.169	26 0.207 22	0.169	26 0.207 22	0.169
36 0.191 176	0.150	27 0.206 126	0.150	27 0.206 126	0.150	27 0.206 126	0.150
125 0.191 167	0.122	124 0.206 12	0.130	124 0.206 12	0.130	124 0.206 12	0.130
5 0.190 12	0.120	33 0.205 167	0.114	33 0.205 167	0.114	33 0.205 167	0.114

Table 54: Analytical results for potassium in precipitations samples.

Potassium in precipitation				Potassium in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 0.204				Theoretical value: 0.238			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:	55	Number of laboratories:	55	Arithmetic mean value:	0.234	Arithmetic mean value:	0.234
Arithmetic mean value:	0.209	Median:	0.205	Median:	0.234	Standard deviation	0.030
Median:	0.205	Standard deviation	0.055	Rel. st. deviation (%)	12.738	Rel. st. deviation (%)	0.030
Standard deviation	0.055	Rel. st. deviation (%)	26.272				
Rel. st. deviation (%)	26.272						
Run 2:				Run 2:			
Number of laboratories:	53	Number of laboratories:	52	Arithmetic mean value:	0.234	Arithmetic mean value:	0.234
Arithmetic mean value:	0.205	Median:	0.205	Median:	0.233	Standard deviation	0.020
Median:	0.205	Standard deviation	0.022	Rel. st. deviation (%)	8.716	Rel. st. deviation (%)	0.020
Standard deviation	0.022	Rel. st. deviation (%)	10.766				
Rel. st. deviation (%)	10.766						
Results in decreasing order:				Results in decreasing order:			
43 < 0.726		43 < 0.726		176	0.330 (*)	31	0.232
176	0.560 (*)	155	0.204	17	0.300 (*)	8	0.231
42	0.284	3	0.203	42	0.284	166	0.231
17	0.250	15	0.203	150	0.279	3	0.230
5	0.235	27	0.203	5	0.275	15	0.230
150	0.233	2	0.201	41	0.273	33	0.230
16	0.232	19	0.200	110	0.260	157	0.230
124	0.230	32	0.200	21	0.255	14	0.227
157	0.230	14	0.199	112	0.250	153	0.224
116	0.227	23	0.199	115	0.250	22	0.222
21	0.220	104	0.199	155	0.249	125	0.221
112	0.220	8	0.197	116	0.247	4	0.220
179	0.220	31	0.197	120	0.247	12	0.220
120	0.218	166	0.193	36	0.246	26	0.218
115	0.217	4	0.190	158	0.246	121	0.218
41	0.216	12	0.190	35	0.244	124	0.218
45	0.213	121	0.190	45	0.244	104	0.213
10	0.211	26	0.188	163	0.242	24	0.211
13	0.210	125	0.186	10	0.240	118	0.210
20	0.210	153	0.183	16	0.240	126	0.210
35	0.210	24	0.182	19	0.240	180	0.207
110	0.210	118	0.180	20	0.240	167	0.201
158	0.210	126	0.180	164	0.240	30	0.200
164	0.210	180	0.174	7	0.238	179	0.197
22	0.209	30	0.170	2	0.235	38	0.190
36	0.208	38	0.160	23	0.234	178	0.118 (*)
163	0.208	167	0.148	27	0.234		
7	0.205	178	0.090 (*)				
33	0.205						
Potassium in precipitation				Potassium in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value: 0.187				Theoretical value: 0.153			
Unit: µg/l				Unit: µg/l			
Run 1:				Run 1:			
Number of laboratories:	55	Number of laboratories:	54	Arithmetic mean value:	0.149	Arithmetic mean value:	0.149
Arithmetic mean value:	0.182	Median:	0.181	Median:	0.147	Standard deviation	0.023
Median:	0.181	Standard deviation	0.032	Rel. st. deviation (%)	15.509	Rel. st. deviation (%)	0.023
Standard deviation	0.032	Rel. st. deviation (%)	17.362				
Rel. st. deviation (%)	17.362						
Run 2:				Run 2:			
Number of laboratories:	51	Number of laboratories:	51	Arithmetic mean value:	0.147	Arithmetic mean value:	0.147
Arithmetic mean value:	0.180	Median:	0.181	Median:	0.147	Standard deviation	0.014
Median:	0.181	Standard deviation	0.017	Rel. st. deviation (%)	9.826	Rel. st. deviation (%)	0.014
Standard deviation	0.017	Rel. st. deviation (%)	9.365				
Rel. st. deviation (%)	9.365						
Results in decreasing order:				Results in decreasing order:			
43 < 0.726		43 < 0.726		176	0.240 (*)	120	0.147
176	0.320 (*)	10	0.181	150	0.206 (*)	2	0.146
17	0.267 (*)	19	0.180	42	0.193	10	0.146
150	0.224	32	0.180	110	0.180	23	0.146
42	0.220	8	0.179	17	0.177	27	0.146
110	0.210	26	0.178	41	0.168	125	0.143
5	0.206	3	0.175	36	0.164	155	0.143
41	0.204	33	0.173	5	0.162	167	0.143
36	0.197	125	0.173	15	0.160	26	0.142
116	0.197	153	0.172	157	0.160	31	0.141
21	0.196	121	0.171	115	0.159	33	0.141
158	0.196	166	0.171	158	0.159	4	0.140
13	0.190	4	0.170	7	0.158	12	0.140
20	0.190	12	0.170	116	0.158	19	0.140
112	0.190	22	0.170	21	0.157	126	0.140
120	0.190	124	0.169	35	0.154	166	0.138
164	0.190	24	0.164	16	0.152	104	0.137
7	0.189	179	0.163	163	0.151	124	0.137
15	0.188	118	0.160	121 < 0.150			
35	0.188	126	0.160	13	0.150	22	0.133
163	0.188	157	0.160	20	0.150	24	0.132
115	0.187	31	0.157	32	0.150	118	0.130
23	0.186	104	0.157	112	0.150	179	0.129
14	0.184	30	0.150	164	0.150	153	0.127
16	0.184	38	0.150	3	0.149	30	0.120
27	0.182	180	0.145	8	0.148	38	0.120
45	0.182	167	0.113 (*)	45	0.148	180	0.118
155	0.182	178	0.091 (*)	14	0.147	178	0.071 (*)
2	0.181						

Table 55: Analytical results for conductivity in precipitations samples.

Conductivity in precipitation				Conductivity in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 27.000				Theoretical value: 30.000			
Unit: $\mu\text{S}/\text{cm}$				Unit: $\mu\text{S}/\text{cm}$			
Run 1:				Run 1:			
Number of laboratories:	53	Number of laboratories:	53	Number of laboratories:	53	Number of laboratories:	53
Arithmetic mean value:	25.010	Arithmetic mean value:	27.332	Arithmetic mean value:	27.332	Arithmetic mean value:	27.332
Median:	26.200	Median:	28.400	Median:	28.400	Median:	28.400
Standard deviation	3.869	Standard deviation	3.945	Standard deviation	3.945	Standard deviation	3.945
Rel. st. deviation (%)	15.468	Rel. st. deviation (%)	14.433	Rel. st. deviation (%)	14.433	Rel. st. deviation (%)	14.433
Run 2:				Run 2:			
Number of laboratories:	52	Number of laboratories:	52	Number of laboratories:	52	Number of laboratories:	52
Arithmetic mean value:	25.440	Arithmetic mean value:	27.801	Arithmetic mean value:	27.801	Arithmetic mean value:	27.801
Median:	26.290	Median:	28.450	Median:	28.450	Median:	28.450
Standard deviation	2.299	Standard deviation	1.994	Standard deviation	1.994	Standard deviation	1.994
Rel. st. deviation (%)	9.037	Rel. st. deviation (%)	7.172	Rel. st. deviation (%)	7.172	Rel. st. deviation (%)	7.172
Results in decreasing order:				Results in decreasing order:			
10	29.100	46	26.120	14	30.720	43	28.200
176	29.000	24	26.000	15	30.200	150	28.200
14	28.280	150	26.000	121	29.800	104	28.000
15	27.900	8	25.900	10	29.700	124	28.000
31	27.300	121	25.800	33	29.700	176	28.000
33	27.300	112	25.700	20	29.600	164	27.870
12	27.200	126	25.600	31	29.500	36	27.700
30	27.200	164	25.540	178	29.500	112	27.700
20	27.100	36	25.100	16	29.400	153	27.540
172	27.100	118	25.000	35	29.370	17	27.400
116	27.000	124	25.000	12	29.300	32	27.150
120	27.000	115	24.300	30	29.200	118	27.000
35	26.960	153	24.170	155	29.140	116	26.710
155	26.930	32	24.000	3	29.100	115	26.500
16	26.900	22	23.500	27	29.100	24	26.000
178	26.900	40	23.000	45	29.100	40	26.000
45	26.800	42	22.800	107	29.000	110	26.000
5	26.700	17	22.650	120	29.000	126	25.700
38	26.600	110	22.400	158	29.000	7	25.540
21	26.510	23	21.700	46	28.870	22	24.800
27	26.500	7	21.460	5	28.800	23	24.800
166	26.500	43	21.300	163	28.800	42	24.500
107	26.400	104	20.000	21	28.740	13	23.700
158	26.400	13	19.700	172	28.700	38	22.900
163	26.400	41	19.570	19	28.600	41	20.880
3	26.380	2	2.670 (*)	166	28.500	2	2.940 (*)
19	26.200			8	28.400		
Conductivity in precipitation				Conductivity in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value:	20.000			Theoretical value:	23.000		
Unit: $\mu\text{S}/\text{cm}$				Unit: $\mu\text{S}/\text{cm}$			
Run 1:				Run 1:			
Number of laboratories:	53	Number of laboratories:	52	Number of laboratories:	52	Number of laboratories:	52
Arithmetic mean value:	19.969	Arithmetic mean value:	22.151	Arithmetic mean value:	22.151	Arithmetic mean value:	22.151
Median:	20.000	Median:	22.745	Median:	22.745	Median:	22.745
Standard deviation	5.326	Standard deviation	3.687	Standard deviation	3.687	Standard deviation	3.687
Rel. st. deviation (%)	26.670	Rel. st. deviation (%)	16.646	Rel. st. deviation (%)	16.646	Rel. st. deviation (%)	16.646
Run 2:				Run 2:			
Number of laboratories:	51	Number of laboratories:	50	Number of laboratories:	50	Number of laboratories:	50
Arithmetic mean value:	19.684	Arithmetic mean value:	22.276	Arithmetic mean value:	22.276	Arithmetic mean value:	22.276
Median:	20.000	Median:	22.745	Median:	22.745	Median:	22.745
Standard deviation	1.397	Standard deviation	1.543	Standard deviation	1.543	Standard deviation	1.543
Rel. st. deviation (%)	7.095	Rel. st. deviation (%)	6.924	Rel. st. deviation (%)	6.924	Rel. st. deviation (%)	6.924
Results in decreasing order:				Results in decreasing order:			
43	52.400 (*)	163	20.000	38	35.700 (*)	12	22.700
38	23.300	21	19.990	126	25.900	163	22.700
3	21.780	116	19.940	121	24.300	107	22.600
121	21.500	166	19.900	14	24.250	166	22.500
14	21.370	19	19.700	15	24.000	150	22.200
8	21.200	12	19.500	8	23.900	110	22.100
46	21.130	107	19.400	116	23.840	112	22.100
15	21.100	112	19.400	33	23.700	32	22.000
33	21.000	150	19.340	20	23.600	176	22.000
20	20.900	36	19.300	178	23.400	153	21.710
178	20.600	32	19.200	31	23.300	164	21.590
30	20.500	104	19.000	3	23.260	35	21.540
31	20.500	176	19.000	27	23.200	118	21.500
27	20.400	7	18.880	30	23.100	7	21.490
126	20.400	110	18.700	155	23.070	36	21.200
10	20.300	153	18.630	24	23.000	43	21.200
172	20.300	118	18.500	104	23.000	115	21.000
155	20.200	115	18.100	120	23.000	46	20.850
16	20.100	164	18.100	124	23.000	22	20.200
45	20.100	13	18.000	172	23.000	42	20.200
158	20.100	23	17.900	16	22.900	23	20.100
35	20.060	22	17.600	45	22.900	17	20.070
17	20.050	42	17.100	5	22.800	13	19.400
5	20.000	40	17.000	19	22.800	40	19.000
24	20.000	41	14.830	158	22.800	41	17.060
120	20.000	2	2.040 (*)	21	22.790	2	2.340 (*)
124	20.000						

Appendix 4

Figures – 28th intercomparison




Figure 18: Youden plot of SO₂-S in absorbing solution.




Figure 19: Youden plot of SO₂-S on impregnated filter.




Figure 20: Youden plot of HNO₃-N on impregnated filter.




Figure 21: Youden plot of NO₂-N in absorbing solution.




Figure 22: Youden plot of NH₃-N on impregnated filter.




Figure 23: Youden plot of SO_4 -S in precipitation.




Figure 24: Youden plot of NO_3 -N in precipitation.




Figure 25: Youden plot of $\text{NH}_4\text{-N}$ in precipitation.




Figure 26: Youden plot of pH in precipitation.




Figure 27: Youden plot of strong acid in precipitation.




Figure 28: Youden plot of Cl in precipitation.




Figure 29: Youden plot of Na in precipitation.





Figure 30: Youden plot of Mg in precipitation.






Figure 32: Youden plot of K in precipitation.




Figure 33: Youden plot of conductivity in precipitation.