



The twenty-fifth intercomparison of analytical methods within EMEP

Hilde Th. Uggerud and Anne-Gunn Hjellbrekke



NILU : EMEP/CCC-Report 7/2008
REFERENCE : O-7729
DATE : DECEMBER 2008

**EMEP Co-operative Programme for Monitoring and Evaluation
of the Long-range Transmission of Air Pollutants
in Europe**

**The twenty-fifth intercomparison of
analytical methods within EMEP**

Hilde Th. Uggerud and Anne-Gunn Hjellbrekke



Norwegian Institute for Air Research
P.O. Box 100, N-2027 Kjeller, Norway

Contents

	Page
1. Introduction.....	5
2. Organisation of the intercomparison	5
3. Data handling	6
3.1 Data analysis.....	6
3.2 Youden plot	6
3.3 Estimating random and systematic errors from laboratory comparisons	7
3.3.1 Estimating random errors	7
3.3.2 Estimating systematic errors.....	9
4. Results	9
4.1 Sulphur dioxide in absorbing solution (A-samples).....	9
4.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples) ...	10
4.3 Nitrogen dioxide in absorbing solution (C-samples).....	10
4.4 Ammonia on impregnated filters (J-samples)	10
4.5 Precipitation (G-samples).....	11
5. Summary.....	11
6. References	11
Appendix 1 Tables	15
Appendix 2 Figures	43

The twenty-fifth intercomparison of analytical methods within EMEP

1. Introduction

40 different laboratories in European countries are performing chemical analysis of air and precipitation samples within EMEP (Co-operative Programme for Monitoring and Evaluation of Long-range Transmission of Air Pollutants in Europe). Since the measurement programme is based on individual national networks, the participating laboratories apply different sampling and analytical methods. Most of the methods used are described in the manual for sampling and chemical analysis (EMEP, 1996).

In order to improve the data comparability and to get a picture of the different laboratories' performance, interlaboratory comparisons are organised by the Chemical Co-ordinating Centre (CCC) at the Norwegian Institute for Air Research (NILU). So far twenty-five intercomparisons have been arranged (Hanssen, 1988, 1990; Hanssen et al., 1983; Hanssen and Ladegård, 1984, 1985, 1987; Hanssen and Skjelmoen, 1992, 1994, 1995, 1996, 1997, 2001; Thrane, 1978, 1980a, 1980b, 1981; Uggerud et al., 2001, 2002, 2003, 2004; Hjellbrekke et al., 2005; Uggerud and Hjellbrekke, 2007).

Since 2000 the laboratory intercomparisons within EMEP have also been open for participation of laboratories from other networks.

This report gives the results of the twenty-fifth interlaboratory test.

2. Organisation of the intercomparison

The samples for the twenty-fifth intercomparison (see Table 2) were prepared and distributed to 89 laboratories in October 2007.

Most of the laboratories had returned their results to the CCC within one month after the deadline given as 15 December 2007. A total of 66 laboratories have returned their results. This includes 32 EMEP-laboratories.

The participating laboratories received the theoretical (expected) values by e-mail 29.01.2008. The laboratories were given the opportunity to compare their results with the expected ones, and give corrected values if obvious mistakes e.g. misprints had occurred. A few corrections were reported. In those cases the corrected value is used in this report. In accordance with the decision of the Steering Body of EMEP, the results are presented in such a way that the different laboratories are identified. Tables 3a and 3b give the names of the participating laboratories together with the numbers used when presenting the results in tables and figures.

Information received on the analytical methods used is given in Tables 4–9.

3. Data handling

The data reported from the participants are presented in Tables 13-28 and Figures 2-17. An overview of all results is presented in Tables 10 and 11.

3.1 Data analysis

The reported values are presented in the tables in decreasing order together with the number of the laboratory. The expected (theoretical) value, the number of results, the arithmetic mean value, the median, the standard deviation and the relative standard deviation in percent are also given. After the first statistical run with all results included, the calculation was repeated with the outliers excluded. The outliers (unused) are defined as the results more than two standard deviations from the mean value in the first run.

The ratio between expected values (theoretical) to reported values, the ratio between measured to calculated conductivity and the ratio between equivalent concentrations of anions to equivalent concentrations cations, are presented in tables.

3.2 Youden plot

The Youden plot is a graphical method to analyse inter-laboratory data where the samples are ordered in pairs with similar concentrations. One plot is made for each pair of samples and gives results for all participating laboratories. The plots visualize both systematic and random errors.

The plot is drawn as a scatter plot where each point represents a pair of concentrations for one laboratory. The expected values for the two samples are drawn as solid blue lines. The arithmetic average of the measured values excluding outliers is drawn as dotted lines. The solid lines divide the plot in four quadrants and a 45° reference line going through the intercept of the solid lines may be added.

If errors are due to random factors, the points will be evenly distributed around the mean value and situated in all four quadrants.

If systematic errors dominate, the results will be close to 45° reference line, and be situated in the upper right quadrant (overestimation) or lower left quadrant (underestimation).

Drawing a line from a given point perpendicular on the 45° reference line gives two line segments, one from the point to the intercept on the reference line (a), and one continuing from the intercept to the point representing the expected values (b). The lengths of these line segments are measures of the random and systematic errors respectively.

Ellipses with radii corresponding to the data quality objectives (DQO, Table 1) are added in each plot. The data points are colour coded depending on the magnitude of errors as given in Table 1.




Figure 1: Youden plot showing concentrations for a pair of samples (green), expected values (blue lines), average of measured values (dotted lines) and random and systematic errors (red lines)

In Figures 18–33 the reported data are presented in Youden plots.

Table 1: Youden plot parameters.

Radii	Components
10%	SO ₂ in abs.sol, NO ₂ in abs.sol.
20%	SO ₂ , HNO ₃ and NH ₃ in impregnated filter
Radii = DQO	Components
10% accuracy or better	SO ₄ ²⁻ , NO ₃ ⁻
15% accuracy or better	NH ₄ ⁺ , Cl ⁻ , Ca ²⁺ , K ⁺ , Mg ²⁺ , Na ⁺ , cond, H ⁺ (from pH)
0.1 units	pH
Criteria	Colour
Within 0.5*DQO	Blue
Within DQO	Green
Within 2*DQO	Orange
> 2*DQO	Red


3.3 Estimating random and systematic errors from laboratory comparisons

Table 12 presents relative random and relative systematic errors obtained by the different laboratories in the analysis of each parameter in the precipitation samples. The calculation method and assumptions used are given in Chapter 3.3.1 and Chapter 3.3.2.

3.3.1 Estimating random errors

Systematic errors or bias in the laboratory analyses give a constant shift in the results from the expected ones at a particular concentration level. It is assumed that laboratories taking part in comparisons will obtain results near the expected

ones when this bias is removed, and that the differences between expected and obtained results more often will be close to zero than not. A triangular distribution, based upon this assumption, can be used to quantify the random errors in the laboratory results (Eurachem/CITAC, 2000).



The triangle distribution is symmetric with a baseline $2a$. The height in the triangle will be $1/a$ when the triangle area equals 1. The standard uncertainty is given by

$$u(x) = \frac{a}{\sqrt{6}} \quad (1)$$

and more than 95 % of the data will be within $\pm 2 \cdot u(x)$. The distance from $-a$ to a (i.e. $2a$) is called the range. When applied on the laboratory comparison results, the range equals the distance between the largest and smallest of the four differences between expected and found concentrations. As long as the bias can be assumed to be constant for the samples in the comparison of a specific component, it cannot have an effect on the distance corresponding to $2a$. The bias may be dependent upon the concentrations, but can be considered approximate constant for the concentrations used here in the comparison of the main components in precipitation, since the differences between the concentrations are small.

L and T represent the laboratories' and the expected concentrations respectively, and D is the difference. The difference for the lowest concentration is

$$D_1 = L_1 - T_1 \quad (2)$$

and the differences are D_1, D_2, D_3, D_4 in increasing order.

The range is $D_4 - D_1$ and the standard uncertainty for the differences $u(D)$ becomes

$$u(D) = \frac{D_4 - D_1}{\sqrt{6}} \quad (3)$$

The average expected concentration T for the four samples is given by

$$T = \frac{T_1 + T_2 + T_3 + T_4}{4} \quad (4)$$

The relative standard uncertainty, RSD, for 4 samples is given by $\frac{u(\bar{C})}{\bar{C}}$, or

$$\text{RSD} = \frac{2 \cdot D_1 - D_1 \cdot 100}{\sqrt{6} \cdot (C_1 + T_2 + T_3 + T_4)} \% \quad (5)$$

and 95 per cent of the laboratory results in this comparison are expected to be within $\pm 2 \cdot \text{RSD}$.

If the data quality objectives (DQO) likewise are looked upon as 95 percentiles, then 95 per cent of the laboratory analytical results should not be more than 10 or 15 per cent from the correct values (10 per cent for S and N containing components and 15 per cent for other components).

Correspondingly, the values $2 \cdot \text{RSD}$ should therefore be less than 10 or 15 per cent in order to comply with the DQO.

3.3.2 Estimating systematic errors

An estimation of bias in single measurements requires a long data series, and four samples as we normally have in laboratory comparison, are merely able to give an indication of the bias or a very coarse estimate.

Coarse estimates have been performed here in the cases where the four samples had similar concentrations and where all four laboratory results were either higher or lower than the expected concentrations. The median of the differences D_i , as defined above, was taken as a measure of the bias, B, in these cases.

$$B = \text{median}[D_i] \quad (6)$$

A relative bias, RB, was also calculated based upon the average expected concentration T , as defined in (4).

$$\text{RB} = \frac{4 \cdot \text{median}[D_i] \cdot 100}{C_1 + T_2 + T_3 + T_4} \% \quad (7)$$

4. Results

4.1 Sulphur dioxide in absorbing solution (A-samples)

Four samples and one blank solution were distributed to the laboratories that use the hydrogen peroxide absorption solution method. The results are given in Table 13, Figure 2. For those laboratories that reported a blank value this has been subtracted from the reported results.

8 laboratories have reported values for SO_2 in absorbing solution. Two laboratories reports systematically results below expected value.

The results are presented in Table 13 and Figure 2.

4.2 Sulphur dioxide and nitric acid on impregnated filter (B-samples)

Five impregnated filter samples (including one blank) for determination of sulphur dioxide were analysed by 20 laboratories. The value reported for the blank filter was subtracted from the other values before the data were used.

The amount of sulphur on the distributed filters corresponds to air concentrations between 0.48-3.20 $\mu\text{g S m}^{-3}$ when 25 m^3 is sampled.

In addition to sulphate, nitrate was added to the same impregnated filters for determination of $\text{HNO}_3\text{-N}$. The value reported for the blank filter was subtracted from the other values before using the data.

The amount of nitrogen on the distributed filters corresponds to air concentrations between 0.39 $\mu\text{g N m}^{-3}$ –1.18 $\mu\text{g N m}^{-3}$ when 25 m^3 sampling volume is used.

Both sulphur dioxide and nitric acid results show prevalence of systematic versus random errors. The systematic error is most clearly for the low concentration samples. More outlying results are reported for sulphur dioxide than nitric acid.

The results are presented in Tables 14 and 15 and Figures 3 and 4.

4.3 Nitrogen dioxide in absorbing solution (C-samples)

The four samples distributed were made to represent both absorption solutions and extracts from iodide-impregnated glass filters. The samples contain known amounts of sodium nitrite diluted in water. In order to assure sample stability and to give the laboratories the opportunity to use the matrix they use in their daily routine, the distributed samples were to be diluted 1:10. The results should be reported as the diluted concentrations.

The 10 times diluted samples correspond to air concentrations between 2.34-5.34 $\mu\text{g NO}_2\text{-N m}^{-3}$, when 70 ml absorbing solution and 1.4 m^3 are used. If 4 ml extraction solution and 0.7 m^3 sampling volume are used, the samples correspond to air concentrations between 0.27-0.61 $\mu\text{g NO}_2\text{-N m}^{-3}$.

14 laboratories have reported results. Two laboratories (lab 3 and 22) report results that deviates more than 20% from expected value. The rest of the reported results are mostly within 10% of expected value.

The results are presented in Table 16 and Figure 5.

4.4 Ammonia on impregnated filters (J-samples)

Six impregnated filters inclusive two unidentified blank filters were sent to 21 laboratories. 17 laboratories have reported their analytical results. The two blank values reported by each laboratory were averaged and subtracted from the other values reported before the data were used. The results are shown in Table 17 and Figure 6.

The amount of nitrogen on the filters correspond to air concentrations between 0.40-1.20 $\mu\text{g N m}^{-3}$, if 25 m^3 sampling volume is used.

17 laboratories have reported their results, which generally are in good agreement with expected value. Only one laboratory reports value outside DQO.

4.5 Precipitation (G-samples)

Four precipitation samples were distributed and 2411 single results from 64 laboratories were reported. 118 results were identified as outliers. This is ~5% of the data, which is slightly higher compared to last year's intercomparison.

Results for sulphate and nitrate are mostly in good agreement with expected value and few results outside DQO are reported. Standard deviation when outliers are excluded varies between 4-6%, which is very good.

Results for ammonium, potassium and chloride shows an improvement compared to last year. Fewer results outside DQO are reported.

Poorer results were reported for magnesium and calcium this round when compared to earlier years. However, only three samples were analysed for Ca in this last round. For magnesium the standard deviation varies between 13-19% and there is an increase in reported results between 15-25% away from expected value. The standard deviation for calcium varies between 16-29% and more results outside DQO are reported compared earlier intercomparisons.

The results are presented in Tables 18-28 and in Figures 7-28.

5. Summary

A total of 66 laboratories participated in the twenty-fifth intercomparison. 32 of these laboratories are within the EMEP network.

As in earlier intercomparisons, outliers are defined as values that deviates more than two standard deviations from the mean value. Outliers occur for all samples and almost all parameters. Out of a total of 2809 single results, 141 are defined as outliers. This is 5% of the reported data, which is slightly higher compared to last year.

6. References

Eurachem/CITAC (2000) Quantifying uncertainty in analytical measurements. 2nd ed. URL: <http://www.measurementuncertainty.org/mu/QUAM2000-1.pdf>.

EMEP (1996) Manual for sampling and chemical analysis. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 1/95) (Newest revision in 2001).

Hanssen, J.E. (1988) The tenth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 3/88).

- Hanssen, J.E. (1990) The eleventh intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/90).
- Hanssen, J.E. and Ladegård, N.E. (1984) The seventh intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/84).
- Hanssen, J.E. and Ladegård, N.E. (1985) The eighth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 4/85).
- Hanssen, J.E. and Ladegård, N.E. (1987) The ninth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 4/87).
- Hanssen, J.E., Ladegård, N.E. and Thrane, K.E. (1983) The sixth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/83).
- Hanssen, J.E. and Skjelmoen, J.E. (1992) The twelfth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 7/92).
- Hanssen, J.E. and Skjelmoen, J.E. (1994) The thirteenth intercomparison of analytical methods within EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/94).
- Hanssen, J.E. and Skjelmoen, J.E. (1995) The fourteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 3/95).
- Hanssen, J.E. and Skjelmoen, J.E. (1996) The fifteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 2/96).
- Hanssen, J.E. and Skjelmoen, J.E. (1997) The sixteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 2/97).
- Hanssen, J.E. and Skjelmoen, J.E. (2001) The seventeenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 10/2001).
- Hjellbrekke, A.-G., Uggerud, H., Hanssen, J.E. and Schaug, J. (2005) The twenty-second intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 8/2005).
- Thrane, K.E. (1978) Report on the first intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/78).

- Thrane, K.E. (1980a) Report on the second and third intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 1/80).
- Thrane, K.E. (1980b) Report on the fourth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 5/80).
- Thrane, K.E. (1981) The fifth intercomparison of analytical methods within the EMEP. Lillestrøm, Norwegian Institute for Air Research (EMEP/CCC-Report 2/81).
- Uggerud, H.T., Hanssen, J.E. and Skjelmoen, J.E. (2001) The eighteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 11/2001).
- Uggerud, H.T., Hanssen, J.E., Schaug, J. and Skjelmoen, J.E. (2002) The nineteenth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 1/2002).
- Uggerud, H.T., Hanssen, J.E., Schaug, J. and Skjelmoen, J.E. (2003) The twentieth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 8/2003).
- Uggerud, H.T., Hanssen, J.E., Schaug, J. and Hjellbrekke, A.-G. (2004) The twenty-first intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 6/2004).
- Uggerud, H. and Hjellbrekke, A.-G. (2007) The twenty-third and twenty-fourth intercomparison of analytical methods within EMEP. Kjeller, Norwegian Institute for Air Research (EMEP/CCC-Report 5/2007).

Appendix 1

Tables

Table 2: Samples distributed for the twenty-second interlaboratory test.

A.	5 synthetic samples for determination of SO ₂ , consisting of 0.3% H ₂ O ₂ absorbing solution and containing different concentrations of sulphuric acid. One of the samples was an unidentified blank.
B.	5 KOH-impregnated Whatman 40 filters, comprising 1 blank and 4 filters to which different amounts of sulphuric acid and nitrate salt have been added.
C.	4 synthetic samples for determination of NO ₂ consisting of sodium nitrite diluted in water.
J.	6 Whatman 40 filters impregnated with 3% oxalic acid, comprising 2 blank and 4 filters to which different amounts of ammonium salt solution have been added.
G.	4 synthetic precipitation samples, containing SO ₄ ²⁻ , NO ₃ ⁻ , NH ₄ ⁺ , H ⁺ , Na ⁺ , Mg ²⁺ and Cl ⁻ , and Ca ²⁺ and K ⁺ .

Table 3a: EMEP laboratories participating in the twenty-second laboratory intercomparison. The numbers in front of the names are used in tables and figures.

Austria	(1)	Umweltbundesamt Zweigstelle Sud, Klagenfurt
Belgia	(2)	Flemish Environment Agency, Antwerpen
Canada	(26)	Meteorological Service of Canada, Toronto
Croatia	(35)	Meteorological and Hydrological Service of Croatia
Czech Republic	(3)	Czech Hydrometeorological Institute, Praha
Denmark	(4)	National Environmental Research Institute. Air Pollution Laboratory
Estonia	(38)	Estonian Environmental Research Centre, Tallinn
Finland	(5)	Finnish Meteorological Institute. Air Quality Department
France	(6)	SGS Multilab, Saint Guenault-Courcouronnes
Georgia	(43)	Centre for Monitoring and Prognostication, Tbilisi
Germany	(7)	IfE Leipzig GmbH, Umweltlabor
Germany	(8)	Umweltbundesamt, Messtelle Schauinsland
Hungary	(10)	Institute for Atmospheric Physics
Iceland	(11)	Ídntæknistofnun Íslands (Technological Inst. of Iceland)
Ireland	(12)	Met. Eirann, Dublin
Italy	(13)	C.N.R. Istituto Inquinamento Atmosferico
Latvia	(33)	Air Pollution Observation Laboratory
Netherlands	(14)	National Institute of Public Health and Environmental Protection (RIVM)
Norway	(15)	Norwegian Institute for Air Research (NILU)
Macedonia	(40)	Hydrometeorological Institute, Skopje
Moldova	(42)	State Hydrometeorological Service, Chisinau
Poland	(16)	Institute of Meteorology and Water Management, Warsaw
Poland	(39)	Environmental Monitoring Laboratory, Institute of Environmental Protection
Portugal	(17)	Laboratorio Santo Andre
Russian Federation	(22)	Institute of Global Climate and Ecology
Serbia and Montenegro	(24)	Rep. Hydrometeorological Institute of Serbia
Slovakia	(31)	Slovak Hydrometeorological Institute
Slovenia	(36)	Hydrometeorological Institute of Slovenia
Spain	(19)	Centro Nacional de Sanidad Ambiental
Sweden	(20)	Swedish Environmental Research Institute (IVL), Gothenburg
Switzerland	(21)	Swiss Federal Laboratories for Materials Testing (EMPA)
United Kingdom	(23)	AEA Technology, National Environmental Technology Centre
United States of America	(27)	Illinois State Water Survey

Table 3b: Participating laboratories outside the EMEP network.

Germany	(102)	Ökologie-Zentrum Universität München
Germany	(104)	Hessige Landwirtschaftliche
Finland	(107)	The Finnish Forest Institute
Germany	(108)	Institut f. Bondenkunde und Standortlehre, Dredsten
Germany	(109)	Bügen-Institute, dep. of Soil Science of Temporal and Boreal Ecosystems
Germany	(110)	Thüringer Landesanstalt für Landwirtschaft (TTL), Jena
Germany	(112)	Niedersächsische Forstliche Versuchsanstalt (NVF)
Germany	(113)	Landesforstanstalt Eberswalde, abt. Waldökologie
Italy	(114)	C.N.R. Istituto Italiano di Idrobiologia
Germany	(115)	Bayerische Landesanstalt f. Wald- und Forstwirtschaft
Switzerland	(116)	Institute for Applied Plant Biology
Germany	(117)	Sächsische Landesanstalt für Forsten, Graupa
Germany	(118)	Forstliche Versuchs-und Forschungsansta
Germany	(120)	Landwirtschaftliche Untersuchungs- und Forschungsanstalt (LUFA)
Germany	(121)	Landeslabor Schleswig-Holstein
Belgium	(124)	Laboratorium voor Bondenkunde, Gent
Germany	(125)	Bayerisches Landesamt für Umweltschutz, Augsburg
Italy	(126)	APPA Laboratorio Biologico Provinciale
Finland	(145)	Tartu Environmental Research, Tartu
Luxembourg	(146)	Cellule de Recherche en Environment et Biotechnologies Public Research Center-Gabriel Lippman
Spain	(150)	Fundación Centro de Estudios ambientales del mediterrain
Belgium	(151)	Laboratoire de l'Unité des Eaux et Forêt (EFOR), Louvain-la-Neuve
Norway	(152)	Norwegian Forest Research Institute, Ås
Slovenia	(153)	Slovenian Forestry Institute, Ljubljana
United Kingdom	(155)	Environmental Research Branch, Forest Research
Hungary	(157)	Ecological Laboratory of Forest research Institute
Japan	(158)	Acid Deposition and Oxidant Research Center (ADOCRC), Niigata
France	(159)	CARSO, Lyon
Ireland	(160)	Coillite Research Laboratory
Thailand	(163)	Environmental Researching and Training Center (ERTC)
Thailand	(164)	Pollution Control Department (PCD)
Viet Nam	(165)	Institute of Meteorology and Hydrology
Poland	(166)	Forest Research Institute, Laboratory of Forest Habitat Chemistry
United Kingdom	(167)	CEH Edinburgh
France	(172)	Soils Agronomy and Spatitization unit INRA

Table 4: Analytical methods used at the participating laboratories for the determination of sulphur dioxide in absorbing solution (A).

Method	Laboratory
1. Ion chromatography	6, 15, 17, 21, 36, 159

Table 5: Analytical methods used at the participating laboratories for the determination of sulphur dioxide on impregnated filters (B).

Method	Laboratory
1. Spectrophotometry	16
2. Ion chromatography	3, 4, 5, 8, 11, 15, 20, 22, 31, 33, 36, 109, 116, 158, 164, 165, 172
3. Capillary Ion Analysis	39

Table 6: Analytical methods used at the participating laboratories for determination of nitric acid on impregnated filters (B).

Method	Laboratory
1. Reduction to nitrite	16
2. Ion chromatography	3, 4, 5, 8, 15, 20, 22, 31, 33, 36, 109, 116, 158, 164, 165, 172
3. Capillary Ion Analysis	39

Table 7: Analytical method for determination of ammonia on impregnated filters (J).

Method	Laboratory
1. Spectrophotometry	4, 10, 15, 20, 33, 39, 116, 172
2. FIA	11
3 Ion chromatography	3, 5, 8, 13, 31, 36, 38, 158, 165

Table 8: Analytical method used for NO₂ in absorbing solution (C).

Method	Laboratory
1. Spectrophotometry	3, 8, 12, 15, 16, 20, 22, 31, 33, 35, 36, 38, 39
2. Ion chromatography	19

Table 9: Analytical methods used for the determination of chemical constituents in precipitation samples.

Lab no.	Network	SO ₄ ²⁻	NH ₄ ⁺	NO ₃ ⁻	Na ⁺	Mg ²⁺	Cl ⁻	Ca ²⁺	K ⁺
1	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
2	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
3	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
4	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AES	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AES
5	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
6	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
7	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
8	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
10	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
11	EMEP	ICP-AES	FIA	FIA	ICP-AES	ICP-AES	FIA	ICP-AES	ICP-AES
13	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
14	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-MS	ICP-MS	Ion chromatography	Ion chromatography	K methods
15	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
16	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
17	EMEP	Ion chromatography	Spectrophotometry, Indophenol	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
19	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
20	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
21	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
22	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
24	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
26	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Ca methods	Flame-AAS
27	EMEP	Ion chromatography	flow injection colorimetry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
31	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
33	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
35	EMEP								
36	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
38	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
39 station	EMEP								
40	EMEP		Spectrophotometry	Spectrophotometry			Spectrophotometry		
43	EMEP	Ion chromatography	photocolorimeter	Ion chromatography	Na methods	titrimetric	Ion chromatography	titrimetric	K methods
102	EMEP	Spectrophotometry	Spectrophotometry	Spectrophotometry	Flame-AES	Flame-AAS	Spectrophotometry	Flame-AAS	Flame-AES
102	EMEP	Spectrophotometry	Spectrophotometry	Spectrophotometry	Flame-AES	Flame-AAS	Spectrophotometry	Flame-AAS	Flame-AES
104	ICP-Forest	Ion chromatography	Spectrophotometry, CFA	Spectrophotometry with CFA	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
107	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
108	EMEP	ICP-AES	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Potentiometric method	ICP-AES	ICP-AES
109	EMEP	ICP-AES	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Potentiometric method	ICP-AES	ICP-AES
110	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
112	ICP-Forest	ICP-AES	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
113	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	ICP-AES	Ion chromatography	Ion chromatography	ICP-AES
114	ICP-Forest	Ion chromatography	Spectrophotometry	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
115	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	Ion chromatography

Table 9, cont.

Lab no.	Network	SO ₄ ²⁻	NH ₄ ⁺	NO ₃ ⁻	Na ⁺	Mg ²⁺	Cl ⁻	Ca ²⁺	K ⁺
116	Other	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AES	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AES
117	ICP-Forest	ICP-AES	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Spectrophotometry	ICP-AES	ICP-AES
118	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
120	ICP-Forest	ICP-AES	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Spectrophotometry	ICP-AES	ICP-AES
121		Spectrophotometry	Spectrophotometry	Spectrophotometry	ICP-AES	ICP-AES	Spectrophotometry	ICP-AES	ICP-AES
124	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
125	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
126	ICP-Forest	Ion chromatography	Spectrophotometry	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
145	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
146	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
150		Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
151	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	ICP-AES	Ion chromatography	ICP-AES	Ion chromatography
152	ICP-Forest	Ion chromatography	Spectrophotometry	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
153	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
155	EMEP	Ion chromatography	Spectrophotometry, CFA	Ion chromatography with suppressor	ICP-AES, with USN	ICP-AES, with USN	Ion chromatography	ICP-AES, with USN	ICP-AES, with USN
157	EMEP	Ion chromatography	Spectrophotometry	Ion chromatography	Flame-AAS	Flame-AAS	Ion chromatography	Flame-AAS	Flame-AAS
158	EANET	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
163	EANET	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
164	EMEP	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
165	EANET	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Flame-AAS	Ion chromatography	Flame-AAS	Ion chromatography
166	ICP-Forest	Ion chromatography	Ion chromatography	Ion chromatography	ICP-AES	ICP-AES	Ion chromatography	ICP-AES	ICP-AES
167	EMEP	Ion chromatography	Spectrophotometry, AMFIA	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography
172		Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography	Ion chromatography

Table 10, cont.

Lab no	SO ₄ ²⁻				NH ₄ ⁺				NO ₃ ⁻				Na ⁺				Mg ²⁺				Cl ⁻				Ca ²⁺				K ⁺				pH				Cond				H ⁺			
	% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value				% deviation from expected value							
	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4	G1	G2	G3	G4
121	-12	-13	-17	-11	-11	-10	-5	-5	7	6	6	4	15	21	15	11	-35	-22	-42	-44	<0,1	<0,1	-1	-24	-16	-30	-15	-3	-37	<0,1	46	0,1	0,04	0,04	0,08	-8	-4	11	5					
124	-2	-4	-3	-3	-25	-25	4	3	-3	-1	-10	-17	-6	-3	5	2	3	4	8	7	-11	-32	-30	-30	-2	-11	-9	-4	-2	10	<0,13	0,02	0,01	-0,01	0,01	1	1	4	4					
125	-1	-1	-2	-2	5	8	11	13	-6	-5	-6	-6	0	2	4	-2	12	36	30	23	11	1	-4	-5	39	66	15	0	-1	1	10													
126	-2																																											
145	-33	-29	-14	-15	10	14	19	6	-29	-21	-12	-13	7	7	-5	1	4	<0,1	28	<0,1	-11	-20	-16	-16	<0,2	25	13	<0,1	3	-3	<0,1													
146	1	1	2	1	-3	-6	-3	-3	0	1	1	-1	0	0	0	-4	-14	-19	-27	-22	1	-1	-1	-4	4	8	1	3	-24	4	10	-0,11	-0,12	-0,15	-0,13	-1	-1	2	1					
150	1	2	2	2	33	39	20	9	-2	-1	-1	0	-2	-2	-2	-5	-3	-4	-6	-5	10	16	-1	-2	-14	-14	-12	-10	-9	-1	8	0,1	0,22	0,22	0,12	-1	-2	-15	0					
151	1	2	4	3	-32	-33	-24	-19	0	3	3	2	5	4	0	-3	-2	-1	-4	-2	-7	-4	-2	-4	1	-1	-1	8	10	13	36	-0,23	-0,212	-0,197	-0,22	-3	-2	0	1					
152	1	0	0	2	0	-3	12	9	-7	-8	-7	-5	-5	-3	0	-3	-7	-10	-8	-7	48	23	-26	-9	5	6	3	2	0	-2	-6	0,04	0,04	0,02	0,03	-15	-16	-4	-7					
153	-5	-4	-4	-4	6	-3	0	-5	-7	-8	-7	-6	-6	5	5	-1	10	139	-4	-3	-4	-14	-11	-18	5	23	3	-15	-7	-12	-31	0,08	0,07	0,09	0,14	-4	-2	3	-1					
155	1	1	0	-1	-41	-10	0	-6	0	2	-2	0	2	2	5	-1	3	0	2	2	-11	-14	-10	-14	-3	4	-1	0	3	2	1	0,04	0,02	0,01	0,01	0	-1	3	3					
157	-1	4	8	5	75	94	45	49	0	5	3	-1	60	37	-66	19	16	19	20	26	300	215	115	83	31	34	44	2	8	-12	-12	-0,1	-0,2	-0,1	-0,2	42	7	5	4					
158	-2	-1	-2	-2	0	2	2	-1	-3	-3	-3	-4	0	-1	0	-3	0	-3	-5	1	21	10	1	-5	2	2	2	-3	-3	-5	-8	0,03	0,03	0,02	0,02	-8	-7	-4	-5					
160					-8	-7	-3	-1	0	-2	0	-1	-1	5	-1	-1	-3	-5	-28	-3					-2	-11	-1	13	13	10	18													
163	3	4	4	3	6	-18	4	1	0	2	3	1	3	5	2	-3					-11	-7	-7	-9	76	34	32	2	-2	3	-2	0,06	0,07	0,04	0,06	-2	-4	2	1					
164	-5	-4	-6	-5	13	12	12	6	-3	-5	-2	-1	3	1	2	-1	-3	-10	-4	-3	-11	-14	-7	-9	-2	-5	-1	7	-7	-4	-12	0,12	0,04	0,03	0,04	-4	-4	-15	-2					
165	11	12	42	-13	-6	-7	-23	-35	2	3	-10	-27	-5	-4	-15	-22	-6	1	-12	-24	51	37	20	13	105	82	39	-25	-15	8	33	0,1	0,16	0,01	0,05	-7	-12	1	-1					
166	-2	-2	1	-1	-13	-3	0	-5	-5	-6	-3	-6	-6	-7	-4	-9	-1	-10	-11	-8	-10	-12	-4	-11	-15	-16	-5	-27	-12	-26	18	0,15	0,11	0,13	0,12	-3	-1	0	0					
167	2	-1	-3	-3	-4	-1	-3	-5	-1	-1	-3	-5	-12	-8	-8	-10	-3	-9	-2	13	-24	-25	-14	-14	-11	-14	-17	-15	-21	-10	-1													
172	6	4	4	6	8	4	5	2	-1	-1	0	0	-12	-13	-13	-19	-21	-27	-27	-26	-10	-6	-9	-2	-12	-4	-16	-1	-13	-1	-23	-0,42	-0,4	-0,45	-0,4	0	1	6	5					

SO₄²⁻ and NO₃⁻ between ± 10 and 20%
 SO₄²⁻ and NO₃⁻ more than + 20%

NH₄⁺, NO₃⁻, Cl⁻, Na⁺, Mg²⁺, Ca²⁺, K⁺ and cond: between ± 15-25%
 NH₄⁺, NO₃⁻, Cl⁻, Na⁺, Mg²⁺, Ca²⁺, K⁺ and cond: between ± 15-25%

pH: between ± 0,1-0,2 pH-units
 pH: more than ± 0,2 pH-units

Table 11: Reported results for filter samples expressed as % deviation from expected value.

Air and aerosols																				
	Absorbing solution				Impregnated filter				Impregnated filter				Absorbing solution				Impregnated filter			
	SO ₂ -S				SO ₂ -S				HNO ₃ -N				NO ₂ -N				NH ₃ -N			
	% deviation form expected value				% deviation form expected value				% deviation form expected value				% deviation form expected value				% deviation form expected value			
	A1	A2	A4	A5	B2	B3	B4	B5	B1	B3	B4	B5	C1	C2	C3	C4	J1	J2	J4	J6
3					8	19	0	-7	-2	-9	-6	-7	-25	-36	-21	-15	-2	0	-5	3
4					2	0	3	-1	2	-1	1	0					-13	-7	-8	-6
5					2	2	8	1	4	-1	2	0					-7	-6	-6	-5
6	0	3	2	3																
8					10		3	-4	13		-2	-4	-6	-7	-13	-15	1	4	4	3
10																	-19	-19	-16	-24
11					-7	4	3	-4												
12													1	9	2	1				
13																	2	14	-5	5
15	-7	-3	-3	-5	-3	-13	-14	-7	-4	-9	-15	-9	-9	-15	-4	-7	-5	1	-2	0
16					-4	3	-2	-2	2	-5	-3	-3	-3	-4	-6	-6				
17	-16	-13	-13	-14																
19	-33	-23	-28	-21									1	0	0	2				
20					-8	2	-1	-4	-6	-8	-9	-10	-5	-4	-4	-3	-8	-6	-4	-9
21	5	4	6	3																
22					-7	-37	8	10	-11	-17	-17	-17	-18	26	-35	-35				
23	7	3	7	14																
31					-8	-5	-4	-6	-2	-6	-3	-1	1	4	0	3	-3	-2	-9	2
33					-3	3	0	0	3	2	-3	123	9	4	5	6	20	0	1	-6
35													-1	-4	0	1				
36	2	2	2	-54	24	26	6	5	2	-3	-4	-3	-5	-2	-3	-1	-4	-8	-6	-7
38													4	6	0	3	4	4	1	10
39					-13	-6	-10	-9	-5	-6	-6	-9	-1	2	3	2	-8	-7	-10	-6
109					-3	6	11	11	8	9	11	11								
116					-4	-11	-14	-9	1	-18	-17	-17					1	1	-2	0
158					20	31	10	11	6	6	5	6					-7	-6	-5	3
159	-2	-3	-2	3																
163					-3	6	11	11	8	9	11	11								
164					-4	-3	0	-2	2	1	0	-1								
165					-20	-25	5	0	13	9	4	4					-6	-6	-3	-4
172					5	10	0	-6	-7	-1	1	-1					8	5	1	9

Results between 10 and 20% or between -10 and -20% from expected value
 more than ± 20% from expected value

Table 12: Relative random and systematic errors obtained by the different laboratories in the analysis of each parameter in the precipitation samples.

Lab. no.	SO ₄ ²⁻		NO ₃ ⁻		NH ₄ ⁺		Mg ²⁺		H ⁺ calc	
	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %
1	1	3	0	1	3	2	0	-5	12	-14
2	1	4	1	3	1	3	1	10	163	84
3	0	1	1	-1	0	8	1	1	2	-4
4	0	0	1	0	1	-2	3	-17	2	-5
5	1	2	2	1	2	-1	1	1	2	-4
6	0	-1	1	-1	2	3	102	10	4	-3
7	2	18	1	-1	1	1	2	3	4	22
8	2	4	1	2	5	-16	351	-3	4	-9
10	1	-7	0	-4	2	-4	4	7	5	-11
11	1	-2	1	2	4	-15	2	0	4	0
13	10	49	1	4	3	4	0	-6	1	0
14	1	-3	1	1	3	5	1	-2	3	-22
15	1	3	2	-1	1	8	3	0	3	-9
16	1	0	2	-1	2	-4	13	1	3	0
17	3	-14	2	-7	4	7	6	-16	26	118
19	3	-8	3	-8	3	11	1	0	6	-27
20	1	0	1	0	4	4	5	21	6	-3
21	1	2	1	1	2	0	2	-1	4	-8
22	1	-5	1	-2	10	51	6	15	3	-3
23	2	-10	1	1	4	6	3	0	3	-5
24	2	-5	4	-8	23	30	6	16	7	-7
26	1	-1	3	3	0	0	1	-1	4	-16
27	0	3	1	1	1	2	2	2	4	-7
31	1	-3	1	-4	3	26	4	-2	3	-11
33	2	-2	2	-6	1	-5	6	17	8	-9
35	0	0	1	-1					2	-18
36	0	0	1	2	2	-7	2	-7	2	-2
38	1	3	1	-4	3	-2	3	-23	11	-29
39									2	0
40			112	269	4	31			32	98
42	2	-1	2	9	15	20	15	48		
43	1	1	2	-1	8	60	66	110		
102	2	1	11	24	7	17	6	18	8	16
104	0	-2	3	-2	5	4	2	-4	14	-9
107	7	24	1	-3	4	-24	6	31	4	-10
108	1	-3	2	-3	4	1	1	-12	17	-25
109	4	-11	2	0			10	16		
110	3	3	3	-6	1	-1	1	-15	7	-24
112	4	-5	2	7	6	0	3	0	5	-15
113	0	-8	2	-3	2	-9	6	-33	3	-13
114	1	0	1	-1	2	-1	13	7	4	2
115	1	12	2	2	3	-4	8	-29	2	-4
116	1	-2	1	-4	4	5	5	-15	12	-5
117	1	-6	2	-4	1	-2	2	-2	33	-14
118	4	-1	5	-3	1	-2	2	-7	1	2
120	1	-8	2	-3	9	-38	2	-8	3	-3
121	2	-13	1	6	1	-7	16	-39	9	-12
124	1	-3	10	-6	9	-5	2	6	3	-3
125	0	-1	2	-5	7	8	3	24		
126	4	1	6	3	2	0	7	-32	3	3
145	12	-24	4	-18	5	9				
146	0	1	1	0	2	-3	4	-22	5	33
150	1	2	1	-1	3	22	1	-5	13	-26
151	1	3	1	3	5	-24	1	-2	23	64
152	1	1	1	-6	7	6	2	-7	4	-8
153	1	-4	2	-7	5	-1	38	6	4	-20
155	1	0	2	0	12	-8	2	2	4	-4
157	3	4	2	2	13	54	6	21	13	30
158	0	-2	1	-3	1	1	2	-1	3	-6
160			1	-1	14	20	8	-11		
163	1	3	1	1	6	3	11	19	6	-13
164	1	-5	1	-2	2	9	2	-4	11	-9
165	17	13	17	-4	22	-14	10	-9	16	-18
166	1	-2	2	-4	3	-5	3	-7	9	-24
167	2	-2	3	-2	3	-3	8	-3		
172	2	5	0	0	1	4	6	-24	42	152

Table 12, cont.

Lab. no.	Na+		Cl-		K+		Cond.	
	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %	Random error %	Systematic error %
1	3	0	8	-2	1	0	7	-4
2	2	-6	5	10	1	16	3	0
3	4	4	4	0	5	-2	3	3
4	3	-5	4	-4	1	3	1	1
5	3	2	5	2	3	3	1	6
6	3	1	3	-18	4	-19	3	-3
7	3	0	3	1	2	-2	3	-1
8	4	-5	2	1	3	-10	2	-1
10	6	-9	11	0	7	10	1	2
11	2	-3	7	1	24	-37	1	4
13	3	0	9	-6	4	4	6	24
14	4	-2	11	-21	2	-2	2	7
15	2	2	3	-2	2	2	2	2
16	4	-2	6	1	5	-3	2	-3
17	12	-26	18	77	47	2	18	-4
19	4	12			3	18	0	3
20	4	3	23	-2	4	-32	2	3
21	4	-4	3	-3	1	0	1	2
22	9	5	5	-37	6	6	3	-13
23	2	-4	2	-5	10	-5	6	-15
24	10	11	10	20	7	16	2	0
26	3	0	2	0	1	0		
27	5	1	10	1	1	0	2	0
31	3	-1	5	-8	7	5	1	5
33	6	-3	2	-5	12	9	4	-3
35	5	-1	1	-4	6	-11		
36	3	-4	5	3	2	-11	3	-9
38	2	-16	13	-14	6	-27	4	-2
39							8	-7
40			17	15			6	-19
42	8	35	12	2	2	46		
43	40	-44	2	0	22	-20		
102	16	-8					1	2
104	3	-10	3	8			17	2
107	2	6	9	1	4	2	0	-4
108	5	-6	58	13	4	-10	4	-7
109	2	-6			9	13		
110	7	-6					2	-3
112	3	1	3	1	3	24	2	-3
113	4	-11	9	-1			6	-9
114	5	-3	3	-3	3	-8	5	2
115	4	-9	12	55	14	-26	3	16
116	6	-3	44	16	1	4	2	3
117	2	18			4	11	19	-4
118	6	-10	7	-30	4	-13	2	-6
120	6	-7	9	4	6	-1		
121	3	16					7	0
124	4	0	15	-29			1	2
125	3	0	6	-2	3	0		
126	4	-13	10	-14	6	-24	1	8
145	4	3	7	-17				
146	2	0	3	-1	16	4	1	0
150	3	-2	6	3	7	-6	5	-2
151	3	2	1	-4	6	12	2	-1
152	2	-3	25	2	2	-1	6	-12
153	4	1	10	-12	5	-14	3	-2
155	3	1	6	-11	2	1	1	1
157	46	28	24	147	8	-3	19	6
158	2	0	8	4	1	-5	3	-6
160	2	-1			3	13		
163	3	2	3	-8	2	0	3	-1
164	1	1	2	-9	7	-6	4	-5
165	12	-10	5	25	21	-6	6	-4
166	4	-5	4	-7	17	-19	2	-1
167	3	-8	3	-17	11	-13		
172	8	-12	3	-5	7	-8	2	3

Table 13: Analytical results for sulphur dioxide in absorbing solution.

SO ₂ -S in absorbing solution				SO ₂ -S in absorbing solution			
Sample no.: A1				Sample no.: A2			
Theoretical value:		2.000		Theoretical value:		1.200	
Unit: ug S/ml				Unit: ug S/ml			
Run 1:				Run 1:			
Number of laboratories:		8		Number of laboratories:		8	
Arithmetic mean value:		1.894		Arithmetic mean value:		1.156	
Median:		1.985		Median:		1.196	
Standard deviation		0.265		Standard deviation		0.111	
Rel. st. deviation (%)		13.987		Rel. st. deviation (%)		9.608	
Run 2:				Run 2:			
Number of laboratories:		7		Number of laboratories:		7	
Arithmetic mean value:		1.972		Arithmetic mean value:		1.188	
Median:		2.010		Median:		1.223	
Standard deviation		0.156		Standard deviation		0.068	
Rel. st. deviation (%)		7.930		Rel. st. deviation (%)		5.704	
Results in decreasing order:				Results in decreasing order:			
23	2.131	159	1.960	21	1.243	15	1.170
21	2.108	15	1.870	23	1.231	159	1.170
36	2.048	17	1.680	6	1.230	17	1.050
6	2.010	19	1.345 (*)	36	1.223	19	0.929 (*)
SO ₂ -S in absorbing solution				SO ₂ -S in absorbing solution			
Sample no.: A4				Sample no.: A5			
Theoretical value:		1.800		Theoretical value:		4.000	
Unit: ug S/ml				Unit: ug S/ml			
Run 1:				Run 1:			
Number of laboratories:		8		Number of laboratories:		8	
Arithmetic mean value:		1.734		Arithmetic mean value:		3.646	
Median:		1.795		Median:		3.955	
Standard deviation		0.210		Standard deviation		0.855	
Rel. st. deviation (%)		12.107		Rel. st. deviation (%)		23.445	
Run 2:				Run 2:			
Number of laboratories:		7		Number of laboratories:		7	
Arithmetic mean value:		1.796		Arithmetic mean value:		3.904	
Median:		1.830		Median:		4.120	
Standard deviation		0.123		Standard deviation		0.479	
Rel. st. deviation (%)		6.874		Rel. st. deviation (%)		12.261	
Results in decreasing order:				Results in decreasing order:			
23	1.925	159	1.760	23	4.551	15	3.790
21	1.909	15	1.750	6	4.140	17	3.440
36	1.838	17	1.560	159	4.140	19	3.149
6	1.830	19	1.298 (*)	21	4.120	36	1.837 (*)

Table 14: Analytical results for sulphur dioxide on impregnated filter.

SO₂-S on impregnated filter
 Sample no.: 1
 Theoretical value: 22.040
 Unit:

Run 1:

Number of laboratories: 20
 Arithmetic mean value: 21.845
 Median: 21.305
 Standard deviation 2.284
 Rel. st. deviation (%) 10.456

Run 2:

Number of laboratories: 19
 Arithmetic mean value: 21.558
 Median: 21.300
 Standard deviation 1.941
 Rel. st. deviation (%) 9.003

Results in decreasing order:

36	27.300 (*)	15	21.300
158	26.390	116	21.260
8	24.200	164	21.260
3	23.901	16	21.130
172	23.040	11	20.410
5	22.510	22	20.400
4	22.480	31	20.375
33	21.333	20	20.300
163	21.310	39	19.100
109	21.310	165	17.590

SO₂-S on impregnated filter
 Sample no.: 3
 Theoretical value: 12.020
 Unit:

Run 1:

Number of laboratories: 19
 Arithmetic mean value: 12.077
 Median: 12.240
 Standard deviation 1.910
 Rel. st. deviation (%) 15.816

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 12.325
 Median: 12.283
 Standard deviation 1.618
 Rel. st. deviation (%) 13.129

Results in decreasing order:

158	15.720	5	12.220
36	15.200	4	11.960
3	14.251	164	11.670
172	13.190	31	11.379
163	12.710	39	11.300
109	12.710	116	10.740
11	12.520	15	10.400
16	12.340	165	8.980
33	12.326	22	7.600 (*)
20	12.240		

SO₂-S on impregnated filter
 Sample no.: 4
 Theoretical value: 64.090
 Unit:

Run 1:

Number of laboratories: 20
 Arithmetic mean value: 64.835
 Median: 65.065
 Standard deviation 4.688
 Rel. st. deviation (%) 7.230

Run 2:

Number of laboratories: 18
 Arithmetic mean value: 65.896
 Median: 65.930
 Standard deviation 3.555
 Rel. st. deviation (%) 5.394

Results in decreasing order:

109	71.010	172	64.270
163	71.010	164	64.110
158	70.400	33	64.050
22	69.300	3	64.011
5	69.260	20	63.180
36	67.800	16	62.950
165	67.530	31	61.405
11	66.090	39	57.900
8	66.000	15	55.400 (*)
4	65.860	116	55.170 (*)

SO₂-S on impregnated filter
 Sample no.: 5
 Theoretical value: 80.120
 Unit:

Run 1:

Number of laboratories: 20
 Arithmetic mean value: 79.710
 Median: 78.480
 Standard deviation 5.344
 Rel. st. deviation (%) 6.704

Run 2:

Number of laboratories: 20
 Arithmetic mean value: 79.710
 Median: 78.480
 Standard deviation 5.344
 Rel. st. deviation (%) 6.704

Results in decreasing order:

163	88.910	16	78.350
109	88.910	20	77.300
158	88.680	8	77.100
22	88.300	11	76.810
36	84.100	31	75.681
5	81.020	172	75.020
165	80.440	3	74.731
33	80.038	15	74.700
4	79.100	39	73.200
164	78.610	116	73.190

Table 15: Analytical results for nitric acid on impregnated filter.

HNO ₃ -N on impregnated filter				HNO ₃ -N on impregnated filter			
Sample no.: 1				Sample no.: 3			
Theoretical value:		9.820		Theoretical value:		18.000	
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories:		19		Number of laboratories:		18	
Arithmetic mean value:		9.956		Arithmetic mean value:		17.500	
Median:		10.000		Median:		17.645	
Standard deviation		0.645		Standard deviation		1.453	
Rel. st. deviation (%)		6.481		Rel. st. deviation (%)		8.305	
Run 2:				Run 2:			
Number of laboratories:		19		Number of laboratories:		18	
Arithmetic mean value:		9.956		Arithmetic mean value:		17.500	
Median:		10.000		Median:		17.645	
Standard deviation		0.645		Standard deviation		1.453	
Rel. st. deviation (%)		6.481		Rel. st. deviation (%)		8.305	
Results in decreasing order:				Results in decreasing order:			
8	11.100	164	9.970	163	19.620	36	17.500
165	11.090	116	9.920	109	19.620	16	17.030
163	10.620	31	9.631	165	19.550	31	16.986
109	10.620	3	9.622	158	19.040	39	16.900
158	10.440	15	9.400	33	18.328	20	16.560
5	10.220	39	9.300	164	18.120	3	16.340
33	10.081	20	9.260	172	17.890	15	16.300
16	10.030	172	9.140	5	17.800	22	14.900
4	10.020	22	8.700	4	17.790	116	14.730
36	10.000						
HNO ₃ -N on impregnated filter				HNO ₃ -N on impregnated filter			
Sample no.: 4				Sample no.: 5			
Theoretical value:		29.450		Theoretical value:		24.540	
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories:		19		Number of laboratories:		19	
Arithmetic mean value:		28.656		Arithmetic mean value:		25.481	
Median:		28.702		Median:		24.260	
Standard deviation		2.340		Standard deviation		7.316	
Rel. st. deviation (%)		8.167		Rel. st. deviation (%)		28.711	
Run 2:				Run 2:			
Number of laboratories:		19		Number of laboratories:		18	
Arithmetic mean value:		28.656		Arithmetic mean value:		23.860	
Median:		28.702		Median:		24.080	
Standard deviation		2.340		Standard deviation		1.953	
Rel. st. deviation (%)		8.167		Rel. st. deviation (%)		8.186	
Results in decreasing order:				Results in decreasing order:			
163	32.620	33	28.572	33	54.657 (*)	36	23.900
109	32.620	16	28.550	109	27.220	16	23.830
158	30.830	36	28.300	163	27.220	8	23.500
165	30.520	3	27.770	158	25.910	3	22.940
5	29.970	39	27.600	165	25.440	39	22.300
172	29.650	20	26.940	5	24.660	15	22.300
4	29.640	15	25.000	4	24.520	20	22.180
164	29.560	116	24.520	31	24.398	116	20.320
8	28.800	22	24.300	164	24.280	22	20.300
31	28.702			172	24.260		

Table 16: Analytical results for nitrogen dioxide in absorbing solution.

NO2-N in absorbing solution				NO2-N in absorbing solution			
Sample no.: 1				Sample no.: 2			
Theoretical value: 0.077				Theoretical value: 0.047			
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories: 14				Number of laboratories: 14			
Arithmetic mean value: 0.074				Arithmetic mean value: 0.046			
Median: 0.075				Median: 0.046			
Standard deviation 0.007				Standard deviation 0.006			
Rel. st. deviation (%) 9.124				Rel. st. deviation (%) 13.849			
Run 2:				Run 2:			
Number of laboratories: 13				Number of laboratories: 13			
Arithmetic mean value: 0.075				Arithmetic mean value: 0.047			
Median: 0.076				Median: 0.047			
Standard deviation 0.005				Standard deviation 0.005			
Rel. st. deviation (%) 6.867				Rel. st. deviation (%) 9.590			
Results in decreasing order:				Results in decreasing order:			
33	0.084	16	0.075	22	0.059	36	0.046
38	0.080	20	0.073	12	0.051	16	0.045
12	0.078	36	0.073	38	0.050	35	0.045
19	0.078	8	0.072	33	0.049	20	0.045
31	0.078	15	0.070	31	0.049	8	0.043
35	0.076	22	0.063	39	0.048	15	0.040
39	0.076	3	0.058 (*)	19	0.047	3	0.030 (*)
NO2-N in absorbing solution				NO2-N in absorbing solution			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.100				Theoretical value: 0.107			
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories: 14				Number of laboratories: 14			
Arithmetic mean value: 0.095				Arithmetic mean value: 0.102			
Median: 0.099				Median: 0.107			
Standard deviation 0.011				Standard deviation 0.011			
Rel. st. deviation (%) 11.467				Rel. st. deviation (%) 11.244			
Run 2:				Run 2:			
Number of laboratories: 13				Number of laboratories: 13			
Arithmetic mean value: 0.097				Arithmetic mean value: 0.105			
Median: 0.100				Median: 0.108			
Standard deviation 0.007				Standard deviation 0.007			
Rel. st. deviation (%) 7.213				Rel. st. deviation (%) 6.775			
Results in decreasing order:				Results in decreasing order:			
33	0.105	36	0.097	33	0.113	36	0.106
39	0.103	15	0.096	38	0.110	20	0.104
12	0.102	20	0.096	31	0.110	16	0.101
35	0.100	16	0.094	39	0.109	15	0.100
38	0.100	8	0.087	19	0.109	3	0.091
19	0.100	3	0.079	12	0.108	8	0.091
31	0.100	22	0.065 (*)	35	0.108	22	0.070 (*)

Table 17: Analytical results for ammonia on impregnated filter. The reported results are corrected for an average blank value (J3 and J5).

NH3-N on impregnated filter ug N/filter				NH3-N on impregnated filter ug			
N/filter				N/filter			
Sample no.: J1				Sample no.: J2			
Theoretical value:		16.040		Theoretical value:		20.050	
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories:		17		Number of laboratories:		17	
Arithmetic mean value:		15.616		Arithmetic mean value:		19.595	
Median:		15.400		Median:		19.641	
Standard deviation		1.365		Standard deviation		1.445	
Rel. st. deviation (%)		8.738		Rel. st. deviation (%)		7.373	
Run 2:				Run 2:			
Number of laboratories:		16		Number of laboratories:		15	
Arithmetic mean value:		15.394		Arithmetic mean value:		19.598	
Median:		15.290		Median:		19.641	
Standard deviation		1.045		Standard deviation		0.894	
Rel. st. deviation (%)		6.786		Rel. st. deviation (%)		4.562	
Results in decreasing order:				Results in decreasing order:			
33	19.170 (*)	15	15.180	13	22.900 (*)	158	18.890
172	17.250	165	15.130	172	21.050	20	18.870
38	16.745	158	14.980	38	20.845	165	18.860
13	16.300	5	14.960	8	20.800	5	18.750
116	16.275	39	14.795	15	20.220	4	18.725
8	16.200	20	14.770	116	20.165	39	18.695
3	15.660	4	14.015	3	20.000	36	18.500
31	15.611	10	13.025	33	19.965	10	16.236 (*)
36	15.400			31	19.641		
NH3-N on impregnated filter ug N/filter				NH3-N on impregnated filter ug			
N/filter				N/filter			
Sample no.: J4				Sample no.: J6			
Theoretical value:		30.080		Theoretical value:		10.030	
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories:		17		Number of laboratories:		17	
Arithmetic mean value:		28.767		Arithmetic mean value:		9.843	
Median:		28.680		Median:		9.990	
Standard deviation		1.449		Standard deviation		0.812	
Rel. st. deviation (%)		5.038		Rel. st. deviation (%)		8.245	
Run 2:				Run 2:			
Number of laboratories:		16		Number of laboratories:		16	
Arithmetic mean value:		28.983		Arithmetic mean value:		9.983	
Median:		28.705		Median:		10.032	
Standard deviation		1.184		Standard deviation		0.591	
Rel. st. deviation (%)		4.086		Rel. st. deviation (%)		5.923	
Results in decreasing order:				Results in decreasing order:			
8	31.200	158	28.590	38	11.045	165	9.670
172	30.450	13	28.500	172	10.950	5	9.540
38	30.445	5	28.250	13	10.500	33	9.450
33	30.420	36	28.200	3	10.370	4	9.435
15	29.440	4	27.785	158	10.350	39	9.395
116	29.425	31	27.481	8	10.300	36	9.300
165	29.130	39	26.995	31	10.241	20	9.110
20	28.730	10	25.326 (*)	116	10.075	10	7.611 (*)
3	28.680			15	9.990		

Table 18: Analytical results for sulphate in precipitations samples.

Sulphate in precipitation				Sulphate in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 1.162				Theoretical value: 1.120			
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories: 63				Number of laboratories: 63			
Arithmetic mean value: 1.162				Arithmetic mean value: 1.120			
Median: 1.162				Median: 1.117			
Standard deviation 0.110				Standard deviation 0.106			
Rel. st. deviation (%) 9.441				Rel. st. deviation (%) 9.457			
Run 2:				Run 2:			
Number of laboratories: 60				Number of laboratories: 60			
Arithmetic mean value: 1.155				Arithmetic mean value: 1.112			
Median: 1.162				Median: 1.115			
Standard deviation 0.061				Standard deviation 0.061			
Rel. st. deviation (%) 5.281				Rel. st. deviation (%) 5.454			
Results in decreasing order:				Results in decreasing order:			
13	1.710 (*)	4	1.162	13	1.640 (*)	6	1.114
107	1.440 (*)	108	1.160	107	1.430 (*)	126	1.110
7	1.313	125	1.153	7	1.320	167	1.109
165	1.289	6	1.152	165	1.251	26	1.108
115	1.280	157	1.150	115	1.250	125	1.106
172	1.240	26	1.150	110	1.170	158	1.105
8	1.230	158	1.148	172	1.170	11	1.102
118	1.220	166	1.144	2	1.165	104	1.100
110	1.200	104	1.140	8	1.160	116	1.097
163	1.200	124	1.140	163	1.160	31	1.094
15	1.200	33	1.140	15	1.160	166	1.093
38	1.200	126	1.140	1	1.160	14	1.090
1	1.200	42	1.139	157	1.160	108	1.085
27	1.195	14	1.130	27	1.150	153	1.080
167	1.188	31	1.128	38	1.150	33	1.080
5	1.187	11	1.128	102	1.147	124	1.080
21	1.186	24	1.120	21	1.146	118	1.080
2	1.185	22	1.119	150	1.144	22	1.074
102	1.184	164	1.110	151	1.142	164	1.070
3	1.184	153	1.110	5	1.142	24	1.070
150	1.183	10	1.098	3	1.141	117	1.059
20	1.182	117	1.097	42	1.138	113	1.045
152	1.180	113	1.082	146	1.135	10	1.045
43	1.180	112	1.070	114	1.130	112	1.040
155	1.175	120	1.070	155	1.128	120	1.040
146	1.173	23	1.054	43	1.125	19	1.024
151	1.172	19	1.048	36	1.124	23	1.019
36	1.170	121	1.030	20	1.122	109	1.000
114	1.170	109	1.000	4	1.122	121	0.970
16	1.167	17	1.000	152	1.120	17	0.970
35	1.166	145	0.780 (*)	35	1.119	145	0.790 (*)
116	1.162			16	1.117		
Sulphate in precipitation				Sulphate in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.712				Theoretical value: 0.886			
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories: 63				Number of laboratories: 63			
Arithmetic mean value: 0.716				Arithmetic mean value: 0.882			
Median: 0.710				Median: 0.874			
Standard deviation 0.072				Standard deviation 0.079			
Rel. st. deviation (%) 10.051				Rel. st. deviation (%) 8.915			
Run 2:				Run 2:			
Number of laboratories: 60				Number of laboratories: 60			
Arithmetic mean value: 0.704				Arithmetic mean value: 0.869			
Median: 0.710				Median: 0.872			
Standard deviation 0.044				Standard deviation 0.045			
Rel. st. deviation (%) 6.256				Rel. st. deviation (%) 5.150			
Results in decreasing order:				Results in decreasing order:			
13	1.010 (*)	152	0.710	13	1.310 (*)	5	0.873
165	1.009 (*)	118	0.710	7	1.074 (*)	16	0.871
7	0.878 (*)	35	0.709	107	1.070 (*)	6	0.871
107	0.854	16	0.707	115	1.000	118	0.870
115	0.820	42	0.705	172	0.940	104	0.870
126	0.770	14	0.703	157	0.930	110	0.870
157	0.770	158	0.700	2	0.920	11	0.869
2	0.750	24	0.700	151	0.917	158	0.868
163	0.740	6	0.699	8	0.916	116	0.866
15	0.740	125	0.698	27	0.911	125	0.864
172	0.740	167	0.692	43	0.910	14	0.864
151	0.739	104	0.690	38	0.910	102	0.862
8	0.734	124	0.690	126	0.910	167	0.861
27	0.732	112	0.690	163	0.910	108	0.860
110	0.730	11	0.690	15	0.910	124	0.860
146	0.725	116	0.686	150	0.908	31	0.857
3	0.723	108	0.685	1	0.906	153	0.850
1	0.723	31	0.684	152	0.900	22	0.844
150	0.723	153	0.680	21	0.897	164	0.840
5	0.723	22	0.677	146	0.896	117	0.832
36	0.721	164	0.670	3	0.896	10	0.824
21	0.720	117	0.665	36	0.887	24	0.820
38	0.720	19	0.662	4	0.883	19	0.819
102	0.717	120	0.650	155	0.881	120	0.810
166	0.717	10	0.645	112	0.880	113	0.803
43	0.717	23	0.645	33	0.880	109	0.800
26	0.713	113	0.633	35	0.880	23	0.799
155	0.711	109	0.630	114	0.880	121	0.790
20	0.711	17	0.623	20	0.879	165	0.773
4	0.711	145	0.610	42	0.875	17	0.770
114	0.710	121	0.590	166	0.874	145	0.750
33	0.710			26	0.874		

Table 19: Analytical results for nitrate in precipitations samples.

Nitrate in precipitation				Nitrate in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 0.300				Theoretical value: 0.294			
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories: 65				Number of laboratories: 65			
Arithmetic mean value: 0.298				Arithmetic mean value: 0.309			
Median: 0.298				Median: 0.292			
Standard deviation 0.023				Standard deviation 0.130			
Rel. st. deviation (%) 7.812				Rel. st. deviation (%) 42.087			
Run 2:				Run 2:			
Number of laboratories: 62				Number of laboratories: 64			
Arithmetic mean value: 0.296				Arithmetic mean value: 0.293			
Median: 0.298				Median: 0.291			
Standard deviation 0.013				Standard deviation 0.020			
Rel. st. deviation (%) 4.345				Rel. st. deviation (%) 6.674			
Results in decreasing order:				Results in decreasing order:			
102	0.407 (*)	20	0.298	40	1.329 (*)	113	0.291
40	0.371 (*)	167	0.298	102	0.391	150	0.291
42	0.332	3	0.297	42	0.330	167	0.291
112	0.330	16	0.296	126	0.320	35	0.290
121	0.321	172	0.296	112	0.320	110	0.290
126	0.320	35	0.296	121	0.313	3	0.290
26	0.313	6	0.295	157	0.310	118	0.290
13	0.313	43	0.294	26	0.308	124	0.290
118	0.310	150	0.293	13	0.306	15	0.290
11	0.308	113	0.292	2	0.305	6	0.290
36	0.308	22	0.291	8	0.304	104	0.290
2	0.308	164	0.290	151	0.304	108	0.290
8	0.307	158	0.290	43	0.303	160	0.289
27	0.305	124	0.290	36	0.303	22	0.285
165	0.305	38	0.290	165	0.302	158	0.285
1	0.304	116	0.288	11	0.302	31	0.282
21	0.303	166	0.285	155	0.300	117	0.281
115	0.302	107	0.285	21	0.300	38	0.280
5	0.302	117	0.284	163	0.300	107	0.280
14	0.302	31	0.283	109	0.300	116	0.280
114	0.301	125	0.283	1	0.299	120	0.280
23	0.301	10	0.283	14	0.298	164	0.280
109	0.300	108	0.280	27	0.298	17	0.280
163	0.300	120	0.280	23	0.297	125	0.278
155	0.300	24	0.280	146	0.296	10	0.277
15	0.300	33	0.280	5	0.296	166	0.276
157	0.300	153	0.280	4	0.295	153	0.270
104	0.300	152	0.280	16	0.294	33	0.270
4	0.300	17	0.277	7	0.294	152	0.270
160	0.300	110	0.270	115	0.294	19	0.267
146	0.299	19	0.267	20	0.294	24	0.260
151	0.299	145	0.213 (*)	114	0.292	145	0.231
7	0.298			172	0.292		
Nitrate in precipitation				Nitrate in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.398				Theoretical value: 0.566			
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories: 65				Number of laboratories: 65			
Arithmetic mean value: 0.412				Arithmetic mean value: 0.568			
Median: 0.397				Median: 0.553			
Standard deviation 0.134				Standard deviation 0.147			
Rel. st. deviation (%) 32.395				Rel. st. deviation (%) 25.791			
Run 2:				Run 2:			
Number of laboratories: 64				Number of laboratories: 64			
Arithmetic mean value: 0.396				Arithmetic mean value: 0.550			
Median: 0.397				Median: 0.553			
Standard deviation 0.020				Standard deviation 0.028			
Rel. st. deviation (%) 5.128				Rel. st. deviation (%) 5.165			
Results in decreasing order:				Results in decreasing order:			
40	1.460 (*)	172	0.397	40	1.710 (*)	26	0.553
102	0.489	150	0.396	42	0.618	3	0.552
112	0.430	6	0.396	121	0.589	10	0.552
42	0.428	114	0.396	2	0.585	22	0.552
13	0.424	7	0.394	115	0.582	20	0.552
121	0.423	43	0.394	13	0.581	6	0.551
8	0.414	107	0.393	112	0.580	16	0.551
11	0.412	22	0.392	151	0.578	38	0.550
2	0.412	164	0.390	8	0.574	109	0.550
163	0.410	104	0.390	11	0.572	15	0.550
157	0.410	155	0.389	27	0.572	108	0.545
36	0.410	166	0.388	163	0.570	113	0.544
151	0.409	158	0.388	36	0.569	158	0.543
27	0.409	31	0.387	1	0.569	116	0.543
115	0.408	10	0.385	102	0.569	31	0.541
21	0.406	167	0.385	150	0.568	104	0.540
26	0.406	113	0.385	155	0.567	167	0.540
1	0.405	116	0.383	172	0.566	152	0.540
16	0.403	117	0.381	21	0.565	110	0.540
23	0.403	118	0.380	14	0.564	166	0.534
146	0.403	33	0.380	23	0.563	117	0.533
4	0.402	110	0.380	157	0.560	17	0.533
35	0.402	38	0.380	120	0.560	125	0.530
5	0.402	125	0.375	164	0.560	153	0.530
14	0.401	152	0.370	35	0.559	33	0.530
15	0.400	153	0.370	146	0.559	118	0.530
120	0.400	24	0.370	160	0.559	126	0.530
126	0.400	17	0.370	4	0.558	19	0.514
109	0.400	19	0.366	107	0.556	24	0.510
108	0.400	124	0.360	43	0.556	145	0.492
160	0.399	165	0.358	7	0.556	124	0.470
3	0.397	145	0.350	114	0.553	165	0.414
20	0.397			5	0.553		

Table 20: Analytical results for ammonium in precipitations sample.

Ammonium in precipitation				Ammonium in precipitation			
Sample no.: 1				Sample no.: 2			
Theoretical value: 0.160				Theoretical value: 0.134			
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories: 64				Number of laboratories: 63			
Arithmetic mean value: 0.167				Arithmetic mean value: 0.141			
Median: 0.160				Median: 0.137			
Standard deviation 0.038				Standard deviation 0.034			
Rel. st. deviation (%) 22.886				Rel. st. deviation (%) 24.203			
Run 2:				Run 2:			
Number of laboratories: 61				Number of laboratories: 59			
Arithmetic mean value: 0.160				Arithmetic mean value: 0.137			
Median: 0.160				Median: 0.136			
Standard deviation 0.026				Standard deviation 0.023			
Rel. st. deviation (%) 16.383				Rel. st. deviation (%) 16.533			
Results in decreasing order:				Results in decreasing order:			
43	0.326 (*)	152	0.160	43	0.270 (*)	5	0.136
157	0.280 (*)	110	0.160	157	0.260 (*)	114	0.135
22	0.264 (*)	112	0.160	22	0.218 (*)	167	0.133
31	0.229	26	0.159	40	0.206	26	0.133
40	0.223	5	0.158	102	0.199	16	0.133
150	0.213	23	0.156	31	0.187	126	0.132
24	0.210	146	0.156	150	0.186	117	0.132
102	0.206	117	0.156	24	0.170	115	0.132
109	0.200	114	0.155	42	0.159	118	0.131
42	0.194	118	0.155	19	0.156	4	0.130
19	0.183	167	0.154	23	0.154	166	0.130
164	0.180	10	0.153	3	0.154	153	0.130
15	0.180	16	0.153	145	0.153	152	0.130
3	0.178	4	0.151	112	0.150	110	0.130
14	0.177	115	0.151	15	0.150	10	0.126
145	0.176	165	0.150	164	0.150	146	0.126
13	0.176	33	0.150	17	0.147	165	0.125
172	0.172	38	0.150	14	0.147	160	0.125
153	0.170	104	0.150	125	0.145	36	0.122
163	0.170	108	0.150	1	0.143	121	0.121
125	0.168	160	0.148	7	0.141	155	0.120
17	0.167	36	0.144	6	0.141	33	0.120
2	0.167	121	0.143	2	0.141	38	0.120
1	0.166	166	0.140	104	0.140	113	0.111
6	0.166	113	0.132	172	0.140	163	0.110
7	0.164	11	0.130	108	0.140	8	0.108
20	0.164	8	0.129	116	0.139	11	0.106
116	0.163	124	0.120	13	0.139	124	0.100
21	0.162	151	0.109	21	0.138	151	0.090
27	0.161	107	0.108	27	0.137	107	0.078
126	0.160	120	0.100	20	0.137	120	0.070 (*)
158	0.160	155	0.094	158	0.137		
Ammonium in precipitation				Ammonium in precipitation			
Sample no.: 3				Sample no.: 4			
Theoretical value: 0.241				Theoretical value: 0.377			
Unit:				Unit:			
Run 1:				Run 1:			
Number of laboratories: 64				Number of laboratories: 64			
Arithmetic mean value: 0.251				Arithmetic mean value: 0.385			
Median: 0.243				Median: 0.377			
Standard deviation 0.040				Standard deviation 0.052			
Rel. st. deviation (%) 15.792				Rel. st. deviation (%) 13.511			
Run 2:				Run 2:			
Number of laboratories: 60				Number of laboratories: 58			
Arithmetic mean value: 0.248				Arithmetic mean value: 0.379			
Median: 0.241				Median: 0.376			
Standard deviation 0.027				Standard deviation 0.030			
Rel. st. deviation (%) 10.906				Rel. st. deviation (%) 7.856			
Results in decreasing order:				Results in decreasing order:			
43	0.380 (*)	7	0.242	157	0.560 (*)	26	0.376
22	0.371 (*)	110	0.240	24	0.540 (*)	6	0.376
157	0.350 (*)	114	0.240	22	0.517 (*)	160	0.375
24	0.330	153	0.240	43	0.500 (*)	14	0.375
40	0.309	166	0.240	42	0.485	158	0.374
31	0.308	38	0.240	40	0.461	118	0.372
42	0.299	21	0.240	31	0.427	117	0.371
150	0.288	26	0.240	125	0.426	1	0.371
145	0.286	5	0.240	150	0.411	110	0.370
19	0.279	112	0.240	152	0.410	4	0.370
102	0.273	108	0.240	17	0.408	21	0.369
164	0.270	126	0.240	102	0.404	5	0.369
152	0.270	155	0.240	19	0.404	16	0.367
125	0.267	4	0.239	116	0.403	114	0.365
116	0.261	118	0.239	20	0.402	146	0.364
104	0.260	117	0.236	164	0.400	113	0.361
3	0.260	146	0.234	145	0.400	33	0.360
15	0.260	160	0.234	3	0.395	153	0.360
109	0.260	167	0.233	23	0.393	112	0.360
17	0.259	115	0.233	124	0.390	166	0.360
20	0.257	10	0.230	15	0.390	167	0.358
23	0.254	33	0.230	108	0.390	10	0.358
172	0.253	121	0.229	104	0.390	115	0.357
13	0.252	16	0.227	126	0.386	121	0.357
2	0.251	36	0.225	2	0.383	155	0.356
14	0.251	113	0.221	172	0.383	36	0.353
163	0.250	107	0.204	27	0.383	11	0.326
124	0.250	11	0.203	109	0.380	8	0.325
6	0.249	8	0.197	163	0.380	107	0.317
27	0.248	165	0.186	38	0.380	151	0.305
158	0.246	151	0.183	7	0.379	120	0.270 (*)
1	0.245	120	0.130 (*)	13	0.378	165	0.246 (*)

Table 21: Analytical results for pH in precipitations samples.

pH in precipitation				pH in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 4.220				Theoretical value: 4.250			
Unit: pH-units				Unit: pH-units			
Run 1:				Run 1:			
Number of laboratories: 59				Number of laboratories: 59			
Arithmetic mean value: 4.238				Arithmetic mean value: 4.259			
Median: 4.260				Median: 4.280			
Standard deviation 0.135				Standard deviation 0.111			
Rel. st. deviation (%) 3.187				Rel. st. deviation (%) 2.599			
Run 2:				Run 2:			
Number of laboratories: 56				Number of laboratories: 56			
Arithmetic mean value: 4.262				Arithmetic mean value: 4.279			
Median: 4.270				Median: 4.280			
Standard deviation 0.082				Standard deviation 0.070			
Rel. st. deviation (%) 1.920				Rel. st. deviation (%) 1.639			
Results in decreasing order:				Results in decreasing order:			
108	4.440	6	4.260	150	4.470	158	4.280
38	4.400	152	4.260	165	4.410	116	4.270
104	4.370	23	4.260	38	4.400	155	4.270
166	4.370	4	4.250	108	4.390	5	4.270
117	4.360	120	4.250	110	4.380	4	4.270
1	4.360	158	4.250	19	4.364	23	4.270
19	4.353	115	4.250	166	4.360	20	4.260
164	4.340	11	4.250	14	4.330	6	4.260
110	4.330	114	4.240	153	4.320	120	4.260
121	4.320	124	4.240	163	4.320	115	4.260
165	4.320	3	4.240	35	4.310	124	4.260
150	4.320	5	4.240	117	4.310	3	4.256
33	4.310	126	4.230	112	4.310	36	4.250
14	4.310	39	4.230	26	4.310	22	4.250
35	4.300	22	4.220	1	4.300	16	4.250
153	4.300	36	4.220	10	4.300	13	4.250
112	4.300	16	4.220	113	4.300	11	4.240
116	4.300	13	4.220	24	4.300	39	4.240
26	4.290	118	4.210	31	4.300	126	4.240
10	4.290	24	4.210	8	4.290	118	4.240
20	4.280	7	4.170	121	4.290	114	4.240
113	4.280	102	4.152	33	4.290	102	4.166
163	4.280	157	4.150	107	4.290	7	4.160
107	4.270	146	4.110	164	4.290	146	4.130
8	4.270	151	3.990	2	4.290	157	4.100
15	4.270	40	3.970	152	4.290	151	4.038
31	4.270	17	3.930 (*)	104	4.280	17	3.920 (*)
27	4.270	172	3.800 (*)	21	4.280	40	3.890 (*)
21	4.270	2	3.630 (*)	27	4.280	172	3.850 (*)
155	4.260			15	4.280		
pH in precipitation				pH in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value: 4.540				Theoretical value: 4.460			
Unit: pH-units				Unit: pH-units			
Run 1:				Run 1:			
Number of laboratories: 59				Number of laboratories: 59			
Arithmetic mean value: 4.544				Arithmetic mean value: 4.461			
Median: 4.570				Median: 4.490			
Standard deviation 0.135				Standard deviation 0.117			
Rel. st. deviation (%) 2.968				Rel. st. deviation (%) 2.623			
Run 2:				Run 2:			
Number of laboratories: 55				Number of laboratories: 56			
Arithmetic mean value: 4.574				Arithmetic mean value: 4.481			
Median: 4.570				Median: 4.490			
Standard deviation 0.074				Standard deviation 0.080			
Rel. st. deviation (%) 1.611				Rel. st. deviation (%) 1.780			
Results in decreasing order:				Results in decreasing order:			
150	4.760	23	4.570	19	4.622	104	4.480
19	4.716	158	4.560	14	4.610	115	4.480
38	4.710	116	4.560	153	4.600	23	4.480
110	4.690	115	4.560	150	4.580	5	4.480
14	4.670	152	4.560	166	4.580	4	4.480
35	4.670	5	4.560	35	4.580	158	4.480
166	4.670	6	4.550	38	4.570	20	4.470
108	4.650	155	4.550	26	4.560	120	4.470
153	4.630	165	4.550	110	4.560	124	4.470
113	4.630	16	4.550	1	4.550	155	4.470
117	4.630	20	4.550	112	4.540	6	4.470
112	4.620	120	4.550	121	4.540	39	4.460
26	4.610	3	4.545	108	4.530	11	4.460
104	4.600	39	4.540	163	4.520	114	4.450
24	4.600	11	4.540	113	4.520	13	4.450
10	4.600	13	4.540	165	4.510	118	4.450
22	4.590	118	4.530	107	4.510	126	4.440
1	4.590	124	4.530	31	4.510	16	4.430
15	4.590	114	4.520	15	4.510	116	4.420
31	4.590	126	4.510	10	4.510	102	4.413
33	4.580	102	4.498	8	4.500	7	4.350
107	4.580	157	4.430	164	4.500	146	4.330
27	4.580	7	4.420	21	4.500	2	4.320
121	4.580	146	4.390	33	4.490	157	4.290
21	4.580	151	4.343	27	4.490	117	4.260
8	4.580	17	4.200 (*)	152	4.490	151	4.240
163	4.580	40	4.180 (*)	22	4.490	40	4.120 (*)
164	4.570	172	4.090 (*)	24	4.490	17	4.080 (*)
36	4.570	2	4.040 (*)	36	4.490	172	4.060 (*)
4	4.570			3	4.490		

Table 22: Analytical results for strong acid calculated from pH.

Strong acid calculated from pH				Strong acid calculated from pH			
Sample no.: G1				Sample no.: G2			
Theoretical value:		60.000		Theoretical value:		57.000	
Unit: µeg/l				Unit: µeg/l			
Run 1:				Run 1:			
Number of laboratories:		59		Number of laboratories:		59	
Arithmetic mean value:		61.517		Arithmetic mean value:		57.231	
Median:		54.954		Median:		52.481	
Standard deviation		30.078		Standard deviation		19.353	
Rel. st. deviation (%)		48.894		Rel. st. deviation (%)		33.815	
Run 2:				Run 2:			
Number of laboratories:		57		Number of laboratories:		56	
Arithmetic mean value:		56.782		Arithmetic mean value:		53.328	
Median:		53.703		Median:		52.481	
Standard deviation		14.573		Standard deviation		9.260	
Rel. st. deviation (%)		25.665		Rel. st. deviation (%)		17.364	
Results in decreasing order:				Results in decreasing order:			
2	234.423 (*)	8	53.703	167	141.254 (*)	102	52.481
167	158.489 (*)	15	53.703	39	128.825 (*)	157	52.481
17	117.490	21	53.703	17	120.226 (*)	21	52.481
39	107.152	31	53.703	150	91.622	8	51.286
150	102.329	27	53.703	155	79.433	151	51.286
145	77.625	104	53.703	145	74.131	104	51.286
155	70.795	158	52.481	7	69.183	120	51.286
43	70.469	112	52.481	43	68.234	163	51.286
7	67.608	20	52.481	11	57.544	33	51.286
117	61.660	10	51.286	117	57.544	2	51.286
24	61.660	26	51.286	113	57.544	31	50.119
36	60.256	110	50.119	125	57.544	1	50.119
13	60.256	152	50.119	172	57.544	24	50.119
16	60.256	115	50.119	13	56.234	112	50.119
22	60.256	35	50.119	22	56.234	10	50.119
125	58.884	33	48.978	36	56.234	26	48.978
172	58.884	14	48.978	16	56.234	35	48.978
121	57.544	146	47.863	3	55.463	116	48.978
5	57.544	120	47.863	114	54.954	110	48.978
3	57.544	164	47.863	6	54.954	158	47.863
113	57.544	109	46.774	118	54.954	152	47.863
11	56.234	163	45.709	20	54.954	14	46.774
118	56.234	19	44.361	121	54.954	165	43.652
114	56.234	1	43.652	5	53.703	19	43.251
157	56.234	116	43.652	115	53.703	109	41.687
4	56.234	102	42.658	23	53.703	107	40.738
151	54.954	165	42.658	153	53.703	38	39.811
153	54.954	38	39.811	4	53.703	164	38.905
6	54.954	107	36.308	27	52.481	146	33.884
23	54.954			15	52.481		
Strong acid calculated from pH				Strong acid calculated from pH			
Sample no.: G3				Sample no.: G4			
Theoretical value:		29.000		Theoretical value:		35.000	
Unit: µeg/l				Unit: µeg/l			
Run 1:				Run 1:			
Number of laboratories:		59		Number of laboratories:		59	
Arithmetic mean value:		30.315		Arithmetic mean value:		36.091	
Median:		26.915		Median:		32.359	
Standard deviation		13.472		Standard deviation		12.704	
Rel. st. deviation (%)		44.441		Rel. st. deviation (%)		35.199	
Run 2:				Run 2:			
Number of laboratories:		55		Number of laboratories:		56	
Arithmetic mean value:		27.035		Arithmetic mean value:		33.629	
Median:		26.915		Median:		32.359	
Standard deviation		4.863		Standard deviation		6.903	
Rel. st. deviation (%)		17.988		Rel. st. deviation (%)		20.527	
Results in decreasing order:				Results in decreasing order:			
2	91.201 (*)	23	26.915	167	87.096 (*)	36	32.359
167	81.283 (*)	4	26.915	17	83.176 (*)	151	32.359
39	66.069 (*)	120	26.303	39	75.858 (*)	3	32.359
17	63.096 (*)	21	26.303	150	57.544	33	32.359
150	45.394	158	26.303	116	54.954	22	32.359
145	40.738	27	26.303	155	51.286	24	32.359
7	38.019	8	26.303	2	47.863	8	31.623
155	37.154	33	26.303	145	46.774	21	31.623
43	31.769	104	26.303	7	44.668	163	31.623
125	30.903	15	25.704	43	38.637	104	30.903
113	30.200	22	25.704	115	38.019	15	30.903
121	29.512	31	25.704	16	37.154	164	30.903
117	29.512	1	25.704	125	36.308	31	30.903
13	28.840	24	25.119	117	35.481	10	30.903
172	28.840	10	25.119	113	35.481	158	30.200
11	28.840	102	25.119	13	35.481	112	30.200
3	28.510	26	24.547	172	34.674	107	29.512
164	28.184	110	23.988	11	34.674	110	28.840
6	28.184	152	23.442	118	33.884	120	28.840
20	28.184	112	23.442	153	33.884	1	28.184
16	28.184	116	23.442	6	33.884	109	27.542
118	28.184	107	22.387	121	33.884	26	27.542
153	28.184	35	21.380	20	33.884	38	26.915
114	27.542	14	21.380	23	33.113	146	26.303
5	27.542	165	21.380	4	33.113	165	26.303
115	27.542	109	20.417	114	33.113	35	26.303
151	27.542	38	19.498	102	33.113	152	25.119
157	27.542	19	19.231	157	33.113	14	24.547
36	26.915	146	17.378	5	33.113	19	23.878
163	26.915			27	32.359		

Table 23: Analytical results for chloride in precipitations samples.

Chloride in precipitation				Chloride in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value:		0.135		Theoretical value:		0.162	
Unit: µg/ml				Unit: µg/ml			
Run 1:				Run 1:			
Number of laboratories:		58		Number of laboratories:		58	
Arithmetic mean value:		0.151		Arithmetic mean value:		0.172	
Median:		0.136		Median:		0.161	
Standard deviation		0.065		Standard deviation		0.061	
Rel. st. deviation (%)		43.399		Rel. st. deviation (%)		35.337	
Run 2:				Run 2:			
Number of laboratories:		57		Number of laboratories:		55	
Arithmetic mean value:		0.144		Arithmetic mean value:		0.160	
Median:		0.136		Median:		0.160	
Standard deviation		0.040		Standard deviation		0.026	
Rel. st. deviation (%)		27.777		Rel. st. deviation (%)		16.167	
Results in decreasing order:				Results in decreasing order:			
157	0.540 (*)	1	0.136	157	0.510 (*)	114	0.160
19	< 0.31			17	0.360 (*)	43	0.160
110	< 0.3			19	< 0.31		
115	< 0.275	2	0.134	115	0.300 (*)	20	0.160
40	0.250	21	0.132	110	< 0.3		
109	< 0.25			109	< 0.25		
17	0.237	114	0.130	165	0.222	146	0.160
20	0.234	15	0.130	116	0.207	15	0.160
165	0.204	38	0.130	152	0.200	21	0.155
24	0.200	153	0.130	120	0.200	151	0.155
152	0.200			40	0.200	35	0.154
102	< 0.19	36	0.130	102	< 0.19		
27	0.186	4	0.130	24	0.190	4	0.154
107	0.173	33	0.130	104	0.190	36	0.154
10	0.164	31	0.128	10	0.188	172	0.152
158	0.163	23	0.127	150	0.188	1	0.151
116	0.163	35	0.126	2	0.185	33	0.150
104	0.160	151	0.125	11	0.181	163	0.150
120	0.160	13	0.124	158	0.179	23	0.150
5	0.155	172	0.121	16	0.177	31	0.148
125	0.150	166	0.121	108	0.175	14	0.143
3	0.150	164	0.120	13	0.172	166	0.143
150	0.149	155	0.120	5	0.170	38	0.140
11	0.147	126	0.120	113	0.170	164	0.140
42	0.146	163	0.120	107	0.167	155	0.140
16	0.141	124	0.120	26	0.167	126	0.140
43	0.140	145	0.120	8	0.165	153	0.140
113	0.140	167	0.103	125	0.164	145	0.129
8	0.139	6	0.101	7	0.163	6	0.124
7	0.139	14	0.101	112	0.163	167	0.122
112	0.138	118	0.080	42	0.162	118	0.110
26	0.137	108	0.075	3	0.161	124	0.110
146	0.136	22	0.057	27	0.161	22	0.074
Chloride in precipitation				Chloride in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value:		0.270		Theoretical value:		0.328	
Unit: µg/ml				Unit: µg/ml			
Run 1:				Run 1:			
Number of laboratories:		61		Number of laboratories:		63	
Arithmetic mean value:		0.268		Arithmetic mean value:		0.320	
Median:		0.264		Median:		0.312	
Standard deviation		0.059		Standard deviation		0.070	
Rel. st. deviation (%)		22.197		Rel. st. deviation (%)		21.839	
Run 2:				Run 2:			
Number of laboratories:		58		Number of laboratories:		59	
Arithmetic mean value:		0.261		Arithmetic mean value:		0.304	
Median:		0.263		Median:		0.310	
Standard deviation		0.033		Standard deviation		0.034	
Rel. st. deviation (%)		12.815		Rel. st. deviation (%)		11.270	
Results in decreasing order:				Results in decreasing order:			
157	0.580 (*)	21	0.262	157	0.600 (*)	11	0.311
17	0.450 (*)	20	0.261	108	0.585 (*)	33	0.310
115	0.380	33	0.260	116	0.522 (*)	114	0.310
24	0.330	125	0.260	17	0.493 (*)	15	0.310
165	0.323	114	0.260	115	0.400	23	0.307
108	0.315	23	0.260	165	0.371	21	0.307
19	< 0.31			40	0.350	1	0.304
40	0.300	120	0.260	36	0.349	20	0.303
110	< 0.3			2	0.349	10	0.301
2	0.299	166	0.259	102	0.343	110	0.300
42	0.296	4	0.258	24	0.340	4	0.300
102	0.294	13	0.252	104	0.340	152	0.300
1	0.291	163	0.250	43	0.328	163	0.300
36	0.288	31	0.250	172	0.322	164	0.300
104	0.280	164	0.250	27	0.322	166	0.293
109	< 0.25			8	0.322	31	0.293
112	0.277	172	0.245	3	0.321	13	0.291
7	0.277	155	0.242	150	0.321	42	0.287
27	0.274	153	0.240	26	0.321	167	0.281
158	0.272	10	0.239	120	0.320	155	0.281
8	0.271	167	0.233	113	0.320	6	0.280
26	0.270	113	0.230	7	0.320	145	0.277
3	0.270	126	0.230	19	0.320	153	0.270
107	0.270	38	0.230	5	0.318	126	0.260
15	0.270	6	0.229	112	0.318	38	0.250
5	0.270	145	0.226	107	0.316	14	0.250
16	0.269	116	0.222	109	< 0.25		
150	0.268	14	0.211	151	0.316	121	0.248
146	0.266	22	0.201	35	0.315	118	0.240
121	0.266	152	0.200	146	0.314	22	0.232
43	0.265	118	0.190	158	0.312	124	0.230
151	0.265	124	0.190	16	0.312	117	0.207
35	0.264	117	0.147 (*)	125	0.312		
11	0.264						

Table 24: Analytical results for sodium in precipitations samples.

Sodium in precipitation				Sodium in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 0.213				Theoretical value: 0.248			
Unit: µg/ml				Unit: µg/ml			
Run 1:				Run 1:			
Number of laboratories: 64				Number of laboratories: 64			
Arithmetic mean value: 0.213				Arithmetic mean value: 0.248			
Median: 0.208				Median: 0.245			
Standard deviation 0.030				Standard deviation 0.028			
Rel. st. deviation (%) 13.835				Rel. st. deviation (%) 11.213			
Run 2:				Run 2:			
Number of laboratories: 61				Number of laboratories: 59			
Arithmetic mean value: 0.209				Arithmetic mean value: 0.246			
Median: 0.207				Median: 0.245			
Standard deviation 0.021				Standard deviation 0.017			
Rel. st. deviation (%) 9.836				Rel. st. deviation (%) 6.981			
Results in decreasing order:				Results in decreasing order:			
157	0.340 (*)	150	0.208	42	0.344 (*)	35	0.245
42	0.306 (*)	6	0.207	157	0.340 (*)	1	0.245
19	0.275 (*)	13	0.207	117	0.311 (*)	150	0.244
117	0.267	33	0.206	121	0.299	11	0.242
22	0.256	120	0.205	19	0.290	152	0.241
102	0.254	4	0.203	24	0.270	116	0.241
121	0.245	21	0.203	107	0.268	124	0.240
27	0.245	152	0.202	145	0.265	114	0.240
107	0.244	165	0.202	3	0.264	109	0.240
3	0.229	10	0.201	6	0.262	23	0.239
145	0.228	166	0.200	153	0.260	115	0.238
116	0.227	153	0.200	110	0.260	120	0.238
5	0.226	124	0.200	163	0.260	33	0.238
151	0.223	109	0.200	22	0.260	165	0.238
163	0.220	114	0.200	15	0.260	8	0.237
112	0.220	110	0.200	160	0.260	4	0.237
16	0.220	43	0.200	151	0.258	21	0.236
15	0.220	36	0.200	13	0.257	108	0.235
24	0.220	35	0.199	20	0.256	36	0.235
164	0.220	11	0.197	5	0.254	10	0.231
155	0.218	8	0.193	26	0.253	118	0.230
1	0.217	2	0.192	155	0.252	166	0.230
7	0.216	118	0.190	125	0.252	102	0.229
26	0.215	115	0.190	14	0.251	2	0.228
158	0.214	23	0.189	27	0.251	167	0.227
146	0.212	167	0.187	164	0.250	104	0.220
125	0.212	172	0.187	112	0.250	113	0.218
108	0.210	113	0.187	16	0.250	172	0.216
20	0.210	104	0.180	7	0.249	126	0.210
31	0.210	126	0.170	146	0.249	38	0.200
160	0.210	17	0.161	31	0.247	43	0.180 (*)
14	0.209	38	0.160	158	0.246	17	0.178 (*)
Sodium in precipitation				Sodium in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value: 0.354				Theoretical value: 0.463			
Unit: µg/ml				Unit: µg/ml			
Run 1:				Run 1:			
Number of laboratories: 64				Number of laboratories: 64			
Arithmetic mean value: 0.343				Arithmetic mean value: 0.439			
Median: 0.350				Median: 0.443			
Standard deviation 0.050				Standard deviation 0.058			
Rel. st. deviation (%) 14.681				Rel. st. deviation (%) 13.104			
Run 2:				Run 2:			
Number of laboratories: 61				Number of laboratories: 61			
Arithmetic mean value: 0.348				Arithmetic mean value: 0.443			
Median: 0.350				Median: 0.444			
Standard deviation 0.028				Standard deviation 0.035			
Rel. st. deviation (%) 8.170				Rel. st. deviation (%) 7.788			
Results in decreasing order:				Results in decreasing order:			
42	0.481 (*)	31	0.350	42	0.619 (*)	31	0.441
117	0.422	11	0.349	24	0.550	26	0.441
121	0.406	150	0.348	157	0.550	16	0.441
24	0.400	23	0.348	121	0.513	109	0.440
19	0.386	14	0.346	117	0.512	150	0.439
20	0.385	8	0.342	19	0.496	22	0.436
22	0.377	116	0.341	107	0.481	33	0.436
35	0.376	36	0.340	20	0.476	14	0.432
33	0.376	166	0.340	124	0.470	116	0.431
107	0.375	16	0.340	145	0.468	114	0.430
13	0.375	2	0.339	6	0.461	2	0.430
124	0.370	145	0.338	15	0.460	126	0.430
155	0.370	21	0.338	164	0.460	115	0.429
153	0.370	4	0.334	153	0.460	36	0.428
125	0.367	110	0.330	160	0.460	4	0.427
6	0.364	108	0.330	35	0.459	8	0.424
3	0.364	109	0.330	155	0.458	21	0.424
5	0.361	167	0.326	13	0.457	110	0.420
164	0.360	104	0.320	125	0.452	166	0.420
27	0.360	102	0.320	27	0.452	108	0.420
15	0.360	120	0.318	151	0.451	167	0.419
114	0.360	10	0.316	5	0.451	104	0.410
163	0.360	113	0.312	3	0.451	120	0.405
112	0.360	115	0.312	112	0.450	10	0.404
1	0.357	118	0.310	163	0.450	113	0.402
158	0.355	38	0.310	152	0.449	118	0.400
146	0.355	172	0.308	158	0.447	38	0.400
151	0.355	165	0.300	23	0.446	172	0.376
152	0.355	126	0.290	1	0.446	102	0.375
7	0.353	17	0.258	146	0.446	165	0.362
26	0.351	43	0.140 (*)	11	0.445	17	0.319 (*)
160	0.350	157	0.120 (*)	7	0.444	43	0.140 (*)

Table 25: Analytical results for magnesium in precipitations samples.

Magnesium in precipitation				Magnesium in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 0.155				Theoretical value: 0.067			
Unit: µg/ml				Unit: µg/ml			
Run 1:				Run 1:			
Number of laboratories: 63				Number of laboratories: 62			
Arithmetic mean value: 0.161				Arithmetic mean value: 0.069			
Median: 0.154				Median: 0.067			
Standard deviation 0.040				Standard deviation 0.024			
Rel. st. deviation (%) 25.014				Rel. st. deviation (%) 33.971			
Run 2:				Run 2:			
Number of laboratories: 61				Number of laboratories: 60			
Arithmetic mean value: 0.155				Arithmetic mean value: 0.066			
Median: 0.153				Median: 0.067			
Standard deviation 0.020				Standard deviation 0.012			
Rel. st. deviation (%) 12.945				Rel. st. deviation (%) 17.842			
Results in decreasing order:				Results in decreasing order:			
6	0.418 (*)	166	0.153	43	0.199 (*)	117	0.067
43	0.248 (*)	151	0.152	153	0.160 (*)	7	0.067
42	0.215	117	0.151	145 <	0.1		
163	0.200	14	0.151	42	0.093	151	0.066
107	0.195	116	0.150	107	0.092	14	0.065
114	0.190	1	0.150	125	0.091	158	0.065
20	0.183	150	0.150	163	0.088	8	0.065
22	0.182	13	0.150	20	0.083	150	0.064
157	0.180	164	0.150	22	0.080	118	0.064
33	0.176	167	0.150	109	0.080	21	0.064
102	0.175	104	0.150	157	0.080	120	0.062
125	0.173	160	0.150	2	0.077	1	0.062
24	0.170	31	0.147	6	0.076	167	0.061
109	0.170	17	0.147	10	0.075	36	0.061
153	0.170	165	0.146	23	0.073	13	0.061
10	0.168	8	0.146	33	0.072	166	0.060
2	0.164	118	0.146	16	0.071	104	0.060
145	0.161	108	0.145	102	0.071	152	0.060
7	0.160	152	0.144	5	0.070	164	0.060
27	0.160	120	0.144	112	0.070	108	0.055
112	0.160	36	0.143	114	0.070	4	0.054
15	0.160	110	0.140	124	0.070	17	0.054
124	0.160	4	0.135	24	0.070	146	0.054
155	0.159	113	0.134	15	0.070	115	0.053
21	0.158	146	0.133	31	0.069	121	0.052
11	0.158	38	0.130	26	0.068	110	0.051
19	0.158	16	0.126	11	0.068	116	0.050
5	0.157	115	0.122	165	0.068	38	0.050
3	0.156	172	0.122	3	0.068	160	0.050
158	0.155	126	0.110	27	0.068	172	0.049
23	0.154	121	0.101	155	0.067	126	0.040
26	0.154			19	0.067	113	0.034
Magnesium in precipitation				Magnesium in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value: 0.083				Theoretical value: 0.103			
Unit: µg/ml				Unit: µg/ml			
Run 1:				Run 1:			
Number of laboratories: 63				Number of laboratories: 62			
Arithmetic mean value: 0.085				Arithmetic mean value: 0.117			
Median: 0.081				Median: 0.102			
Standard deviation 0.031				Standard deviation 0.112			
Rel. st. deviation (%) 36.391				Rel. st. deviation (%) 95.460			
Run 2:				Run 2:			
Number of laboratories: 62				Number of laboratories: 61			
Arithmetic mean value: 0.081				Arithmetic mean value: 0.103			
Median: 0.081				Median: 0.102			
Standard deviation 0.015				Standard deviation 0.019			
Rel. st. deviation (%) 19.068				Rel. st. deviation (%) 18.529			
Results in decreasing order:				Results in decreasing order:			
43	0.293 (*)	164	0.080	8	0.970 (*)	26	0.102
42	0.121	104	0.080	42	0.167	117	0.101
107	0.115	151	0.080	43	0.147	14	0.101
125	0.108	153	0.080	109	0.140	151	0.101
145	0.106	15	0.080	107	0.135	104	0.100
20	0.103	112	0.080	157	0.130	153	0.100
102	0.102	158	0.079	125	0.127	112	0.100
33	0.101	1	0.078	20	0.126	160	0.100
109	0.100	8	0.078	102	0.121	164	0.100
163	0.100	150	0.078	163	0.120	15	0.100
24	0.100	13	0.077	145 <	0.1		
157	0.100	31	0.077	33	0.120	1	0.099
22	0.099	36	0.077	24	0.120	150	0.098
6	0.094	152	0.076	22	0.118	13	0.097
2	0.093	118	0.076	167	0.116	118	0.096
124	0.090	120	0.075	2	0.114	152	0.096
114	0.090	166	0.074	6	0.113	166	0.095
10	0.087	165	0.073	124	0.110	120	0.095
7	0.085	108	0.070	114	0.110	36	0.094
155	0.085	116	0.067	10	0.109	108	0.090
5	0.084	110	0.066	7	0.107	116	0.089
27	0.083	4	0.066	27	0.106	110	0.088
19	0.083	17	0.063	155	0.105	4	0.085
16	0.083	172	0.061	16	0.105	17	0.081
11	0.082	146	0.061	31	0.105	146	0.080
3	0.082	160	0.060	23	0.105	38	0.080
23	0.082	38	0.060	3	0.104	165	0.078
26	0.082	115	0.055	158	0.104	172	0.076
21	0.081	126	0.050	5	0.104	115	0.072
167	0.081	121	0.048	11	0.103	126	0.070
14	0.081	113	0.048	19	0.103	113	0.067
117	0.081			21	0.103	121	0.058

Table 26: Analytical results for calcium in precipitations samples.

Calcium in precipitation		Calcium in precipitation	
Sample no.: G1		Sample no.: G2	
Theoretical value: -999.000		Theoretical value: 0.153	
Unit:		Unit: µg/ml	
Run 1:		Run 1:	
		Number of laboratories: 62	
		Arithmetic mean value: 0.180	
		Median: 0.152	
		Standard deviation 0.127	
		Rel. st. deviation (%) 70.358	
Run 2:		Run 2:	
		Number of laboratories: 60	
		Arithmetic mean value: 0.160	
		Median: 0.152	
		Standard deviation 0.046	
		Rel. st. deviation (%) 28.496	
Results in decreasing order:		Results in decreasing order:	
		17 1.023 (*) 6 0.152	
		43 0.526 (*) 14 0.152	
		165 0.314 120 0.152	
		163 0.270 23 0.152	
		107 0.267 5 0.152	
		108 0.255 112 0.150	
		42 0.248 15 0.150	
		113 0.238 164 0.150	
		125 0.212 11 0.150	
		157 0.200 124 0.150	
		109 0.200 160 0.150	
		10 0.185 155 0.148	
		8 0.178 36 0.147	
		2 0.174 4 0.147	
		26 0.171 31 0.142	
		16 0.166 38 0.140	
		20 0.166 110 0.140	
		19 0.166 126 0.140	
		33 0.163 118 0.140	
		24 0.160 167 0.136	
		152 0.160 172 0.135	
		153 0.160 13 0.131	
		102 < 0.16	
		146 0.159 150 0.131	
		117 0.159 166 0.130	
		35 0.157 1 0.130	
		158 0.156 121 0.129	
		27 0.156 104 0.110	
		3 0.155 115 0.109	
		7 0.154 114 0.100	
		151 0.154 22 0.094	
		21 0.153 116 0.008	
Calcium in precipitation		Calcium in precipitation	
Sample no.: G3		Sample no.: G4	
Theoretical value: 0.179		Theoretical value: 0.243	
Unit: µg/ml		Unit: µg/ml	
Run 1:		Run 1:	
Number of laboratories: 63		Number of laboratories: 64	
Arithmetic mean value: 0.192		Arithmetic mean value: 0.255	
Median: 0.181		Median: 0.241	
Standard deviation 0.061		Standard deviation 0.068	
Rel. st. deviation (%) 31.610		Rel. st. deviation (%) 26.838	
Run 2:		Run 2:	
Number of laboratories: 59		Number of laboratories: 60	
Arithmetic mean value: 0.185		Arithmetic mean value: 0.246	
Median: 0.180		Median: 0.241	
Standard deviation 0.036		Standard deviation 0.041	
Rel. st. deviation (%) 19.663		Rel. st. deviation (%) 16.497	
Results in decreasing order:		Results in decreasing order:	
43 0.481 (*) 15 0.180		109 0.600 (*) 3 0.241	
107 0.350 (*) 112 0.180		107 0.438 (*) 117 0.241	
165 0.326 (*) 16 0.180		43 0.401 (*) 151 0.240	
42 0.303 151 0.177		42 0.389 164 0.240	
125 0.297 3 0.177		157 0.350 155 0.240	
108 0.270 11 0.176		165 0.338 160 0.240	
109 0.250 36 0.173		113 0.337 11 0.239	
163 0.240 4 0.173		163 0.320 23 0.237	
157 0.240 172 0.171		108 0.310 31 0.237	
20 0.227 164 0.170		24 0.280 35 0.236	
145 0.223 38 0.170		125 0.280 6 0.233	
153 0.220 120 0.166		145 0.275 4 0.233	
113 0.219 23 0.163		33 0.267 36 0.231	
33 0.213 13 0.162		10 0.263 38 0.230	
10 0.203 31 0.161		126 0.260 110 0.230	
24 0.200 110 0.160		20 0.257 166 0.230	
8 0.194 126 0.160		2 0.256 13 0.228	
146 0.193 118 0.160		16 0.254 120 0.227	
19 0.192 124 0.160		19 0.254 1 0.226	
152 0.190 160 0.160		7 0.253 124 0.220	
		102 < 0.160	
		2 0.188 1 0.159	
		7 0.188 150 0.154	
		117 0.188 167 0.154	
		155 0.186 104 0.150	
		26 0.185 166 0.150	
		21 0.185 17 0.142	
		158 0.182 114 0.140	
		5 0.182 22 0.134	
		6 0.182 121 0.125	
		27 0.182 115 0.123	
		35 0.181 116 0.043 (*)	
		14 0.181	
		26 0.252 150 0.213	
		8 0.252 104 0.210	
		152 0.250 118 0.210	
		112 0.250 102 0.208	
		153 0.250 121 0.206	
		15 0.250 172 0.205	
		5 0.248 167 0.201	
		158 0.247 114 0.190	
		146 0.246 22 0.188	
		14 0.245 17 0.172	
		27 0.243 115 0.152	
		21 0.242 116 0.118 (*)	

Table 27: Analytical results for potassium in precipitations samples

Potassium in precipitation				Potassium in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 0.177				Theoretical value: 0.204			
Unit: µg/ml				Unit: µg/ml			
Run 1:				Run 1:			
Number of laboratories: 62				Number of laboratories: 63			
Arithmetic mean value: 0.172				Arithmetic mean value: 0.196			
Median: 0.176				Median: 0.201			
Standard deviation 0.022				Standard deviation 0.030			
Rel. st. deviation (%) 12.943				Rel. st. deviation (%) 15.252			
Run 2:				Run 2:			
Number of laboratories: 60				Number of laboratories: 59			
Arithmetic mean value: 0.172				Arithmetic mean value: 0.196			
Median: 0.176				Median: 0.201			
Standard deviation 0.019				Standard deviation 0.023			
Rel. st. deviation (%) 11.262				Rel. st. deviation (%) 11.918			
Results in decreasing order:				Results in decreasing order:			
42	0.249 (*)	172	0.176	17	0.280 (*)	22	0.200
112	0.210	26	0.176	42	0.277 (*)	124	0.200
19	0.206	31	0.175	112	0.240	163	0.200
11	0.201	7	0.173	24	0.240	158	0.197
160	0.200	158	0.172	10	0.237	33	0.194
2	0.200	14	0.172	160	0.230	120	0.190
10	0.199	13	0.172	2	0.228	114	0.190
151	0.192	121	0.171	19	0.228	164	0.190
107	0.191	3	0.171	117	0.227	153	0.190
164	0.190	124	0.170	151	0.224	36	0.186
		102	< 0.17				
24	0.190	108	0.165	157	0.220	108	0.185
22	0.188	23	0.163	5	0.214	150	0.185
5	0.187	36	0.163	155	0.210	8	0.181
17	0.186	16	0.160	107	0.210	104	0.180
116	0.184	114	0.160	145	0.210	166	0.180
117	0.183	150	0.159	109	0.210	118	0.180
146	0.182	8	0.156	4	0.210	172	0.178
109	0.180	35	0.155	110	0.210	113	0.175
157	0.180	6	0.154	15	0.210	165	0.173
110	0.180	118	0.150	116	0.209	35	0.173
120	0.180	43	0.150	13	0.208	23	0.170
						102	< 0.17
152	0.180	167	0.150	31	0.207	6	0.167
15	0.180	153	0.150	21	0.206	167	0.161
4	0.180	115	0.141	14	0.204	126	0.160
163	0.180	104	0.140	152	0.203	115	0.157
27	0.178	126	0.140	27	0.203	146	0.155
1	0.178	113	0.133	1	0.202	20	0.152
21	0.178	165	0.132	125	0.202	38	0.150
33	0.178	166	0.130	16	0.202	11	0.139
125	0.177	38	0.130	26	0.201	121	0.129 (*)
155	0.177	20	0.124 (*)	3	0.201	43	0.120 (*)
				7	0.201		
Potassium in precipitation				Potassium in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value: 0.136				Theoretical value: 0.102			
Unit: µg/ml				Unit: µg/ml			
Run 1:				Run 1:			
Number of laboratories: 62				Number of laboratories: 57			
Arithmetic mean value: 0.133				Arithmetic mean value: 0.105			
Median: 0.134				Median: 0.104			
Standard deviation 0.022				Standard deviation 0.023			
Rel. st. deviation (%) 16.327				Rel. st. deviation (%) 21.722			
Run 2:				Run 2:			
Number of laboratories: 59				Number of laboratories: 54			
Arithmetic mean value: 0.132				Arithmetic mean value: 0.106			
Median: 0.134				Median: 0.104			
Standard deviation 0.018				Standard deviation 0.018			
Rel. st. deviation (%) 13.908				Rel. st. deviation (%) 17.089			
Results in decreasing order:				Results in decreasing order:			
42	0.202 (*)	23	0.134	42	0.172 (*)	27	0.104
				102	< 0.17		
112	0.180 (*)	1	0.134	121	0.149	107	0.103
19	0.173	27	0.134	112	0.140	14	0.103
33	0.171	17	0.133	109	0.140	155	0.103
109	0.170	152	0.133	24	0.140	26	0.102
102	< 0.17						
2	0.162	145	0.132	151	0.139	5	0.101
22	0.155	14	0.131	165	0.136	167	0.101
151	0.154	16	0.130	33	0.130	163	0.100
124	0.150	114	0.130	2	0.129	15	0.100
24	0.150	120	0.130	19	0.128	43	0.100
						104	< 0.1
						113	< 0.1
						121	< 0.1
						145	< 0.1
160	0.150	164	0.130	31	0.126	21	0.099
117	0.149	158	0.129	117	0.122	152	0.096
31	0.148	35	0.127	160	0.120	158	0.094
165	0.147	8	0.125	166	0.120	35	0.091
13	0.146	167	0.122	3	0.114	8	0.091
10	0.145	113	0.121	146	0.112	118	0.090
146	0.142	153	0.120	125	0.112	157	0.090
116	0.142	157	0.120	13	0.111	164	0.090
15	0.140	118	0.120	4	0.110	114	0.090
163	0.140	110	0.120	150	0.110	108	0.090
4	0.140	36	0.116	120	0.110	36	0.087
155	0.139	108	0.110	22	0.110	126	0.080
				110	< 0.11		
26	0.137	6	0.102	115	0.109	172	0.079
125	0.137	38	0.100	10	0.108	6	0.078
						17	< 0.077
107	0.135	104	0.100	116	0.105	153	0.070
5	0.135	43	0.100	23	0.105	38	0.070
150	0.135	126	0.100	16	0.105	20	0.055 (*)
172	0.134	166	0.100	7	0.104	11	0.036 (*)
7	0.134	20	0.097	1	0.104		
3	0.134	115	0.093				
21	0.134	11	0.085 (*)				

Table 28: Analytical results for conductivity in precipitations samples..

Conductivity in precipitation				Conductivity in precipitation			
Sample no.: G1				Sample no.: G2			
Theoretical value: 31.000				Theoretical value: 30.000			
Unit: µS/cm				Unit: µS/cm			
Run 1:				Run 1:			
Number of laboratories: 56				Number of laboratories: 56			
Arithmetic mean value: 30.295				Arithmetic mean value: 29.180			
Median: 30.150				Median: 29.400			
Standard deviation 2.806				Standard deviation 2.502			
Rel. st. deviation (%) 9.263				Rel. st. deviation (%) 8.575			
Run 2:				Run 2:			
Number of laboratories: 55				Number of laboratories: 52			
Arithmetic mean value: 30.045				Arithmetic mean value: 29.218			
Median: 30.100				Median: 29.400			
Standard deviation 2.116				Standard deviation 1.549			
Rel. st. deviation (%) 7.044				Rel. st. deviation (%) 5.303			
Results in decreasing order:				Results in decreasing order:			
157	44.000 (*)	2	30.100	13	36.950 (*)	27	29.400
13	35.800	166	30.100	115	34.750 (*)	150	29.400
115	35.650	117	30.000	157	32.000	151	29.350
14	32.700	153	29.910	126	31.900	153	29.350
5	32.600	107	29.900	5	31.400	8	29.200
126	32.400	164	29.800	31	31.230	107	29.000
104	32.000	114	29.800	14	31.000	16	28.900
11	32.000	112	29.500	19	30.900	6	28.900
19	31.800	6	29.500	11	30.700	121	28.900
31	31.700	7	29.400	172	30.400	110	28.900
3	31.500	110	29.400	102	30.400	112	28.800
20	31.400	16	29.300	20	30.360	164	28.800
102	31.400	33	29.200	3	30.300	163	28.700
10	31.400	118	29.000	124	30.300	33	28.600
124	31.400	17	29.000	114	30.300	1	28.400
21	31.380	165	28.900	21	30.280	118	28.000
155	31.150	38	28.700	10	30.200	158	28.000
15	31.100	121	28.600	116	30.140	39	27.700
116	31.100	158	28.600	15	30.100	108	27.500
172	31.000	36	27.990	24	30.000	36	27.200
4	30.900	108	27.500	7	30.000	113	27.200
24	30.800	40	27.330	104	30.000	165	26.400
150	30.800	1	26.900	2	29.800	22	26.200
146	30.800	22	26.800	4	29.800	117	26.000
27	30.600	152	26.500	155	29.750	23	25.600
8	30.400	113	26.100	166	29.600	152	25.100
163	30.300	39	25.490	146	29.600	40	24.000 (*)
151	30.200	23	24.900	38	29.400	17	19.000 (*)
Conductivity in precipitation				Conductivity in precipitation			
Sample no.: G3				Sample no.: G4			
Theoretical value: 19.000				Theoretical value: 24.000			
Unit: µS/cm				Unit: µS/cm			
Run 1:				Run 1:			
Number of laboratories: 56				Number of laboratories: 56			
Arithmetic mean value: 19.529				Arithmetic mean value: 24.245			
Median: 19.400				Median: 24.200			
Standard deviation 2.199				Standard deviation 2.056			
Rel. st. deviation (%) 11.259				Rel. st. deviation (%) 8.481			
Run 2:				Run 2:			
Number of laboratories: 53				Number of laboratories: 53			
Arithmetic mean value: 19.284				Arithmetic mean value: 24.144			
Median: 19.400				Median: 24.200			
Standard deviation 1.302				Standard deviation 1.232			
Rel. st. deviation (%) 6.753				Rel. st. deviation (%) 5.102			
Results in decreasing order:				Results in decreasing order:			
104	29.000 (*)	8	19.400	117	32.000 (*)	146	24.200
13	27.550 (*)	27	19.400	13	29.750 (*)	166	24.100
115	22.700	146	19.350	115	27.100	24	24.100
14	21.100	6	19.300	14	26.100	8	24.000
3	21.030	17	19.250	126	26.100	150	24.000
121	21.000	1	19.200	5	25.800	17	24.000
126	21.000	165	19.100	116	25.640	7	23.900
5	20.700	166	19.080	31	25.630	165	23.800
114	20.700	151	19.050	11	25.300	33	23.800
20	20.410	112	18.970	121	25.100	6	23.800
31	20.400	110	18.800	15	25.100	1	23.700
102	20.300	38	18.800	172	25.100	110	23.700
116	20.260	16	18.600	19	25.000	153	23.700
172	20.200	7	18.500	157	25.000	112	23.570
11	20.200	39	18.420	3	24.990	16	23.500
15	20.100	152	18.200	20	24.950	164	23.500
2	20.100	158	18.200	21	24.940	107	23.200
157	20.000	107	18.100	114	24.900	108	23.100
24	20.000	108	18.100	102	24.900	118	23.000
21	19.820	118	18.000	124	24.900	104	23.000
33	19.800	117	18.000	10	24.800	158	22.700
124	19.800	113	17.800	4	24.700	39	22.520
19	19.800	36	17.700	155	24.660	113	22.300
4	19.700	22	16.800	27	24.400	152	22.300
155	19.620	23	16.500	2	24.400	36	22.290
10	19.600	164	16.100	151	24.350	22	20.900
153	19.530	150	16.060	38	24.300	23	20.600
163	19.400	40	15.000 (*)	163	24.200	40	16.330 (*)

Appendix 2

Figures




Figure 2: Youden plot of SO₂-S in absorbing solution.




Figure 3: Youden plot of SO₂-S on impregnated filter.




Figure 4: Youden plot of HNO₃-N on impregnated filter.




Figure 5: Youden plot of NO₂-N in absorbing solution.




Figure 6: Youden plot of NH₃-N on impregnated filter.




Figure 7: Youden plot of SO_4 -S in precipitation.




Figure 8: Youden plot of $\text{NO}_3\text{-N}$ in precipitation.




Figure 9: Youden plot of $\text{NH}_4\text{-N}$ in precipitation.




Figure 10: Youden plot of pH in precipitation.




Figure 11: Youden plot of strong acid in precipitation.




Figure 12: Youden plot of Cl in precipitation.




Figure 13: Youden plot of Na in precipitation.





Figure 14: Youden plot of Mg in precipitation.




Figure 15: Youden plot of Ca in precipitation.





Figure 16: Youden plot of K in precipitation.





Figure 17: Youden plot of conductivity in precipitation.